



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center - WO66-G609
Silver Spring, MD 20993-0002

Arthrex Inc.
Mr. David L. Rogers
Project Manager, Regulatory Affairs
1370 Creekside Boulevard
Naples, Florida 34108-1945

May, 16, 2017

Re: K171020

Trade/Device Name: Arthrex Knotless SutureTak Anchor
Regulation Number: 21 CFR 888.3040
Regulation Name: Smooth or threaded metallic bone fixation fastener
Regulatory Class: Class II
Product Code: MBI
Dated: April 4, 2017
Received: April 5, 2017

Dear Mr. Rogers:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-

related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely,

Mark N. Melkerson -S

Mark N. Melkerson
Director
Division of Orthopedic Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

510(k) Summary: K171020

Date Prepared	March 15, 2017
Submitter	Arthrex Inc. 1370 Creekside Boulevard Naples, FL 34108-1945
Contact Person	David L Rogers Project Manager, Regulatory Affairs 1-239-643-5553, ext. 71924 david.rogers@arthrex.com
Name of Device	Arthrex Knotless SutureTak Anchor
Common Name	Suture Anchor
Product Code	MBI
Classification Name	21 CFR 888.3040: Smooth or threaded metallic bone fixation fastener
Regulatory Class	II
Predicate Device	K063478: DePUY Mitek VERSALOK Anchor
Purpose of Submission	This traditional 510(k) premarket notification is submitted to obtain knee joint capsule closure indications for the Arthrex Knotless SutureTak originally cleared under K120155.
Device Description	The Arthrex Knotless SutureTak Anchor is a non-absorbable, “tap-in” suture anchor with a ribbed profile and a proximally placed external suture eyelet. The anchor is manufactured of Polyetheretherketone (PEEK) and is preloaded with Arthrex Suture and is offered preassembled on a driver. The anchor was originally cleared under K120155 for indications in the foot, ankle, knee, hand, wrist, elbow, shoulder, and hip.
Indications for Use	The Arthrex Knotless SutureTak Anchor is intended to be used for suture or tissue fixation in the foot, ankle, knee, hand, wrist, elbow, shoulder, and hip. Specific indications are listed below: • Elbow: Biceps Tendon Reattachment, Ulnar or Radial Collateral Ligament Reconstruction • Shoulder: Rotator Cuff Repairs, Bankart Repair, SLAP Lesion Repair, Biceps Tenodesis, Acromio-Clavicular Separation Repair, Deltoid Repair, Capsular Shift or Capsulolabral Reconstruction • Hand/Wrist: Scapholunate Ligament Reconstruction, Carpal Ligament Reconstructions, Repair/Reconstruction of collateral ligaments, Repair of Flexor and Extensor Tendons at the PIP, DIP, and MCP joints for all digits, Digital Tendon Transfers • Foot/Ankle: Lateral Stabilization, Medial Stabilization, Achilles Tendon Repair, Metatarsal Ligament Repair, Hallux Valgus reconstruction, Digital Tendon Transfers, Mid-foot Reconstruction • Knee: Medial Collateral Ligament Repair, Lateral Collateral Ligament Repair, Patellar Tendon Repair, Posterior Oblique Ligament Repair, Iliotibial Band Tenodesis, Joint Capsule Closure • Hip: Capsular Repair, Acetabular Labral repair
Performance Data	Pull-out testing was conducted to demonstrate that the Arthrex Knotless SutureTak Anchor performs statistically equivalent to the predicate device.

	Bacterial endotoxin per EP 2.6.14/USP <85> was conducted to demonstrate that the device meets pyrogen limit specifications.
Conclusion	The Arthrex Knotless SutureTak Anchor is substantially equivalent to the predicate device in which the basic design features and intended uses are the same. Any differences between the proposed device and the predicate device are considered minor and do not raise questions concerning safety or effectiveness. Based on the indications for use, technological characteristics, and the summary of data submitted, Arthrex Inc. has determined that the proposed device is substantially equivalent to the currently marketed predicate device.