



Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Center – WO66-G609  
Silver Spring, MD 20993-0002

Shenzhen Mindray Bio-Medical Electronics Co., Ltd.  
% Mr. Chao Ma  
Engineer of Technical Regulation Department  
Keji 12th Road South  
Hi-tech Industrial Park, Nanshan  
Shenzhen 518057 Guangdong  
P R CHINA

July 10, 2017

Re: K171034  
Trade/Device Name: M9/M9CV/M9T/M8 Elite Diagnostic Ultrasound System  
Regulation Number: 21 CFR 892.1550  
Regulation Name: Ultrasonic pulsed doppler imaging system  
Regulatory Class: II  
Product Code: IYN, IYO, ITX  
Dated: June 19, 2017  
Received: June 21, 2017

Dear Mr. Ma:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638 2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, “Misbranding by reference to premarket notification” (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH’s Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

A handwritten signature in blue ink that reads "Michael D. O'Hara". The signature is written over a large, light blue watermark of the letters "FDA".

For

Robert Ochs, Ph.D.  
Director  
Division of Radiological Health  
Office of In Vitro Diagnostics  
and Radiological Health  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)

K171034

Device Name

M9/M9CV/M9T/M8 Elite Diagnostic Ultrasound System

Indications for Use (Describe)

The M9/M9CV/M9T/M8 Elite Diagnostic Ultrasound System is applicable for adults, pregnant women, pediatric patients and neonates. It is intended for use in fetal, abdominal, Intra-operative, pediatric, small organ(breast, thyroid, testes), neonatal and adult cephalic, trans-rectal, trans-vaginal, musculo-skeletal(conventional, superficial), adult and pediatric cardiac, trans-esoph.(Cardiac), peripheral vessel , urology exams.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

Diagnostic Ultrasound Indications For Use Format

System: M9/M9CV/M9T/M8 Elite Diagnostic Ultrasound System

Transducer: N/A

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation							
General (Track 1 Only)	Specific (Track 1 & 3)	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Combined (specify)	Other (specify)
Ophthalmic	Ophthalmic								
Fetal Imaging & Other	Fetal	P	P	P		P	P	P	Note 1, 2,4,5,8,9
	Abdominal	P	P	P	P	P	P	P	Note 1, 2,4,5,8,9,10
	Intra-operative (Specify*)	N	N	N		N	N	N	Note1,2,8
	Intra-operative (Neuro)								
	Laparoscopic								
	Pediatric	P	P	P	P	P	P	P	Note 1, 2,4,5,8
	Small Organ (Specify**)	P	P	P		P	P	P	Note 1,2,5,6,8
	Neonatal Cephalic	P	P	P	P	P	P	P	Note 1,2,4,5,8
	Adult Cephalic	P	P	P	P	P	P	P	Note 1, 2,4,5,8
	Trans-rectal	P	P	P		P	P	P	Note 1,2,4,5,8
	Trans-vaginal	P	P	P		P	P	P	Note 1, 2,4,5,8
	Trans-urethral								
	Trans-esoph. (non-Card.)								
	Musculo-skeletal (Conventional)	P	P	P		P	P	P	Note 1,2,4,5,6,8
	Musculo-skeletal (Superficial)	P	P	P		P	P	P	Note 1,2,5,6,8
Intravascular									
Cardiac	Cardiac Adult	P	P	P	P	P	P	P	Note 1,2,3,4,5,7,8
	Cardiac Pediatric	P	P	P	P	P	P	P	Note 1,2,3,4,5,8
	Intravascular (Cardiac)								
	Trans-esoph. (Cardiac)	P	P	P	P	P	P	P	Note 1, 3,4,8
	Intra-cardiac								
Peripheral vessel	Peripheral vessel	P	P	P		P	P	P	Note 1,2,4,5,8
	Other (Specify***)	P	P	P		P	P	P	Note 1,2,4,5,8

N=new indication; P=previously cleared by FDA(K152543); E=added under Appendix E

Additional comments: Combined modes--B+M、PW+B、Color + B、Power + B、PW +Color+ B、Power + PW +B.

\*Intraoperative includes abdominal, thoracic, and vascular etc.

\*\*Small organ-breast, thyroid, testes.

\*\*\*Other use includes Urology.

Note 1: Tissue Harmonic Imaging. The feature does not use contrast agents.

Note 2: iScape

Note 3:TDI

Note 4: Color M

Note5: Biopsy Guidance

Note6: Elastography

Note7: Contrast imaging (Contrast agent for LVO)

Note8: Smart 3D

Note9: 4D(Real-time 3D)

Note10: Contrast imaging (Contrast agent for Liver)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostics and Radiological Health (OIR)

Prescription USE (Per 21 CFR 801.109)

System: M9/M9CV/M9T/M8 Elite Diagnostic Ultrasound System

Transducer: C11-3s

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation							
General (Track 1 Only)	Specific (Track 1 & 3)	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Combined (specify)	Other (specify)
Ophthalmic	Ophthalmic								
Fetal Imaging & Other	Fetal								
	Abdominal	P	P	P		P	P	P	Note 1,2,4,5,8
	Intra-operative (Specify*)								
	Intra-operative (Neuro)								
	Laparoscopic								
	Pediatric	P	P	P		P	P	P	Note 1,2,4,5,8
	Small Organ (Specify**)								
	Neonatal Cephalic	P	P	P		P	P	P	Note 1,2,4,5,8
	Adult Cephalic								
	Trans-rectal								
	Trans-vaginal								
	Trans-urethral								
	Trans-esoph. (non-Card.)								
	Musculo-skeletal (Conventional)								
Musculo-skeletal (Superficial)									
Intravascular									
Cardiac	Cardiac Adult								
	Cardiac Pediatric	P	P	P		P	P	P	Note 1,2,4,5,8
	Intravascular (Cardiac)								
	Trans-esoph. (Cardiac)								
	Intra-cardiac								
Peripheral vessel	Peripheral vessel	P	P	P		P	P	P	Note 1,2,4,5,8
	Other (Specify***)								

N=new indication; P=previously cleared by FDA(K152543); E=added under Appendix E

Additional comments: Combined modes--B+M、PW+B、Color + B、Power + B、PW +Color+ B、Power + PW +B.

\*Intraoperative includes abdominal, thoracic, and vascular etc.

\*\*Small organ-breast, thyroid, testes.

\*\*\*Other use includes Urology.

Note 1: Tissue Harmonic Imaging. The feature does not use contrast agents.

Note 2: iScape

Note 3:TDI

Note 4: Color M

Note5: Biopsy Guidance

Note6: Elastography

Note7: Contrast imaging (Contrast agent for LVO)

Note8: Smart 3D

Note9: 4D(Real-time 3D)

Note10: Contrast imaging (Contrast agent for Liver)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostics and Radiological Health (OIR)

Prescription USE (Per 21 CFR 801.109)

System: M9/M9CV/M9T/M8 Elite Diagnostic Ultrasound System

Transducer: C5-1s

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation							
General (Track 1 Only)	Specific (Track 1 & 3)	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Combined (specify)	Other (specify)
Ophthalmic	Ophthalmic								
Fetal Imaging & Other	Fetal	P	P	P		P	P	P	Note 1,2,4,5,8
	Abdominal	P	P	P		P	P	P	Note 1,2,4,5,8,10
	Intra-operative (Specify*)								
	Intra-operative (Neuro)								
	Laparoscopic								
	Pediatric	P	P	P		P	P	P	Note 1,2,4,5,8
	Small Organ (Specify**)								
	Neonatal Cephalic								
	Adult Cephalic								
	Trans-rectal								
	Trans-vaginal								
	Trans-urethral								
	Trans-esoph. (non-Card.)								
	Musculo-skeletal (Conventional)	P	P	P		P	P	P	Note 1,2,4,5,8
	Musculo-skeletal (Superficial)								
Intravascular									
Cardiac	Cardiac Adult								
	Cardiac Pediatric								
	Intravascular (Cardiac)								
	Trans-esoph. (Cardiac)								
	Intra-cardiac								
Peripheral vessel	Peripheral vessel	P	P	P		P	P	P	Note 1,2,4,5,8
	Other (Specify***)								

N=new indication; P=previously cleared by FDA(K152543); E=added under Appendix E

Additional comments: Combined modes--B+M、PW+B、Color + B、Power + B、PW +Color+ B、Power + PW +B.

\*Intraoperative includes abdominal, thoracic, and vascular etc.

\*\*Small organ-breast, thyroid, testes.

\*\*\*Other use includes Urology.

Note 1: Tissue Harmonic Imaging. The feature does not use contrast agents.

Note 2: iScape

Note 3:TDI

Note 4: Color M

Note5: Biopsy Guidance

Note6: Elastography

Note7: Contrast imaging (Contrast agent for LVO)

Note8: Smart 3D

Note9: 4D(Real-time 3D)

Note10: Contrast imaging (Contrast agent for Liver)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostics and Radiological Health (OIR)

Prescription USE (Per 21 CFR 801.109)

**010-4**

System: M9/M9CV/M9T/M8 Elite Diagnostic Ultrasound System

Transducer: L12-4s

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation							
General (Track 1 Only)	Specific (Track 1 & 3)	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Combined (specify)	Other (specify)
Ophthalmic	Ophthalmic								
Fetal Imaging & Other	Fetal								
	Abdominal	P	P	P		P	P	P	Note 1,2,5,8
	Intra-operative (Specify*)								
	Intra-operative (Neuro)								
	Laparoscopic								
	Pediatric	P	P	P		P	P	P	Note 1,2,5,8
	Small Organ (Specify**)	P	P	P		P	P	P	Note 1,2,5,6,8
	Neonatal Cephalic	P	P	P		P	P	P	Note 1,2,5,8
	Adult Cephalic								
	Trans-rectal								
	Trans-vaginal								
	Trans-urethral								
	Trans-esoph. (non-Card.)								
	Musculo-skeletal (Conventional)	P	P	P		P	P	P	Note 1,2,5,6,8
Musculo-skeletal (Superficial)	P	P	P		P	P	P	Note 1,2,5,6,8	
Intravascular									
Cardiac	Cardiac Adult								
	Cardiac Pediatric								
	Intravascular (Cardiac)								
	Trans-esoph. (Cardiac)								
	Intra-cardiac								
Peripheral vessel	Peripheral vessel	P	P	P		P	P	P	Note 1,2,5,8
	Other (Specify***)								

N=new indication; P=previously cleared by FDA(K152543); E=added under Appendix E

Additional comments: Combined modes--B+M, PW+B, Color + B, Power + B, PW +Color+ B, Power + PW +B.

\*Intraoperative includes abdominal, thoracic, and vascular etc.

\*\*Small organ-breast, thyroid, testes.

\*\*\*Other use includes Urology.

Note 1: Tissue Harmonic Imaging. The feature does not use contrast agents.

Note 2: iScape

Note 3:TDI

Note 4: Color M

Note5: Biopsy Guidance

Note6: Elastography

Note7: Contrast imaging (Contrast agent for LVO)

Note8: Smart 3D

Note9: 4D(Real-time 3D)

Note10: Contrast imaging (Contrast agent for Liver)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostics and Radiological Health (OIR)

Prescription USE (Per 21 CFR 801.109)

System: M9/M9CV/M9T/M8 Elite Diagnostic Ultrasound System

Transducer: P7-3Ts

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation							
General (Track 1 Only)	Specific (Track 1 & 3)	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Combined (specify)	Other (specify)
Ophthalmic	Ophthalmic								
Fetal Imaging & Other	Fetal								
	Abdominal								
	Intra-operative (Specify*)								
	Intra-operative (Neuro)								
	Laparoscopic								
	Pediatric								
	Small Organ (Specify**)								
	Neonatal Cephalic								
	Adult Cephalic								
	Trans-rectal								
	Trans-vaginal								
	Trans-urethral								
	Trans-esoph. (non-Card.)								
	Musculo-skeletal (Conventional)								
Musculo-skeletal (Superficial)									
Intravascular									
Cardiac	Cardiac Adult								
	Cardiac Pediatric								
	Intravascular (Cardiac)								
	Trans-esoph. (Cardiac)	P	P	P	P	P	P	P	Note 1, 3,4
	Intra-cardiac								
Peripheral vessel	Peripheral vessel								
	Other (Specify***)								

N=new indication; P=previously cleared by FDA(K152543); E=added under Appendix E

Additional comments: Combined modes--B+M, PW+B, Color + B, Power + B, PW +Color+ B, Power + PW +B.

\*Intraoperative includes abdominal, thoracic, and vascular etc.

\*\*Small organ-breast, thyroid, testes.

\*\*\*Other use includes Urology.

Note 1: Tissue Harmonic Imaging. The feature does not use contrast agents.

Note 2: iScape

Note 3:TDI

Note 4: Color M

Note5: Biopsy Guidance

Note6: Elastography

Note7: Contrast imaging (Contrast agent for LVO)

Note8: Smart 3D

Note9: 4D(Real-time 3D)

Note10: Contrast imaging (Contrast agent for Liver)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostics and Radiological Health (OIR)

Prescription USE (Per 21 CFR 801.109)



System: M9/M9CV/M9T/M8 Elite Diagnostic Ultrasound System

Transducer: SP5-1s

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation							
General (Track 1 Only)	Specific (Track 1 & 3)	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Combined (specify)	Other (specify)
Ophthalmic	Ophthalmic								
Fetal Imaging & Other	Fetal								
	Abdominal	P	P	P	P	P	P	P	Note 1,2,4,5,8
	Intra-operative (Specify*)								
	Intra-operative (Neuro)								
	Laparoscopic								
	Pediatric	P	P	P	P	P	P	P	Note 1,2,4,5,8
	Small Organ (Specify**)								
	Neonatal Cephalic	P	P	P	P	P	P	P	Note 1,2,4,5,8
	Adult Cephalic	P	P	P	P	P	P	P	Note 1,2,4,5,8
	Trans-rectal								
	Trans-vaginal								
	Trans-urethral								
	Trans-esoph. (non-Card.)								
	Musculo-skeletal (Conventional)								
Musculo-skeletal (Superficial)									
Intravascular									
Cardiac	Cardiac Adult	P	P	P	P	P	P	P	Note 1,2,3,4,5,7,8
	Cardiac Pediatric	P	P	P	P	P	P	P	Note 1,2,3,4,5,8
	Intravascular (Cardiac)								
	Trans-esoph. (Cardiac)								
	Intra-cardiac								
Peripheral vessel	Peripheral vessel								
	Other (Specify***)								

N=new indication; P=previously cleared by FDA(K152543); E=added under Appendix E

Additional comments: Combined modes--B+M, PW+B, Color + B, Power + B, PW +Color+ B, Power + PW +B.

\*Intraoperative includes abdominal, thoracic, and vascular etc.

\*\*Small organ-breast, thyroid, testes.

\*\*\*Other use includes Urology.

Note 1: Tissue Harmonic Imaging. The feature does not use contrast agents.

Note 2: iScape

Note 3:TDI

Note 4: Color M

Note5: Biopsy Guidance

Note6: Elastography

Note7: Contrast imaging (Contrast agent for LVO)

Note8: Smart 3D

Note9: 4D(Real-time 3D)

Note10: Contrast imaging (Contrast agent for Liver)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostics and Radiological Health (OIR)

Prescription USE (Per 21 CFR 801.109)

System: M9/M9CV/M9T/M8 Elite Diagnostic Ultrasound System

Transducer: P10-4s

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation							
General (Track 1 Only)	Specific (Track 1 & 3)	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Combined (specify)	Other (specify)
Ophthalmic	Ophthalmic								
Fetal Imaging & Other	Fetal								
	Abdominal	P	P	P	P	P	P	P	Note 1,4,8
	Intra-operative (Specify*)								
	Intra-operative (Neuro)								
	Laparoscopic								
	Pediatric	P	P	P	P	P	P	P	Note 1,4,8
	Small Organ (Specify**)								
	Neonatal Cephalic	P	P	P	P	P	P	P	Note 1,4,8
	Adult Cephalic								
	Trans-rectal								
	Trans-vaginal								
	Trans-urethral								
	Trans-esoph. (non-Card.)								
	Musculo-skeletal (Conventional)								
Musculo-skeletal (Superficial)									
Intravascular									
Cardiac	Cardiac Adult								
	Cardiac Pediatric	P	P	P	P	P	P	P	Note 1, 3,4,8
	Intravascular (Cardiac)								
	Trans-esoph. (Cardiac)								
	Intra-cardiac								
Peripheral vessel	Peripheral vessel								
	Other (Specify***)								

N=new indication; P=previously cleared by FDA(K152543); E=added under Appendix E

Additional comments: Combined modes--B+M、PW+B、Color + B、Power + B、PW +Color+ B、Power + PW +B.

\*Intraoperative includes abdominal, thoracic, and vascular etc.

\*\*Small organ-breast, thyroid, testes.

\*\*\*Other use includes Urology.

Note 1: Tissue Harmonic Imaging. The feature does not use contrast agents.

Note 2: iScape

Note 3:TDI

Note 4: Color M

Note5: Biopsy Guidance

Note6: Elastography

Note7: Contrast imaging (Contrast agent for LVO)

Note8: Smart 3D

Note9: 4D(Real-time 3D)

Note10: Contrast imaging (Contrast agent for Liver)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostics and Radiological Health (OIR)

Prescription USE (Per 21 CFR 801.109)

**010-8**

System: M9/M9CV/M9T/M8 Elite Diagnostic Ultrasound System

Transducer: L14-6Ns

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation							
General (Track 1 Only)	Specific (Track 1 & 3)	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Combined (specify)	Other (specify)
Ophthalmic	Ophthalmic								
Fetal Imaging & Other	Fetal								
	Abdominal	P	P	P		P	P	P	Note 1,2,5,8
	Intra-operative (Specify*)								
	Intra-operative (Neuro)								
	Laparoscopic								
	Pediatric	P	P	P		P	P	P	Note 1,2,5,8
	Small Organ (Specify**)	P	P	P		P	P	P	Note 1,2,5,6,8
	Neonatal Cephalic	P	P	P		P	P	P	Note 1,2,5,8
	Adult Cephalic								
	Trans-rectal								
	Trans-vaginal								
	Trans-urethral								
	Trans-esoph. (non-Card.)								
	Musculo-skeletal (Conventional)	P	P	P		P	P	P	Note 1,2,5,6,8
Musculo-skeletal (Superficial)	P	P	P		P	P	P	Note 1,2,5,6,8	
Intravascular									
Cardiac	Cardiac Adult								
	Cardiac Pediatric								
	Intravascular (Cardiac)								
	Trans-esoph. (Cardiac)								
	Intra-cardiac								
Peripheral vessel	Peripheral vessel	P	P	P		P	P	P	Note 1,2,5,8
	Other (Specify***)								

N=new indication; P=previously cleared by FDA(K152543); E=added under Appendix E

Additional comments: Combined modes--B+M, PW+B, Color + B, Power + B, PW +Color+ B, Power + PW +B.

\*Intraoperative includes abdominal, thoracic, and vascular etc.

\*\*Small organ-breast, thyroid, testes.

\*\*\*Other use includes Urology.

Note 1: Tissue Harmonic Imaging. The feature does not use contrast agents.

Note 2: iScape

Note 3:TDI

Note 4: Color M

Note5: Biopsy Guidance

Note6: Elastography

Note7: Contrast imaging (Contrast agent for LVO)

Note8: Smart 3D

Note9: 4D(Real-time 3D)

Note10: Contrast imaging (Contrast agent for Liver)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostics and Radiological Health (OIR)

Prescription USE (Per 21 CFR 801.109)

System: M9/M9CV/M9T/M8 Elite Diagnostic Ultrasound System

Transducer: V11-3Ws

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation							
General (Track 1 Only)	Specific (Track 1 & 3)	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Combined (specify)	Other (specify)
Ophthalmic	Ophthalmic								
Fetal Imaging & Other	Fetal	P	P	P		P	P	P	Note 1,2,4,5,8
	Abdominal								
	Intra-operative (Specify*)								
	Intra-operative (Neuro)								
	Laparoscopic								
	Pediatric								
	Small Organ (Specify**)								
	Neonatal Cephalic								
	Adult Cephalic								
	Trans-rectal	P	P	P		P	P	P	Note 1,2,4,5,8
	Trans-vaginal	P	P	P		P	P	P	Note 1,2,4,5,8
	Trans-urethral								
	Trans-esoph. (non-Card.)								
	Musculo-skeletal (Conventional)								
Musculo-skeletal (Superficial)									
Intravascular									
Cardiac	Cardiac Adult								
	Cardiac Pediatric								
	Intravascular (Cardiac)								
	Trans-esoph. (Cardiac)								
	Intra-cardiac								
Peripheral vessel	Peripheral vessel								
	Other (Specify***)	P	P	P		P	P	P	Note 1,2,4,5,8

N=new indication; P=previously cleared by FDA(K152543); E=added under Appendix E

Additional comments: Combined modes--B+M, PW+B, Color + B, Power + B, PW +Color+ B, Power + PW +B.

\*Intraoperative includes abdominal, thoracic, and vascular etc.

\*\*Small organ-breast, thyroid, testes.

\*\*\*Other use includes Urology.

Note 1: Tissue Harmonic Imaging. The feature does not use contrast agents.

Note 2: iScape

Note 3:TDI

Note 4: Color M

Note5: Biopsy Guidance

Note6: Elastography

Note7: Contrast imaging (Contrast agent for LVO)

Note8: Smart 3D

Note9: 4D(Real-time 3D)

Note10: Contrast imaging (Contrast agent for Liver)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE NEEDED)

Concurrence of CDRH, Office of Device Evaluation(ODE)

Prescription USE (Per 21 CFR 801.109)

**010-10**

System: M9/M9CV/M9T/M8 Elite Diagnostic Ultrasound System

Transducer: CW2s

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation							
General (Track 1 Only)	Specific (Track 1 & 3)	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Combined (specify)	Other (specify)
Ophthalmic	Ophthalmic								
Fetal Imaging & Other	Fetal								
	Abdominal								
	Intra-operative (Specify*)								
	Intra-operative (Neuro)								
	Laparoscopic								
	Pediatric					P			
	Small Organ (Specify**)								
	Neonatal Cephalic								
	Adult Cephalic					P			
	Trans-rectal								
	Trans-vaginal								
	Trans-urethral								
	Trans-esoph. (non-Card.)								
	Musculo-skeletal (Conventional)								
Musculo-skeletal (Superficial)									
Intravascular									
Cardiac	Cardiac Adult					P			
	Cardiac Pediatric					P			
	Intravascular (Cardiac)								
	Trans-esoph. (Cardiac)								
	Intra-cardiac								
Peripheral vessel	Peripheral vessel					P			
	Other (Specify***)								

N=new indication; P=previously cleared by FDA(K152543); E=added under Appendix E

Additional comments: Combined modes--B+M, PW+B, Color + B, Power + B, PW +Color+ B, Power + PW +B.

\*Intraoperative includes abdominal, thoracic, and vascular etc.

\*\*Small organ-breast, thyroid, testes.

\*\*\*Other use includes Urology.

Note 1: Tissue Harmonic Imaging. The feature does not use contrast agents.

Note 2: iScape

Note 3:TDI

Note 4: Color M

Note5: Biopsy Guidance

Note6: Elastography

Note7: Contrast imaging (Contrast agent for LVO)

Note8: Smart 3D

Note9: 4D(Real-time 3D)

Note10: Contrast imaging (Contrast agent for Liver)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE NEEDED)

Concurrence of CDRH, Office of Device Evaluation(ODE)

Prescription USE (Per 21 CFR 801.109)

System: M9/M9CV/M9T/M8 Elite Diagnostic Ultrasound System

Transducer: CW5s

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation							
General (Track 1 Only)	Specific (Track 1 & 3)	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Combined (specify)	Other (specify)
Ophthalmic	Ophthalmic								
Fetal Imaging & Other	Fetal								
	Abdominal								
	Intra-operative (Specify*)								
	Intra-operative (Neuro)								
	Laparoscopic								
	Pediatric				P				
	Small Organ (Specify**)								
	Neonatal Cephalic								
	Adult Cephalic								
	Trans-rectal								
	Trans-vaginal								
	Trans-urethral								
	Trans-esoph. (non-Card.)								
	Musculo-skeletal (Conventional)								
Musculo-skeletal (Superficial)									
Intravascular									
Cardiac	Cardiac Adult								
	Cardiac Pediatric				P				
	Intravascular (Cardiac)								
	Trans-esoph. (Cardiac)								
	Intra-cardiac								
Peripheral vessel	Peripheral vessel								
	Other (Specify***)								

N=new indication; P=previously cleared by FDA(K152543); E=added under Appendix E

Additional comments: Combined modes--B+M, PW+B, Color + B, Power + B, PW +Color+ B, Power + PW +B.

\*Intraoperative includes abdominal, thoracic, and vascular etc.

\*\*Small organ-breast, thyroid, testes.

\*\*\*Other use includes Urology.

Note 1: Tissue Harmonic Imaging. The feature does not use contrast agents.

Note 2: iScape

Note 3:TDI

Note 4: Color M

Note5: Biopsy Guidance

Note6: Elastography

Note7: Contrast imaging (Contrast agent for LVO)

Note8: Smart 3D

Note9: 4D(Real-time 3D)

Note10: Contrast imaging (Contrast agent for Liver)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE NEEDED)

Concurrence of CDRH, Office of Device Evaluation(ODE)

Prescription USE (Per 21 CFR 801.109)

**010-12**

System: M9/M9CV/M9T/M8 Elite Diagnostic Ultrasound System

Transducer: C6-2Gs

Intended Use: Diagnostic Ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation							
General (Track 1 Only)	Specific (Track 1 & 3)	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Combined (specify)	Other (specify)
Ophthalmic	Ophthalmic								
Fetal Imaging & Other	Fetal	P	P	P		P	P	P	Note1,2,4,5,8
	Abdominal	P	P	P		P	P	P	Note1,2,4,5,8,10
	Intra-operative (Specify*)								
	Intra-operative (Neuro)								
	Laparoscopic								
	Pediatric	P	P	P		P	P	P	Note1,2,4,5,8
	Small Organ (Specify**)								
	Neonatal Cephalic								
	Adult Cephalic								
	Trans-rectal								
	Trans-vaginal								
	Trans-urethral								
	Trans-esoph. (non-Card.)								
	Musculo-skeletal (Conventional)								
	Musculo-skeletal (Superficial)								
Intravascular									
Cardiac	Cardiac Adult								
	Cardiac Pediatric								
	Intravascular (Cardiac)								
	Trans-esoph. (Cardiac)								
	Intra-cardiac								
Peripheral vessel	Peripheral vessel								
	Other (Specify***)								

N=new indication; P=previously cleared by FDA(K152543); E=added under Appendix E

Additional comments: Combined modes--B+M、PW+B、Color + B、Power + B、PW +Color+ B、Power + PW +B.

\*Intraoperative includes abdominal, thoracic, and vascular etc.

\*\*Small organ-breast, thyroid, testes.

\*\*\*Other use includes Urology.

Note 1: Tissue Harmonic Imaging. The feature does not use contrast agents.

Note 2: iScape

Note 3:TDI

Note 4: Color M

Note5: Biopsy Guidance

Note6: Elastography

Note7: Contrast imaging (Contrast agent for LVO)

Note8: Smart 3D

Note9: 4D(Real-time 3D)

Note10: Contrast imaging (Contrast agent for Liver)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE NEEDED)

Concurrence of CDRH, Office of Device Evaluation(ODE)

Prescription USE (Per 21 CFR 801.109)

System: M9/M9CV/M9T/M8 Elite Diagnostic Ultrasound System

Transducer: L14-6Ws

Intended Use: Diagnostic Ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation							
General (Track 1 Only)	Specific (Track 1 & 3)	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Combined (specify)	Other (specify)
Ophthalmic	Ophthalmic								
Fetal Imaging & Other	Fetal								
	Abdominal	P	P	P		P	P	P	Note 1,2,5,8
	Intra-operative (Specify*)								
	Intra-operative (Neuro)								
	Laparoscopic								
	Pediatric	P	P	P		P	P	P	Note 1,2,5,8
	Small Organ (Specify**)	P	P	P		P	P	P	Note 1,2,5,6,8
	Neonatal Cephalic	P	P	P		P	P	P	Note 1,2,5,8
	Adult Cephalic								
	Trans-rectal								
	Trans-vaginal								
	Trans-urethral								
	Trans-esoph. (non-Card.)								
	Musculo-skeletal (Conventional)	P	P	P		P	P	P	Note 1,2,5,8
Musculo-skeletal (Superficial)	P	P	P		P	P	P	Note 1,2,5,8	
Intravascular									
Cardiac	Cardiac Adult								
	Cardiac Pediatric								
	Intravascular (Cardiac)								
	Trans-esoph. (Cardiac)								
	Intra-cardiac								
Peripheral vessel	Peripheral vessel	P	P	P		P	P	P	Note 1,2,5,8
	Other (Specify***)								

N=new indication; P=previously cleared by FDA(K152543); E=added under Appendix E

Additional comments: Combined modes--B+M, PW+B, Color + B, Power + B, PW +Color+ B, Power + PW +B.

\*Intraoperative includes abdominal, thoracic, and vascular etc.

\*\*Small organ-breast, thyroid, testes.

\*\*\*Other use includes Urology.

Note 1: Tissue Harmonic Imaging. The feature does not use contrast agents.

Note 2: iScape

Note 3:TDI

Note 4: Color M

Note5: Biopsy Guidance

Note6: Elastography

Note7: Contrast imaging (Contrast agent for LVO)

Note8: Smart 3D

Note9: 4D(Real-time 3D)

Note10: Contrast imaging (Contrast agent for Liver)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE NEEDED)

Concurrence of CDRH, Office of Device Evaluation(ODE)

Prescription USE (Per 21 CFR 801.109)

**010-14**



System: M9/M9CV/M9T/M8 Elite Diagnostic Ultrasound System

Transducer: P7-3s

Intended Use: Diagnostic Ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation							
General (Track 1 Only)	Specific (Track 1 & 3)	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Combined (specify)	Other (specify)
Ophthalmic	Ophthalmic								
Fetal Imaging & Other	Fetal								
	Abdominal	P	P	P	P	P	P	P	Note 1, 2,4,8
	Intra-operative (Specify*)								
	Intra-operative (Neuro)								
	Laparoscopic								
	Pediatric	P	P	P	P	P	P	P	Note 1, 2,4,8
	Small Organ (Specify**)								
	Neonatal Cephalic	P	P	P	P	P	P	P	Note 1, 2,4,8
	Adult Cephalic	P	P	P	P	P	P	P	Note 1, 2,4,8
	Trans-rectal								
	Trans-vaginal								
	Trans-urethral								
	Trans-esoph. (non-Card.)								
	Musculo-skeletal (Conventional)	P	P	P	P	P	P	P	Note 1, 2,4,8
Musculo-skeletal (Superficial)									
Intravascular									
Cardiac	Cardiac Adult	P	P	P	P	P	P	P	Note 1, 2,3,4,8
	Cardiac Pediatric	P	P	P	P	P	P	P	Note 1, 2,3,4,8
	Intravascular (Cardiac)								
	Trans-esoph. (Cardiac)								
	Intra-cardiac								
Peripheral vessel	Peripheral vessel								
	Other (Specify***)								

N=new indication; P=previously cleared by FDA(K152543); E=added under Appendix E

Additional comments: Combined modes--B+M、PW+B、Color + B、Power + B、PW +Color+ B、Power + PW +B.

\*Intraoperative includes abdominal, thoracic, and vascular etc.

\*\*Small organ-breast, thyroid, testes.

\*\*\*Other use includes Urology.

Note 1: Tissue Harmonic Imaging. The feature does not use contrast agents.

Note 2: iScape

Note 3:TDI

Note 4: Color M

Note5: Biopsy Guidance

Note6: Elastography

Note7: Contrast imaging (Contrast agent for LVO)

Note8: Smart 3D

Note9: 4D(Real-time 3D)

Note10: Contrast imaging (Contrast agent for Liver)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE NEEDED)

Concurrence of CDRH, Office of Device Evaluation(ODE)

Prescription USE (Per 21 CFR 801.109)

System: M9/M9CV/M9T/M8 Elite Diagnostic Ultrasound System

Transducer: L10-3s

Intended Use: Diagnostic Ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation							
General (Track 1 Only)	Specific (Track 1 & 3)	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Combined (specify)	Other (specify)
Ophthalmic	Ophthalmic								
Fetal Imaging & Other	Fetal								
	Abdominal	P	P	P		P	P	P	Note 1,2,8
	Intra-operative (Specify*)								
	Intra-operative (Neuro)								
	Laparoscopic								
	Pediatric	P	P	P		P	P	P	Note 1,2,8
	Small Organ (Specify**)	P	P	P		P	P	P	Note 1,2,6,8
	Neonatal Cephalic	P	P	P		P	P	P	Note 1,2,8
	Adult Cephalic								
	Trans-rectal								
	Trans-vaginal								
	Trans-urethral								
	Trans-esoph. (non-Card.)								
	Musculo-skeletal (Conventional)	P	P	P		P	P	P	Note 1,2,8
Musculo-skeletal (Superficial)	P	P	P		P	P	P	Note 1,2,8	
Intravascular									
Cardiac	Cardiac Adult								
	Cardiac Pediatric								
	Intravascular (Cardiac)								
	Trans-esoph. (Cardiac)								
	Intra-cardiac								
Peripheral vessel	Peripheral vessel	P	P	P		P	P	P	Note 1,2,8
	Other (Specify***)								

N=new indication; P=previously cleared by FDA(K152543); E=added under Appendix E

Additional comments: Combined modes--B+M、PW+B、Color + B、Power + B、PW +Color+ B、Power + PW +B.

\*Intraoperative includes abdominal, thoracic, and vascular etc.

\*\*Small organ-breast, thyroid, testes.

\*\*\*Other use includes Urology.

Note 1: Tissue Harmonic Imaging. The feature does not use contrast agents.

Note 2: iScape

Note 3:TDI

Note 4: Color M

Note5: Biopsy Guidance

Note6: Elastography

Note7: Contrast imaging (Contrast agent for LVO)

Note8: Smart 3D

Note9: 4D(Real-time 3D)

Note10: Contrast imaging (Contrast agent for Liver)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE NEEDED)

Concurrence of CDRH, Office of Device Evaluation(ODE)

Prescription USE (Per 21 CFR 801.109)

System: M9/M9CV/M9T/M8 Elite Diagnostic Ultrasound System

Transducer: L16-4Hs

Intended Use: Diagnostic Ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation							
General (Track 1 Only)	Specific (Track 1 & 3)	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Combined (specify)	Other (specify)
Ophthalmic	Ophthalmic								
Fetal Imaging & Other	Fetal								
	Abdominal	N	N	N		N	N	N	Note1,2,8
	Intra-operative (Specify*)	N	N	N		N	N	N	Note1,2,8
	Intra-operative (Neuro)								
	Laparoscopic								
	Pediatric	N	N	N		N	N	N	Note1,2,8
	Small Organ (Specify**)	N	N	N		N	N	N	Note1,2,8
	Neonatal Cephalic	N	N	N		N	N	N	Note1,2,8
	Adult Cephalic								
	Trans-rectal								
	Trans-vaginal								
	Trans-urethral								
	Trans-esoph. (non-Card.)								
	Musculo-skeletal (Conventional)	N	N	N		N	N	N	Note1,2,8
Musculo-skeletal (Superficial)	N	N	N		N	N	N	Note1,2,8	
Intravascular									
Cardiac	Cardiac Adult								
	Cardiac Pediatric								
	Intravascular (Cardiac)								
	Trans-esoph. (Cardiac)								
	Intra-cardiac								
Peripheral vessel	Peripheral vessel	N	N	N		N	N	N	Note1,2,8
	Other (Specify***)								

N=new indication; P=previously cleared by FDA; E=added under Appendix E

Additional comments: Combined modes--B+M, PW+B, Color + B, Power + B, PW +Color+ B, Power + PW +B.

\*Intraoperative includes abdominal, thoracic, and vascular etc.

\*\*Small organ-breast, thyroid, testes.

\*\*\*Other use includes Urology.

Note 1: Tissue Harmonic Imaging. The feature does not use contrast agents.

Note 2: iScape

Note 3:TDI

Note 4: Color M

Note5: Biopsy Guidance

Note6: Elastography

Note7: Contrast imaging (Contrast agent for LVO)

Note8: Smart 3D

Note9: 4D(Real-time 3D)

Note10: Contrast imaging (Contrast agent for Liver)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE NEEDED)

Concurrence of CDRH, Office of Device Evaluation(ODE)

Prescription USE (Per 21 CFR 801.109)

System: M9/M9CV/M9T/M8 Elite Diagnostic Ultrasound System

Transducer: P8-3Ts

Intended Use: Diagnostic Ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation							
General (Track 1 Only)	Specific (Track 1 & 3)	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Combined (specify)	Other (specify)
Ophthalmic	Ophthalmic								
Fetal Imaging & Other	Fetal								
	Abdominal								
	Intra-operative (Specify*)								
	Intra-operative (Neuro)								
	Laparoscopic								
	Pediatric								
	Small Organ (Specify**)								
	Neonatal Cephalic								
	Adult Cephalic								
	Trans-rectal								
	Trans-vaginal								
	Trans-urethral								
	Trans-esoph. (non-Card.)								
	Musculo-skeletal (Conventional)								
Musculo-skeletal (Superficial)									
Intravascular									
Cardiac	Cardiac Adult								
	Cardiac Pediatric								
	Intravascular (Cardiac)								
	Trans-esoph. (Cardiac)	N	N	N	N	N	N	N	Note 1,3,4
	Intra-cardiac								
Peripheral vessel	Peripheral vessel								
	Other (Specify***)								

N=new indication; P=previously cleared by FDA; E=added under Appendix E

Additional comments: Combined modes--B+M, PW+B, Color + B, Power + B, PW +Color+ B, Power + PW +B.

\*Intraoperative includes abdominal, thoracic, and vascular etc.

\*\*Small organ-breast, thyroid, testes.

\*\*\*Other use includes Urology.

Note 1: Tissue Harmonic Imaging. The feature does not use contrast agents.

Note 2: iScape

Note 3:TDI

Note 4: Color M

Note5: Biopsy Guidance

Note6: Elastography

Note7: Contrast imaging (Contrast agent for LVO)

Note8: Smart 3D

Note9: 4D(Real-time 3D)

Note10: Contrast imaging (Contrast agent for Liver)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE NEEDED)

Concurrence of CDRH, Office of Device Evaluation(ODE)

Prescription USE (Per 21 CFR 801.109)

System: M9/M9CV/M9T/M8 Elite Diagnostic Ultrasound System

Transducer: L20-5s

Intended Use: Diagnostic Ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation							
General (Track 1 Only)	Specific (Track 1 & 3)	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Combined (specify)	Other (specify)
Ophthalmic	Ophthalmic								
Fetal Imaging & Other	Fetal								
	Abdominal								
	Intra-operative (Specify*)								
	Intra-operative (Neuro)								
	Laparoscopic								
	Pediatric								
	Small Organ (Specify**)	P	P	P		P	P	P	Note 1,2,8
	Neonatal Cephalic								
	Adult Cephalic								
	Trans-rectal								
	Trans-vaginal								
	Trans-urethral								
	Trans-esoph. (non-Card.)								
	Musculo-skeletal (Conventional)		P	P	P		P	P	P
Musculo-skeletal (Superficial)		P	P	P		P	P	P	Note 1,2,6,8
Intravascular									
Cardiac	Cardiac Adult								
	Cardiac Pediatric								
	Intravascular (Cardiac)								
	Trans-esoph. (Cardiac)								
	Intra-cardiac								
Peripheral vessel	Peripheral vessel	P	P	P		P	P	P	Note 1,2,8
	Other (Specify***)								

N=new indication; P=previously cleared by FDA(K161525); E=added under Appendix E

Additional comments: Combined modes--B+M, PW+B, Color + B, Power + B, PW +Color+ B, Power + PW +B.

\*Intraoperative includes abdominal, thoracic, and vascular etc.

\*\*Small organ-breast, thyroid, testes.

\*\*\*Other use includes Urology.

Note 1: Tissue Harmonic Imaging. The feature does not use contrast agents.

Note 2: iScape

Note 3:TDI

Note 4: Color M

Note5: Biopsy Guidance

Note6: Elastography

Note7: Contrast imaging (Contrast agent for LVO)

Note8: Smart 3D

Note9: 4D(Real-time 3D)

Note10: Contrast imaging (Contrast agent for Liver)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE NEEDED)

Concurrence of CDRH, Office of Device Evaluation(ODE)

Prescription USE (Per 21 CFR 801.109)

System: M9/M9CV/M9T/M8 Elite Diagnostic Ultrasound System

Transducer: D7-2s

Intended Use: Diagnostic Ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation							
General (Track 1 Only)	Specific (Track 1 & 3)	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Combined (specify)	Other (specify)
Ophthalmic	Ophthalmic								
Fetal Imaging & Other	Fetal	N	N	N		N	N	N	Note 1,2,4,9
	Abdominal	N	N	N		N	N	N	Note 1,2,4,9
	Intra-operative (Specify*)								
	Intra-operative (Neuro)								
	Laparoscopic								
	Pediatric								
	Small Organ (Specify**)								
	Neonatal Cephalic								
	Adult Cephalic								
	Trans-rectal								
	Trans-vaginal								
	Trans-urethral								
	Trans-esoph. (non-Card.)								
	Musculo-skeletal (Conventional)								
	Musculo-skeletal (Superficial)								
Intravascular									
Cardiac	Cardiac Adult								
	Cardiac Pediatric								
	Intravascular (Cardiac)								
	Trans-esoph. (Cardiac)								
	Intra-cardiac								
Peripheral vessel	Peripheral vessel								
	Other (Specify***)								

N=new indication; P=previously cleared by FDA; E=added under Appendix E

Additional comments: Combined modes--B+M, PW+B, Color + B, Power + B, PW +Color+ B, Power + PW +B.

\*Intraoperative includes abdominal, thoracic, and vascular etc.

\*\*Small organ-breast, thyroid, testes.

\*\*\*Other use includes Urology.

Note 1: Tissue Harmonic Imaging. The feature does not use contrast agents.

Note 2: iScape

Note 3:TDI

Note 4: Color M

Note5: Biopsy Guidance

Note6: Elastography

Note7: Contrast imaging (Contrast agent for LVO)

Note8: Smart 3D

Note9: 4D(Real-time 3D)

Note10: Contrast imaging (Contrast agent for Liver)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE NEEDED)

Concurrence of CDRH, Office of Device Evaluation(ODE)

Prescription USE (Per 21 CFR 801.109)

**010-20**

System: M9/M9CV/M9T/M8 Elite Diagnostic Ultrasound System

Transducer: L14-5sp

Intended Use: Diagnostic Ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation							
General (Track 1 Only)	Specific (Track 1 & 3)	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Combined (specify)	Other (specify)
Ophthalmic	Ophthalmic								
Fetal Imaging & Other	Fetal								
	Abdominal	P	P	P		P	P	P	Note 1,2,8
	Intra-operative (Specify*)	P	P	P		P	P	P	Note 1,2,8
	Intra-operative (Neuro)								
	Laparoscopic								
	Pediatric	P	P	P		P	P	P	Note 1,2,8
	Small Organ (Specify**)	P	P	P		P	P	P	Note 1,2,8
	Neonatal Cephalic	P	P	P		P	P	P	Note 1,2,8
	Adult Cephalic								
	Trans-rectal								
	Trans-vaginal								
	Trans-urethral								
	Trans-esoph. (non-Card.)								
	Musculo-skeletal (Conventional)	P	P	P		P	P	P	Note 1,2,8
Musculo-skeletal (Superficial)	P	P	P		P	P	P	Note 1,2,8	
Intravascular									
Cardiac	Cardiac Adult								
	Cardiac Pediatric								
	Intravascular (Cardiac)								
	Trans-esoph. (Cardiac)								
	Intra-cardiac								
Peripheral vessel	Peripheral vessel	P	P	P		P	P	P	Note 1,2,8
	Other (Specify***)								

N=new indication; P=previously cleared by FDA(K161525); E=added under Appendix E

Additional comments: Combined modes--B+M、PW+B、Color + B、Power + B、PW +Color+ B、Power + PW +B.

\*Intraoperative includes abdominal, thoracic, and vascular etc.

\*\*Small organ-breast, thyroid, testes.

\*\*\*Other use includes Urology.

Note 1: Tissue Harmonic Imaging. The feature does not use contrast agents.

Note 2: iScape

Note 3:TDI

Note 4: Color M

Note5: Biopsy Guidance

Note6: Elastography

Note7: Contrast imaging (Contrast agent for LVO)

Note8: Smart 3D

Note9: 4D(Real-time 3D)

Note10: Contrast imaging (Contrast agent for Liver)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE NEEDED)

Concurrence of CDRH, Office of Device Evaluation(ODE)

Prescription USE (Per 21 CFR 801.109)

# 510(K) SUMMARY

This summary of 510(k) safety and effectiveness information is being submitted in accordance with the requirements of SMDA 1990 and 21 CFR §807.92(c).

The assigned 510(k) number is: K171034.

## **1. Submitter:**

Shenzhen Mindray Bio-medical Electronics Co., LTD  
Mindray Building, Keji 12th Road South, Hi-tech Industrial Park, Nanshan,  
Shenzhen, 518057, P. R. China

Tel: +86 755 8188 6573

Fax: +86 755 2658 2680

### **Contact Person:**

Ma Chao

Shenzhen Mindray Bio-medical Electronics Co., LTD  
Mindray Building, Keji 12th Road South, Hi-tech Industrial Park,  
Nanshan, Shenzhen, 518057, P. R. China

**Date Prepared:** March 29, 2017

## **2. Device Name:** M9/M9T/M9CV/M8 Elite Diagnostic Ultrasound System

### **Classification**

Regulatory Class: II

Review Category: Tier II

21 CFR 892.1550 Ultrasonic Pulsed Doppler Imaging System (IYN)

21 CFR 892.1560 Ultrasonic Pulsed Echo Imaging System (IYO)

21 CFR 892.1570 Diagnostic Ultrasound Transducer (ITX)

## **3. Device Description:**

M9/M9T/M9CV/M8 Elite Diagnostic Ultrasound System is a general purpose, software controlled, ultrasonic diagnostic system. Its function is to acquire and display ultrasound data in B-Mode, M-Mode, PW-Mode, CW-Mode, Color-Mode, Power/Dirpower Mode, THI, iScape mode, TDI mode, Color M, Elastography, 3D/4D mode, Contrast imaging (Contrast agent for Liver), LVO or the combined



mode (i.e. B/M-Mode, B/PW-mode, B/PW/Color).

This system is a Track 3 device that employs an array of probes that include linear array, convex array and phased array.

#### **4. Intended Use:**

The M9/M9CV/M9T/M8 Elite Diagnostic Ultrasound System is applicable for adults, pregnant women, pediatric patients and neonates. It is intended for use in fetal, abdominal, Intra-operative, pediatric, small organ(breast, thyroid, testes), neonatal and adult cephalic, trans-rectal, trans-vaginal, musculo-skeletal(conventional, superficial), adult and pediatric cardiac, trans-esoph.(Cardiac), peripheral vessel , urology exams.

#### **5. Summary of Modifications and New Added Features**

This submission device is a modification to M9/M9CV/M9T/M8 Elite Diagnostic Ultrasound System previously cleared in K152543.

The following is a brief overview of the modifications and new added features.

- New added transducers  
L16-4Hs, P8-3Ts, L20-5s, D7-2s and L14-5sp
- New added needle-guided bracket  
CIVCO 658-001, CIVCO 610-543 and CIVCO 610-1274
- New added model  
M8 Elite
- New added features  
4D Function  
Smart 3D  
Smart Volume  
iPage  
Contrast imaging (Contrast agent for Liver)  
Healthy fetus weight algorithm to obstetric measurement  
FG and GA formula of fetal foot length  
MAPSE measurement in the M mode  
LV and RV of FAC calculation

#### **6. Comparison with Predicate Devices:**

M9/M9T/M9CV/M8 Elite Diagnostic Ultrasound System is comparable with and substantially equivalent to these predicate devices:

Predicate Device	Manufacturer	Model	510(k) Control Number
------------------	--------------	-------	-----------------------

1	Mindray	M9/M9CV/M9T (Main predicate device)	K152543
2	Mindray	DC-8	K170277
3	Mindray	Resona 7	K162267

Compared to the predicate devices M9/M9CV/M9T (K152543):

- M9/M9T/M9CV/M8 Elite has the same technological characteristics, are comparable in key safety and effectiveness features, and have the same intended uses and basic operating modes. All systems transmit ultrasonic energy into patients, perform post processing of received echoes to generate onscreen display of anatomic structures and fluid flow within the body. All systems allow for specialized measurements of structures and flow, and calculations.
- The transducers are same except for adding L16-4Hs, P8-3Ts, L20-5s, D7-2s and L14-5sp. All of the new added probes are substantial equivalent with the predicate DC-8 (K170277).
- The patient contact materials of the new added transducers are tested under ISO 10993-1.
- The newly added needle-guided bracket that was purchased from CIVCO MEDICAL INSTRUMENTS CO., INC. be cleared along with CIVCO 658-001 (K093713), CIVCO 610-543 (K970514) and CIVCO 610-1274 (K970514).
- It has the system acoustic power level which is below FDA limits and it is designed in compliance with same FDA recognized standards.
- M9/M9T/M9CV/M8 Elite has the same capability in term of measurements and calculation functions except adding healthy fetus weight algorithm, FG and GA formula of fetal foot length, MAPSE measurement in the M mode and LV and RV of FAC calculation. New added features are substantial equivalent with the predicates Resona 7 (K162267).
- M9/M9T/M9CV/M8 Elite has the same performance specification as the predicate device M9/M9T/M9CV (K152543) except adding 4D Function, Smart 3D, Smart Volume, iPage and Contrast imaging (Contrast agent for Liver). New added functions are substantial equivalent with the predicates DC-8 (K170277).

## **7. Non-clinical Tests:**

M9/M9T/M9CV/M8 Elite Diagnostic Ultrasound System has been evaluated for acoustic output, biocompatibility, cleaning and disinfection effectiveness as well as thermal, electrical and mechanical safety, and has been designed to conform with applicable medical safety standards. This device has been tested and evaluated under the following standards:

- UD 2 Acoustic Output Measurement Standard for Diagnostic Ultrasound Equipment.

- UD 3 Standard for Real Time Display of Thermal and Mechanical Acoustic Output Indices on Diagnostic Ultrasound Equipment
- AAMI/ANSI ES60601-1 Medical electrical equipment - Part 1: General requirements for basic safety and essential performance.
- AAMI/ANSI/IEC 60601-1-2 Medical electrical equipment - Part 1-2: General requirements for basic safety and essential performance - Collateral standard: Electromagnetic compatibility - Requirements and tests (Edition 3)
- IEC 60601-1-6 Medical electrical equipment - part 1-6: general requirements for basic safety and essential performance - collateral standard: usability
- IEC 60601-2-37 Medical electrical equipment - Part 2-37: Particular requirements for the basic safety and essential performance of ultrasonic medical diagnostic and monitoring equipment
- ISO 14971 Medical devices - Application of risk management to medical devices
- ISO 10993-1 Biological evaluation of medical devices -- Part 1: Evaluation and testing within a risk management process
- IEC 62304 Medical device software - Software life cycle processes
- IEC 62366 Medical devices - Application of usability engineering to medical devices

These non-clinical tests relied on in this premarket notification submission can support the determination of substantial equivalence of the subject device.

## **8. Clinical Studies**

Not applicable. The subject of this submission, M9/M9T/M9CV/M8 Elite Diagnostic Ultrasound System, does not require clinical studies to support substantial equivalence.

### **Conclusion:**

Intended uses and other key features are consistent with traditional clinical practices, FDA guidelines and established methods of patient examination. The design, development and quality process of the manufacturer confirms with 21 CFR 820, ISO 9001 and ISO 13485 quality systems. The device conforms to applicable medical device safety standards. Therefore, the M9/M9T/M9CV/M8 Elite Diagnostic Ultrasound System is substantially equivalent with respect to safety and effectiveness to devices currently cleared for market.