



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center - WO66-G609
Silver Spring, MD 20993-0002

September 6, 2017

Katalyst Inc.
Amish Patel
Chief Product Officer
316 Occidental Avenue
South B300
Seattle, Washington 98104

Re: K171035
Trade/Device Name: Katalyst Mark 1 Muscle Stimulation System
Regulation Number: 21 CFR 890.5850
Regulation Name: Powered Muscle Stimulator
Regulatory Class: Class II
Product Code: NGX
Dated: August 1, 2017
Received: August 3, 2017

Dear Mr. Patel:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education (DICE) at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education (DICE) at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely,

Vivek J. Pinto -S

for Carlos L. Peña, PhD, MS
Director
Division of Neurological
and Physical Medicine Devices
Office of Device Evaluation
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K171035

Device Name

Katalyst Mark 1 Muscle Stimulation System

Indications for Use (Describe)

The Katalyst Mark 1 is intended to stimulate healthy muscles in order to improve or facilitate muscle performance.

It is not intended to be used in conjunction with therapy or treatment of medical diseases or medical conditions of any kind. None of the training programs or operational parameters are designed to target injured or ailing muscles and its use on such muscles is contraindicated.

The Katalyst Mark 1 is Rx only.

Type of Use (Select one or both, as applicable)

☒ Prescription Use (Part 21 CFR 801 Subpart D)

☐ Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(K) SUMMARY

This summary is being submitted in accordance with the requirements of 21 CFR 807.92.

1.1 Submitter

Katalyst Inc.
316 Occidental Avenue S.
B300
Seattle, WA 98104

Contact: Amish Patel, Chief Product Officer
Tel: 406-616-3671
Email: devices@katalyst-fitness.com

1.2 Date Prepared

September 6, 2017

1.3 Device Name and Classification Information:

Trade/Proprietary Name: Katalyst Mark 1 Muscle Stimulation System
Common Name: Powered Muscle Stimulator for Muscle Conditioning
Classification Name: Powered Muscle Stimulator
Classification: 21 CFR 890.5850, Class II
Product Code: NGX
Panel: 89, Physical Medicine

1.4 Predicate Devices

Primary Predicate: Compex® Sport Plus (K083140)

Secondary Predicate: E-Fit EF-1280 (K133225)

1.5 Device Description

The Katalyst Mark 1 Muscle Stimulation System consists of the following three main components:

Katalyst Suit	Worn by the client, the Katalyst Suit is a compression textile with embedded electrodes connected by a cable harness. When connected to the Impulse Pack it provides localized EMS impulses to major muscle groups at controllable intensities.
----------------------	---

Impulse Pack	A powered muscle stimulation device specifically designed to mate with the Katalyst Suit through a single port connector. It delivers electrical impulses through a battery powered regulated booster. The Impulse Pack is operator controlled wirelessly via the Katalyst Application over Bluetooth.
Katalyst Application	Controlled by the Operator, the Katalyst Application is run on a Microsoft Surface Pro 3 running Windows 10 Professional. It is used to configure and run a Katalyst training, including the setting of independent intensities for each muscle group channel.

Accessories included: cable harnesses and an Impulse Pack charging cable.

1.6 Indications for Use:

The Katalyst Mark 1 is intended to stimulate healthy muscles in order to improve or facilitate muscle performance.

It is not intended to be used in conjunction with therapy or treatment of medical diseases or medical conditions of any kind. None of the training programs or operational parameters are designed to target injured or ailing muscles and its use on such muscles is contraindicated.

The Katalyst Mark 1 is Rx only.

1.7 Contraindications

The Katalyst Mark 1 should only be operated by a certified Katalyst Operator. Trainings should not be used by anyone under 18 years of age, with a cardiac pacemaker, with known heart conditions, with a history of cardiac arrhythmia, who is pregnant, with an implanted defibrillator, or with other implanted metallic or electronic device. Such use could cause cardiac arrhythmia, electric shock, burns, electrical interference, or death. Other contraindications are provided in the operator guide.

1.8 Basic Device Characteristics - Comparison with Predicate Devices

Characteristic	New Device (Katalyst Mark 1 Muscle Stimulation System)	Primary Predicate Compex Sport Plus (K083140)	Secondary Predicate: E-Fit EF-1280 (K133225)	Similar/ Different
Device Name, Model	Katalyst Mark 1 Muscle Stimulation System	Compex Sport Plus (currently marketed	E-Fit EF-1280	K083140 - Different

		as Compex Sport Elite)		K133225 - Different
Manufacturer	Katalyst Inc.	DJO LLC	FIT PRO, LLC	K083140 - Different K133225 - Different
Indications for Use	The Katalyst Mark 1 is intended to stimulate healthy muscles in order to improve or facilitate muscle performance. It is not intended to be used in conjunction with therapy or treatment of medical diseases or medical conditions of any kind. None of the training programs or operational parameters are designed to target injured or ailing muscles and its use on such muscles is contraindicated. The Katalyst Mark 1 is Rx only.	The Compex® Sport Plus is intended to stimulate healthy muscles in order to improve or facilitate muscle performance. The Compex® Sport Plus is not intended to be used in conjunction with therapy or treatment of medical diseases or medical conditions of any kind. None of the Compex® Sport Plus training programs is designed for injured or ailing muscles and its use on such muscles is contraindicated. The Compex® Sport Plus electrical impulses allow the triggering of action potentials on motoneurons of motor nerves (excitations). These excitations of motoneurons are transmitted to the muscle fibers via the	E-fit EF-1280 is a machine with electronic muscle stimulation based on EMS technology. Regarding its use, the device is specifically designed as an addition to other sports and for training muscles. It must be used for only healthy muscles and clients, not for rehabilitation purposes. The E-Fit EF-1280 intended to stimulate healthy muscles in order to improve or facilitate muscle performance. The E-Fit EF-1280 is not intended to be used in conjunction with therapy or treatment of medical diseases or medical conditions of any kind. None	K083140 - Similar K133225 - Similar

		motor endplate where they generate mechanical muscle fiber responses that correspond to muscle work. Depending on the parameters of the electrical impulses (pulse frequency, duration of contraction, duration of rest, total session duration), different types of muscle work can be imposed on the stimulated muscles. The various types of muscle work that the Compex® Sport Plus can impose on the stimulated muscles are able to improve or facilitate muscle performance. The Compex® Sport Plus may therefore be considered a technique of muscle training	of the E-Fit EF-1280 training programs is designed for injured or ailing muscles and its use on such muscles is contraindicated. The E-Fit EF-1280 electrical impulses allow the triggering of action potentials on motoneurons of motor nerves (excitations). These excitations of motoneurons are transmitted to the muscle fibers via the motor endplate where they generate mechanical muscle fiber responses that correspond to muscle work. Depending on the parameters of the electrical impulses (pulse frequency, duration of contraction, duration of rest, total session duration), different types of muscle work can be imposed	
--	--	---	--	--

			on the stimulated muscles. The various types of muscle work that the E-Fit EF-1280 can impose on the stimulated muscles are able to improve or facilitate muscle performance. The E-Fit EF-1280 may therefore be considered a technique of muscle training.	
Connection of device to electrodes	The Katalyst Impulse Pack unit is connected to the electrodes via a cable electrode harness. The connection to the Impulse Pack is via a ROHS compliant standard DB-25 Female connector, and snapping onto the electrodes using standard 4mm SNAP connectors integrated in the electrodes themselves. The SNAP headers are combined and over molded in the lead cables.	With 6-pole cables including female custom SNAP plugged on the custom Compex female SNAP assembled in the electrode. Entire electronic circuit for four (4) Stimulation Channels and User Interface is combined into same casing, connected to the electrodes with 6-pole cables.	The EF-1280 unit is connected to the electrodes via a cable electrode harness which is connected to the EMS device with a physical connection. The connection to the EMS device is via a plastic 12pin waterproof ip68 Connector and the connection to the electrodes is via SNAP connectors directly onto the outer vest and leg straps.	K083140 - Similar K133225 - Similar
Power Source(s)	Li-Po Battery 3.8V, 7.79Wh (2050mAh)	Rechargeable Ni-Mh Battery 4.6V (4 cells AA=R6)	Lead Acid Battery	K083140 - Different K133225 -

	IEC 62133: 2012 (2nd Edition)		12V, 3400mAH	Different Difference does not adversely affect the safety and effectiveness of the device
Charging System	AC/DC wall plug in, distributed directly to the device. 18.0[v], 1.33[A]	AC/DC 9[v] 0.4[A] Distributed directly to the device	AC/DC wall plug in, distributed directly to the device.	K083140 - Different K133225 - Different Difference does not adversely affect the safety and effectiveness of the device
Waveform	Bipolar Square	Bipolar Square	Bipolar Square	K083140 – Similar K133225 - Similar
Channels	10	4	12	K083140 - Different K133225 - Similar
Operation Modes	3	9	5+5	K083140 - Similar K133225 - Different
Charging Isolation	Yes	Yes	Yes	K083140 - Similar K133225 - Similar
Rechargeable Integrated Battery	Yes	Yes	Yes	K083140 - Similar K133225 - Similar
Display / Control Interface	Yes; (LCD Display on Microsoft Surface Pro 3 PC)	Yes, Integrated (LCD)	LCD 2x40 character LCD display with	K083140 - Different K133225 -

			LED backlight. The rotary encoder allows for a quick setup and ease of use	Different See section 1.10 Differences below for justification of differences
Per Channel Control	Yes. Using Windows 10 control application	Yes, Using integrated buttons.	Yes. Using rotary encoder knobs for each channel.	K083140 - Similar K133225 - Similar
Electrode Placement	Integrated into wearable body suit and positioned over corresponding muscle group positions. Cannot be modified/alterd.	Not defined. Placement of electrodes is dictated by the operator.	Placed into outer layer, appropriately positioned over corresponding muscle group positions. Cannot be modified/alterd.	K083140 - Different K133225 - Similar
Electrode Leads	21 CFR 898 compliant.	21 CFR 898 compliant.	21 CFR 898 compliant.	K083140 - Similar K133225 - Similar
Electrode Interface	4mm Standard SNAP Connector.	4mm Standard SNAP Connector.	4mm Standard SNAP Connector.	K083140 - Similar K133225 - Similar
Electrodes	Only 2 per channel; Max 20.	2 per channel; Max 8.	2 per channel; Max 24.	K083140 - Similar K133225 - Similar
Transthoracic stimulation	Yes	No	Yes	K083140 - Different K133225 - Similar electrode placement and safety mitigations as K133225
Wireless	Yes. Utilizing	No	No	K083140 -

Communication with Control Interface	Bluetooth Smart (BT LE 4.0).			Different K133225 - Different See section 1.10 Differences below for justification of differences
Mobile / Form Factor	Yes (tablet, handheld device and wearable suit).	Yes (pocket device with travel case)	Portable with difficulty, no mobile device, its intended use requires the qualified and trained operator.	K083140 – Similar K133225 - Similar
Channel Control Interface	Touch User Interface provided by Katalyst Application on Surface Pro 3 PC	Physical Buttons and Display on unit.	Using rotary encoder knobs for each channel and master rotary encoder for all channels.	K083140 – Different K133225 - Different See section 1.10 Differences below for justification of differences
Instant Shutoff	Physical Button	Physical Button	Physical Button	K083140 - Similar K133225 - Similar
Microprocessor Controlled?	Yes	Yes	Yes	K083140 - Similar K133225 - Similar
Automatic overload trip?	Yes	Yes	Yes	K083140 - Similar K133225 - Similar
Automatic No-Load Trip?	Yes	Yes	Yes	K083140 - Similar K133225 - Similar

Automatic Shutoff?	On/Off Switch	On/Off Switch	On/Off Switch	K083140 - Similar K133225 - Similar
Patient Override Control?	Yes	Yes	Yes	K083140 - Similar K133225 - Similar
Interface Indicators	On/Off - Yes (on Impulse Pack via LED and in application) Low Battery - Yes (on Impulse Pack) Voltage/Current Level - Yes (in application). Charging - yes on Impulse Pack. Duration of training- yes (in application)	On/Off - yes Low Battery - yes Voltage/Current Level - yes Charging - yes Duration of training training- yes.	On/Off - yes Low Battery - yes Voltage/Current Level - yes Charging - yes Duration of training training- yes.	K083140 - Similar K133225 - Similar
Duration in minutes	40 (shown on display)	55	30 minutes	K083140 - Similar K133225 - Similar

Table 10: Device Characteristic Comparisons

1.9 Output Specification - Comparison with Predicate Devices

Characteristic	New Device (Katalyst Mark 1 Muscle Stimulation System)	Primary Predicate: Compex Sport Plus (K083140)	Secondary Predicate: E- Fit EF-1280 (K133225)	Similar/ Different
Waveform	Symmetrical Biphasic	Symmetrical Biphasic	Symmetrical Biphasic	K083140 - Similar K133225 -

				Similar
Shape	Rectangular	Rectangular	Rectangular	K083140 - Similar K133225 - Similar
Maximum Output Voltage (+/- 10%)	60 V @ 500 Ω 96 V @ 2 k Ω 114.2 V @ 10 k Ω	60 V @ 500 Ω 165 V @ 2 k Ω 165 V @ 10 k Ω	36 V @ 500 Ω	K083140 - Similar K133225 - Different
Maximum Output Current (+/- 10%)	120 mA @ 500 Ω 48 mA @ 2 k Ω 11 mA @ 10 k Ω	120 mA @ 500 Ω 82 mA @ 2 k Ω 16 mA @ 10 k Ω	72 mA @ 500 Ω	K083140 - Similar K133225 - Different
Positive Pulse Width	75 μ s or 175 μ s	200 μ s to 400 μ s	100 μ s to 500 μ s	K083140 - Different K133225 - Different Difference does not adversely affect the safety and effectiveness of the device
Negative Pulse Width	75 μ s or 175 μ s	200 μ s to 400 μ s	100 μ s to 500 μ s	K083140 - Different K133225 - Different Difference does not adversely affect the safety and effectiveness of the device
Frequency	85Hz or 100Hz	1 to 120Hz	5 to 120Hz	K083140 - Similar K133225 - Different
For interferential modes only:	No	No	No	K083140 - Similar K133225 - Similar
-Beat Frequency	No	No	No	K083140 -

				Similar K133225 - Similar
For multiphasic waveforms only:	Yes	Yes	Yes	K083140 - Similar K133225 - Similar
- Symmetrical phases?	Yes	Yes	Yes	K083140 - Similar K133225 - Similar
- Phase Duration (include units) (state range, if applicable) (both phases, if asymmetrical)	75µs Symmetrical or 175µSec Symmetrical	70-200µs Symmetrical	100 - 500µs Symmetrical	K083140 - Similar K133225 - Different
Net Charge [µC/pulse]	0.5 [µC] @ 500Ω Excitation pulse fully compensated	0 [µC] @ 500Ω Excitation pulse fully compensated	0 [µC] @ 500Ω Excitation pulse fully compensated	K083140 - Similar K133225 - Similar
Maximum Phase Charge [µC]	21µC @ 500 Ω	48µC @ 500Ω	36µC @ 500 Ω	K083140 - Similar K133225 - Similar
Maximum Current (RMS) Density [mA/cm²]	2.4mA/cm ² @500Ω	1.49mA/cm ² @ 500Ω	0.85mA/cm ² @ 500Ω	K083140 - Different K133225 - Different See section 1.10 Differences below for justification of differences
Maximum Power Density [mW/cm²]	5.04mW/cm ² @ 500 Ω	27.6mW/cm ² @500Ω	6.3mW/cm ² @ 500 Ω	K083140 - Similar K133225 - Similar
Burst Mode (i.e., pulse trains) a. Pulses per burst b. Bursts per second	N/A	Unknown	Unknown	K083140 - Unknown K133225 - Unknown

c. Burst duration (seconds) d. Duty Cycle [Line (b) x Line (c)]				
ON Time (seconds)	1 or 4, 1 second intervals (software adjustable modes)	Not Adjustable	0.1 to 60 seconds, 0.1 second intervals (adjustable within range)	K083140 - Different K133225 - Different Difference does not adversely affect the safety and effectiveness of the device
OFF Time (seconds)	1 or 4, 1 second intervals (software adjustable modes)	Not Adjustable	0.1 to 60 seconds, 0.1 second intervals (adjustable within range)	K083140 - Different K133225 - Different Difference does not adversely affect the safety and effectiveness of the device
Treatment Training Time	1-40 minutes, 1-minute intervals (software adjustable within range)	Not Adjustable	1-30 minutes, 1-minute intervals (adjustable within range)	K083140 - Different K133225 - Similar Difference does not adversely affect the safety and effectiveness of the device
Ramp up time per impulse	0 seconds or 300ms (software adjustable within range)	Not Adjustable	0 ms - 5000 ms, 100ms intervals (adjustable within range)	K083140 - Different K133225 - Different Difference

				does not adversely affect the safety and effectiveness of the device
--	--	--	--	--

Table 11: Output Specification Comparisons

1.10 Differences

Fundamentally, when we compare the core design, output specifications, operating behavior, intended use and general modes of operation we can conclude there is little difference between the Compex Sport Plus, EF-1280, and the Katalyst Mark 1 Muscle Stimulator.

The following table outlines the essential differences with justifications for how the differences do not adversely impact the safety and effectiveness of the subject device.

Difference	Additional Details	Justification for how the difference does not adversely impact safety and effectiveness
Katalyst Mark 1's use of wireless communication between the controller device and the impulse generator.	In order to better utilize modern interfaces alongside with strong and robust platforms. Katalyst utilizes a Bluetooth Smart connection between its interface controller and the impulse pack.	System validation testing scenarios covering mitigation of wireless risks in accordance with FDA's "<i>Radio Frequency Wireless Technology in Medical Devices - Guidance for Industry and Food and Drug Administration Staff</i>" were added to our full system testing protocol to ensure safe and effective use In addition, this is similar to the PowerDot Pd-01 Muscle Stimulator (with PowerDot Mobile Application) (K150078)

Using a tablet / touch interface to control channel impulse intensities.	The touch application is used to control the impulse intensities wirelessly via Bluetooth.	Impulse Pack firmware and Katalyst Application software documentation, consistent with a Moderate level of concern, is provided with this 510(k). System Validation Testing proves that all software requirement specifications were met and all software hazards were mitigated to Accepted risk level In addition, this is similar to the PowerDot Pd-01 Muscle Stimulator (with PowerDot Mobile Application) (K150078)
The Katalyst Mark 1 has a current density (RMS) greater than 2mA/cm²	As outlined in IEC 60601-2-10 additional warnings and guidance are included in the Operator's Guide to properly advise operators on proper procedures to increase to intensities.	The Katalyst Mark 1 has been tested and conformed to the IEC 60601-2-10:2012 "Medical electrical equipment - Part 2-10: Particular requirements for the basic safety and essential performance of nerve and muscle stimulators." Edition 2.0 standard. Furthermore, the subject device has complied with the recommendations set forth in FDA's "Guidance Document for Powered Muscle Stimulator 510(k)s"

		issued on June 9, 1999 In addition, Compex Sport (K011880) has a current density of 3.84mA/cm² which includes similar warnings and guidance.
--	--	---

1.11 Non-Clinical Tests

The following non-clinical testing was provided in this 510(k):

Biocompatibility Testing

Skin contacting electrodes pads have been tested to ISO 10993-10:2009 and ISO 10993-5:2009 standard under GLP.

Software Verification and Validation

Impulse Pack firmware and Katalyst Application software documentation, consistent with a Moderate level of concern, is provided with this 510(k).

System Validation Testing proves that all software requirement specifications were met and all software hazards were mitigated to Accepted risk level.

Electrical Safety and Electromagnetic Compatibility Testing

The Katalyst Suit and the Impulse Pack have been designed to comply and tested for compatibility with the applicable clauses of the following FDA-recognized standards:

- IEC/EN 60601-1:2005 “Medical Electrical Equipment - Part 1: General Requirements for Safety” Edition 3.1
- IEC 60601-2-10:2012 “Medical electrical equipment - Part 2-10: Particular requirements for the basic safety and essential performance of nerve and muscle stimulators.” Edition 2.0
- IEC/EN 60601-1-2:2014 “Medical electrical equipment - Part 1-2: General requirements for basic safety and essential performance - Collateral standard: Electromagnetic Disturbances. Edition 4.0

Battery Testing

Lithium-Polymer battery, used in Impulse Pack was tested by battery manufacturer for compliance with IEC 62133 Edition 2.0.

Engineering Bench Testing

In addition to the full system validation testing, the 510(k) also included testing in accordance with the recommendations of FDA's "*Guidance Document for Powered Muscle Stimulator 510(k)s*" issued on June 9, 1999. Oscilloscope tracings were obtained of the device output waveforms under maximum supported voltage and pulse widths under loads of 500 Ω , 2 k Ω and 10 k Ω . Also, several system validation testing scenarios covering mitigation of wireless risks in accordance with FDA's "*Radio Frequency Wireless Technology in Medical Devices - Guidance for Industry and Food and Drug Administration Staff*" were added to our full system testing protocol.

1.12 Conclusion

Test results demonstrate the Katalyst Mark 1 Muscle Stimulation System is substantially equivalent to the predicate devices.