



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center – WO66-G609
Silver Spring, MD 20993-0002

DRTECH Corporation
% Choul-Woo Shin
Vice President
Suite No. 2, 3 Floor, 29, Dunchon-daero541 beon-gil, Jungwon-gu
Seongnam-si, Gyeonggi-Do 13230
REPUBLIC OF KOREA

May 15, 2017

Re: K171137
Trade/Device Name: EVS 2430W, EVS 2430GW
Regulation Number: 21 CFR 892.1680
Regulation Name: Stationary x-ray system
Regulatory Class: II
Product Code: MQB
Dated: April 14, 2017
Received: April 17, 2017

Dear Choul-Woo Shin:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638 2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, “Misbranding by reference to premarket notification” (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH’s Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

A handwritten signature in blue ink that reads "Michael D. O'Hara". The signature is written over a large, semi-transparent watermark of the FDA logo.

For

Robert Ochs, Ph.D.
Director
Division of Radiological Health
Office of In Vitro Diagnostics
and Radiological Health
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K171137

Device Name

EVS 2430W, EVS 2430GW

Indications for Use (Describe)

The EVS 2430W and EVS 2430GW Digital X-ray detector is indicated for digital imaging solution designed for providing general radiographic diagnosis of human anatomy. This device is intended to replace film or screen based radiographic systems in all general purpose diagnostic procedures. This device is not intended for mammography applications.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary

[As required by 21 CFR 807.92]

This 510(k) summary of safety and effectiveness information is prepared in accordance with 21 CFR 807.92

1. Date Prepared [21 CFR 807.92(a) (1)]

04/14/2017

2. Submitter's Information [21 CFR 807.92(a) (1)]

- Name of Sponsor: DRTECH Corporation
- Address: Suit No. 2, 3 Floor, 29, Dunchon-daero541 beon-gil, Jungwon-gu, Seongnam-si, Gyeonggi-do, 13230, Republic of Korea
- Contact Name: Choul-Woo Shin
- Telephone No.: + 82-31-779-7783
- Fax No.: + 82-31-779-7790
- Email Address : ymkim@drtech.co.kr
- Registration Number: 3005172103
- Name of Manufacturer: Same as Sponsor

3. Trade Name, Common Name, Classification [21 CFR 807.92(a) (2)]

- Trade Name: EVS 2430W, EVS 2430GW
- Common Name: Digital Flat Panel X-ray Detector
- Classification Name: Stationary X-ray System
- Classification Panel: Radiology
- Classification Regulation: 21 CFR 892.1680
- Product Code: MQB
- Device Class: II

4. Identification of Predicate Device(s) [21 CFR 807.92(a) (3)]

- 510(k) Number: K162552
- Applicant: DRTECH Corporation
- Trade Name: EVS 3643, EVS 3643G
- Classification Name: Stationary X-ray System
- Classification Panel: Radiology
- Classification Regulation: 21 CFR 892.1680
- Product Code: MQB
- Device Class: II

5. Description of the Modified Device [21 CFR 807.92(a) (4)]

<Modification>

- Addition of EVS 2430(G)W: The differences between the subject device and the predicate device are the detector size, pixel pitch and performance (Resolution, DQE) of their detectors. These EVS 2430W and EVS 2430GW differ in terms of the layers of their scintillators as shown in the following table:

Model Name	Scintillator layer
EVS 2430W	CsI (Cesium Iodide)
EVS 2430GW	GoS (Gadolinium Oxysulfide)

The EVS 2430(G)W is a wired/wireless flat-panel type digital X-ray detector that captures projection radiographic images in digital format within seconds, eliminating the need for an entire x-ray film or an image plate as an image capture medium. EVS 2430(G)W differs from traditional X-ray systems in that, instead of exposing a film and chemically processing it to create a hard copy image, a device called a Detector is used to capture the image in electronic form.

6. Intended Use [21 CFR 807.92(a)(5)]

The intended use has not changed as a result of the modification and is as follows:

The EVS 2430W and EVS 2430GW Digital X-ray detector is indicated for digital imaging solution designed for providing general radiographic diagnosis of human anatomy. This device is intended to replace film or screen based radiographic systems in all general purpose diagnostic procedures. This device is not intended for mammography applications.

7. Technological Characteristics [21 CFR 807.92(a)(6)]

The EVS 2430(G)W Detector is an indirect conversion device in the form of a square plate in which converts the incoming X-rays into visible light. This visible light is then collected by an optical sensor, which generates an electric charges representation of the spatial distribution of the incoming X-ray quanta.

The charges are converted to a modulated electrical signal through thin film transistors. The amplified signal is converted to a voltage signal and is then converted from an analog to digital signal which can be transmitted to a viewed image print out, transmitted to remote viewing or stored as an electronic data file for later viewing.

Comparisons with the predicate, devices show the technological characteristics of the EVS 2430(G)W to be same to the predicate devices. The EVS 2430(G)W is functionally identical to the predicate devices.

The subject device integrates with X-ray system through three modes such as Auto Trigger mode (AT), Sync. Trigger mode and USB SW mode. The Auto Trigger Mode (AT) is available for acquiring images without any connection to X-ray generator. A generator interface cable is not required. The Sync. Trigger mode is the most common and recommended exposure mode at B2B scopes. User can achieve the high quality images with Sync. Tigger Mode. The USB SW Mode is the most common and recommended exposure mode at a retrofit scope. User can achieve high quality images with USB SW Mode.

8. Hardware and Software Requirements

- X-ray System Requirement

Contents		Requirements	
Generator	Power frequency	30kHz ~ 240kHz	
	KV	40kVp ~ 150kVp	
	mA Range	10mA ~ 1000mA	
	Exposure Time	0.001~10sec	
	mAs Range	0.1~1000mAs	
	Accuracy	± 5%	
Bucky	Operating Type	moving	Stepping Motor
			Spring
	Static(Fixed)		CAM Motor type
	Trey size (mm)	267.5 mm x 327.5 mm x 15.6 mm or higher	
Grid	Ratio	5:1, 6:1, 8:1, 10:1, 12:1, 15:1	
	Line	85 ~ 215 Line	
	SID	100 ~ 180 cm	

* The EVS 2430(G)W detector is not compatible with trays that are less than 267.5 mm x 327.5 mm x 15.6 mm.

- Software Requirement

The EVS 2430(G)W detector is compatible with Econsole1 (K152172).

9. Substantial Equivalence [21 CFR 807.92(b)]

Parameter	Subject Device	Predicate Device
510(K) Number	Not Known	K162552
Manufacturer	DRTECH Corporation	DRTECH Corporation
Model Name	EVS 2430G, EVS2430GW	EVS 3643, EVS 3643G
Classification Name	Stationary X-ray System	
Classification Panel	Radiology	
Classification Regulation	21 CFR 892.1680	
Product Code	MQB	
Device Class	Class II	
Intended Use	The EVS 2430W, EVS 2430GW Digital X-ray detector is indicated for digital imaging solution designed for providing general radiographic diagnosis of human anatomy. This device is intended to replace film or screen based radiographic systems in all general purpose diagnostic procedures. This device is not intended for mammography applications	The EVS 3643, EVS 3643G Digital X-ray detector is indicated for digital imaging solution designed for providing general radiographic diagnosis of human anatomy. This device is intended to replace film or screen based radiographic systems in all general purpose diagnostic procedures. This device is not intended for mammography applications

Design	Panel Shape	Rectangular Panel	Rectangular Panel
	Detector Size	240 x 300 mm	360 x 430 mm
	Pixel Pitch	76 μ m	140 μ m
Materials Scintillator		TFT –amorphous Silicon	TFT –amorphous Silicon
		EVS 2430W: CsI	EVS 3643: CsI
		EVS 2430GW: GoS	EVS 3643G: GoS
Performance	DQE	EVS 2430W: 45% at 1.0 lp/mm 30% at 3.0 lp/mm	EVS 3643: 28% at 1.0 lp/mm 13% at 3.0 lp/mm
		EVS 2430GW: 25% at 1.0 lp/mm 5% at 3.0 lp/mm	EVS 3643G: 22% at 1.0 lp/mm 5% at 3.0 lp/mm
	MTF	EVS 2430W: 35% at 2.0 lp/mm	EVS 3643: 35% at 2.0 lp/mm
		EVS 2430GW: 35% at 2.0 lp/mm	EVS 3643G: 30% at 2.0 lp/mm
	Resolution	3072 x 3840	2560 x 3072
	Wireless Charging	available	available
Anatomical Sites		General Radiography	General Radiography
Power Supply		100~240V~, 50/60 Hz	100~240V~, 50/60 Hz
Communication Method		Wire	Wire
		Wireless <ul style="list-style-type: none"> • IEEE 802.11a//g/n (2.4 GHz / 5 GHz) • Security: WEP/WPA/WPA2 	Wireless <ul style="list-style-type: none"> • IEEE 802.11a//g/n (2.4 GHz / 5 GHz) • Security: WEP/WPA/WPA2

When compared to the predicate device (K162552), the EVS 2430(G)W presented in this submission has the same:

- Intended Use
- Technological characteristics
- Operating principle
- Materials Scintillator
- Design features
- Performance (MTF)

A few differences are as follows

- Image Size
- Performance (Resolution, DQE)

There are no significant differences between the EVS 2430(G)W and the predicate devices that would adversely affect the use of the product. It is substantially equivalent to these devices in design, function, materials, operational principles and intended use. The difference in the sizes of the detectors only result in the difference in the area that the X-ray images can be taken. Therefore, the difference in the sizes of the detectors does not affect the safety and effectiveness. Also, even though the resolution and DQE of the predicate device and the subject device are different, the results of the clinical image evaluation proved the clinical effectiveness of the subject device.

10. Summary of Non-Clinical Data [21 CFR 807.92(b)(1)]

The non-clinical performance testing constrains that the main physical values for comparison of X-ray devices like DQE and MTF are basically equal or higher the predicate device as following table:

Parameter	Modified Device	Predicate Device (K162552)	Remark
DQE	EVS 2430W: 45% at 1.0 lp/mm 30% at 3.0 lp/mm	EVS 3643: 28% at 1.0 lp/mm 13% at 3.0 lp/mm	Different
	EVS 2430GW: 25% at 1.0 lp/mm 5% at 3.0 lp/mm	EVS 3643G: 22% at 1.0 lp/mm 5% at 3.0 lp/mm	Similar
MTF	EVS 2430W: 35% at 2.0 lp/mm	EVS 3643: 35% at 2.0 lp/mm	Similar
	EVS 2430GW: 35% at 2.0 lp/mm	EVS 3643G: 30% at 2.0 lp/mm	

The EVS 2430(G)W complies with the following international and FDA-recognized consensus standards:

- AAMI ANSI ES60601-1: Medical Electrical Equipment -- Part 1: General Requirements for Basic Safety And Essential Performance (IEC 60601- 1:2005, Mod)
- IEC 60601-1-2: Medical Electrical Equipment - Part 1-2: General Requirements For Basic Safety And Essential Performance – Collateral Standard: Electromagnetic Compatibility - Requirements And Tests (Edition 3)
- ISO 14971: Medical Devices - Application of Risk Management to Medical Devices. (General I (QS/RM))
- IEC 62220-1: Medical electrical equipment - Characteristics of digital X-ray imaging devices - Part 1: Determination of the detective quantum efficiency
- NEMA PS 3.1 - 3.20: Digital Imaging and Communications in Medicine (DICOM) Set

The EVS 2430(G)W software complies with the following FDA Guidance documents:

- “Content of Premarket Submissions for Management of Cybersecurity in Medical Devices” and
- “Guidance for the Content of Premarket Submissions for Software Contained in Medical Devices”

11. Summary of Clinical Data [21 CFR 807.92(b)(2)]

A single-blinded concurrence study according to CDRH's Guidance for the Submission of 510(k)'s for Solid State X-ray Imaging Devices was conducted, and the study confirmed that the x-ray detectors EVS 2430(G)W provide images of equivalent diagnostic capability to the predicate devices (K162552) and its results demonstrate substantial equivalence.

Clinical images were provided; these images were not necessary to establish substantial equivalence based on the modifications to the device (note X-ray detector technology that is identical to the predicate device) but they provide further evidence in addition to the laboratory performance data to show that the subject device works as intended.

12. Conclusion [21 CFR 807.92(b)(3)]

The EVS 2430(G)W is substantially equivalent to the currently marketed and predicate device (EVS 3643(G)(K162552) in terms of design features, fundamental scientific technology, indications for use, and safety and effectiveness.

Additionally, Substantial equivalence was demonstrated through the non-clinical performance, which complied with the requirements specified in the international and FDA-recognized consensus standards, IEC60601-1, IEC 60601-1-2, IEC62220-1 and IEC 62133 and the clinical test, which complied with the requirements specified in the CDRH's Guidance for the Submission of 510(k)'s for Solid State X-ray Imaging Devices.

The results of these tests demonstrate that EVS 2430(G)W meets the acceptance criteria and is adequate for this intended use. The comparison of technological characteristics, non-clinical performance data, safety testing, and clinical image concurrence data demonstrate that the device is as safe and effectiveness.