



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

May 19, 2017

Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Center - WO66-G609  
Silver Spring, MD 20993-0002

Stryker Craniomaxillofacial  
Jonathan Schell  
Staff Regulatory Affairs Specialist  
750 Trade Centre Way, Suite 200  
Portage, Michigan 49002

Re: K171152

Trade/Device Name: Stryker Universal Neuro III System AXS Screw  
Regulation Number: 21 CFR 882.5360  
Regulation Name: Cranioplasty Plate Fastener  
Regulatory Class: Class II  
Product Code: HBW  
Dated: April 19, 2017  
Received: April 20, 2017

Dear Mr. Schell:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in

the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely,

**Michael J. Hoffmann -S**

for Carlos L. Peña, PhD, MS  
Director  
Division of Neurological  
and Physical Medicine Devices  
Office of Device Evaluation  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)

K171152

Device Name

Stryker Universal Neuro III System AXS Screw

Indications for Use (Describe)

The Stryker Universal Neuro III System is intended for reconstruction, stabilization and/or rigid fixation of non load-bearing bony areas subsequent to craniotomy, craniectomy and cranial fractures in adults and adolescents (age 12 and higher).

Type of Use (Select one or both, as applicable)

☒ Prescription Use (Part 21 CFR 801 Subpart D)

☐ Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

## Section 5. 510(k) Summary

This section provides a summary of 510(k) information in accordance with the requirements of 21 CFR 807.92.

### I. SUBMITTER

510(k) Owner: Stryker Leibinger GmbH & Co. KG  
Boetzinger Strasse 41  
D-79111 Freiburg, Germany

Submitter/ Contact Person: Jonathan Schell  
Staff Regulatory Affairs Specialist  
Stryker Craniomaxillofacial  
750 Trade Centre Way  
Portage, MI 49002  
Phone: 269-389-5596  
Fax: 877-648-7114

Date prepared: May 12, 2017

### II. DEVICE

Trade Name: Stryker Universal Neuro III System AXS screw

Common or Usual name: Cranioplasty plate fastener

Classification name: Cranioplasty Plate Fastener 21 CFR §882.5360

Regulatory Class: Class II

Product Code: HBW

### III. PREDICATE DEVICE

Predicate: Stryker Universal Neuro III System AXS Screw – K151387

Reference Devices: K120352 Stryker QuikFlap Sterile Procedure Pack, and K131775 Universal Neuro 3 System

#### IV. DEVICE DESCRIPTION

The Stryker Universal Neuro III System (UN III) consists of an assortment of different bone plates and screws. The predicate device, the UN III AXS screw, is a part of the overall UN III System, and was cleared in K151387. Here, this special 510(k) is submitted to show the ability to market the UN III AXS screw sterile in two different sterile packaging configurations which are:

1. UN III AXS Screws packed in sterile procedure packs, and
2. UN III AXS Screws packed in a sterile screw cartridge.

The specific devices included are:

| Description                             | Packaging Configuration Type       |
|-----------------------------------------|------------------------------------|
| 2-Hole Plate Set, SD Screw              | UN III AXS Sterile Procedure Pack  |
| 2-Hole Plate Set, ST Screw              | UN III AXS Sterile Procedure Pack  |
| 2-Hole Plate Set, Low Profile with Tab  | UN III AXS Sterile Procedure Pack  |
| 2-Hole Plate / Burr Hole Cover 14mm Set | UN III AXS Sterile Procedure Pack  |
| 2-Hole Plate / Burr Hole Cover 20mm Set | UN III AXS Sterile Procedure Pack  |
| Burr Hole Cover 14mm Set                | UN III AXS Sterile Procedure Pack  |
| UN III AXS Screw, ST, 1.5x4mm           | UN III AXS Sterile Screw Cartridge |
| UN III AXS Screw, ST, 1.5x5mm           | UN III AXS Sterile Screw Cartridge |
| UN III AXS Screw, ST, 1.5x6mm           | UN III AXS Sterile Screw Cartridge |
| UN III AXS Screw, SD, 1.5x3mm           | UN III AXS Sterile Screw Cartridge |
| UN III AXS Screw, SD, 1.5x4mm           | UN III AXS Sterile Screw Cartridge |
| UN III AXS Screw, SD, 1.5x5mm           | UN III AXS Sterile Screw Cartridge |
| UN III AXS Screw, Emergency, 1.7x4mm    | UN III AXS Sterile Screw Cartridge |

#### V. INDICATIONS FOR USE

The Stryker Universal Neuro III System is intended for reconstruction, stabilization and/or rigid fixation of non load-bearing bony areas subsequent to craniotomy, craniectomy and cranial fractures in adults and adolescents (age 12 and higher).

Offering the AXS screw sterile does not alter the Indications for Use statement for the proposed device. The Indications for Use is identical to the predicate device from the Stryker Universal Neuro III System (K151387).

## VI. COMPARISON OF TECHNOLOGICAL CHARACTERISTICS WITH THE PREDICATE DEVICE

The subject device is compared to its predicate device for substantial equivalence of technological characteristics based on the following criteria:

- A. Principle of Operation
- B. Technological Characteristics

### A. Principle of Operation

The basic operational principle of the UN III AXS screw is unchanged from the predicate and remains the same.

### B. Technological Characteristics

The technological characteristics remain the same as the predicate and are unchanged.

## VII. PERFORMANCE DATA

In Table 5.1 below, the relevant topics, as given in "Submission and Review of Sterility Information in Premarket Notification (510(k)) Submissions for Devices Labeled as Sterile Guidance for Industry and Food and Drug Administration Staff (2016)," are referenced.

**Table 5-1: Reference of testing**

|                                                       |                                                                                                |
|-------------------------------------------------------|------------------------------------------------------------------------------------------------|
| Category A Sterilization Methods – Established Method | Gamma Radiation                                                                                |
| Sterilization Chamber:                                | Terminally sterilized in final packaging.                                                      |
| SAL                                                   | SAL $1 \times 10^{-6}$ for all products                                                        |
| Pyrogenicity                                          | Acceptance Criteria met                                                                        |
| Packaging Description                                 | Sterile Barrier Systems:<br>Procedure Packs: Flexible bag (sealed)<br>Screws: Blister (sealed) |

### Process Validations

Overall the subject device is unchanged and has the same implant material and manufacturing process, and the same duration or location of contact to the patient as the predicate device. The only difference between subject device and predicate device lies in the sterile packaging process required with the two sterile barrier systems referenced above.

To evaluate the influence of the changes compared to the predicate device, and based on the Risk Analysis performed on the modification of providing the UN III AXS screw sterile, the following validations were performed and referenced in support of substantial equivalence:

- Process Cleaning Validation (Bioburden level, Process residuals, Endotoxin Level)
- Sealing Validation of seals closing the sterile barrier systems
- Process Sterilization Validation Gamma Radiation with  $VD_{max}25$  Cycle

A Biocompatibility Assessment was performed to support substantial equivalence respecting the additional manufacturing process steps due to the sterile packaging compared to the predicate device. Substantial equivalence was proven.

### **Performance Bench Testing**

The following performance bench tests were completed and referenced.

- Transport safety of product and sterile barrier system including weight and compression test
- Shelf life and integrity of sterile barrier system
- Handling test of overall packaging functionality

The subject device met all pre-defined acceptance criteria. Overall, the results of the performance bench tests support the substantial equivalence of the subject device to the predicate device.

### **Animal Testing**

Animal testing was not required as a basis for substantial equivalence.

### **Clinical Testing**

Clinical testing was not required as a basis for substantial equivalence.

## **VIII. CONCLUSIONS**

The results of the non-clinical data demonstrate that the modified UN III AXS sterile screws perform as intended in the specified use conditions. According to the comparison based on the requirements of 21 CFR 807.87 and the information provided herein, it is concluded that the information included in this submission supports substantial equivalence.