



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center - WO66-G609
Silver Spring, MD 20993-0002

Altus Partners, LLC
Mark Melton
Senior Engineer
1340 Enterprise Drive
West Chester, Pennsylvania 19380

August 29, 2017

Re: K171329

Trade/Device Name: Altus Spine Titanium Interbody Fusion System
Regulation Number: 21 CFR 888.3080
Regulation Name: Intervertebral Body Fusion Device
Regulatory Class: Class II
Product Code: MAX
Dated: July 12, 2017
Received: August 4, 2017

Dear Mr. Melton:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-

related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education (DICE) at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education (DICE) at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely,

Mark N. Melkerson -S

Mark N. Melkerson
Director
Division of Orthopedic Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K171329

Device Name

Altus Spine Titanium Interbody Fusion System

Indications for Use (Describe)

The Altus Spine Titanium Interbody Fusion System is indicated for use with autogenous bone graft in skeletally mature patients with degenerative disc disease (“DDD”) at one or two contiguous spinal levels from L2-S1. DDD is defined as discogenic back pain with degeneration of the disc confirmed by history and radiographic studies. These patients should have had six months of non-operative treatment. These DDD patients may have had a previous non-fusion spinal surgery at the involved spinal level(s), and may have up to Grade 1 spondylolisthesis or retrolisthesis at the involved level(s).

The Altus Spine Titanium Interbody Fusion System is to be combined with cleared supplemental fixation systems, such as the Altus Spine Pedicle Screw System.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

“An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number.”

8. 510(k) Summary

SUBMITTER: Altus Partners, LLC
1340 Enterprise Drive
West Chester, PA 19380
Phone: 610-355-4156
Fax: 610-300-3049

CONTACT PERSON: Mark Melton
mmelton@altus-spine.com

DATE PREPARED: August 28, 2017

COMMON NAME: Interbody Fusion Device

PROPRIETARY NAME: Altus Spine Titanium Interbody Fusion System

PRIMARY PREDICATE DEVICES: Altus Spine Titanium Interbody Fusion System (K170512)

ADDITIONAL PREDICATE DEVICES: Altus Spine Interbody Fusion System (K160976)

CLASSIFICATION NAME: 21 CFR §888.3080 Intervertebral Body Fusion Device

PRODUCT CODES: MAX

DEVICE CLASS: Class II

MATERIAL: Titanium Alloy that conforms to ASTM F136

DEVICE DESCRIPTION:

The Altus Spine Titanium Interbody Fusion System implants are available in a variety of different shapes and sizes to suit the individual pathology and anatomical conditions of the patient.

The Altus Spine Titanium Interbody Fusion System implants are made of titanium alloy (Ti-6Al-4V) that conforms to ASTM F136.

The Altus Spine Titanium Interbody Fusion System has a hollow chamber to permit packing with autogenous bone graft to facilitate fusion. The superior and inferior surfaces of the construct have a pattern of teeth to provide increased stability and to help prevent movement of the device.

INDICATIONS FOR USE:

The Altus Spine Titanium Interbody Fusion System is indicated for use with autogenous bone graft in skeletally mature patients with degenerative disc disease (“DDD”) at one or two contiguous spinal levels from L2-S1. DDD is defined as discogenic back pain with degeneration of the disc confirmed by history and radiographic studies. These patients should have had six months of non-operative treatment. These DDD patients may have had a previous non-fusion spinal surgery at the involved spinal level(s), and may have up to Grade 1 spondylolisthesis or retrolisthesis at the involved level(s).

The Altus Spine Titanium Interbody Fusion System is to be combined with cleared supplemental fixation systems, such as the Altus Spine Pedicle Screw System.

SUMMARY OF TECHNOLOGICAL CHARACTERISTICS:

The Altus Spine Titanium Interbody Fusion System is a modification to the predicate (K170512). The modification includes larger graft windows and an additional implant footprint. The Altus Spine Interbody Fusion System and the predicate (K073502) share the same indications for use and surgical technique. The design is essentially the same fundamental technology with minor dimensional changes.

SUMMARY OF NON-CLINAL TESTS SUBMITTED:

Engineering analysis was presented to demonstrate that the Altus Spine Titanium Interbody Fusion System does not present a new worst case in performance and is substantially equivalent to the predicates.

SUBSTANTIAL EQUIVALENCE CONCLUSION:

The Altus Spine Titanium Interbody Fusion System is the same as the predicate (K170512) in regards to indications for use and surgical technique.

Altus Spine has determined that the modification of the Altus Spine Titanium Interbody Fusion System do not alter the system function, strength and stability. Therefore, the Altus Spine Titanium Interbody Fusion System is substantially equivalent to the predicate devices, and raises no new questions of safety or effectiveness.