



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center - WO66-G609
Silver Spring, MD 20993-0002

June 1, 2017

Xenacor, Inc
Mr. Spencer Walker
Director, Regulatory Affairs
630 Komas Dr. Suite 200
Salt Lake City, Utah 84108

Re: K171344

Trade/Device Name: Xenacor Xenoscope Laparoscopic System
Regulation Number: 21 CFR 876.1500
Regulation Name: Endoscope and Accessories
Regulatory Class: Class II
Product Code: GCJ, HET
Dated: May 5, 2017
Received: May 8, 2017

Dear Mr. Walker:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-

related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely,

**Jennifer R.
Stevenson -S**

For Binita S. Ashar, M.D., M.B.A., F.A.C.S.
Director
Division of Surgical Devices
Office of Device Evaluation
Center for Devices and Radiological Health

Enclosure

Xenacor®, Inc.
Special 510(k) Premarket Notification
Xenoscope Laparoscopic System

Statement of Indications for Use

Indications for Use

510(k) Number (if known): K171344

Device Name: Xenacor® Xenoscope™ Laparoscopic System

The Xenoscope™ is intended to be used in diagnostic and therapeutic procedures for endoscopy and endoscopic surgery within the thoracic and peritoneal cavities including the female reproductive organs.

Prescription Use
(Per 21 CFR 801.109)

OR Over-The-Counter Use

(PLEASE DO NOT WRITE BELOW THIS LINE – CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office Of Device Evaluation (ODE)

Xenacor®, Inc.
Special 510(k) Premarket Notification
Xenoscope Laparoscopic System

510(k) Summary

Submitter: Xenacor®, Inc

Contact Person: Spencer Walker, MSC - Director Regulatory Affairs
630 Komas Dr. Suite 200
Salt Lake City, UT 84108
(801) 581-5080

Date Prepared: May 05, 2017

Trade Name: Xenacor® Xenoscope™ Laparoscopic System

Classification Name: Endoscope/ Ureteroscope and Accessories
21 CFR §876.1500, Product Code GCJ, FGB, Class II

Gynecologic Laparoscope and Accessories
21 CFR §884.1720, Product Code HET, Class II

Predicate Device(s):

- K161838 – Xenacor®, Xenoscope™ Laparoscopic System

Device Description:

The Xenoscope™ System contains two separate functioning components. First, the Xenoscope™ Laparoscope Device is a 0° or 30°, 30 Fr (10mm), 10 - 36 cm long, single-use, high-definition video image. For certain procedures the shorter 10 cm laparoscope is preferred. Likewise, for other procedures, the longer 36 cm laparoscope is preferred. Except for the length difference, the scientific principles, materials of construction and design are otherwise identical. Second, the Xenoscope™ Dongle (now termed as the Xenobox™), which converts the camera image onto the video display screen for the surgeon to use, is provided with the Xenoscope™. Together, the Xenoscope™ and the Xenobox™ comprise the Xenoscope™ Laparoscopic System and work synergistically together.

Indications For Use:

The Xenoscope™ is intended to be used in diagnostic and therapeutic procedures for endoscopy and endoscopic surgery within the thoracic and peritoneal cavities including the female reproductive organs.

Comparative Analysis:

It has been demonstrated that the Xenoscope™ Laparoscopic System is comparable to the predicate device in intended use, fundamental scientific technology, design, principles of operation and functional performance evaluations. The Xenoscope™

Xenacor®, Inc.
Special 510(k) Premarket Notification
Xenoscope Laparoscopic System

Laparoscopic System has been fully assessed within the Xenacor® Risk Management and Design Controls systems. All necessary verification steps met pre-determined acceptance criteria to confirm safety and efficacy.

Functional/Safety Testing:

The following functional tests were performed. All data met pre-determined acceptance criteria.

- **Biocompatibility** – Per ISO 10993-1 for External communicating device, direct tissue contact, duration \leq 24 hours.
- **Software Verification** – Verification of modifications for image rotation feature of the 30° scope. The Xenoscope™ Laparoscope software level of concern is not changed.
- **Shaft Bending** – Inspection of the laparoscope shaft to confirm functionality after bending the shaft.

Conclusion:

The Xenoscope™ Laparoscopic System is substantially equivalent to the cited predicate device. Additionally, the Xenoscope™ Laparoscopic System met all acceptance criteria to confirm safety and efficacy.