



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center - WO66-G609
Silver Spring, MD 20993-0002

RaySearch Laboratories AB (publ)
% Mr. David Hedfors
Quality and Regulatory Affairs Director
Sveavägen 44
Stockholm, 111 34 Sweden

June 27, 2017

Re: K171536

Trade/Device Name: RayStation 6.1
Regulation Number: 21 CFR 892.5050
Regulation Name: Medical charged-particle radiation therapy system
Regulatory Class: II
Product Code: MUJ
Dated: May 21, 2017
Received: May 25, 2017

Dear David Hedfors:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-

related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely,

 For

Robert Ochs, Ph.D.

Director

Division of Radiological Health

Office of In Vitro Diagnostics

and Radiological Health

Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

Device Name
RayStation 6.1

Indications for Use (Describe)

RayStation is a software system designed for treatment planning and analysis of radiation therapy. The treatment plans provide treatment unit set-up parameters and estimates of dose distributions expected during the proposed treatment, and may be used to administer treatments after review and approval by the intended user.

The system functionality can be configured based on user needs.

The intended users of RayStation shall be clinically qualified radiation therapy staff trained in using the system.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

Document ID and Title	Version:
RSL-D-RS-6.1 Traditional 510(k) Submission RayStation 6.1	1.0

5. 510(k) Summary

5.1 510(k) owner

RaySearch Laboratories AB (publ)
Sveavägen 44
111 34 Stockholm
Sweden

Tel: +46 8 510 530 00

5.2 Contact person

David Hedfors
Quality and Regulatory Affairs Director
RaySearch Laboratories AB (publ)
Email: david.hedfors@raysearchlabs.com
Tel: +46 722 366 110

5.3 Preparation date

May 18th, 2017

5.4 Trade name

The trade name is RayStation.

The trade name and version number are written together, i.e. “RayStation 6.1” to easily distinguish the submitted device from the predicate device RayStation 6.0.

5.5 Common name

Radiation therapy treatment planning system

5.6 Classification name

Medical charged-particle radiation therapy system (21 CFR 892.5050, Product Code MUJ)

5.7 Predicate device

RayStation 6.0 K170355

5.8 Device description

RayStation 6.1 is a radiation therapy treatment planning system, i.e. a software program for planning and analysis of radiation therapy treatment plans. Functionality includes fusion capabilities (CT, PET and MRI), contouring, collapsed cone convolution dose computation and 4D data compatibility, as well as unique features such as multi-criteria optimization, dose tracking, treatment adaptation and deformable registration, all available on one platform.

The main workflow, creating a treatment plan from imported patient image data, is described below:

Flow of Events

<i>User</i>	<i>System</i>
1. The user launches RayStation 6.1	
2. The user imports a patient and case with CT images through DICOM	
	3. The system imports the data and checks consistency of in-data
4. The user enters the Structure Definition module and creates ROIs using the contouring tools	
	5. The system adds the ROIs to the patient case
6. The user enters the Plan Setup module and creates a plan and a treatment setup with specified machine, treatment energy and	

Document ID and Title	Version:
RSL-D-RS-6.1 Traditional 510(k) Submission RayStation 6.1	1.0

<p>delivery type</p> <p>7. The user specifies beam configuration including isocenter, dose grid and fluence grid resolution</p> <p>8. The system adds the plan and treatment setup to the patient case</p> <p>9. The user enters the Plan Optimization module and optimizes the plan parameters</p> <p>10. The system generates a deliverable plan</p> <p>11. The system displays the plan as</p> <ul style="list-style-type: none"> - 2D and 3D dose and patient displays - DVH curves - Plan data (beams, segments etc.) <p>12. The user reviews the plan</p> <p>13. The user enters the Plan Evaluation module and evaluates the plan</p> <p>14. The user approves and exports the plan together with dose, structure sets and images</p> <p>15. The system exports the plan and patient data to a DICOM server</p>

5.9 Intended use

RayStation is a software system designed for treatment planning and analysis of radiation therapy. The treatment plans provide treatment unit set-up parameters and estimates of dose distributions expected during the proposed treatment, and may be used to administer treatments after review and approval by the intended user.

The system functionality can be configured based on user needs.

The intended users of RayStation shall be clinically qualified radiation therapy staff trained in using the system.

The intended use for RayStation 6.1 is the same as for the predicate device RayStation 6.0.

5.10 Technological characteristics summary

The technological characteristics are the same for RayStation 6.1 as for the predicate device RayStation 6.0.

Comparing RayStation 6.1 with RayStation 6.0, the newer version includes usability, computational speed and connectivity improvements. Both versions are built on the same software platform and share design to a high degree. Both versions have been developed under the same quality system meeting the same requirements for safety and effectiveness.

5.11 Assessment of non-clinical performance data

The test specification of RayStation 6.1 is a further developed version of the test specification of RayStation 6.0. This is supported by the requirements specification, for which the same is true. The successful verification and validation of RayStation 6.1 therefore support the substantial equivalence of the above RayStation versions.

5.12 Test conclusion

The summary of the performed non-clinical tests shows that RayStation 6.1 is as safe and effective, and performs as well as the predicate device.