



Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Center - WO66-G609  
Silver Spring, MD 20993-0002

ConMed Corporation  
Ms. Diana L. Nader-Martone  
Regulatory Affairs Specialist  
525 French Road  
Utica, New York 13502

August 29, 2017

Re: K171592

Trade/Device Name: CrossFT™ Knotless Biocomposite Suture Anchor with Disposable Driver

Regulation Number: 21 CFR 888.3030

Regulation Name: Single/multiple component metallic bone fixation appliances and accessories

Regulatory Class: Class II

Product Code: MAI

Dated: May 30, 2017

Received: May 31, 2017

Dear Ms. Nader-Martone:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the [Federal Register](#).

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply

with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education (DICE) at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education (DICE) at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely,

**Mark N. Melkerson -S**

Mark N. Melkerson  
Director  
Division of Orthopedic Devices  
Office of Device Evaluation  
Center for Devices and  
Radiological Health

Enclosure

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
Food and Drug Administration

Form Approved: OMB No. 0910-0120  
Expiration Date: January 31, 2017  
See PRA Statement below.

## Indications for Use

510(k) Number (if known)

K171592

Device Name

CrossFT™ Knotless Biocomposite Suture Anchor with Disposable Driver

Indications for Use (Describe)

The CrossFT™ Knotless Biocomposite Suture Anchor with Disposable Driver is intended to reattach soft tissue to bone in the following orthopedic surgical procedures:

Procedure	Joint
Rotator Cuff Repair	Shoulder
Achilles Repair	Ankle
Bicep Tenodesis	Shoulder
Gluteus Medius Repair	Hip
Medial Patellofemoral Ligament (MPFL)	Knee

The device may be used in either arthroscopic or open surgical procedures. After the suture is anchored to the bone, it may be used to reattach soft tissue, such as ligaments, tendons, or joint capsules to the bone. The suture anchor system thereby stabilizes the damaged soft tissue, in conjunction with appropriate postoperative immobilization, throughout the healing period.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
PRAStaff@fda.hhs.gov

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

---

510(k) SUMMARY

In accordance with the requirements of the Safe Medical Device Act of 1990 and 21 CFR 807.92, CONMED Corporation is hereby submitting the 510(k) Summary of Safety and Effectiveness for 510(k) Number K171592.

I. SUBMITTER

CONMED Corporation  
525 French Road  
Utica, New York 13502

Phone: 727-399-5425  
Fax: 727-399-5264

Contact Person: Diana L. Nader-Martone  
Date Prepared: May 30, 2017

II. DEVICE NAME

Device Name:	CrossFT™ Knotless Biocomposite Suture Anchor with Disposable Driver
Common Name:	Bioabsorbable Suture Anchor
Classification Name:	Fastener, fixation, biodegradable, soft tissue
Regulatory Class:	Class II, per 21 CFR Part 888. 3030
Product Codes:	MAI

III. PREDICATE DEVICE

Device Name:	CrossFT™ Knotless Biocomposite Suture Anchor with Disposable Driver
Company Name:	ConMed Corporation
510(k) #:	K170501

IV. DEVICE DESCRIPTION

The CrossFT™ Knotless Biocomposite Suture Anchors with Disposable Driver are sterile, single use devices. The CrossFT™ Knotless Biocomposite Suture Anchors with Disposable Driver are manufactured from polylactide copolymer (96L/4D PLA) and  $\beta$ -Tricalcium Phosphate ( $\beta$ -TCP). The anchors are provided sterile, single use and preloaded on a disposable driver. The anchors are available in three sizes and nine configurations.

V. INTENDED USE/ INDICATIONS FOR USE

The CrossFT™ Knotless Biocomposite Suture Anchor with Disposable Driver is intended to reattach soft tissue to bone in the following orthopedic surgical procedures:

<u>Procedure</u>	<u>Joint</u>
Rotator cuff repair	Shoulder
Achilles Repair	Ankle
Bicep Tenodesis	Shoulder
Gluteus Medius Repair	Hip
Medial Patellofemoral Ligament (MPFL)	Knee

The device may be used in either arthroscopic or open surgical procedures. After the suture is anchored to the bone, it may be used to reattach soft tissue, such as ligaments, tendons, or joint capsules to the bone. The suture anchor system thereby stabilizes the damaged soft tissue, in conjunction with appropriate postoperative immobilization, throughout the healing period.

VI. COMPARISON OF THE TECHNOLOGICAL CHARACTERISTICS WITH THE PREDICATE DEVICE

The following table represents a summary of the technological characteristics between the proposed and the predicate device.

	CrossFT™ Knotless Biocomposite Suture Anchor with Disposable Driver Proposed	CrossFT™ Knotless Biocomposite Suture Anchor with Disposable Driver Predicate												
Device Description	Same	The CrossFT™ Knotless Biocomposite Suture Anchor with Disposable Drivers are sterile, single use devices. The CrossFT™ Knotless Biocomposite Suture Anchor with Disposable Drivers are manufactured from polylactide copolymer (96L/4D PLA) and β-Tricalcium Phosphate (β-TCP). The anchors are provided sterile, single use and preloaded on a disposable driver. The anchors are available in three sizes and nine configurations.												
Intended Use	The CrossFT™ Knotless Biocomposite Suture Anchor with Disposable Driver is intended to reattach soft tissue to bone in the following orthopedic surgical procedures: <table border="1"> <thead> <tr> <th><u>Procedure</u></th> <th><u>Joint</u></th> </tr> </thead> <tbody> <tr> <td>Rotator Cuff Repair</td> <td>Shoulder</td> </tr> <tr> <td>Achilles Repair</td> <td>Ankle</td> </tr> <tr> <td>Bicep Tenodesis</td> <td>Shoulder</td> </tr> <tr> <td>Gluteus Medius Repair</td> <td>Hip</td> </tr> <tr> <td>Medial Patellofemoral Ligament (MPFL)</td> <td>Knee</td> </tr> </tbody> </table>	<u>Procedure</u>	<u>Joint</u>	Rotator Cuff Repair	Shoulder	Achilles Repair	Ankle	Bicep Tenodesis	Shoulder	Gluteus Medius Repair	Hip	Medial Patellofemoral Ligament (MPFL)	Knee	The CrossFT™ Knotless Biocomposite Suture Anchor with Disposable Driver is intended to reattach soft tissue to bone in orthopedic surgical procedures.
<u>Procedure</u>	<u>Joint</u>													
Rotator Cuff Repair	Shoulder													
Achilles Repair	Ankle													
Bicep Tenodesis	Shoulder													
Gluteus Medius Repair	Hip													
Medial Patellofemoral Ligament (MPFL)	Knee													
Indication for Use	Same	The device may be used in either arthroscopic or open surgical procedures. After the suture is anchored to the bone, it may be used to reattach soft tissue, such as ligaments, tendons, or joint capsules to the bone. The suture anchor system thereby stabilizes the damaged soft tissue, in conjunction with appropriate postoperative immobilization, throughout the healing period.												

	CrossFT™ Knotless Biocomposite Suture Anchor with Disposable Driver Proposed	CrossFT™ Knotless Biocomposite Suture Anchor with Disposable Driver Predicate
Contraindications	Same	<ul style="list-style-type: none"> <li>*Pathological conditions of bone which would adversely affect the CrossFT™ Knotless Biocomposite Suture Anchor with Disposable Driver with Disposable Driver.</li> <li>* Pathological conditions in the soft tissue to be repaired or reconstructed which would adversely affect suture fixation.</li> <li>* Physical conditions that would eliminate, or tend to eliminate, adequate implant support or retard healing.</li> <li>* Conditions which tend to limit the patient's ability or willingness to restrict activities or follow directions during the healing period.</li> <li>* Attachment of artificial ligaments or other implants.</li> <li>* Foreign body sensitivity, known or suspected allergies to implant and/ or instrument materials.</li> <li>* This device is not approved for screw attachment or fixation to the posterior elements (pedicles) of the cervical, thoracic or lumbar spine.</li> </ul>
Components	Same	<ul style="list-style-type: none"> <li>Bioabsorbable anchor</li> <li>Disposable Driver</li> <li>Suture and/or Suture Tape</li> <li>Threader</li> </ul>

## VII. PERFORMANCE DATA

Testing has been completed to demonstrate that the CrossFT™ Knotless Biocomposite Suture Anchor with Disposable Driver performs as intended and is substantially equivalent to the predicate device. The bacterial endotoxin testing was conducted and met the endotoxin limits. The test parameters utilized were supported by clinical data in the literature to support the specific indications and corresponding anatomic locations in this submission.

Completed testing includes the following:

### Verification Testing

- Reliability
- Ultimate Fixation Strength
- Cyclic
- Sterilization
- Pyrogenicity
- Biocompatibility
- Shelf-life

### Validation Testing

- User Validation
- Packaging
- Transportation

## VIII. CONCLUSION

The CrossFT™ Knotless Biocomposite Suture Anchor with Disposable Driver is identical in design, materials, principles of operation, and technical characteristics to the predicate, and substantially equivalent in intended use to the predicate. Based upon the findings of our performance testing, the difference regarding intended use presents no new issues of

---

safety and efficacy, and the proposed, CrossFT™ Knotless Biocomposite Suture Anchor with Disposable Driver is substantially equivalent to the CrossFT™ Knotless Biocomposite Suture Anchor with Disposable Driver (K170501).