



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center - WO66-G609
Silver Spring, MD 20993-0002

PRECIFIT MEDICAL LTD
% Mr. Kellen Hills
Quality and Regulatory Consultant
Orchid Design
4600 East Shelby Drive, Suite 1
Memphis, Tennessee 38118

July 25, 2017

Re: K171630
Trade/Device Name: LumFuse-TP
Regulation Number: 21 CFR 888.3080
Regulation Name: Intervertebral Body Fusion Device
Regulatory Class: Class II
Product Code: MAX
Dated: May 22, 2017
Received: June 2, 2017

Dear Mr. Hills:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR

Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely,

Mark N. Melkerson -S

Mark N. Melkerson
Director
Division of Orthopedic Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K171630

Device Name

Lumfuse-TP

Indications for Use (Describe)

The Lumfuse-TP cage is intended to be used in spinal fusion procedures for patients diagnosed with Degenerative Disc Disease (DDD) at one or two contiguous levels from L2 to S1. DDD patients may also have up to Grade 1 Spondylolisthesis or retrolisthesis at the involved levels. DDD is defined as discogenic back pain with degeneration of the disc confirmed by history and radiographic studies. When used for these indications, the Lumfuse-TP cage is intended for use with supplemental fixation systems cleared for use in the lumbar spine. These patients should be skeletally mature and have had six months of nonoperative treatment. The Lumfuse-TP cage is intended to be used with autograft and/or allogenic bone graft comprised of cancellous and/or corticocancellous bone graft when the subject device is used as an adjunct to fusion.

Type of Use (Select one or both, as applicable)

☒ Prescription Use (Part 21 CFR 801 Subpart D)

☐ Over-The-Counter Use (21 CFR 801 Subpart C)

PLEASE DO NOT WRITE BELOW THIS LINE – CONTINUE ON A SEPARATE PAGE IF NEEDED.

FOR FDA USE ONLY

Concurrence of Center for Devices and Radiological Health (CDRH) (Signature)

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

“An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number.”

510(k) Summary

[As Required by 21 CFR 807.92]

- (a)(1) Submitted By: PRECIFIT MEDICAL LTD
 Address: 951 Aviation Pkwy Ste 100,
 Morrisville, NORTH CAROLINA, 27560
 Phone: 901-433-1990
 Date: July 21, 2017
 Contact Persons
 Primary: Kellen Hills (Orchid Design Consulting)
 Secondary: Eric Wu (Precifit Medical Ltd)
- (a)(2) Proprietary Name: Lumfuse-TP
 Common Name: Lumbar interbody fusion device, interbody cage
 Classification Name and Reference: 21 CFR 888.3080: Intervertebral Fusion Device With Bone Graft, Lumbar
 Product Code: MAX
- (a)(3) Predicate Devices:
 Primary: Medtronic® CAPSTONE (K151128);
- (a)(4) Device Description:
 The Lumfuse-TP cage consists of PEEK cages of various lengths and heights, which can be inserted between two lumbar or lumbosacral vertebral bodies to give support and correction during lumbar interbody fusion surgeries. The hollow geometry of the implants allows them to be packed with autograft and/or allogenic bone graft comprised of cancellous and/or corticocancellous bone graft. The devices are made from PEEK radiolucent material (ASTM F2026) with embedded tantalum x-ray markers (ASTM F560 or ISO 13782).
 The purpose of this submission is to gain initial marketing authorization in the United States.
- (a)(5) Indications for Use:
 The Lumfuse-TP cage is intended to be used in spinal fusion procedures for patients diagnosed with Degenerative Disc Disease (DDD) at one or two contiguous levels from L2 to S1. DDD patients may also have up to Grade 1 Spondylolisthesis or retrolisthesis at the involved levels. DDD is defined as discogenic back pain with degeneration of the disc confirmed by history and radiographic studies. When used for these indications, the Lumfuse-TP cage is intended for use with supplemental fixation systems cleared for use in the lumbar spine. These patients should be skeletally mature and have had six months of nonoperative treatment. The Lumfuse-TP cage is intended to be used with autograft and/or allogenic bone graft comprised of cancellous and/or corticocancellous bone graft when the subject device is used as an adjunct to fusion.
- (a)(6) Comparison of Technological Characteristics:
 The Lumfuse-TP has the same intended use and is similar in basic shape, material and performance characteristics to the predicate device. The technological characteristics do not raise any new questions of safety and efficacy.
- (b)(1) Non-clinical testing:
 The worst case devices were evaluated for mechanical performance in static and dynamic axial compression and compression-shear per ASTM F2077-14 and subsidence per ASTM F2267-04 (2011). End-user cleaning and sterilization as well as implantable device biocompatibility endpoints were also evaluated.

(b)(2) Clinical testing:

Clinical testing was not required to demonstrate substantial equivalence in this premarket notification.

(b)(3) Conclusions:

Based on the information provided in this premarket notification, we believe that the subject Lumfuse-TP demonstrates substantial equivalence to the identified predicate device.