



Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Center – WO66-G609  
Silver Spring, MD 20993-0002

GE HUALUN MEDICAL SYSTEMS CO. LTD.  
% Ms. Lifeng Wang  
Regulatory Affairs Leader  
No1 Yong Chang North Road  
Beijing Economic Technological Development Zone  
Beijing 100176 Beijing  
CHINA

July 12, 2017

Re: K171800  
Trade/Device Name: OEC Elite™ MiniView™  
Regulation Number: 21 CFR 892.1650  
Regulation Name: Image-intensified fluoroscopic x-ray system  
Regulatory Class: II  
Product Code: OXO, JAA  
Dated: June 15, 2017  
Received: June 16, 2017

Dear Ms. Wang:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638 2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, “Misbranding by reference to premarket notification” (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH’s Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

 For

Robert Ochs, Ph.D.  
Director  
Division of Radiological Health  
Office of In Vitro Diagnostics  
and Radiological Health  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)

K171800

Device Name

OEC Elite™ MiniView™

Indications for Use (Describe)

The OEC Elite MiniView (mobile mini C-Arm) is designed to provide physicians with real time general fluoroscopic visualization of patients of all ages. It is intended to aid physicians and surgeons during diagnostic or therapeutic treatment/surgical procedures of the limbs/extremities and shoulders including, but not limited to, orthopedics and emergency medicine.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
PRASStaff@fda.hhs.gov

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*



**GE Healthcare**

510(k) Premarket Notification Submission- OEC Elite™ MiniView™

**510(k) SUMMARY OF SAFETY AND EFFECTIVENESS**

In accordance with the requirements of 21 CFR 807.92 the following summary of information is provided:

**Date:** June 16, 2017  
**Submitter:** GE HUALUN MEDICAL SYSTEMS CO. Ltd.  
No.1 Yong Chang North Road, Beijing Economic  
Technological Development Zone  
Beijing 100176,China

**Manufacturer/  
Manufacturing Location** GE HUALUN MEDICAL SYSTEMS CO. Ltd.  
No.1 Yong Chang North Road, Beijing Economic  
Technological Development Zone  
Beijing 100176,China

**Primary Contact Person:** Lifeng Wang  
Regulatory Affairs Manager  
GE Healthcare  
+86 10 57083145  
[Lifeng.Wang@ge.com](mailto:Lifeng.Wang@ge.com)

**Secondary Contact Person:** Helen Peng  
Regulatory Affairs Director  
GE Healthcare  
+1 262 424-8222  
[Hong.Peng@ge.com](mailto:Hong.Peng@ge.com)

**Device Trade Name:** OEC Elite™ MiniView™

**Common/Usual Name:** Fluoroscopic X-Ray System, Mobile  
Mini Mobile C-Arm, Mini C-Arm

**Classification Names:** Image-intensified fluoroscopic x-ray system

**Device Class** Class II

**Classification regulation:** 21CFR 892.1650

**Primary Product Code:** OXO

**Secondary Product Code:** JAA

**Predicate Device(s):** K160131, OEC Elite MiniView



## GE Healthcare

### 510(k) Premarket Notification Submission- OEC Elite™ MiniView™

#### **Device Description:**

The OEC Elite™ MiniView™ is a mobile fluoroscopic mini C-arm system that provides fluoroscopic images of patients of all ages during diagnostic, treatment, and surgical procedures of the shoulders, limbs, and extremities. The system consists of a C-arm attached to an image processing workstation. A CsI(Tl) -CMOS flat panel detector and the identical X-ray source monoblock are used for image acquisition.

The C-arm supports the high-voltage generator, X-ray tube, X-ray controls, collimator, and the FPD. The C-arm is capable of performing linear (vertical, horizontal, orbital) and rotational motions that allow the user to position the X-Ray imaging components at various angles and distances with respect to the patient extremity anatomy to be imaged. The C and support arm are mechanically balanced allowing for ease of movement and capable of being “locked” in place using an electronically controlled braking system. The workstation is a stable mobile platform that supports the C-arm, image display monitor(s), image processing equipment/software, recording devices, data input/output devices and power control systems.

The primary technology change for the subject device compared to the unmodified predicate OEC Elite MiniView, was to add an optional wireless footswitch which provides identical functionalities as the wired footswitch to control X-ray on and off.

#### **Intended Use:**

The OEC Elite™ MiniView™ (mobile mini C-arm) is intended to provide fluoroscopic images of the patient during diagnostic or therapeutic treatment/surgical procedures of the limbs/extremities and shoulders.

#### **Indications for Use**

The OEC Elite™ MiniView™ (mobile mini C-Arm) is designed to provide physicians with real time general fluoroscopic visualization of patients of all ages. It is intended to aid physicians and surgeons during diagnostic or therapeutic treatment/surgical procedures of the limbs/extremities and shoulders including, but not limited to, orthopedics and emergency medicine.

#### **Technology:**

The modified OEC Elite™ MiniView™ employs the same fundamental scientific technology as the unmodified predicate device. The proposed device will provide an optional off-the-shelf wireless foot switch uses 2.4GHz radio frequency technology with proprietary private protocols. The introduction of this wireless footswitch does not change any existing hardware or software of the predicated device. The new option is equivalent to the wired footswitch on the predicate in that the functionalities are identical. The wireless option provides placement flexibility and reduces cable clutter.



## GE Healthcare

510(k) Premarket Notification Submission- OEC Elite™ MiniView™

### **Determination of Substantial Equivalence:**

#### Summary of Non-Clinical Testing:

The modified OEC Elite™ MiniView™ system with the wireless footswitch was developed under the GE Healthcare's design controls processes and overall quality management system.

The system has been tested by an NRTL and certified compliant with the IEC 60601-1 Ed. 3 series, including IEC60601-2-54 and IEC 60601-2-43. All applicable 21CFR Subchapter J performance standards are met.

Risk management activities including using risk analysis to identify any potential issues incorporating the wireless footswitch were performed. These issues were reviewed and mitigated with inherent safe designs and labeling. The mitigations were verified and validated as a part of the design verification and validation testing that has been executed with acceptable results. This testing also includes the EMC and coexistence testing per the "Radio Frequency Wireless Technology in Medical Devices" Guidance for Industry and Food and Drug Administration Staff issued on August 14, 2013 (Wireless Guidance).

#### Clinical Testing:

The subject of this premarket submission, OEC Elite MiniView with the addition of the wireless footswitch, did not require clinical studies to support substantial equivalence.

Design verification and validation testing was performed to confirm that the safety and effectiveness of the devices has not been affected. The test plans and results have been executed with acceptable results.

### **Conclusion:**

The OEC Elite MiniView incorporates a wireless footswitch for the user to initiate and terminate x-ray exposures as an option. The addition of the optional wireless footswitch to the system does not raise new questions on safety and effectiveness. The modified subject device has the same technological characteristics and perform as well as the unmodified predicate device and other similar devices that incorporate a wireless footswitch currently on the market.

Based on the successful verification and validation testing, additional engineering bench testing per the FDA Wireless guidance, conformance to standards, and development under GE's Quality Management System, we believe that the modified OEC MiniView is of comparable type and substantially equivalent to the predicate device OEC Elite MiniView (K160131), and therefore is safe and effective for its intended use.