



October 5, 2017
Renovo, Inc.
% Mr. Robert Packard
President
Medical Device Academy, Inc.
345 Lincoln Hill Rd.
Shrewsbury, Vermont 05738

Re: K172093
Trade/Device Name: Reprocessed Covidien Trocar
Regulation Number: 21 CFR 876.1500
Regulation Name: Endoscope and Accessories
Regulatory Class: Class II
Product Code: NLM
Dated: October 3, 2017
Received: October 3, 2017

Dear Mr. Packard:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820);

and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education (DICE) at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education (DICE) at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely,

**Jennifer R.
Stevenson -S3**

For Binita S. Ashar, M.D., M.B.A., F.A.C.S.

Director

Division of Surgical Devices

Office of Device Evaluation

Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K172093

Device Name

Reprocessed Covidien Trocars

Indications for Use (Describe)

The Reprocessed Covidien Trocars are indicated for use in a variety of gynecologic, general, thoracic, and urologic endoscopic procedures to create and maintain a port of entry.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) SUMMARY

This summary of 510(k) safety and effectiveness information is submitted in accordance with the requirements of 21 CFR §807.92:

I. SUBMITTER

ReNovo, Inc.
Mark K. Wells
340 SW Columbia St.
Bend OR, 97702 USA
Tel: +1.541.422.8880
Fax: +1.541.422.8881

II. CONTACT PERSON

Robert Packard, President of Medical Device Academy
345 Lincoln Hill Rd.
Shrewsbury, VT 05738 USA
Tel: +1.802.281.4318
Email: rob@13485cert.com

Date Prepared: October 2, 2017

III. DEVICE

Name of Device:	Reprocessed Covidien Trocar
Classification Name:	Laparoscope, General & Plastic Surgery, Reprocessed
Regulation:	21 CFR 876.1500
Regulatory Class:	Class II
Product Classification Code:	NLM

IV. PREDICATE (OEM) DEVICE

Predicate Manufacturer:	Covidien (Original 510(k) was issued to United States Surgical, a division of Tyco Healthcare Group LP
Predicate Trade Name:	autosuture™ Modified VERSAPORT™ Trocar with Fixation Sleeve
Predicate 510(k):	K062326

No reference devices were used in this submission.

V. DEVICE DESCRIPTION

A trocar is a sterile device used in a variety of endoscopic procedures to establish and maintain a port of entry for surgical tools. Trocars consists of three primary components; the cannula which allows access into the abdominal cavity during the procedure; the seal, located at the top of the cannula allows instruments to pass through the cannula while preventing loss of pneumoperitoneum when instruments are inserted and withdrawn from the port; and the obturator, which is a mechanism that allows the cannula to penetrate the abdomen. There are also various configurations of diameter and length, but all models have the same basic design, materials, and technological characteristics.

VI. INDICATIONS FOR USE

Reprocessed Covidien Trocars are intended for use in a variety of gynecologic, general, thoracic and urologic endoscopic procedures to create and maintain a port of entry.

VII. COMPARISON OF TECHNOLOGICAL CHARACTERISTICS WITH THE OEM PREDICATE DEVICE

The fundamental technological characteristics of the subject device are identical to the predicate (OEM).

The following characteristics were compared between the subject device and the OEM device in order to demonstrate substantial equivalence:

- Indications for Use – The predicate and subject device have the same indications for use, both are indicated for “use in a variety of gynecologic, general, thoracic and urologic endoscopic procedures to create and maintain a port of entry.”
- Materials – The materials in the subject device are identical to the predicate with exception to the lubricant. The lubricants are not identical, but are equivalent in functional and biocompatible performance. .
- Design – The predicate and subject devices have identical designs.
- Energy Source – Neither the predicate nor the subject device require power or an energy source.
- Performance Testing – The subject devices are reprocessed versions of the predicate (OEM) devices. The OEM devices are performance-tested cleared medical devices; therefore the performance testing consisted of side-by-side testing of the reprocessed devices to an identical new OEM device. This testing included seal drag, seal leakage, insertion, fixation, and removal testing.
- Cleaning Validation – While it is very important to test the performance of the reprocessed device to ensure it performs as new; an important part of the risk management of a reprocessed device is the need to ensure cleanliness. A cleaning validation was performed to ensure that the reprocessing steps thoroughly cleaned and sanitized the devices so that they were as clean as a new OEM device.

VIII. PERFORMANCE DATA

The following performance data was provided in support of the substantial equivalence determination.

Performance Testing

- Functional
 - Seal Drag
 - Seal Leakage
 - Insertion, Fixation, and Removal Force Test
- Product Stability
 - Shelf Life Testing 1-year Accelerated Aging

Cleaning Validation

- Residual Protein and Hemoglobin
- Visual Inspection
- Cleaning Performance Qualification

Sterilization and Packaging

- EtO Sterilization Testing
- Simulated Shipment Testing

Biocompatibility Testing

- Cytotoxicity
- Irritation and Sensitization
- Acute Systemic
- EtO Residuals Testing
- Material Mediated Pyrogenicity

Electrical Safety and EMC Testing

Electrical safety and EMC testing were not applicable.

Animal Testing

Animal performance testing was not required to demonstrate safety and effectiveness.

Clinical Testing

Clinical testing was not required to demonstrate the safety and effectiveness of the Reprocessed Covidien Trocar. Substantial equivalency is based upon benchtop performance testing.

IX. **CONCLUSIONS**

Side by side testing ensured that the reprocessed trocars performed equivalently to the same device new from the OEM. Cleaning validations ensured that the devices are reprocessed and cleaned to the condition of a new OEM device. Therefore, based on a comparison of technological characteristics, indications for use, and performance comparison data, it can be concluded that the proposed Reprocessed Covidien Trocars are substantially equivalent to the OEM predicate device.