



December 1, 2017

Beckman Coulter, Inc
Shannon Popson
Staff Regulatory Affairs
1584 Enterprise Blvd
West Sacramento, CA 95691

Re: K172255

Trade/Device Name: MicroScan Dried Gram Negative MIC/Combo Panels with
Ceftolozane/tazobactam (0.25/4 – 16/4 µg/mL)

Regulation Number: 21 CFR 866.1640

Regulation Name: Antimicrobial susceptibility test powder

Regulatory Class: Class II

Product Code: LTT, LRG, JWY

Dated: July 25, 2017

Received: July 26, 2017

Dear Ms. Popson:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801 and Part 809); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/>) and CDRH Learn (<http://www.fda.gov/Training/CDRHLearn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<http://www.fda.gov/DICE>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Ribhi Shawar -S For

Uwe Scherf, M.Sc., Ph.D.

Director

Division of Microbiology Devices

Office of In Vitro Diagnostics

and Radiological Health

Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)
K172255

Device Name
MicroScan Dried Gram-Negative MIC/Combo Panels with Ceftolozane/Tazobactam (0.25/4 – 16/4 µg/mL)

Indications for Use (Describe)

The MicroScan Dried Gram-Negative MIC/Combo Panel is used to determine quantitative and/or qualitative antimicrobial agent susceptibility of colonies grown on solid media of rapidly growing aerobic and facultative anaerobic gram-negative bacilli. After inoculation, panels are incubated for 16 – 20 hours at 35°C +/- 1°C in a non-CO2 incubator, and read either visually or with MicroScan instrumentation, according to the Package Insert.

This particular submission is for the addition of the antimicrobial ceftolozane/tazobactam at concentrations of 0.25/4 to 16/4 µg/mL to the test panel.

Ceftolozane/tazobactam has been shown to be active both clinically and in vitro against the following organisms: *Enterobacter cloacae*, *Escherichia coli*, *Klebsiella oxytoca*, *Klebsiella pneumoniae*, *Proteus mirabilis*, and *Pseudomonas aeruginosa*

In vitro data is available for the following organisms, but their clinical significance is unknown: *Citrobacter koseri*, *Morganella morganii*, *Proteus vulgaris*, *Providencia rettgeri*, *Providencia stuartii*, *Serratia liquefaciens*, and *Serratia marcescens*

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

“An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number.”



510(k) Summary

510(k) Submission Information:

Device Manufacturer: Beckman Coulter
Contact name: Shannon Popson, Staff Regulatory Affairs
Phone: 916-374-3330
Fax: 916-374-2119
Date prepared: October 23, 2017
Product Name: Microdilution Minimum Inhibitory Concentration (MIC) Panels
Trade Name: MicroScan Dried Gram-Negative MIC/Combo Panels with Ceftolozane/Tazobactam (0.25/4 – 16/4 µg/mL)
Intended Use: To determine antimicrobial agent susceptibility
Classification: Class II
Product Code: LTT
510(k) Notification: New antimicrobial agent – Ceftolozane/Tazobactam
Predicate device: MicroScan Dried Gram-Negative MIC/Combo Panels – Imipenem (K162740)

510(k) Summary:

MicroScan Dried Gram-Negative MIC/Combo Panels are designed for use in determining quantitative and/or qualitative antimicrobial agent susceptibility of colonies grown on solid media of rapidly growing aerobic and facultative anaerobic gram-negative bacilli.

The antimicrobial susceptibility tests are miniaturizations of the broth dilution susceptibility test that have been diluted in broth and dehydrated. Various antimicrobial agents are diluted in broth to concentrations bridging the range of clinical interest. Panels are rehydrated with water after inoculation with a standardized suspension of the organism. After incubation in a non-CO₂ incubator for 16-20 hours, the minimum inhibitory concentration (MIC) for the test organism is read by determining the lowest antimicrobial concentration showing inhibition of growth.

The proposed MicroScan Dried Gram-Negative MIC/Combo Panel demonstrated substantially equivalent performance when compared with a CLSI frozen Reference Panel, as defined in the FDA document "Class II Special Controls Guidance Document: Antimicrobial Susceptibility Test (AST) Systems; Guidance for Industry and FDA", dated August 28, 2009. The Premarket Notification (510[k]) presents data in support of the MicroScan Dried Gram-Negative MIC/Combo Panel with ceftolozane/tazobactam.

The external evaluations were conducted with fresh, recent and stock Efficacy isolates and stock Challenge strains. The external evaluations were designed to confirm the acceptability of the proposed Dried Gram-Negative Panel by comparing its performance with a CLSI frozen Reference panel. The Dried Gram-Negative Panel demonstrated acceptable performance with an overall Essential Agreement of 92.8% for ceftolozane/tazobactam when compared with the frozen Reference panel.

Inoculum and instrument reproducibility testing demonstrated acceptable reproducibility and precision with ceftolozane/tazobactam, regardless of which inoculum method (i.e., Turbidity or Prompt™), or instrument (autoSCAN-4 instrument or WalkAway system) was used.

Quality Control testing demonstrated acceptable results for ceftolozane/tazobactam.

Beckman Coulter, the stylized logo, MicroScan, autoSCAN, and WalkAway are trademarks of Beckman Coulter, Inc. Beckman Coulter and the stylized logo are registered with the USPTO.

Prompt™ is a trademark of 3M Company, St. Paul, MN USA