



October 17, 2018

Aesculap, Inc.
Kathy Racosky
Senior Regulatory Affairs Specialist
3773 Corporate Parkway
Center Valley, Pennsylvania 18034

Re: K172709

Trade/Device Name: ELAN 4 Tools
Regulation Number: 21 CFR 882.4310
Regulation Name: Powered Simple Cranial Drills, Burrs, Trephines, And Their Accessories
Regulatory Class: Class II
Product Code: HBE
Dated: September 7, 2017
Received: September 8, 2017

Dear Ms. Racosky:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part

801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/CombinationProducts/GuidanceRegulatoryInformation/ucm597488.htm>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/>) and CDRH Learn (<http://www.fda.gov/Training/CDRHLearn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<http://www.fda.gov/DICE>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Matthew C. Krueger -S

for Carlos L. Peña, PhD, MS
Director
Division of Neurological
and Physical Medicine Devices
Office of Device Evaluation
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)
K172709

Device Name
Aesculap ELAN 4 Tools

Indications for Use (Describe)

Aesculap ELAN 4 Tools are intended for high speed cutting, sawing, and drilling of bone in the fields of Spine, ENT, Neuro, and Maxillofacial Surgery

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

B. 510(k) SUMMARY (as required by 21 CFR 807.92)

*Aesculap ELAN 4 Tools
October 15, 2018*

COMPANY: Aesculap[®], Inc.
3773 Corporate Parkway
Center Valley, PA 18034
Establishment Registration Number: 2916714

CONTACT: Kathy A. Racosky
610-984-9291 (phone)
610-791-6882 (fax)
kathy.racosky@aesculapimplants.com

TRADE NAME: ELAN 4 Tools

COMMON NAME: Drills, Burrs, Trephines & Accessories (Simple, Powered)

CLASSIFICATION: Class II

CLASSIFICATION NAME: Powered simple cranial drills, burrs, trephines and their accessories

REGULATION NUMBER: 882.4310

PRODUCT CODE: HBE

SUBSTANTIAL EQUIVALENCE

The ELAN 4 Tools are substantially equivalent to the primary predicate, Anspach Dissection Tools (K113476) and the reference predicate, Aesculap ELAN 4 Electro Motor System (K152960).

DEVICE DESCRIPTION

The ELAN 4 Tools are cutting devices designed for use with the ELAN 4 Electro Motor System. The ELAN 4 Tools are designed for cutting, sawing and drilling of bone. The ELAN 4 Tools have an attachment mechanism designed specifically for the type of motors and attachments with which they will be used. The ELAN 4 Tools are manufactured from stainless steel. The ELAN 4 Tools are components of the ELAN 4 Electro Motor System.

INDICATIONS FOR USE

The ELAN 4 Tools are intended for high speed cutting, sawing, and drilling of bone in the fields of Spine, ENT, Neuro, and Maxillofacial Surgery.

TECHNOLOGICAL CHARACTERISTICS (compared to predicate)

The ELAN 4 Tools are substantially equivalent to the predicate, Anspach Dissection Tools (K113476) and the reference predicate ELAN 4 Electro Motor System (K152960). The subject device is shown to be substantially equivalent and has the same performance characteristics to its primary predicate device through comparison in design, principles of operation, intended use, and materials. The subject device has the exact same indications as the reference predicate. A comparison table summarizing these characteristics can be found below.

The ELAN 4 Tools offers similar components when compared to the primary predicate, Anspach Dissection Tools (K113475). The only difference is the subject device is also offered in a blade type design. Bench testing determined that the blade design does not raise new questions of safety and effectiveness. Similar to the devices that are subject to this submission, the reference predicate, ELAN 4 Electro Motor System (K152960) also has the same indications for use.

	Aesculap Inc. ELAN 4 Tools	Anspach Dissection Tool	Aesculap, Inc. ELAN 4 Electro Motor System
K#	K172709	K113476 - Primary	K152960
Indications	The ELAN Tools are intended for high speed cutting, sawing, and drilling of bone in the fields of Spine, ENT, Neuro, and Maxillofacial Surgery.	Dissection tools are intended for cutting and shaping bone including spine and cranium.	The ELAN 4 Electro Motor System is intended for high speed cutting, sawing, and drilling of bone in the fields of Spine, ENT, Neuro, and Maxillofacial Surgery.
Regulation number	882.4310	882.4310	882.4310
Product Code	HBE	HBE	HBE
Regulatory Panel	Neurology	Neurology	Neurology
Tools			
Burrs & drills			
--Size	0.6 mm to 9 mm	0.5 mm to 24.5 mm	N/A
--Type	fluted, ball, conical, oval, barrel, acorn, pin, twist, drills, cutters, and discs	fluted, ball, conical, oval, barrel, acorn, pin, twist, drills, cutters, and discs	N/A
Blades	Sagittal & reciprocating	N/A	N/A
Materials	Stainless Steel	Stainless Steel, diamond	N/A
Attachment mechanism	Yes	Yes	N/A
Sterile Single Use	Yes	Yes	N/A
Sterilization	GAMMA irradiation	unknown	N/A

PERFORMANCE DATA**Biocompatibility Testing**

Biocompatibility testing in accordance to ISO 10991-1 and Use of International Standard ISO 10993-1, "Biological evaluation of medical devices – Part 1: Evaluation and testing within a risk management process", Guidance for Industry and Food and Drug Administration Staff has been performed to demonstrate that the ELAN 4 Tools are substantially equivalent to other predicate devices.

Biocompatibility testing within this submission includes Cytotoxicity and Hemolysis testing. The biocompatibility test results yielded a non-toxic response. Other biocompatibility endpoints were leveraged through a risk assessment.

Sterilization Validation

Product sterilization is validated in accordance with the ANSI/AAMI/ ISO 11137-1:2006 - Sterilization of health care products -- Radiation -- Part 1: Requirements for development, validation and routine control of a sterilization process for medical devices and ISO 11137-2:2006 - Sterilization of health care products -- Radiation -- Part 2: Establishing the sterilization dose.

Packaging is validated per ISO 11607. Accelerated aging data (5 years) for ELAN 4 Tools has been generated to support this submission.

Bacterial Endotoxin testing per Endotoxin ISO 10993-5 met the acceptance criterion.

Bench Testing

The following bench testing was performed to demonstrate that the ELAN 4 Tools perform as intended and are substantially equivalent to the predicate devices.

Test	Test Summary	Results
Rotating performance test for burrs and disc	Design not suitable, suitable for speeds up to 100,000 rpm	Pass
Functional testing of the tool adaptation with small blade	Distal tool adaptation and burr release mechanism, locks during use	Pass
Measurement of oscillation frequency and fatigue test (small blade)	Downforce oscillation output frequencies 0 to 20,000 rpm at the distal tool adaptation.	Pass
Measurement of oscillation frequency and fatigue test (large blade)	Distal tool adaptation and burr release mechanism, locks during use. Downforce oscillation output frequencies 0 to 20,000 rpm at the distal tool adaptation.	Pass
Measurement of temperature	With constant pressure and regular irrigation, evaluated bone temperature increase after milling, drilling or cutting	Pass

CONCLUSION:

The biocompatibility and bench testing results along with a comparison between the technology, materials and intended use for the ELAN 4 Tools and the predicate devices demonstrate that the ELAN 4 Tools are as safe, as effective, and perform as well as the predicate devices. The biocompatibility and bench testing results and technological comparison demonstrates that the ELAN 4 Tools are substantially equivalent to the predicate devices.