



December 14, 2017

ALPINION MEDICAL SYSTEMS Co., Ltd.
% Boyeon CHO
Quality Management Representative (QMR)
1FL and 6FL, Verdi Tower, 72, Digital-ro 26-gil, Guro-gu
Seoul, 08393
REPUBLIC OF KOREA

Re: K172732
Trade/Device Name: E-CUBE 8
Regulation Number: 21 CFR 892.1550
Regulation Name: Ultrasonic pulsed doppler imaging system
Regulatory Class: II
Product Code: IYN, IYO, ITX
Dated: November 27, 2017
Received: November 29, 2017

Dear Boyeon CHO:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/>) and CDRH Learn (<http://www.fda.gov/Training/CDRHLearn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<http://www.fda.gov/DICE>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,



Michael D. O'Hara For

Robert Ochs, Ph.D.
Director
Division of Radiological Health
Office of In Vitro Diagnostics
and Radiological Health
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K172732

Device Name
E-CUBE 8

Indications for Use (Describe)

The device is intended for use by a qualified physician for the evaluation of soft tissue and blood flow in the clinical applications; Fetal; Abdominal (renal & GYN/pelvic); Pediatric, Small Organ (breast, testes, thyroid); Neonatal Cephalic, Adult Cephalic; Trans-rectal, Trans-vaginal, Musculo-skeletal (Conventional); Musculo-skeletal (Superficial); Cardiac (adult& pediatric); Peripheral Vascular (PV); and Urology (including prostate).

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRAStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

Diagnostic Ultrasound Indications for Use

E-CUBE 8 Ultrasound System

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation								
	B	M	PWD	CWD	Color Doppler	Power Doppler	Tissue Harmonic Imaging	Combined* (Specify)	Other** (Specify)
Ophthalmic									
Fetal	N	N	N		N	N	N	N	N
Abdominal	N	N	N		N	N	N	N	N
Intra-operative (Specify)									
Intra-operative (Neuro)									
Laparoscopic									
Pediatric	N	N	N		N	N	N	N	N
Small Organ (breast, testes, thyroid)	N	N	N		N	N	N	N	
Neonatal Cephalic	N	N	N		N	N	N	N	
Adult Cephalic	N	N	N		N	N	N	N	
Trans-rectal	N	N	N		N	N	N	N	
Trans-vaginal	N	N	N		N	N	N	N	
Trans-urethral									
Trans-esoph. (non-Card.)									
Musculo-skeletal (Conventional)	N	N	N		N	N	N	N	
Musculo-skeletal (Superficial)	N	N	N		N	N	N	N	
Intravascular									
Cardiac Adult	N	N	N	N	N	N	N	N	
Cardiac Pediatric	N	N	N	N	N	N	N	N	
Intravascular (Cardiac)									
Trans-esoph. (Cardiac)									
Intra-cardiac									
Peripheral vessel	N	N	N	N	N	N	N	N	
Urology (including prostate)	N	N	N		N	N	N	N	N

N = new indication; P = previously cleared by FDA; E = added under appendix

* Combined: B/Color Doppler, B/PWD, B/Color Doppler/PWD; **Other: 3D, 4D

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In-Vitro Diagnostic Devices (OIVD)

Prescription User (Per 21 CFR 801.109)

Diagnostic Ultrasound Indications for Use

E-CUBE 8 with C1-6CT Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation								
	B	M	PWD	CWD	Color Doppler	Power Doppler	Tissue Harmonic Imaging	Combined* (Specify)	Other** (Specify)
Ophthalmic									
Fetal	N	N	N		N	N	N	N	
Abdominal	N	N	N		N	N	N	N	
Intra-operative (Specify)									
Intra-operative (Neuro)									
Laparoscopic									
Pediatric	N	N	N		N	N	N	N	
Small Organ (breast, testes, thyroid)									
Neonatal Cephalic									
Adult Cephalic									
Trans-rectal									
Trans-vaginal									
Trans-urethral									
Trans-esoph. (non-Card.)									
Musculo-skeletal (Conventional)	N	N	N		N	N	N	N	
Musculo-skeletal (Superficial)	N	N	N		N	N	N	N	
Intravascular									
Cardiac Adult									
Cardiac Pediatric									
Intravascular (Cardiac)									
Trans-esoph. (Cardiac)									
Intra-cardiac									
Peripheral vessel	N	N	N		N	N	N	N	
Urology (including prostate)	N	N	N		N	N	N	N	

N = new indication; P = previously cleared; E = added under appendix

* Combined: B/Color Doppler, B/PWD, B/Color Doppler/PWD; **Other: 3D, 4D

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In-Vitro Diagnostic Devices (OIVD)

Prescription User (Per 21 CFR 801.109)

Diagnostic Ultrasound Indications for Use

E-CUBE 8 with C5-8NT Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation								
	B	M	PWD	CWD	Color Doppler	Power Doppler	Tissue Harmonic Imaging	Combined* (Specify)	Other** (Specify)
Ophthalmic									
Fetal									
Abdominal	P	P	P		P	P	P	P	
Intra-operative (Specify)									
Intra-operative (Neuro)									
Laparoscopic									
Pediatric	P	P	P		P	P	P	P	
Small Organ (breast, testes, thyroid)									
Neonatal Cephalic	N	N	N		N	N	N	N	
Adult Cephalic									
Trans-rectal									
Trans-vaginal									
Trans-urethral									
Trans-esoph. (non-Card.)									
Musculo-skeletal (Conventional)									
Musculo-skeletal (Superficial)									
Intravascular									
Cardiac Adult	P	P	P		P	P	P	P	
Cardiac Pediatric	N	N	N		N	N	N	N	
Intravascular (Cardiac)									
Trans-esoph. (Cardiac)									
Intra-cardiac									
Peripheral vessel	N	N	N		N	N	N	N	
Urology (including prostate)									

N = new indication; P = previously cleared by FDA K153424; E = added under appendix

* Combined: B/Color Doppler, B/PWD, B/Color Doppler/PWD; **Other: 3D, 4D

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In-Vitro Diagnostic Devices (OIVD)

Prescription User (Per 21 CFR 801.109)

Diagnostic Ultrasound Indications for Use

E-CUBE 8 L3-12T with Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation								
	B	M	PWD	CWD	Color Doppler	Power Doppler	Tissue Harmonic Imaging	Combined* (Specify)	Other** (Specify)
Ophthalmic									
Fetal									
Abdominal	N	N	N		N	N	N	N	
Intra-operative (Specify)									
Intra-operative (Neuro)									
Laparoscopic									
Pediatric	P	P	P		P	P	P	P	
Small Organ (breast, testes, thyroid)	P	P	P		P	P	P	P	
Neonatal Cephalic	N	N	N		N	N	N	N	
Adult Cephalic									
Trans-rectal									
Trans-vaginal									
Trans-urethral									
Trans-esoph. (non-Card.)									
Musculo-skeletal (<i>Conventional</i>)	P	P	P		P	P	P	P	
Musculo-skeletal (<i>Superficial</i>)	P	P	P		P	P	P	P	
Intravascular									
Cardiac Adult									
Cardiac Pediatric									
Intravascular (Cardiac)									
Trans-esoph. (Cardiac)									
Intra-cardiac									
Peripheral vessel	P	P	P		P	P	P	P	
Urology (including prostate)									

N = new indication; P = previously cleared by FDA K153424; E = added under appendix

* Combined: B/Color Doppler, B/PWD, B/Color Doppler/PWD; **Other: 3D, 4D

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In-Vitro Diagnostic Devices (OIVD)

Prescription User (Per 21 CFR 801.109)

Diagnostic Ultrasound Indications for Use

E-CUBE 8 with L3-12H Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation								
	B	M	PWD	CWD	Color Doppler	Power Doppler	Tissue Harmonic Imaging	Combined* (Specify)	Other** (Specify)
Ophthalmic									
Fetal									
Abdominal	N	N	N		N	N	N	N	
Intra-operative (Specify)									
Intra-operative (Neuro)									
Laparoscopic									
Pediatric	N	N	N		N	N	N	N	
Small Organ (breast, testes, thyroid)	P	P	P		P	P	P	P	
Neonatal Cephalic	N	N	N		N	N	N	N	
Adult Cephalic									
Trans-rectal									
Trans-vaginal									
Trans-urethral									
Trans-esoph. (non-Card.)									
Musculo-skeletal (Conventional)	P	P	P		P	P	P	P	
Musculo-skeletal (Superficial)	P	P	P		P	P	P	P	
Intravascular									
Cardiac Adult									
Cardiac Pediatric									
Intravascular (Cardiac)									
Trans-esoph. (Cardiac)									
Intra-cardiac									
Peripheral vessel	P	P	P		P	P	P	P	
Urology (including prostate)									

N = new indication; P = previously cleared by FDA K161439; E = added under appendix

* Combined: B/Color Doppler, B/PWD, B/Color Doppler/PWD; **Other: 3D, 4D

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)Concurrence of CDRH, Office of In-Vitro Diagnostic Devices (OIVD)
Prescription User (Per 21 CFR 801.109)

Diagnostic Ultrasound Indications for Use
E-CUBE 8 L3-12H^{WD} with Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation								
	B	M	PWD	CWD	Color Doppler	Power Doppler	Tissue Harmonic Imaging	Combined* (Specify)	Other** (Specify)
Ophthalmic									
Fetal	N	N	N		N	N	N	N	
Abdominal	N	N	N		N	N	N	N	
Intra-operative (Specify)									
Intra-operative (Neuro)									
Laparoscopic									
Pediatric	N	N	N		N	N	N	N	
Small Organ (breast, testes, thyroid)	P	P	P		P	P	P	P	
Neonatal Cephalic									
Adult Cephalic									
Trans-rectal									
Trans-vaginal									
Trans-urethral									
Trans-esoph. (non-Card.)									
Musculo-skeletal (Conventional)	P	P	P		P	P	P	P	
Musculo-skeletal (Superficial)	P	P	P		P	P	P	P	
Intravascular									
Cardiac Adult									
Cardiac Pediatric									
Intravascular (Cardiac)									
Trans-esoph. (Cardiac)									
Intra-cardiac									
Peripheral vessel	P	P	P		P	P	P	P	
Urology (including prostate)									

N = new indication; P = previously cleared by FDA K161439; E = added under appendix

* Combined: B/Color Doppler, B/PWD, B/Color Doppler/PWD; **Other: 3D, 4D

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In-Vitro Diagnostic Devices (OIVD)

Prescription User (Per 21 CFR 801.109)

Diagnostic Ultrasound Indications for Use

E-CUBE 8 with SP3-8T Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation								
	B	M	PWD	CWD	Color Doppler	Power Doppler	Tissue Harmonic Imaging	Combined* (Specify)	Other** (Specify)
Ophthalmic									
Fetal	N	N	N		N	N	N	N	
Abdominal	P	P	P		P	P	P	P	
Intra-operative (Specify)									
Intra-operative (Neuro)									
Laparoscopic									
Pediatric	P	P	P		P	P	P	P	
Small Organ (breast, testes, thyroid)									
Neonatal Cephalic	N	N	N		N	N	N	N	
Adult Cephalic	P	P	P		P	P	P	P	
Trans-rectal									
Trans-vaginal									
Trans-urethral									
Trans-esoph. (non-Card.)									
Musculo-skeletal (Conventional)									
Musculo-skeletal (Superficial)									
Intravascular									
Cardiac Adult	P	P	P	P	P	P	P	P	
Cardiac Pediatric	P	P	P	P	P	P	P	P	
Intravascular (Cardiac)									
Trans-esoph. (Cardiac)									
Intra-cardiac									
Peripheral vessel									
Urology (including prostate)									

N = new indication; P = previously cleared by FDA K153424; E = added under appendix

* Combined: B/Color Doppler, B/PWD, B/Color Doppler/PWD; **Other: 3D, 4D

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In-Vitro Diagnostic Devices (OIVD)

Prescription User (Per 21 CFR 801.109)

Diagnostic Ultrasound Indications for Use

E-CUBE 8 with P1-5CT Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation								
	B	M	PWD	CWD	Color Doppler	Power Doppler	Tissue Harmonic Imaging	Combined* (Specify)	Other** (Specify)
Ophthalmic									
Fetal	N	N	N		N	N	N	N	
Abdominal	N	N	N		N	N	N	N	
Intra-operative (Specify)									
Intra-operative (Neuro)									
Laparoscopic									
Pediatric	N	N	N		N	N	N	N	
Small Organ (breast, testes, thyroid)									
Neonatal Cephalic									
Adult Cephalic	N	N	N		N	N	N	N	
Trans-rectal									
Trans-vaginal									
Trans-urethral									
Trans-esoph. (non-Card.)									
Musculo-skeletal (Conventional)									
Musculo-skeletal (Superficial)									
Intravascular									
Cardiac Adult	N	N	N	N	N	N	N	N	
Cardiac Pediatric	N	N	N	N	N	N	N	N	
Intravascular (Cardiac)									
Trans-esoph. (Cardiac)									
Intra-cardiac									
Peripheral vessel	N	N	N		N	N	N	N	
Urology (including prostate)									

N = new indication; P = previously cleared by FDA; E = added under appendix

* Combined: B/Color Doppler, B/PWD, B/Color Doppler/PWD; **Other: 3D, 4D

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In-Vitro Diagnostic Devices (OIVD)

Prescription User (Per 21 CFR 801.109)

Diagnostic Ultrasound Indications for Use

E-CUBE 8 with VC1-6T Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation								
	B	M	PWD	CWD	Color Doppler	Power Doppler	Tissue Harmonic Imaging	Combined* (Specify)	Other** (Specify)
Ophthalmic									
Fetal	N	N	N		N	N	N	N	N
Abdominal	N	N	N		N	N	N	N	N
Intra-operative (Specify)									
Intra-operative (Neuro)									
Laparoscopic									
Pediatric	N	N	N		N	N	N	N	N
Small Organ (breast, testes, thyroid)									
Neonatal Cephalic									
Adult Cephalic									
Trans-rectal									
Trans-vaginal									
Trans-urethral									
Trans-esoph. (non-Card.)									
Musculo-skeletal (Conventional)									
Musculo-skeletal (Superficial)									
Intravascular									
Cardiac Adult									
Cardiac Pediatric									
Intravascular (Cardiac)									
Trans-esoph. (Cardiac)									
Intra-cardiac									
Peripheral vessel									
Urology (including prostate)	N	N	N		N	N	N	N	N

N = new indication; P = previously cleared; E = added under appendix

* Combined: B/Color Doppler, B/PWD, B/Color Doppler/PWD; **Other: 3D, 4D

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In-Vitro Diagnostic Devices (OIVD)

Prescription User (Per 21 CFR 801.109)

Diagnostic Ultrasound Indications for Use

E-CUBE 8 with EV3-10T Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation								
	B	M	PWD	CWD	Color Doppler	Power Doppler	Tissue Harmonic Imaging	Combined* (Specify)	Other** (Specify)
Ophthalmic									
Fetal	N	N	N		N	N	N	N	
Abdominal									
Intra-operative (Specify)									
Intra-operative (Neuro)									
Laparoscopic									
Pediatric									
Small Organ (breast, testes, thyroid)									
Neonatal Cephalic									
Adult Cephalic									
Trans-rectal	P	P	P		P	P	P	P	
Trans-vaginal	P	P	P		P	P	P	P	
Trans-urethral									
Trans-esoph. (non-Card.)									
Musculo-skeletal (Conventional)									
Musculo-skeletal (Superficial)									
Intravascular									
Cardiac Adult									
Cardiac Pediatric									
Intravascular (Cardiac)									
Trans-esoph. (Cardiac)									
Intra-cardiac									
Peripheral vessel	N	N	N		N	N	N	N	
Urology (including prostate)	P	P	P		P	P	P	P	

N = new indication; P = previously cleared by FDA K153424; E = added under appendix

* Combined: B/Color Doppler, B/PWD, B/Color Doppler/PWD; **Other: 3D, 4D

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In-Vitro Diagnostic Devices (OIVD)

Prescription User (Per 21 CFR 801.109)

Diagnostic Ultrasound Indications for Use

E-CUBE 8 with EC3-10T Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation								
	B	M	PWD	CWD	Color Doppler	Power Doppler	Tissue Harmonic Imaging	Combined* (Specify)	Other** (Specify)
Ophthalmic									
Fetal	N	N	N		N	N	N	N	
Abdominal									
Intra-operative (Specify)									
Intra-operative (Neuro)									
Laparoscopic									
Pediatric									
Small Organ (breast, testes, thyroid)									
Neonatal Cephalic									
Adult Cephalic									
Trans-rectal	P	P	P		P	P	P	P	
Trans-vaginal	P	P	P		P	P	P	P	
Trans-urethral									
Trans-esoph. (non-Card.)									
Musculo-skeletal (Conventional)									
Musculo-skeletal (Superficial)									
Intravascular									
Cardiac Adult									
Cardiac Pediatric									
Intravascular (Cardiac)									
Trans-esoph. (Cardiac)									
Intra-cardiac									
Peripheral vessel	N	N	N		N	N	N	N	
Urology (including prostate)	P	P	P		P	P	P	P	

N = new indication; P = previously cleared by FDA K153424; E = added under appendix

* Combined: B/Color Doppler, B/PWD, B/Color Doppler/PWD; **Other: 3D, 4D

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In-Vitro Diagnostic Devices (OIVD)

Prescription User (Per 21 CFR 801.109)

Diagnostic Ultrasound Indications for Use

E-CUBE 8 with E3-10 Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation								
	B	M	PWD	CWD	Color Doppler	Power Doppler	Tissue Harmonic Imaging	Combined* (Specify)	Other** (Specify)
Ophthalmic									
Fetal	P	P	P		P	P	P	P	
Abdominal									
Intra-operative (Specify)									
Intra-operative (Neuro)									
Laparoscopic									
Pediatric									
Small Organ (breast, testes, thyroid)									
Neonatal Cephalic									
Adult Cephalic									
Trans-rectal	P	P	P		P	P	P	P	
Trans-vaginal	P	P	P		P	P	P	P	
Trans-urethral									
Trans-esoph. (non-Card.)									
Musculo-skeletal (<i>Conventional</i>)									
Musculo-skeletal (<i>Superficial</i>)									
Intravascular									
Cardiac Adult									
Cardiac Pediatric									
Intravascular (Cardiac)									
Trans-esoph. (Cardiac)									
Intra-cardiac									
Peripheral vessel	N	N	N		N	N	N	N	
Urology (including prostate)	P	P	P		P	P	P	P	

N = new indication; P = previously cleared by FDA **K132687**; E = added under appendix

* Combined: B/Color Doppler, B/PWD, B/Color Doppler/PWD; **Other: 3D, 4D

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In-Vitro Diagnostic Devices (OIVD)

Prescription User (Per 21 CFR 801.109)

Diagnostic Ultrasound Indications for Use

E-CUBE 8 with CW2.0 Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation								
	B	M	PWD	CWD	Color Doppler	Power Doppler	Tissue Harmonic Imaging	Combined* (Specify)	Other** (Specify)
Ophthalmic									
Fetal									
Abdominal									
Intra-operative (Specify)									
Intra-operative (Neuro)									
Laparoscopic									
Pediatric									
Small Organ (breast, testes, thyroid)									
Neonatal Cephalic									
Adult Cephalic									
Trans-rectal									
Trans-vaginal									
Trans-urethral									
Trans-esoph. (non-Card.)									
Musculo-skeletal (Conventional)									
Musculo-skeletal (Superficial)									
Intravascular									
Cardiac Adult				P					
Cardiac Pediatric				P					
Intravascular (Cardiac)									
Trans-esoph. (Cardiac)									
Intra-cardiac									
Peripheral vessel									
Urology (including prostate)									

N = new indication; P = previously cleared by FDA K161439; E = added under appendix

* Combined: B/Color Doppler, B/PWD, B/Color Doppler/PWD; **Other: 3D, 4D

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In-Vitro Diagnostic Devices (OIVD)

Prescription User (Per 21 CFR 801.109)

Diagnostic Ultrasound Indications for Use

E-CUBE 8 with CW5.0 Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation								
	B	M	PWD	CWD	Color Doppler	Power Doppler	Tissue Harmonic Imaging	Combined* (Specify)	Other** (Specify)
Ophthalmic									
Fetal									
Abdominal									
Intra-operative (Specify)									
Intra-operative (Neuro)									
Laparoscopic									
Pediatric									
Small Organ (breast, testes, thyroid)									
Neonatal Cephalic									
Adult Cephalic									
Trans-rectal									
Trans-vaginal									
Trans-urethral									
Trans-esoph. (non-Card.)									
Musculo-skeletal (Conventional)									
Musculo-skeletal (Superficial)									
Intravascular									
Cardiac Adult				P					
Cardiac Pediatric				P					
Intravascular (Cardiac)									
Trans-esoph. (Cardiac)									
Intra-cardiac									
Peripheral vessel				N					
Urology (including prostate)									

N = new indication; P = previously cleared by FDA K161439; E = added under appendix

* Combined: B/Color Doppler, B/PWD, B/Color Doppler/PWD; **Other: 3D, 4D

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In-Vitro Diagnostic Devices (OIVD)

Prescription User (Per 21 CFR 801.109)

510(k) Summary

In accordance with 21CFR807.92, the following summary of information is provided;

Date September 8th, 2017

Submitter: ALPINION MEDICAL SYSTEMS Co., Ltd.
Address: 1FL and 6FL, Verdi Tower, 72, Digital-ro 26-gil, Guro-gu,
Seoul, Republic of Korea

Primary Contact Person Boyeon CHO
Quality Management Representative(QMR)
Address: 1FL and 6FL, Verdi Tower, 72, Digital-ro 26-gil, Guro-gu,
Seoul, Republic of Korea
Phone: +82 70 7465 2104
Fax: +82 2 851 5595
Email: qa_ra@alpinion.com

Secondary Contact Person JULIAN LEE
Address: 21312 30th Dr SE Ste 100 Bothell, WA 98021, United States
Phone: 425 949 1059
Fax: 425 949 4910
Email: julian.lee@alpinionusa.com

Device Trade Name: E-CUBE 8

Common/ Usual Name: Ultrasonic Pulsed Doppler Imaging System

Classification Names System, Imaging, Pulsed Doppler Ultrasonic

Product Code: Ultrasonic Pulsed Doppler Imaging System, 21CFR 892.1550 90-IYN
Ultrasonic Pulsed Echo Imaging System, 21CFR 892.1560, 90-IYO
Diagnostic Ultrasound Transducer, 21CFR 892.1570, 90-ITX

Primary Predicate Device K161439 E-CUBE 11 Diagnostic Ultrasound System

Predicate Devices K153424 E-CUBE i7 Diagnostic Ultrasound System
K132687 E-CUBE 7 Diagnostic Ultrasound System
K163691 HS70A Diagnostic Ultrasound System
K163120 EPIQ 5 / EPIQ 7 Diagnostic Ultrasound System

Device Description: E-CUBE 8 product is an ultrasound imaging system for medical diagnosis. This innovative system platform provides optimal patient diagnosis workflow with the wide flat panel display, ergonomic control panel with easy user interface, optimal image quality.

1. Signal Mode:

2D(B) mode, Harmonic mode (HAR), M mode, Color M mode, Anatomical M mode, Color Flow Doppler(CF) Mode, Power Doppler(PD) Mode, Directional PD mode, Pulsed Wave Doppler(PWD) Mode, Continuous Wave Doppler(CWD) Mode, High PRF Doppler mode, Tissue Doppler Imaging(TDI) Mode, 3D/4D mode

2. Combination Mode:

B/Color Doppler, B/PWD, B/Color Doppler/PWD

Acoustic output track:

Track 3

Types of transducers compatible with the device :

	C1-6CT	C5-8NT	P1-5CT	SP3-8T	VC1-6T
Transducer type	Convex	Convex	Sector Phased	Sector Phased	Volume Convex
Applicable frequency	1~6MHz	5~8MHz	1~5MHz	3~8MHz	1~6MHz
Intended usage	Fetal, Abdominal, Pediatric, Musculo-skeletal, Urology	Fetal, Pediatric, Cardiac Adult, Cardiac Pediatric	Abdominal, Pediatric, Adult Cephalic, Cardiac Adult, Cardiac Pediatric	Abdominal, Pediatric, Adult Cephalic, Cardiac Adult, Cardiac Pediatric	Fetal, Abdominal, Pediatric, Urology
Applicable mode	B/M/PWD/ Color Doppler/ Power Doppler/ Tissue Harmonic Imaging/ Combined	B/M/PWD/ Color Doppler/ Power Doppler/ Tissue Harmonic Imaging/ Combined	B/M/PWD/ CWD Color Doppler/ Power Doppler/ Tissue Harmonic Imaging/ Combined	B/M/PWD/ CWD Color Doppler/ Power Doppler/ Tissue Harmonic Imaging/ Combined	B/M/PWD/ Color Doppler/ Power Doppler/ Tissue Harmonic Imaging/ Combined/ Others
Scanning Depth (mm)	300	140	300	300	300
Total number of element	128	128	64	64	128
Predicate	New (SC1-6H:K161439, C5-1:K163120)	Previously cleared (K153424)	New (SP1-5T:K153424, S5-1:K163120)	Previously cleared (K153424)	New (VC1-6:K132687)

	L3-12T	L3-12H	L3-12HWD	EC3-10T	EV3-10T
Transducer type	Linear	Linear	Linear	Endocavity	Endocavity
Applicable frequency	3~12MHz	3~12MHz	3~12MHz	3~10MHz	3~10MHz
Intended usage	Pediatric, Small Organ, Musculo-skeletal, Peripheral vessel	Pediatric, Small Organ, Musculo-skeletal, Peripheral vessel	Pediatric, Small Organ, Musculo-skeletal, Peripheral vessel	Fetal, Trans-rectal, Trans-vaginal, Urology	Fetal, Trans-rectal, Trans-vaginal, Urology
Applicable mode	B/M/PWD/ Color Doppler/ Power Doppler/ Tissue Harmonic Imaging/ Combined	B/M/PWD/ Color Doppler/ Power Doppler/ Tissue Harmonic Imaging/ Combined	B/M/PWD/ Color Doppler/ Power Doppler/ Tissue Harmonic Imaging/ Combined	B/M/PWD/ Color Doppler/ Power Doppler/ Tissue Harmonic Imaging/ Combined	B/M/PWD/ Color Doppler/ Power Doppler/ Tissue Harmonic Imaging/ Combined
Scanning Depth (mm)	100	100	100	100	100
Total number of element	128	192	192	128	128
Predicate	Previously cleared (K153424)	Previously cleared (K161439)	Previously cleared (K161439)	Previously cleared (K153424)	Previously cleared (K153424)

	E3-10	CW2.0	CW5.0
Transducer type	Endocavity	Pencil Doppler	Pencil Doppler
Applicable frequency	3~10MHz	2.0MHz	5.0MHz
Intended usage	Fetal, Trans-rectal, Trans-vaginal, Urology	Cardiac Adult, Cardiac Pediatric	Cardiac Adult, Cardiac Pediatric
Applicable mode	B/M/PWD/ Color Doppler/ Power Doppler/ Tissue Harmonic Imaging/ Combined	CWD	CWD
Scanning Depth (mm)	140	N/A	N/A
Total number of element	128	2 (TX1, RX1)	2 (TX1, RX1)
Predicate	Previously cleared (K132687)	Previously cleared (K161439)	Previously cleared (K161439)

Indications For Use: The device is intended for use by a qualified physician for the evaluation of soft tissue and blood flow in the clinical applications; Fetal; Abdominal (renal & GYN/pelvic); Pediatric, Neonatal Cephalic, Small Organ (breast, testes, thyroid); Adult Cephalic; Trans-rectal, Trans-vaginal, Musculo-skeletal (Conventional); Musculo-skeletal (Superficial); Cardiac (adult& pediatric); Peripheral Vascular (PV); and Urology (including prostate).

510(k) E-CUBE 8

Determination of
Substantial Equivalence: Comparison table with Predicate devices:

Model	Proposed E-CUBE 8 ALPINION Medical Systems Co., Ltd.	Predicate E-CUBE 11 ALPINION Medical Systems Co., Ltd.	Predicate E-CUBE i7 ALPINION Medical Systems Co., Ltd.	Predicate E-CUBE 7 ALPINION Medical Systems Co., Ltd.	Predicate HS70A Samsung Medison co., ltd	Predicate EPIQ 5 and EPIQ 7 Philips Ultrasound, Inc.
Feature	-	K161439	K153424	K132687	K163691	K163120
Indications for Use						
- Fetal	√	√	√	√	√	√
- Abdominal (Renal&GYN/Pelvic)	√	√	√	√	√	√
- Pediatric	√	√	√	√	√	√
- Small Organ (breast, testes, thyroid)	√	√	√	√	√	√
- Neonatal Cephalic	√				√	√
- Adult Cephalic	√	√	√		√	√
- Trans-rectal	√	√	√	√	√	√
- Trans-vaginal	√	√	√	√	√	√
- Musculo-skeletal (Conventional)	√	√	√	√	√	√
- Musculo skeletal (Superficial)	√	√	√	√	√	√
- Cardiac (Adult)	√	√	√	√	√	√
- Cardiac (Pediatric)	√	√	√	√	√	√
- Peripheral Vessel	√	√	√	√	√	√
- Urology (including prostate)	√	√	√	√	√	√
Imaging modes						
- 2D(B) mode	√	√	√	√	√	√
- Harmonic mode	√	√	√	√	√	√
- M mode	√	√	√	√	√	√
- Color M mode	√	√	√		√	√
- Anatomical M mode	√	√	√		√	√
- Color Flow Doppler (CF) mode	√	√	√	√	√	√
- Power Doppler (PD) mode	√	√	√	√	√	√
- Directional PD mode	√	√	√		√	√

510(k) E-CUBE 8

- Pulsed wave Doppler (PWD) mode	√	√	√	√	√	√
- Continuous wave Doppler (CWD) mode	√	√	√		√	√
- High PRF Doppler mode	√	√	√		√	√
- Tissue Doppler imaging (TDI) mode	√	√	√		√	√
- 3D/4D mode	√	√		√	√	√
Imaging modes						
- 2D(B) mode	√	√	√	√	√	√
- Harmonic mode	√	√	√	√	√	√
- M mode	√	√	√	√	√	√
- Color M mode	√	√	√		√	√
- Anatomical M mode	√	√	√		√	√
- Color Flow Doppler (CF) mode	√	√	√	√	√	√
- Power Doppler (PD) mode	√	√	√	√	√	√
- Directional PD mode	√	√	√		√	√
- Pulsed wave Doppler (PWD) mode	√	√	√	√	√	√
- Continuous wave Doppler (CWD) mode	√	√	√		√	√
- High PRF Doppler mode	√	√	√		√	√
- Tissue Doppler imaging (TDI) mode	√	√	√		√	√
- 3D/4D mode	√	√		√	√	√
Imaging Functions						
- Xpeed™	√	√	√	√	√	√
- Full SRI™	√	√	√	√	√	√
- Spatial Compounding Image (SCI)	√	√	√	√	√	√
- Frequency Compounding image (FCI)	√	√		√	√	
- Panoramic	√	√	√	√	√	√
- Stress Echo	√	√		√	√	√
- Cube Strain™	√	√		√	√	√
- Live HQ™	√	√		√	√	√
- Needle Vision™ / Needle Vision™ Plus	√	√	√	√	√	√
- Elastography	√	√			√	√

- Cube view™	√	√	√	√		
- Volume Advance™						
• Free Angle MSV	√				√	
• AnySlice™	√				√	
• Volume Analysis	√				√	
Thermal, Mechanical and Electrical safety						
NEMA UD2, UD3	√	√	√	√	√	√
AIUM Medical Ultrasound Safety	√	√	√	√	√	√
IEC 60601-1	√	√	√	√	√	√
IEC 60601-1-2	√	√	√	√	√	√
IEC 60601-2-37	√	√	√	√	√	√

Summary of Non-Clinical Tests:

E-CUBE 8 has been evaluated for biocompatibility, acoustic output as well as thermal, electrical, electromagnetic, and mechanical safety, and has been found to conform to applicable medical device safety standards. E-CUBE 8 and its application comply with voluntary standards as detailed in this premarket submission.

- ◆ IEC60601-1:2005(Third Edition)+CORR.1:2006+CORR.2:2007+A1:2012, Medical Electrical Equipment – Part 1: General Requirements for Safety
- ◆ IEC60601-1-2 Edition 3:2007-03, Medical Electrical Equipment – Part 1-2: General Requirements for Safety – Collateral Standard: Electromagnetic Compatibility Requirements and Tests
- ◆ IEC60601-2-37 Edition 2.0:2007, Medical Electrical Equipment – Part 2-37: Particular Requirements for the Safety of Ultrasonic Medical Diagnostic and Monitoring Equipment
- ◆ AAMI/ANSI/ISO10993-1:2009(R)2013, Biological Evaluation of Medical Devices - Part 1: Evaluation and Testing
- ◆ AAMI/ANSI/ISO14971:2007/(R)2010, Medical devices-Application of risk management to medical devices
- ◆ AIUM MUS, Third edition, Medical Ultrasound Safety
- ◆ NEMA UD 2-2004(R2009), Acoustic Output Measurement Standard for Diagnostic Ultrasound Equipment
- ◆ NEMA UD 3-2004(R2009), Standard for Real Time Display of Thermal and Mechanical Acoustic Output Indices on Diagnostic ultrasound Equipment

The following quality management system measures were applied to the development of E-CUBE 8:

- ◆ Medical Device Risk Management
- ◆ Requirements Reviews
- ◆ Design Reviews
- ◆ Component Verification
- ◆ Integration Review (System Verification)
- ◆ Performance Testing (System Verification)
- ◆ Safety Testing (Compliance Test)
- ◆ Design Validation

Transducer materials and other patient contact materials are biocompatible.

Summary of Clinical Tests:

The subject of this premarket submission, E-CUBE 8, did not require clinical studies to support substantial equivalence.

Discussion:

E-CUBE 8 was compared with the predicate devices. The subject device is in conformance with applicable safety standards.

Therefore, the differences between E-CUBE 8 and the predicate devices would not affect the safety, effectiveness and essential performance.

Conclusion: ALPINION MEDICAL SYSTEMS Co., Ltd. considers E-CUBE 8 to be as safe, as effective. Performance, technology and software are substantially equivalent to the predicate devices.

ALPINION MEDICAL SYSTEMS Co., Ltd. will update and include in this summary any other information deemed reasonably necessary by the FDA or the requirements will be published in guidance documents.