



February 1, 2018

Beijing HangTian KaDi Technology R&D Institute
% Diana Hong
General Manager
Mid-Link Consulting Co., Ltd
P.O. Box 120-119
Shanghai, 200120 China

Re: K172789

Trade/Device Name: Specimen Retrieval Bag
Regulation Number: 21 CFR 876.1500
Regulation Name: Endoscope and Accessories
Regulatory Class: Class II
Product Code: GCJ
Dated: January 5, 2018
Received: January 11, 2018

Dear Diana Hong:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good

manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/>) and CDRH Learn (<http://www.fda.gov/Training/CDRHLearn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<http://www.fda.gov/DICE>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

**Jennifer R.
Stevenson -S3**

For Binita S. Ashar, M.D., M.B.A., F.A.C.S.
Director
Division of Surgical Devices
Office of Device Evaluation
Center for Devices and Radiological Health

Enclosure

DEPARTMENT OF HEALTH AND HUMAN SERVICES
Food and Drug Administration

Form Approved: OMB No. 0910-0120
Expiration Date: January 31, 2017
See PRA Statement below.

Indications for Use

510(k) Number (if known)

K172789

Device Name

Specimen Retrieval Bag

Indications for Use (Describe)

The device is indicated for use as a receptacle for the collection and extraction of tissue, organs and calculi during general and laparoscopic procedures.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary

This 510(k) Summary is being submitted in accordance with requirements of Title 21, CFR Section 807.92.

The assigned 510(k) Number: K172789

1. Date of Preparation: 01/05/2018
2. Sponsor Identification

Beijing HangTian KaDi Technology R&D Institute

Room 301-08, 09, 10, 11, Third floor, Building No.13, No.15 Jing Sheng Nan Er Street, TongZhou District, 101102 Beijing, PEOPLE'S REPUBLIC OF CHINA

Establishment Registration Number: Not yet registered

Contact Person: Liying Zhang

Position: General Manager

Tel: +86-10-56407751

Fax: +86-10-56407795-803

Email: zly1110@139.com

3. Designated Submission Correspondent

Ms. Diana Hong (Primary Contact Person)

Ms. Ying Xu (Alternative Contact Person)

Mid-Link Consulting Co., Ltd

P.O. Box 120-119, Shanghai, 200120, China

Tel: +86-21-22815850,

Fax: 3609253199

Email: info@mid-link.net

4. Identification of Proposed Device

Trade Name: Specimen Retrieval Bag

Common Name: Specimen Retrieval Bag

Regulatory Information

Classification Name: Laparoscope, General& Plastic Surgery

Classification: II;

Product Code: GCJ

Regulation Number: CFR 876.1500

Review Panel: General& Plastic Surgery

Intended Use Statement:

The device is indicated for use as a receptacle for the collection and extraction of tissue, organs and calculi during general and laparoscopic procedures.

Device Description

The proposed device, Specimen Retrieval Bag, is comprised of retracted bag, opening support, outer introducer shaft, outer introducer handle, inner introducer shaft, inner introducer handle, string and pull loop. The device is available in a series of models. The difference between each model is the retracted bag size and volume. The device is provided sterile and single use.

5. Identification of Predicate Device

510(k) Number: K100959

Product Name: Specimen Retrieval System

6. Non-Clinical Test Conclusion

Non clinical tests were conducted to verify that the proposed device met all design specifications as was Substantially Equivalent (SE) to the predicate device. The test results demonstrated that the proposed device complies with the following standards:

- ISO 10993-1:2009 Biological evaluation of medical devices - Part 1: Evaluation and testing within a risk management process
- ISO 10993-5:2009 Biological Evaluation of Medical Devices- Part 5: Tests For In Vitro Cytotoxicity.

- ISO 10993-7:2008 Biological Evaluation of Medical Devices-Part 7: Ethylene Oxide Sterilization Residuals.
- ISO 10993-10:2010 Biological Evaluation of Medical Devices- Part 10: Tests for Irritation and Skin Sensitization.
- ASTM F88/F88M-15 Standard Test Method for Seal Strength of Flexible Barrier Materials.
- ASTM F 1929-15 Standard Test Method for Detecting Seal Leaks in Porous Medical Packaging by Dye Penetration

The performance tests were performed on both proposed device and predicate device. The test items include follow items

- Retracted Bag Peeling Force Test
- String Tensile Strength Test
- Force to Withdraw Retracted Bag Test
- Force to Push out Retracted Bag Test
- Penetration Force Test
- Leak Test

7. Clinical Test Conclusion

No clinical study is included in this submission.

8. Substantially Equivalent (SE) Comparison

Table 1 Comparison of Technology Characteristics

Item	Proposed Device	Predicate Device K100959
Product Code	GCJ	Same
Regulation Number	21 CFR 876.1500	Same
Class	CLASS II	Same
Intended Use	The specimen retrieval bag is indicated for use as a receptacle for the collection and extraction of tissue, organs and calculi during general and laparoscopic surgical procedures.	Same
Configuration	Retracted Bag	Same
	Outer Introducer Shaft	
	Outer Introducer Handle	
	Inner Introducer Shaft	
	Inner Introducer Handle	
	String	
	Opening Support	
	Pull loop	
Single Use	Single Use	Same
Label/Labeling	Complied with 21 CFR part 801	Same
Sterilization	EO sterilization	Irradiation Sterilization
Biocompatibility		
Cytotoxicity	No cytotoxicity	Conform with ISO 10993 requirements
Skin Sensitization	No skin sensitization	
Irritation	No irritation	

9. Substantially Equivalent (SE) Conclusion

Based on the comparison and analysis above, the proposed devices are determined to be Substantially Equivalent (SE) to the predicate devices.