



December 20, 2017  
Paragon 28  
Eric Lintula  
Director of Regulatory Affairs  
4B Inverness Ct. E  
Ste 280  
Englewood, Colorado 80112

Re: K172886  
Trade/Device Name: Breakaway Screw System  
Regulation Number: 21 CFR 888.3030  
Regulation Name: Single/Multiple Component Metallic Bone Fixation Appliances And Accessories  
Regulatory Class: Class II  
Product Code: HRS, HWC, HTN  
Dated: November 3, 2017  
Received: November 14, 2017

Dear Eric Lintula:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820);

and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/>) and CDRH Learn (<http://www.fda.gov/Training/CDRHLearn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<http://www.fda.gov/DICE>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Vincent J. Devlin -S

for

Mark N. Melkerson  
Director  
Division of Orthopedic Devices  
Office of Device Evaluation  
Center for Devices and  
Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)

K172886

Device Name

Breakaway Screw System

Indications for Use (Describe)

The BABY GORILLA®/GORILLA® Bone Plates and Bone Screws of the BABY GORILLA®/GORILLA® Plating System are indicated for use in stabilization and fixation of fractures or osteotomies; intra and extra articular fractures, joint depression, and multi-fragmentary fractures; revision procedures, joint fusion and reconstruction of small bones of the toes, feet and ankles including the distal tibia, talus, and calcaneus. The system can be used in both adult and pediatric patients.

In addition, the non-locking, titanium screws and washers are indicated for use in bone reconstruction, osteotomy, arthrodesis, joint fusion, fracture repair and fracture fixation, appropriate for the size of the device

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

**Section 8: 510(k) Summary**

<b>Date:</b>	9/18/17
<b>Sponsor:</b>	Paragon 28, Inc. 4B Inverness Ct. E., STE 280 Englewood, Colorado 80112 Phone: (888) 728-1888 Fax: (888) 728-1220
<b>Sponsor contact:</b>	Eric Lintula, Director of Regulatory Affairs
<b>Trade Names:</b>	Breakaway Screw System
<b>Regulatory Class:</b>	Class II
<b>Regulation, Product Code, Classification, and Common Name:</b>	888.3030, HRS, Single/multiple component metallic bone fixation appliances and accessories, bone plate system  888.3040, HWC, Smooth or threaded metallic bone fixation fastener, bone screw  888.3030, HTN, Single/multiple component metallic bone fixation accessories, washer
<b>Device Description:</b>	The BABY GORILLA®/GORILLA® implants are lower extremity fixation systems. Gorilla Plates are offered in “mini” and “standard” set sizes in a variety of shapes based upon the anatomical fixation required. Screws are also offered in “mini” and “standard” sets and, in addition, in locking and non-locking versions. Size-matched washers are available for use with the non-locking screws when the latter are used for fixation without the plates.
<b>Materials:</b>	The BABY GORILLA®/GORILLA® implants are manufactured from medical grade titanium (per ASTM F67), stainless steel (per ASTM F138), and titanium alloy (per ASTM F136).
<b>Indications for Use:</b>	The BABY GORILLA®/GORILLA® Bone Plates and Bone Screws of the BABY GORILLA®/GORILLA® Plating System are indicated for use in stabilization and fixation of fractures or osteotomies; intra and extra articular fractures, joint depression, and multi-fragmentary fractures; revision procedures, joint fusion and reconstruction of small bones of the toes, feet and ankles including the distal tibia, talus, and calcaneus. The system can be used in both adult and pediatric patients.  In addition, the non-locking, titanium screws and washers are indicated for use in bone reconstruction,

	osteotomy, arthrodesis, joint fusion, fracture repair and fracture fixation, appropriate for the size of the device
<b>Primary Predicate:</b>	K140397, Gorilla Plating System (formerly Paralock Plating System)
<b>Performance Data:</b>	<p>Engineering analysis is presented to provide evidence that the original testing and subsequent performance is not adversely affected by the modified screw geometry.</p> <p>The results of the analysis demonstrated the modified design is substantially equivalent to the predicate device.</p>
<b>Technological Characteristics:</b>	<p>The modified Gorilla screw possesses the same technological characteristics as the predicate devices. These include:</p> <ul style="list-style-type: none"><li>• performance,</li><li>• basic design,</li><li>• material, manufacturing and</li><li>• sizes (dimensions are comparable to those offered by the predicate systems).</li></ul> <p>Therefore, the fundamental scientific technology of the modified Gorilla Screw is similar to previously cleared devices.</p>
<b>Conclusion:</b>	The modified Gorilla Screw possesses indications for use and technological characteristics the same as the predicate device. Therefore, the modified Gorilla Screw is substantially equivalent to the predicate.