



Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Center – WO66-G609  
Silver Spring, MD 20993-0002

November 22, 2017

Instrumentation Laboratory Co.  
Shajunath Nirupama  
Regulatory Affairs Specialist I  
180 Hartwell Road  
Bedford, MA 01730

Re: K172903

Trade/Device Name: HemosIL D-Dimer HS 500  
Regulation Number: 21 CFR 864.7320  
Regulation Name: Fibrinogen/fibrin degradation products assay  
Regulatory Class: Class II  
Product Code: DAP  
Dated: October 26, 2017  
Received: October 27, 2017

Dear Ms. Nirupama:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21

CFR Part 807); labeling (21 CFR Parts 801 and 809); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

**Leonthena R. Carrington -S**

Lea Carrington

Director

Division of Immunology and Hematology Devices

Office of In Vitro Diagnostics and Radiological

Health

Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)

K172903

Device Name

HemosIL, D-Dimer HS 500

Indications for Use (Describe)

HemosIL D-Dimer HS 500 is an automated latex enhanced immunoassay for the quantitative determination of D-Dimer in human citrated plasma on the ACL TOP® Family and ACL TOP Family 50 Series Systems for use, in conjunction with a clinical pretest probability (PTP) assessment model to exclude venous thromboembolism (VTE) in outpatients suspected of deep venous thrombosis (DVT) and pulmonary embolism (PE).

For in vitro diagnostic use.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

## 510(k) Summary

The submission meets the criteria for a Special 510(k) under the FDA guidance “The New 510(k) Paradigm - Alternate Approaches to Demonstrating Substantial Equivalence in Premarket Notifications”.

|  |   |
|--|---|
| <b>Submitter's Information</b>                   | Instrumentation Laboratory (IL) Co.<br>180 Hartwell Road<br>Bedford, MA 01730, USA  |
| <b>Contact Person</b>                            | Shajunath Nirupama,<br>Regulatory Affairs Specialist I<br>Phone: 781-861-4083<br>Fax: 781-861-4207<br>Email: snirupama@ilww.com   |
| <b>Preparation Date</b>                          | November 17, 2017   |
| <b>Device Trade Name</b>                         | HemosIL D-Dimer HS 500  |
| <b>Regulatory Information</b>                    | Classification: Class II<br>Regulation No.: 21 CFR 864.7320<br>Common Name: Fibrinogen and Fibrin split products, Antigen, Antiserum, Control<br>Panel: Hematology (81)<br>Product Code: DAP  |
| <b>Predicate Device</b>                          | HemosIL D-Dimer HS 500 510(k) No: K090264   |
| <b>Device Indications for Use / Intended Use</b> | HemosIL D-Dimer HS 500 is an automated latex enhanced immunoassay for the quantitative determination of D-Dimer in human citrated plasma on the ACL TOP® Family and ACL TOP Family 50 Series Systems for use, in conjunction with a clinical pretest probability (PTP) assessment model to exclude venous thromboembolism (VTE) in outpatients suspected of deep venous thrombosis (DVT) and pulmonary embolism (PE).<br>For <i>in vitro</i> diagnostic use.  |
| <b>Device Description</b>                        | The D-Dimer HS 500 Latex Reagent is a suspension of polystyrene latex particles of uniform size coated with the F(ab') <sub>2</sub> fragment of a monoclonal antibody highly specific for the D-Dimer domain included in fibrin soluble derivatives. The use of the F(ab') <sub>2</sub> fragment allows a more specific D-Dimer detection avoiding the interference of some endogenous factors like the Rheumatoid Factor. When a plasma containing D-Dimer is mixed with the Latex Reagent and the Reaction Buffer included in the HemosIL D-Dimer HS 500 kit, the coated latex particles agglutinate. The degree of agglutination is directly proportional to the concentration of D-Dimer in the sample and is determined by measuring the decrease of transmitted light caused by the aggregates (turbidimetric immunoassay). |

**Comparison to Predicate:**

This Special 510(k) is being submitted to add general information from peer-reviewed published literature to the Summary and Principle section of the HemosIL D-Dimer HS 500 insert sheet regarding the association of patient age with D-Dimer levels.

The submission meets the criteria for a Special 510(k) based on the following:

- ) No change in indications for use or intended use
- ) No change in operating principle
- ) No change to labeled performance claims, including no change to the assay cut-off
- ) No change to stability claims or to storage instructions
- ) No change to reagent preparation
- ) No change to specimen collection and preparation
- ) No change to formulation or materials
- ) No change to data reduction software
- ) No change to test parameters
- ) No change to calibration
- ) No change to quality controls

Following is a description of the similarities and differences between the currently marketed HemosIL D-Dimer HS 500 (K090264) and HemosIL D-Dimer HS 500 with the insert sheet modifications:

| <i>Similarities</i> |   |                        |
|---------------------|---|------------------------|
| <b>Item</b>         | <b>Predicate (K090264)</b>  | <b>Modified Device</b> |
| Indications for Use | HemosIL D-Dimer HS 500 is an automated latex enhanced immunoassay for the quantitative determination of D-Dimer in human citrated plasma on the ACL TOP® Family and ACL TOP Family 50 Series Systems for use, in conjunction with a clinical pretest probability (PTP) assessment model to exclude venous thromboembolism (VTE) in outpatients suspected of deep venous thrombosis (DVT) and pulmonary embolism (PE).<br>For <i>in vitro</i> diagnostic use | Same                   |
| Analyte             | D-Dimer   | Same                   |
| Methodology         | Latex-enhanced immuoturbidimetric assay   | Same                   |
| Analyzers           | ACL TOP Family and ACL TOP Family 50 Series   | Same                   |
| Sample Type         | Citrated Plasma   | Same                   |
| Cut-off             | 500 ng/mL   | Same                   |
| Linearity           | 215 - 128000 ng/mL  | Same                   |
| Detection Limit     | 203 ng/mL   | Same                   |
| Performance Claims  | No change to labeled performance claims   |                        |

**Comparison to Predicate (Cont.):**

| <i>Differences</i>                                     |  |   |
|--|--|---|
| Item   | Predicate (K090264)  | Modified Device   |
| Insert Sheet   | Current insert language  | Insert revisions in <i>italic</i>   |
| Summary and Principle<br>(Insert Section)              | <p>Elevated levels of D-Dimer are found in clinical conditions such as deep vein thrombosis (DVT), pulmonary embolism (PE) and disseminated intravascular coagulation (DIC).</p> <p>D-Dimer levels also rise during normal pregnancy but very high levels are associated with complications.</p>   | <p>Elevated levels of D-Dimer are found in clinical conditions such as deep vein thrombosis (DVT), pulmonary embolism (PE) and disseminated intravascular coagulation (DIC).</p> <p>D-Dimer levels also rise <i>with age</i>, and during normal pregnancy but very high levels are associated with complications.</p> <p><i>Further, age-adjusted cut-off values for DVT and PE suspicion have been shown to increase specificity of D-Dimer and reduce the number of unnecessary imaging studies in patient populations greater than 50 years.</i></p> <p>NOTE: <i>Bibliography in insert updated with applicable supporting references.</i></p> |
| Limitations/Interfering Substances<br>(Insert Section) | <p>Specimens from patients who have received preparation of mouse monoclonal antibody for diagnosis or therapy may contain human anti-mouse antibody (HAMA). The presence of HAMA may cause an over-estimation of results in immunoassays that utilize mouse monoclonal antibodies. The Reaction Buffer contains a blocking agent against HAMA to minimize this interference on the assay results.</p> | <p>Specimens from patients who have received preparation of mouse monoclonal antibody for diagnosis or therapy may contain human anti-mouse antibody (HAMA). The presence of HAMA may cause an over-estimation of results in immunoassays that utilize mouse monoclonal antibodies. The Reaction Buffer contains a blocking agent against HAMA to minimize this interference on the assay results.</p> <p><i>The performance of this assay has not been validated for use with age-adjusted cut-off values.</i></p>   |

**Conclusion:**

HemosIL D-Dimer HS 500, with the modified Summary and Principle and Limitations/Interfering Substances insert sections, is substantially equivalent to the legally marketed predicate device FDA cleared under K090264.