



Institut Straumann AG
% Jennifer Jackson
Director, Regulatory Affairs
Straumann USA, LLC
60 Minuteman Road
Andover, Massachusetts 01801

November 14, 2017

Re: K173410

Trade/Device Name: n!ce for Planmill
Regulation Number: 21 CFR 872.6660
Regulation Name: Porcelain Powder For Clinical Use
Regulatory Class: Class II
Product Code: EIH
Dated: October 30, 2017
Received: November 1, 2017

Dear Jennifer Jackson:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820);

and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/>) and CDRH Learn (<http://www.fda.gov/Training/CDRHLearn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<http://www.fda.gov/DICE>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Mary S. Runner -S

for Tina Kiang, Ph.D.
Acting Director
Division of Anesthesiology,
General Hospital, Respiratory,
Infection Control, and Dental Devices
Office of Device Evaluation
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

Click or tap here to enter text.

Device Name

n!ce® for Planmill

Indications for Use (Describe)

Once finalized into a suitable design, the n!ce® glass-ceramic blocks are indicated for use as inlays, onlays, veneers, partial crowns and crowns.

Type of Use (Select one or both, as applicable)

☒ Prescription Use (Part 21 CFR 801 Subpart D)

☐ Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

5. 510(k) Summary K173410

Submitter: Straumann USA, LLC (on behalf of Institut Straumann AG)
60 Minuteman Road
Andover, MA 01810
Registration No.: 1222315 Owner/Operator No.: 9005052

Contact Person: Jennifer M. Jackson, MS
Director of Regulatory Affairs and Quality
+1 (978) 747-2509

**Prepared By &
Secondary Contact:** Shokoufeh Khodabandeh
Regulatory Affairs and Compliance Manager
Institut Straumann AG
+41 61 965 1260

Date Prepared: November 10, 2017
Product Code(s): EIH

Device Class: II

**Classification
Panel:** Dental
**Classification
Name:** Porcelain powder for clinical use (21 CFR 872.6660)
Common Name: Porcelain powder for clinical use (21 CFR 872.6660)

Proprietary Name: n!ce® for Planmill
**Predicate
Device(s):** K160262 n!ce Glass Ceramic Blocks; Manufacturer Straumann

**Reference
Device(s):** K170420; n!ce for Armann Girrback; Institut Straumann AG
Device Description: Straumann® n!ce™ glass ceramic is a proprietary lithium disilicate (Li₂O-SiO₂) glass ceramic material intended to be milled to produce prosthetic restorations for natural and endosseous dental implant abutment borne teeth. The material is suitable for use in inlays, onlays, veneers, copings and monolithic crown restorations.

The mandrel bonded to the blocks is compatible with Planmill milling equipment.

Intended Use The n!ce® glass ceramic is intended to be used to manufacture ceramic restorations for the restoration of natural teeth or on top of abutments.

Indications For Use:

Once finalized into a suitable design, the n!ce® glass-ceramic blocks are indicated for use as inlays, onlays, veneers, partial crowns and crowns.

Materials:

lithium aluminosilicate ceramic reinforced with lithium disilicate glass-ceramic material (exact same material as previously cleared under K160262)

Technological Characteristics:

A comparison of the relevant technological characteristics between the subject and the primary predicate is provided in the table that follows.

Feature	Subject Device n!ce Blocks for Planmill	Primary Predicate Devices Straumann n!ce Glass Ceramic Blocks (K160262)	Equivalence discussions
Indication for use	Once finalized into a suitable design, the n!ce® glass-ceramic blocks are indicated for use as inlays, onlays, veneers, partial crowns and crowns.		Identical
Block Chemical Composition	lithium aluminosilicate ceramic reinforced with lithium disilicate glass-ceramic material	lithium disilicate – lithium aluminosilicate glass-ceramic material	Identical Same material with a name change
Crystallization State as Supplied	Fully crystallized		Identical
Esthetic Characteristics	Translucency: High Translucency (HT) Low Translucency (LT) Shades: HT/LT: 6 A-C Color Uniformity: Homogenous Fluorescence: Present		Identical
Block Dimensions	C14 (12.4 x 14.5 x 18.0 mm)		Identical

Feature	Subject Device n!ce Blocks for Planmill	Primary Predicate Devices Straumann n!ce Glass Ceramic Blocks (K160262)	Equivalence discussions
Mandrel Design	The mandrel is compatible with material holders of Planmill mills.	The mandrel is compatible with material holders of Sirona CEREC and inLab mills and other third-party mills.	Equivalent Both the subject and predicate designs can be effectively processed in the mills corresponding with the mandrel design.
Mandrel Material	AlSi1Sn1MgBi Alloy		Identical
Minimum Wall Thickness of Milled Restoration	1.0 mm		Identical
Flexural Strength	Meets ISO 6872 requirements for a Type II, Class 2 dental ceramic material.		Identical
Radioactivity	Meets ISO 6872 requirements for a Type II, Class 2 dental ceramic material		Identical
Chemical Solubility	Meets ISO 6872 requirements for a Type II, Class 2 dental ceramic material.		Identical
Coefficient of Thermal Expansion (CTE) 100- 500°C	HT: $7.1 \times 10^{-6}/K$ LT: $7.2 \times 10^{-6}/K$		Identical
Glass Transition Temperature (T_g)	HT: 497°C LT: 491°C		Identical

Performance Data: Test data to support the evaluation of the subject n!ce® Glass-Ceramic Blocks has been submitted or included by reference as follows:

- Product performance testing per ISO 6872, *Dentistry—Ceramic materials*
- ISO 7991, *Glass—Determination of coefficient of mean linear thermal expansion.*
- Biocompatibility assessment as follows:
 - o Evaluation per ISO 10993-1, *Biological evaluation of medical devices — Part 1: Evaluation and testing within a risk management process.*
 - o Cytotoxicity assessment per ISO 10993-5, *Biological evaluation of medical devices — Part 5: Tests for in vitro cytotoxicity.*
 - o Chemical characterization per ISO 10993-18, *Biological evaluation of medical devices—Part 18: Chemical characterization of materials.*
- Transport and package testing per ISTA 2A and the standards referenced therein.
- Evaluation of shelf life per ASTM F1980, *Standard Guide for Accelerated Aging of Sterile Barrier Systems for Medical Devices.*
- Simulated Use Validation including a machinability assessment
- Mandrel-block adhesion

No animal or human clinical studies were conducted.

Conclusions:

Based upon our assessment of the design and applicable performance data, the subject devices have been determined to be substantially equivalent to the identified predicate devices.