



February 9, 2018

Becton, Dickinson and Company
Monica Giguere
Regulatory Affairs Project Manager
7 Loveton Circle
Sparks, Maryland 21152

Re: K173523

Trade/Device Name: BD Phoenix Automated Microbiology System - GN Meropenem-vaborbactam
(0.125/8-32/8 ug/mL)

Regulation Number: 21 CFR 866.1645

Regulation Name: Fully automated short-term incubation cycle antimicrobial susceptibility system

Regulatory Class: Class II

Product Code: LON

Dated: November 10, 2017

Received: November 14, 2017

Dear Ms. Giguere:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801 and Part 809); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR

Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/>) and CDRH Learn (<http://www.fda.gov/Training/CDRHLearn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<http://www.fda.gov/DICE>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

 Ribhi Shavar -S

For

Uwe Scherf, M. Sc., Ph.D.

Director

Division of Microbiology Devices

Office of In Vitro Diagnostics

and Radiological Health

Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)
K173523

Device Name
BD Phoenix™ Automated Microbiology System - GN Meropenem-vaborbactam (0.125/8-32/8µg/mL)

Indications for Use (Describe)

The BD Phoenix™ Automated Microbiology System is intended for in vitro quantitative determination of antimicrobial susceptibility by minimal inhibitory concentration (MIC) of most Gram-negative aerobic and facultative anaerobic bacteria isolates from pure culture for Enterobacteriaceae and Non-Enterobacteriaceae and most Gram-positive bacteria isolates from pure culture belonging to the genera Staphylococcus, Enterococcus and Streptococcus.

Meropenem-vaborbactam has been shown to be active in vitro against most strains of microorganisms listed below, as described in the FDA-approved package insert for this antimicrobial agent.

Active In Vitro and in Clinical Infections Against

Enterobacter cloacae species complex
Escherichia coli
Klebsiella pneumoniae

Active In Vitro but clinical significance is unknown

Citrobacter freundii
Citrobacter koseri
Enterobacter aerogenes
Klebsiella oxytoca
Morganella morganii
Proteus mirabilis
Providencia spp.
Serratia marcescens

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

“An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number.”

510(k) SUMMARY

- SUBMITTED BY:** Becton, Dickinson and Company
7 Loveton Circle MC 694
Sparks, MD 21152
Phone: 410-316-4827
Fax: 410-316-4499
- CONTACT NAME:** Monica Giguere, RAC
Regulatory Affairs Project Manager
- DATE PREPARED:** November 10, 2017
- DEVICE TRADE NAME:** BD Phoenix™ Automated Microbiology System – GN Meropenem-vaborbactam (0.125/8-32/8 µg/mL)
- DEVICE COMMON NAME:** Antimicrobial susceptibility test system-short incubation
- DEVICE CLASSIFICATION:** 21 CFR 866.1645
Fully Automated Short-Term Incubation Cycle
Antimicrobial Susceptibility System.
(Product Code LON)
- PREDICATE DEVICES:** BD Phoenix™ Automated Microbiology System – GN Tigecycline (0.25-16 µg/mL)
- INTENDED USE:** The BD Phoenix Automated Microbiology System is intended for the *in vitro* rapid identification (ID) and quantitative determination of antimicrobial susceptibility by minimal inhibitory concentration (MIC) of Gram Negative aerobic and facultative anaerobic bacteria belonging to the family *Enterobacteriaceae* and non-*Enterobacteriaceae*.
- DEVICE DESCRIPTION:**
- The BD Phoenix Automated Microbiology System (Phoenix System) is an automated system for the rapid identification (ID) and antimicrobial susceptibility testing (AST) of clinically relevant bacterial isolates. The system includes the following components:
- BD Phoenix instrument and software.
 - BD Phoenix panels containing biochemicals for organism ID testing and antimicrobial agents for AST determinations.
 - BD Phoenix ID Broth used for performing ID tests and preparing AST Broth inoculum.
 - BD Phoenix AST Broth used for performing AST tests only.
 - BD Phoenix AST Indicator solution added to the AST Broth to aid in bacterial growth determination.

The Phoenix panel is a sealed and self-inoculating molded polystyrene tray with 136 micro-wells containing dried reagents. Organisms for susceptibility testing must be a pure culture and preliminarily identified as a Gram-negative or Gram-positive isolate. Phoenix panels are inoculated with a specified organism density and placed into the instrument. Inoculum for use with the Phoenix system may be prepared either manually or may be automated using the BD Phoenix™ AP System.

The Phoenix AST method is a broth based microdilution test. The Phoenix System utilizes a redox indicator for the detection of organism growth in the presence of an antimicrobial agent. Measurements of changes to the indicator as well as bacterial turbidity are used in the determination of bacterial growth. Each AST panel configuration contains several antimicrobial agents with a wide range of two-fold doubling dilution concentrations.

The instrument houses the panels where they are continuously incubated at a nominal temperature of $35^{\circ} \pm 1^{\circ}\text{C}$. The instrument takes readings of the panels every 20 minutes. The readings are interpreted to give an identification of the isolate, minimum inhibitory concentration (MIC) values and category interpretations, S, I, R or N (susceptible, intermediate, resistant or not susceptible).

DEVICE COMPARISON:

The BD Phoenix™ Automated Microbiology System demonstrated substantially equivalent performance for inoculum prepared manually and inoculum prepared with the BD Phoenix™ AP instrument when compared with the CLSI reference broth microdilution method. This premarket notification provides data supporting the use of the BD Phoenix™ Automated Microbiology System Gram negative ID/AST or AST only Phoenix panels with this antimicrobial agent.

SUMMARY OF SUBSTANTIAL EQUIVALENCE TESTING:

The BD Phoenix™ Automated Microbiology System has demonstrated substantially equivalent performance when compared to the CLSI reference broth microdilution method (AST panels prepared according to CLSI M7). The system has been evaluated as defined in the FDA guidance document, “Class II Special Controls Guidance Document: Antimicrobial Susceptibility Test (AST) Systems; Guidance for Industry and FDA”, August 28, 2009. Shelf life (stability data) for the drug is being collected and will be maintained on file at BD as indicated in the guidance document.

Site Reproducibility

Intra- and inter-site reproducibility of this antimicrobial agent in the BD Phoenix System was evaluated at three sites using a panel of Gram-negative isolates. Each site tested the isolates in triplicate on three different days using one lot of Gram Negative Phoenix panels containing this antimicrobial agent and associated reagents. The results of the study demonstrate that for this antimicrobial agent and the Gram-negative organisms tested there was an overall reproducibility across test sites of greater than 95% (+/- 1 dilution) agreement when compared to the test mode.

Clinical Studies

Clinical, stock and challenge isolates were tested across multiple geographically diverse sites across the United States to demonstrate the performance of the Phoenix antimicrobial susceptibility test with a Gram Negative Phoenix Panel containing this antimicrobial agent. Phoenix System results for Challenge set isolates were compared to the expected results. Phoenix System results for clinical isolates were compared to the results obtained from the CLSI reference broth microdilution method.

The performance of the Phoenix System was assessed by calculating Essential Agreement (EA) and Category Agreement (CA) to expected/reference results for all isolates tested. Essential Agreement (EA) occurs when the BD Phoenix™ Automated Microbiology System agrees exactly or within \pm one two-fold dilution to the reference result. Category Agreement (CA) occurs when the BD Phoenix™ Automated Microbiology System agrees with the reference method with respect to the FDA categorical interpretive criteria (susceptible, intermediate, resistant or not susceptible).

The following table summarizes the performance for Clinical and Challenge isolates tested in this study.

Performance of BD Phoenix System for Meropenem-vaborbactam - GN

Antimicrobial	Concentration	EA (n)	EA (%)	CA (n)	CA (%)
Meropenem-vaborbactam	0.125/8-32/8 μ g/mL	1141	98.9	1141	99.7

Conclusions Drawn from Substantial Equivalence Studies

The data collected from the substantial equivalence studies demonstrate that testing on the BD Phoenix™ Automated Microbiology System with this antimicrobial agent is substantially equivalent as outlined in the FDA draft guidance document, “Class II Special Controls Guidance Document: Antimicrobial Susceptibility Test (AST) Systems; Guidance for Industry and FDA”, August 28, 2009. Technological characteristics of this system are substantially equivalent to those used in the VITEK® system, which received approval by the FDA under PMA number N50510.