



July 23, 2018

GBS Commonwealth Co., Ltd.
Jimmy Kim
Regulatory Affairs
#1007-1, WOOLIM Lion's Valley B, 168 Gasan Digital 1-ro
Geumcheon-gu, Seoul
SOUTH KOREA

Re: K173645

Trade/Device Name: JASPER Spinal Fixation System
Regulation Number: 21 CFR 888.3070
Regulation Name: Thoracolumbosacral pedicle screw system
Regulatory Class: Class II
Product Code: NKB
Dated: June 12, 2018
Received: June 18, 2018

Dear Mr. Kim:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820);

and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/>) and CDRH Learn (<http://www.fda.gov/Training/CDRHLearn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<http://www.fda.gov/DICE>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,


Ronald P. Jean -S

for Mark N. Melkerson
Director
Division of Orthopedic Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K173645

Device Name

JASPER Spinal Fixation System

Indications for Use (Describe)

The JASPER Spinal Fixation System is non-cervical spinal fixation devices intended for use as posterior pedicle screw fixation systems (T1-S2/ilium), or as an anterolateral fixation system (T8-L5). All components in the system are limited to skeletally mature patients. These devices are indicated as an adjunct to fusion for all of the following indications regardless of the intended use:

The JASPER Spinal Fixation System is indicated for the following:

- Degenerative Disc Disease (defined as discogenic back pain with degeneration of the disc confirmed by history and radiographic studies)
- Spondylolisthesis
- Trauma (i.e., fracture or dislocation)
- Deformities or curvatures (i.e., scoliosis, kyphosis, and/or lordosis)
- Tumor
- Stenosis
- Pseudoarthrosis
- Failed previous fusion

The JASPER Spinal Fixation System is a pedicle screw system indicated for the treatment of severe spondylolisthesis (Grade 3 and 4) of the L5-S1 vertebra in skeletally mature patients receiving fusion by autogenous bone graft having implants attached to the lumbar and sacral spine (L3 to sacrum) with removal of the implants after the attainment of a solid fusion.

In addition, the JASPER Spinal Fixation System is intended to provide immobilization and stabilization of spinal segments in skeletally mature patients as an adjunct to fusion in the treatment of the following acute and chronic instabilities or deformities of the thoracic, lumbar and sacral spine: degenerative spondylolisthesis with objective evidence of neurological impairment, fracture, dislocation, scoliosis, kyphosis, spinal tumor and failed previous fusion (pseudoarthrosis).

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."



510(k) SUMMARY

The following 510(k) summary is being submitted as required by 21 CFR 807.92(a):

1. Device Identification

Submitter: GBS Commonwealth Co., Ltd.
#1007-1, 168 Gasan Digital 1-ro, Geumcheon-Gu Seoul,
South Korea
Phone. 82-2-6925-4469
e-mail: Jimmy.kim@gbscommonwealth.com

Contact Person: Jimmy Kim

Date prepared Nov, 22, 2017

Trade Name	JASPER Spinal Fixation System
Classification	21 CFR 888.3070 Thoracolumbosacral pedicle screw system, Class II Product Code: NKB

2. Purpose of 510(k)

The GBS Commonwealth Co. Ltd., here by submits this traditional 510(k): for Initial product Introduction of JASPER Spinal Fixation System

3. Predicate or legally marketed devices which are substantially equivalent

- 1) Primary Predicate Device: K120270 VENUS Spinal Fixation System
- 2) Additional Predicate : K162801 CastleLoc Spinal Fixation System

4. Description of the Device

The JASPER Spinal Fixation System is a top-loading multiple component, posterior(thoraco-lumbar) spinal fixation system which consists of screws, hooks, rods, set screws, cross links, rod connectors and iliac connectors for spinal deformity system. The JASPER Spinal Fixation System will allow surgeons to build a spinal implant construct to stabilize and promote spinal fusion.



Materials : All products are made of titanium alloy(Ti-6Al-4V ELI, ASTM F136)and CoCrMo alloy(Cobalt-28Chromium-6Molybdenum, ASTM F1537) approved for medical use.

5. Indication for Use

The JASPER Spinal Fixation System is non-cervical spinal fixation devices intended for use as posterior pedicle screw fixation systems (T1-S2/ilium), or as an anterolateral fixation system (T8-L5). All components in the system are limited to skeletally mature patients. These devices are indicated as an adjunct to fusion for all of the following indications regardless of the intended use:

The JASPER Spinal Fixation System is indicated for the following:

- Degenerative Disc Disease (defined as discogenic back pain with degeneration of the disc confirmed by history and radiographic studies)
- Spondylolisthesis
- Trauma (i.e., fracture or dislocation)
- Deformities or curvatures (i.e., scoliosis, kyphosis, and/or lordosis)
- Tumor
- Stenosis
- Pseudoarthrosis
- Failed previous fusion

The JASPER Spinal Fixation System is a pedicle screw system indicated for the treatment of severe spondylolisthesis (Grade 3 and 4) of the L5-S1 vertebra in skeletally mature patients receiving fusion by autogenous bone graft having implants attached to the lumbar and sacral spine (L3 to sacrum) with removal of the implants after the attainment of a solid fusion.

In addition, the JASPER Spinal Fixation System is intended to provide immobilization and stabilization of spinal segments in skeletally mature patients as an adjunct to fusion in the treatment of the following acute and chronic instabilities or deformities of the thoracic, lumbar and sacral spine: degenerative spondylolisthesis with objective evidence of neurological impairment, fracture, dislocation, scoliosis, kyphosis, spinal tumor and failed previous fusion (pseudoarthrosis).



6. Comparison of the technological characteristics of the subject and predicate devices

The JASPER Spinal Fixation System is considered substantially equivalent to legally marketed devices VENUS Spinal Fixation System and CastleLoc Spinal Fixation System. They are similar in design, material, scientific technologies and indications for use.

7. Performance Testing

Bench testing results demonstrate that JASPER Spinal Fixation System performs equivalently to the predicates in static compression bending, static torsion and fatigue compression bending construct testing (in accordance with ASTM F1717-15) and static and torsional gripping capacity and transverse moment testing (in accordance with ASTM F1798)

8. Conclusion

The JASPER Spinal Fixation System is substantially equivalent to legally marketed predicates.