



April 6, 2018

Intuitive Surgical, Inc.
Ms. Gayle Perry
Senior Regulatory Affairs Engineer
1266 Kifer Road
Sunnyvale, California 94086

Re: K180033

Trade/Device Name: 8mm Monopolar Curved Scissors
Regulation Number: 21 CFR 876.1500
Regulation Name: Endoscope and Accessories
Regulatory Class: Class II
Product Code: NAY
Dated: March 9, 2018
Received: March 12, 2018

Dear Ms. Perry:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820);

and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/>) and CDRH Learn (<http://www.fda.gov/Training/CDRHLearn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<http://www.fda.gov/DICE>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

**Jennifer R.
Stevenson -S3**

For Binita S. Ashar, M.D., M.B.A., F.A.C.S.
Director
Division of Surgical Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K180033

Device Name

8mm Monopolar Curved Scissors

Indications for Use (Describe)

The Intuitive Surgical Endoscopic Instrument Control System is intended to assist in the accurate control of Intuitive Surgical Endoscopic Instruments including rigid endoscopes, blunt and sharp endoscopic dissectors, scissors, scalpels, ultrasonic shears, forceps/pick-ups, needle holders, endoscopic retractors, stabilizers, electrocautery and accessories for endoscopic manipulation of tissue, including grasping, cutting, blunt and sharp dissection, approximation, ligation, electrocautery, suturing, and delivery and placement of microwave and cryogenic ablation probes and accessories, during urologic surgical procedures, general laparoscopic surgical procedures, gynecologic laparoscopic surgical procedures, transoral otolaryngology robotic surgical procedures restricted to benign and malignant tumors classified as T1 and T2, and for benign base of tongue resection procedures, general thoracoscopic surgical procedures, and thoracoscopically assisted cardiotomy procedures. The system can also be employed with adjunctive mediastinotomy to perform coronary anastomosis during cardiac revascularization. The system is indicated for adult and pediatric use (except for transoral otolaryngology robotic surgical procedures). It is intended to be used by trained physicians in an operating room environment in accordance with the representative, specific procedures set forth in the Professional Instructions for Use.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number.

8mm Monopolar Curved Scissors

510(k) Summary

510(k) Owner:	Intuitive Surgical, Inc. 1266 Kifer Road Sunnyvale, CA 94086
Contact:	Gayle Perry Senior Regulatory Affairs Engineer Phone Number: 408-523-7252 Fax Number: 408-523-8907 Email: gayle.perry@intusurg.com
Date Summary Prepared:	January 3, 2018
Trade Name:	8mm Monopolar Curved Scissors
Common Name:	Endoscope and accessories
Classification:	Class II 21 CFR 876.1500, Endoscope and Accessories
Product Codes:	NAY
Classification Advisory Committee:	General and Plastic Surgery
Predicate Device:	K050369 (use with the IS2000 system), with additional indications under K081137 (use with the IS3000 system), K123329 (updated TORS indication) and updated reprocessing instructions under K170644

Device Description

The Intuitive Surgical *EndoWrist* 8mm Monopolar Curved Scissors instrument is used with the Intuitive Surgical IS2000 *da Vinci S* Surgical System or IS3000 *da Vinci Si* Surgical System for cutting, cauterizing, coagulation, manipulating and blunt dissection of tissue. The instrument consists of the housing, shaft, wrist, and tip. The shaft and wrist allow for different axes of rotation, and the instrument tip is used to interact with the patient tissue. This instrument is reusable and is provided non-sterile. The instrument is used with a single use tip cover accessory.

Intended Use/Indications for Use:

EndoWrist Instruments, including scissors, scalpels, forceps, needle drivers and electrocautery are intended for endoscopic manipulation of tissue, including: grasping, cutting, blunt and sharp dissection, approximation, ligation, electrocautery and suturing.

The Intuitive Surgical Endoscopic Instrument Control System is intended to assist in the accurate control of Intuitive Surgical Endoscopic Instruments including rigid endoscopes, blunt and sharp endoscopic dissectors, scissors, scalpels, ultrasonic shears, forceps/pick-ups, needle holders, endoscopic retractors, stabilizers, electrocautery and accessories for endoscopic manipulation of tissue, including grasping, cutting, blunt and sharp dissection, approximation, ligation, electrocautery, suturing, and delivery and placement of microwave and cryogenic ablation probes and accessories, during urologic surgical procedures, general laparoscopic surgical procedures, gynecologic laparoscopic surgical procedures, transoral otolaryngology surgical procedures restricted to benign and malignant tumors classified as T1 and T2, and for benign base of tongue resection procedures, general thoracoscopic surgical procedures, and thoracoscopically assisted cardiomy procedures. The system can also be employed with adjunctive mediastinotomy to perform coronary anastomosis during cardiac revascularization. The system is indicated for adult and pediatric use (except for transoral otolaryngology surgical procedures). It is intended to be used by trained physicians in an operating room environment in accordance with the representative, specific procedures set forth in the Professional Instructions for Use.

Technological Characteristics:

The subject 8mm Monopolar Curved Scissors is very similar to its predicate device originally cleared under K050369 for use with the IS2000 *da Vinci S* System, and subsequently cleared under K081137 for use with the IS3000 *da Vinci Si* System, with additional indications cleared under K123329 and reprocessing instructions under K170644. It has the same intended use, same fundamental scientific technology, and similar technological characteristics as the predicate device. Modifications consist of:

- material changes to the main tube and the extension tube,
- addition of an adhesive between the main tube and extension tube,
- geometrical changes to the main tube/extension tube interface.

Performance Data:

In accordance with the Design Control process, risk analysis was conducted to evaluate the impact of modifications to the predicate device. Design verification and design validation testing were conducted on the subject device to confirm that the design outputs meet design input requirements and that the device is safe and effective for its intended use.

Design Verification:

The bench testing with the subject device was performed on an IS3000 *da Vinci Si* Surgical System. The design verification summarized in this submission verifies mechanical and labeling requirements for the subject instrument, such as:

- instrument reliability and durability,
- leakage,

- roll, pitch, and yaw range of motion,
- jaw close and open positions,
- friction,
- electrical safety,
- instrument labeling,
- compatibility with the tip cover accessory,
- compatibility with system software.

Design Validation:

The design validation summarized in this submission validates functional and interaction (compatibility) requirements for the subject device. Design validation addresses how the features of the instrument meet the user needs and intended use as documented in the product requirements document. The safety and efficacy of the instruments was assessed in representative simulated clinical settings that utilized a porcine model to evaluate applicable requirements through normal and expected worst case clinical use. Representative tissue types were used, as appropriate, for evaluating applicable requirements.

Summary:

Based on the intended use, indications for use, technological characteristics, and performance data, the subject 8mm Monopolar Curved Scissors is substantially equivalent to the predicate device.