



February 12, 2018

Arthrex Inc.
Rebecca R. Homan
Regulatory Affairs Associate
1370 Creekside Boulevard
Naples, Florida 34108

Re: K180118

Trade/Device Name: Arthrex NanoSuture Anchor
Regulation Number: 21 CFR 888.3040
Regulation Name: Smooth or threaded metallic bone fixation fastener
Regulatory Class: Class II
Product Code: MBI
Dated: January 11, 2018
Received: January 16, 2018

Dear Ms. Homan:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820);

and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/>) and CDRH Learn (<http://www.fda.gov/Training/CDRHLearn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<http://www.fda.gov/DICE>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Katherine D. Kavlock -S

for
Mark N. Melkerson
Director
Division of Orthopedic Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

DEPARTMENT OF HEALTH AND HUMAN SERVICES
Food and Drug Administration

Form Approved: OMB No. 0910-0120
Expiration Date: January 31, 2017
See PRA Statement below.

Indications for Use

510(k) Number (if known)

K180118

Device Name

Arthrex NanoSuture Anchor

Indications for Use (Describe)

The Arthrex NanoSuture Anchor is intended to be used for suture or tissue fixation in the foot/ankle, knee, hand/wrist, elbow, and shoulder. Specific indications are listed below:

- Elbow: Biceps Tendon Reattachment, Ulnar or Radial Collateral Ligament Reconstruction
- Shoulder: Rotator Cuff Repair, Bankart Repair, SLAP Lesion Repair, Biceps Tenodesis, Acromio-Clavicular Separation Repair, Deltoid Repair, Capsular Shift or Capsulolabral Reconstruction
- Hand/Wrist: Scapholunate Ligament Reconstruction, Repair/Reconstruction of collateral ligaments, Repair of Flexor and Extensor Tendons at the PIP, DIP and MCP joints for all digits, digital tendon transfers, Carpal Ligament Reconstruction and Carpometacarpal joint arthroplasty (basal thumb joint arthroplasty)
- Foot/Ankle: Lateral Stabilization, Medial Stabilization, Achilles Tendon Repair, Metatarsal Ligament Repair, Hallux Valgus reconstruction, digital tendon transfers, Mid-foot reconstruction
- Knee: Medial Collateral Ligament Repair, Lateral Collateral Ligament Repair, Patellar Tendon Repair, Posterior Oblique Ligament Repair, Iliotibial Band Tenodesis

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary or 510(k) Statement

Date Prepared	February 8, 2018
Submitter	Arthrex Inc. 1370 Creekside Boulevard Naples, FL 34108-1945
Contact Person	Rebecca R. Homan Regulatory Affairs Associate 1-239-643-5553, ext. 73429 rebecca.homan@arthrex.com
Name of Device	Arthrex NanoSuture Anchor
Common Name	Screw, fixation, bone
Product Code	MBI
Classification Name	21 CFR 888.3040: Fastener, Fixation, Nondegradable, Soft Tissue
Regulatory Class	II
Predicate Device	K112237: Arthrex MicroSuture Anchors
Purpose of Submission	This Special 510(k) premarket notification is submitted to obtain clearance for a line extension to the Arthrex MicroSuture Anchors family cleared under K112237.
Device Description	The Arthrex NanoSuture Anchor is a fully threaded suture anchor pre-loaded with Arthrex suture on a disposable inserter. The anchor is manufactured from titanium and measures 1.7 mm in diameter and 5 mm in length.
Indications for Use	<p>The Arthrex NanoSuture Anchor is intended to be used for suture or tissue fixation in the foot/ankle, knee, hand/wrist, elbow, and shoulder. Specific indications are listed below:</p> <p><i>Elbow:</i> Biceps Tendon Reattachment, Ulnar or Radial Collateral Ligament Reconstruction</p> <p><i>Shoulder:</i> Rotator Cuff Repair, Bankart Repair, SLAP Lesion Repair, Biceps Tenodesis, Acromio-Clavicular Separation Repair, Deltoid Repair, Capsular Shift or Capsulolabral Reconstruction</p> <p><i>Hand/Wrist:</i> Scapholunate Ligament Reconstruction, Repair/Reconstruction of collateral ligaments, Repair of Flexor and Extensor Tendons at the PIP, DIP and MCP joints for all digits, digital tendon transfers, Carpal Ligament Reconstruction and Carpometacarpal joint arthroplasty (basal thumb joint arthroplasty)</p> <p><i>Foot/Ankle:</i> Lateral Stabilization, Medial Stabilization, Achilles Tendon Repair, Metatarsal Ligament Repair, Hallux Valgus reconstruction, digital tendon transfers, Mid-foot reconstruction</p> <p><i>Knee:</i> Medial Collateral Ligament Repair, Lateral Collateral Ligament Repair, Patellar Tendon Repair, Posterior Oblique Ligament Repair, Iliotibial Band Tenodesis</p>
Performance Data	<p>Pullout testing was conducted to demonstrate that the proposed Arthrex NanoSuture Anchor perform statistically equivalent to the predicate.</p> <p>Bacterial endotoxin per EP 2.6.14/USP <85> was conducted to demonstrate that the device meets pyrogen limit specifications.</p>
Conclusion	The Arthrex NanoSuture Anchor is substantially equivalent to the predicate device in which the basic design features and intended uses are the same. Any

differences between the proposed device and the predicate device are considered minor and do not raise questions concerning safety or effectiveness.

Based on the indications for use, technological characteristics, and the summary of data submitted, Arthrex Inc. has determined that the proposed device is substantially equivalent to the currently marketed predicate device.
