



March 29, 2018

Corindus, Inc.
Robert Lavado
Manager, Regulatory Affairs
309 Waverley Oaks Road
Waltham, Massachusetts 02452

Re: K180517
Trade/Device Name: CorPath GRX System
Regulation Number: 21 CFR 870.1290
Regulation Name: Steerable Catheter Control System
Regulatory Class: Class II
Product Code: DXX
Dated: February 26, 2018
Received: February 27, 2018

Dear Robert Lavado:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good

manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/>) and CDRH Learn (<http://www.fda.gov/Training/CDRHLearn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<http://www.fda.gov/DICE>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Nicole G. Ibrahim -S

for Bram D. Zuckerman, M.D.
Director
Division of Cardiovascular Devices
Office of Device Evaluation
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K180517

Device Name

CorPath GRX System

Indications for Use (Describe)

The CorPath GRX System is intended for use in the remote delivery and manipulation of guidewires and rapid exchange catheters, and remote manipulation of guide catheters during percutaneous coronary and vascular procedures.

Type of Use (Select one or both, as applicable)

☒ Prescription Use (Part 21 CFR 801 Subpart D)

☐ Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(K) SUMMARY

Submitter's Name and Address:	Corindus, Inc. 309 Waverley Oaks Road Suite 105 Waltham, MA 02452
Establishment Registration Number:	3007822508
Date of Summary:	March 28, 2018
Contact Person:	Robert Lavado, Manager, Regulatory Affairs
Telephone Number:	(508) 653-3335 x211
Fax Number:	(508) 653-3355
Name of the Device:	CorPath GRX System
Common Name:	CorPath GRX System
Regulatory Status and Regulation Number:	Class II 21 CFR 870.1290
Classification Name:	System, Catheter Control, Steerable
Device Classification:	Product Code: DXX: Steerable Catheter Control System.
Indications for Use:	The CorPath GRX System is intended for use in the remote delivery and manipulation of guidewires and rapid exchange catheters, and remote manipulation of guide catheters during percutaneous coronary and vascular procedures.
Identification of the Legally Marketed Device (Predicate Device):	CorPath GRX System Device Class: II Product Code: DXX Regulation Number: 21 CFR 870.1290 510(k) Number: K173288
Device Description:	The CorPath GRX System is intended to allow physicians to deliver and manipulate commercially available guidewires, rapid exchange catheters and guide catheters

during percutaneous coronary and vascular procedures. During the use of the CorPath GRX System, the physician maneuvers interventional devices using intuitive controls under independent angiographic fluoroscopy visual guidance using computer controlled movements while in a seated position away from the radiation source.

The CorPath GRX System is composed of the following two functional sub-units:

1. Bedside Unit – Which consists of the Extended Reach Arm, Robotic Drive and Single-use Cassette
2. Remote Workspace – Which consists of the Control Console, angiographic monitor(s), hemodynamic monitors, X-ray foot pedal, and optional Interventional Cockpit.

Commercially available guidewires, rapid exchange catheters, and guide catheters are loaded into the Single-use Cassette. By using the joysticks or the Control Console touch screen, the physician can control the Robotic Drive to advance, retract, and rotate the guidewire, advance and retract the rapid exchange catheter, and advance, retrace, and rotate the guide catheter. The Robotic Drive and Control Console communicate via a single communication cable.

**Substantial
Equivalence:**

The product subject of this premarket notification is substantially equivalent in design and functionality to the CorPath GRX System (**K173288**, cleared February 15, 2018).

The modified CorPath GRX System and the predicate CorPath GRX System have the same technological characteristics and functionality. The changes are limited to a new bedrail connection design for the Extended Reach Arm component. There have been no changes to the modified CorPath GRX System with respect to packaging, sterilization, or method of action.

Verification/validation testing of the CorPath GRX System has been conducted to demonstrate the modified CorPath GRX Systems is substantially equivalent to the predicate devices. Tests conducted were identified on the basis of risk analysis activities performed to evaluate the impact of the modification on the device/components.

Specifically, the following non-clinical laboratory tests were performed to determine substantial equivalence:

- Performance Testing – Extended Reach Arm
- IEC 60601-1 – Safety Testing

All testing has demonstrated that the device is

substantially equivalent to the predicate devices.

Performance:

The determination of substantial equivalence for this device was based on a detailed device description and non-clinical laboratory testing. The testing demonstrated that the device can be considered substantially equivalent to the predicate devices.

Conclusion:

Based on the bench testing conducted, it is concluded that the CorPath GRX System is substantially equivalent to the predicate device, the CorPath GRX System (**K173288**, cleared February 15, 2018).