



November 8, 2018

IntroMedic Co., Ltd.
Jinyoung Lee
Director
Suite 1105, 1106 E&C Venture Dream Tower 6-Cha,
Guro-dong, Guro-gu
Seoul, 08375
Korea

Re: K180732
Trade/Device Name: MiroCam® Capsule Endoscope System
Regulation Number: 21 CFR§ 876.1300
Regulation Name: Ingestible Telemetric Gastrointestinal Capsule Imaging System
Regulatory Class: II
Product Code: NEZ
Dated: September 19, 2018
Received: September 24, 2018

Dear Jinyoung Lee:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies.

You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/CombinationProducts/GuidanceRegulatoryInformation/ucm597488.htm>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/>) and CDRH Learn (<http://www.fda.gov/Training/CDRHLearn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<http://www.fda.gov/DICE>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Jeffrey W. Cooper -
S

for
Benjamin R. Fisher, Ph.D.
Director
Division of Reproductive, Gastro-Renal,
and Urological Devices
Office of Device Evaluation
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K180732

Device Name

MiroCam® Capsule Endoscope System

Indications for Use (Describe)

The MiroCam® Capsule Endoscope System is intended for visualization of the small bowel mucosa.

- It may be used in the visualization and monitoring of lesions that may indicate Crohn's disease not detected by upper and lower endoscopy.
- It may be used in the visualization and monitoring of lesions that may be a source of obscure bleeding (either overt or occult) not detected by upper and lower endoscopy.
- It may be used in the visualization and monitoring of lesions that may be potential causes of iron deficiency anemia (IDA) not detected by upper and lower endoscopy.

It may be used as a tool in the detection of abnormalities of the small bowel and this device is indicated for adults and children from two years of age.

The Suspected GI Bleeding Indicator (SGIB) is intended to mark frames of the video suspected of containing blood or red areas.

Type of Use (Select one or both, as applicable)

☒ Prescription Use (Part 21 CFR 801 Subpart D)

☐ Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

5. 510(k) Summary

This summary of 510(k) safety and effectiveness information is being submitted in accordance with requirements of 21 CFR Part 807.92.

And the subject device had no prior submissions.

1. Company and Correspondent making the submission

Date prepared	November 6 th , 2018 th
Name	IntroMedic Co., Ltd.
Address	Suite 1105, 1106, E&C Venture Dream Tower 6-Cha, 197-28 Guro-dong, Guro-gu, Seoul, 08375, Korea
Telephone	+82-2-801-9300
Fax	+82-2-801-9330
Contact	Jinyoung Lee
Internet	http://www.intromedic.com

2. Device

Proprietary Name	MiroCam® Capsule Endoscope System
Common Name	Capsule Imaging System
Classification Name	21 CFR 876.1300 (Product Code NEZ) Ingestible telemetric gastrointestinal capsule imaging system
Regulatory Class	II

3. Predicate Device

Manufacturer	IntroMedic Co., Ltd.
Device	MiroCam® Capsule Endoscope System
510(K) Number	K143663
Manufacturer	IntroMedic Co., Ltd.
Device	MiroCam® Capsule Endoscope System
510(K) Number	K170438

Manufacturer	Given Imaging Ltd.
Device	Given PillCam SB3 Capsule endoscopy system
510(K) Number	K123864 (Reference device)

4. Description

The MiroCam® Capsule Endoscope System is comprised of the following core components:

- MiroCam® Capsule: MC1600-B
- MiroCam® Receiver: MR2000
- Data Cable / Data Belt (Optional): MR1000-D / MR1000-D(S) or MR1000-D(M) or MR1000-D(L)
- MiroView™ Software: MiroView™ U 4.0

The general usage workflow of the MiroCam® system is as follows, the MiroCam® capsule captures images of the GI tract, which are sent via Human Body Communication to sensor pads which are affixed to the patient body. The sensor pads are connected to the receiver by the data cables. The image data is stored on the receiver for the duration of the patient procedure. After removing the receiver set from the patient body, the receiver is connected via USB to the Commercial PC and the image data is uploaded. Following upload, the physician (Gastroenterologist) reviews the patient image data for suspected abnormalities of the small bowel. The key system components are explained in detail below.

MiroCam® Capsule

Functionally, the MiroCam® capsule captures imaging via a CMOS imaging sensor for at least 12 hours at the rate of 6 images per second. Twelve white LEDs flash in concert with the imaging sensor. A sensor PCB links the imaging sensor to the two gold electrodes of the capsule, from which a weak current is emitted containing the image data via Human Body Communication technology. The weak current is then passed through the bodily tissue and fluid to be picked up by the sensor pads. Human Body Communications is uni-directional from the capsule to the sensor pads.

Physically, the capsule is 24.5mm in length, and 10.8mm in diameter. The exterior of the capsule is composed of biocompatible materials, capable of withstanding potential bite forces and exposure to fluids in the GI tract.

MiroCam® Receiver

The image data from the capsule is saved to the MiroCam® Receiver via the sensor pads and data cables. The sensor pads are standard ECG type sensor pads (3M Red Dot), which are affixed to the patient's abdomen according to the location guide. The 9 sensor pads are attached to the 9 leads of the data cable,

and the data cable is subsequently attached to the Receiver. The Receiver sits in a shoulder pouch, which is put on the patient for the duration of the procedure. A waist pouch is used to organize the data cables. The Receiver can store images for up to 12 hours. Using LCD display it shows icons that indicate the state of receiver, real time image view and preview. User can capture image for the real time images are displayed. Three LEDs also represent the receiver state by switching its power. And MiroCam® receiver play some sounds for notifying to user. To reduce the battery consumption of MiroCam® receiver backlight brightness is controlled. User can control MiroCam® receiver by using three buttons.

The dimensions of the receiver are 145mm in height, 82.5mm in width, and 29.8mm in length. The units weight 330 grams, which is data cable. The battery is lithium ion, and is unique to the Receiver. The battery is rechargeable, and can be charged by placing in the battery charger provided with the Receiver. The receiver uses a storage flash for storing the images.

MiroView™ U 4.0

The MiroView™ U 4.0 helps the user to upload the image data from the Receiver and review the image data with displaying.

MiroView™ U 4.0 will be deployed as three types of software: MiroView™ Server, MiroView™ Operator and MiroView™ Client.

Server PC includes two types of S/W: MiroView™ Server and Operator.

MiroView™ Server and MiroView™ Operator will be deployed on a Server PC. MiroView™ Client S/W will be deployed on reviewer's PCs. 10 Client S/Ws will be able to connect to one Server PC via Network (LAN) at a time.

MiroView™ Server has functionalities: Data Storage, PACS connectivity.

MiroView™ Operator has functionalities: Receiver Management, Case Management and Account Management.

MiroView™ Client has functionalities: List Mode (Case Data Management), Review Mode (Video Display, Image Capture, Annotation Box, etc.), Report Mode (User Comment, Auto Disease Name Check, etc), Export Mode (JPG, AVI, MiroView™ file format)

MiroView™ U 4.0 provides some kinds of Modes and Views to help the user review patient video data quickly and conveniently; however, IntroMedic, Co., Ltd. recommends that the doctor review the patient

video in the Normal Mode / Single View to prevent from a wrong diagnosis.

MiroView™ U 4.0 is a new version of MiroView™ U 3.0, which has new updated features from the MiroView™ U 3.0. The features are described below:

- Compatibility with New Receiver MR2000: MiroView™ U Operator 4.0
- UI Function for Reviewing Patient Data of Dual Tip Capsule: MiroView™ U Client 4.0
- Support of H.264 Codec: MiroView™ U Common 4.0
- Express Play ver 3.0: MiroView™ U Client 4.0
- Supplementary Play: MiroView™ U Client 4.0

5. Indications for Use

The MiroCam® Capsule Endoscope System is intended for visualization of the small bowel mucosa.

- It may be used in the visualization and monitoring of lesions that may indicate Crohn's disease not detected by upper and lower endoscopy.
- It may be used in the visualization and monitoring of lesions that may be a source of obscure bleeding (either overt or occult) not detected by upper and lower endoscopy.
- It may be used in the visualization and monitoring of lesions that may be potential causes of iron deficiency anemia (IDA) not detected by upper and lower endoscopy.

It may be used as a tool in the detection of abnormalities of the small bowel and this device is indicated for adults and children from two years of age.

The Suspected GI Bleeding Indicator (SGIB) is intended to mark frames of the video suspected of containing blood or red areas.

6. Technological Characteristics and Substantial Equivalence

The characteristics of the IntroMedic Co., Ltd. MiroCam® Capsule Endoscope System are substantially equivalent to the following current legally marketed predicate devices based on indications for use, typical clinical use, and operational and fundamental technological characteristics

- | | |
|---|---------|
| ● MiroCam® Capsule Endoscope System by IntroMedic Co., Ltd. | K143663 |
| ● MiroCam® Capsule Endoscope System by IntroMedic Co., Ltd. | K170438 |
| ● Given PillCam® SB3 Capsule endoscopy system by Given Imaging Ltd. | K123864 |
| (Reference device) | |

Table 5-1. Side-by-Side Comparison of the MiroCam® Capsule Endoscope System
with the Predicate Device

Type	Characteristics	MiroCam® Capsule Endoscope System Subject device	MiroCam® Capsule Endoscope System K143663	MiroCam® Capsule Endoscope System K170438	Given® PillCam® SB3 Capsule endoscopy system K123864 (Reference device)
Capsul	Indications for Use	The MiroCam® Capsule Endoscope System is intended for visualization of the small bowel mucosa. - It may be used in the visualization and monitoring of lesions that may indicate Crohn's disease not detected by upper and lower endoscopy. - It may be used in the visualization and monitoring of lesions that may be a source of obscure bleeding (either overt or occult) not detected by upper and lower endoscopy. - It may be used in the visualization and monitoring of lesions that may be potential causes of iron deficiency anemia (IDA) not detected by upper and lower endoscopy. It may be used as a tool in the detection of abnormalities of the small bowel and this device is indicated for adults and children from two years of age. The Suspected GI Bleeding Indicator (SGIB) is intended to mark frames of the video suspected of containing blood or red areas.	MiroCam® Capsule Endoscope System is intended for visualization of the small bowel mucosa. It may be used as a tool in the detection of abnormalities of the small bowel in adults and children from two years of age. The Suspected GI Bleeding Indicator (SGIB) is intended to mark frames of the video suspected of containing blood or red areas.	The MiroCam® Capsule Endoscope System is intended for visualization of the small bowel mucosa. - It may be used in the visualization and monitoring of lesions that may indicate Crohn's disease not detected by upper and lower endoscopy. - It may be used in the visualization and monitoring of lesions that may be a source of obscure bleeding (either overt or occult) not detected by upper and lower endoscopy. - It may be used in the visualization and monitoring of lesions that may be potential causes of iron deficiency anemia (IDA) not detected by upper and lower endoscopy. It may be used as a tool in the detection of abnormalities of the small bowel and this device is indicated for adults only. The Suspected GI Bleeding Indicator (SGIB) is intended to mark frames of the video suspected of containing blood or red areas.	The PillCam SB capsule is intended for visualization of the small bowel mucosa. ● It may be used in the visualization and monitoring of lesions that may indicate Crohn's disease not detected by upper and lower endoscopy. ● It may be used in the visualization and monitoring of lesions that may be a source of obscure bleeding (either overt or occult) not detected by upper and lower endoscopy. ● It may be used in the visualization and monitoring of lesions that may be potential causes of iron deficiency anemia (IDA) not detected by upper and lower endoscopy. The suspected Blood Indicator (SBI) feature is intended to mark frames of the video suspected of containing blood or red areas. The PillCam SB capsule may be used as a tool in the detection of abnormalities of the small bowel and is intended for use in adults and children from two years of age.
	Manufacturer	IntroMedic Co., Ltd.	IntroMedic Co., Ltd.	IntroMedic Co., Ltd.	Given Imaging Ltd.
	Model Name	MC1600-B	MC1200-B	MC2000-B	PillCam SB3

e	Size	Length: 24.5mm Diameter: 10.8mm	Length: 24.5mm Diameter: 10.8mm	Length: 30.1mm Diameter: 10.8mm	Length: 26.2mm Diameter: 11.4mm
	Weight	3.25g±0.1g	3.25g	3.5g±0.1g	3.0±0.1g
	Material	Human Compliance Plastic	Human Compliance Plastic	Human Compliance Plastic	Biocompatible plastic
	Dome Material	COP (Cyclo Olefin Polymers)	COP (Cyclo Olefin Polymers)	COP (Cyclo Olefin Polymers)	Polycarbonate
	Light	6 White LEDs	6 White LEDs	12 White LEDs (each 6 White LEDs)	4 White LEDs
	LED Size	2.0mm(L)*1.25mm(W)* 0.8mm(H)	2.0mm(L)*1.25mm(W)* 0.8mm(H)	2.0mm(L)*1.25mm(W)* 0.8mm(H)	
	LED Viewing Angle	130°	130°	130°	
	Exposure Time	0.5 ~ 24 ms	0.5~24 ms	0.5 ~ 24 ms	
	Field of View	170°	170°	170°	156°
	Image Sensor	CMOS	CMOS	CMOS	CMOS
	Depth of Field	Length: 3cm	Length: 3cm	Length: 3cm	Length: 3cm
	Enlargement Ratio	1:8	1:8	1:8	
	Detectable Range	Under 0.1mm	Under 0.1mm	Under 0.1mm	At least 0.07mm
	Sampling Ratio	6 frames per second	3 frames per second	6 frames per second (each 3 frames per head)	2 or 2~6 frames per second
	Pixel size of Sensor	6.0 μm*6.0 μm	6.0 μm*6.0 μm	6.0 μm*6.0 μm	
	Working Time	12 Hours	12 Hours	12 Hours	Over 8 Hours
	Chemical Safety	Safe in pH=2 ~ pH=8	Safe in pH=2 ~ pH=8	Safe in pH=2 ~ pH=8	Safe in pH=2 ~ pH=8
	Battery Type	Silver Oxide Cell	Silver Oxide Cell	Silver Oxide Cell	Silver Oxide Cell
	Optimum working distance	0mm	0mm	0mm	
	Direction of View	0°	0°	0°	
	Resolution	320x320	320x320	320x320	
	Distortion (%)	-14.5	-14.5	-14.5	
	Lens Uniformity (%)	79.39	79.39	79.39	
	Shelf life	18 months from date manufacture	18 months from date manufacture	12 months from date manufacture	
	Operation Temp.	20 ~ 40 °C	20 ~ 40 °C	20 ~ 40 °C	20 ~ 40 °C
	Storage Temp.	0 ~ 50 °C	0 ~ 50 °C	0 ~ 50 °C	0 ~ 25 °C
	Transmission Methods	HBC (Human Body Communication)	HBC (Human Body Communication)	HBC (Human Body Communication)	RF
Receiver	Model Name	MR2000	MR1100	MR2000	DR3
	Recording Time	12 Hours	12 Hours	12 Hours	Up to 15 hours at LCD off
	Weight	330g (Include Battery)	350g (Include Battery)	330g (Include Battery)	500g (Include Battery Pack)
	Operation Voltage	3.65Vdc, 0.45A	3.7Vdc, 0.45A	3.65Vdc, 0.45A	3.5~4.2Vdc, 0.15~0.5A
	Battery Type	Lithium Ion Battery, 3.65V, 10,500mA	Lithium Ion Battery, 3.7V, 10,400mA	Lithium Ion Battery, 3.65V, 10,500mA	Lithium Ion Battery, 3.8V, 8,800mAh
	Operation Temp.	0 ~ 40 °C	0 ~ 40 °C	0 ~ 40 °C	0 ~ 40 °C
	Storage Temp.	0 ~ 55 °C	0 ~ 55 °C	0 ~ 55 °C	0 ~ 55 °C
	Category	Internal power supply, Type BF	Internal power supply, Type BF	Internal power supply, Type BF	Internal power supply, Type BF
	Real Time View	LCD	USB, Wifi	LCD	LCD
	Method of Data Communication	Data Cable or Data Belt (Optional)	Data Cable or Data Belt (Optional)	Data Cable or Data Belt (Optional)	Sensor Array or Sensor Belt

	Characteristics	IntroMedic Co., Ltd.	IntroMedic Co., Ltd.	IntroMedic Co., Ltd.	Given Imaging Ltd.
Software	Software	MiroView™ U 4.0	MiroView™ U 3.0 MiroView™ RTV (Real Time Viewer Application-only USB) MiroView™ RTV-I (Real Time Viewer Application-only Wifi)	MiroView™ U 4.0	RAPID 8.0
	Language	English, Spanish, Portuguese, Danish, Chinese, French, Italian, Korean, Dutch, Russian, Swedish	English, Spanish, Portuguese, Danish, Chinese, French, Italian, Korean, Dutch, Russian, Swedish	English, Spanish, Portuguese, Danish, Chinese, French, Italian, Korean, Dutch, Russian, Swedish	English, French, German, Italian, Spanish, Portuguese, Dutch, Swedish, Finish, Danish, Chinese-Mandarin, Korean, Russian, Greek
	Data Export	JPEG Image, AVI Video Clip, PDF Report Data, ODF (Open Data Format), EXMIF (Exchangeable Medical Information Format)	JPEG Image, AVI Video Clip, PDF Report Data, ODF (Open Data Format), EXMIF (Exchangeable Medical Information Format)	JPEG Image, AVI Video Clip, PDF Report Data, ODF (Open Data Format), EXMIF (Exchangeable Medical Information Format)	JPEG Images, (MPEG) Video clips, grml (Given proprietary) files, PDF Reports, generic XML-format, Capsule Endoscopy report data
	Data Display	Single and Multi Image, Time Bar, Diagnosis Data, Dual Bar	Single and Multi Image, Time Bar, Diagnosis Data	Single and Multi Image, Time Bar, Diagnosis Data, Dual Bar	Single and Multi Image, Time Bar, Colorbar with region specific color and other diagnostic data
	Event Marker	Small Image with Explanation	Small Image with Explanation	Small Image with Explanation	Annotated thumbnails
	Display Ratio	3~10,000fps	3~10,000fps	3~10,000fps	5~80fps
	Display Mode	Single View, Dual View, Quad View, Range View, Map View	Single View, Dual View, Quad View, Range View, Map View	Single View, Dual View, Quad View, Range View, Map View	Single View, Dual View, Quad View, Mosaic View, Dual Head View
	Running Mode	Normal Mode, Express, SGIB	Normal Mode, Express, SGIB	Normal Mode, Express, SGIB	Normal, Auto, Quick View, SBI

For the predicate device and the subject device, this has all the same performances except Receiver and Software.

[Capsule]

The Subject device and the predicate device which is MC1200-B is identical in every aspect such optical system, raw material, mechanical design, working principle, usage and etc. except the frame rates. For the frame rates of MC1200-B and MC1600-B, it is 3 frames per second and 6 frames per second respectively. What the frame rates is different is that this is not the hardware change, only the frame rate can be adjusted by changing the [0] bit value of address 0x04 in the EEPROM of the capsule as shown in the table below during the manufacturing process.

ADDRESS (HEX)	REGISTER NAME	R/W	DEFAULT (HEX)	DESCRIPTION
0x00	PRODUCT ID0	R	0x52	Product ID0
0x01	PRODUCT ID1	R	0x01	Product ID1
0x02	PRODUCT ID2	R	0x01	Product ID2
0x03	PRODUCT ID3	R	0x01	Product ID3
0x04	CONTROL_FLAG	R/W	0x40	[7] : convolution mode [6:5] : tx mode 00 -> Direct 01 -> BPSK 10 -> DBPSK [4] : sel_tx [3] : Dual [2] : AE disable [1] : EFIC enable [0] : 2X mode

How to change the value is as follows;

0100 0000
 4 0

For 3 frames per second that means MC1200-B, set the [0] bit value indicated in red to “0” as above.

0100 0001
 4 0

For 6 frames per second that means MC1600-B, set the [0] bit value indicated in red to “1” as above.

And, MC1200-B and MC1600-B are indicated for adults and children from two years of age since the two products is identical in size, which it means that the ease of device ingestion is proven.

In addition, for the other predicate device which is MC2000-B, it is identical to MC1600-B in “Indication for Use” because two products have the same frame rates as 6 frames for second.

[Receiver]

The subject device same as the predicate device MR2000 (K170438).

[Software]

The subject device same as the predicate device MiroView U 4.0 (K170438).

7. Performance Testing

The MiroCam® Capsule Endoscope System performance testing was done for the submission of predicate device but it was not included when Sensor Bench testing, MTF Measurement testing (including the predicate device, K143663), Minimum Distinguishable Contrast Value and Depth of Field testing (including the predicate device, K143663), Geometric distortion testing (including the predicate device, K143663), Field of View testing (including the predicate device, K143663), Color Reproduction testing (including the predicate device, K143663), Testing optical spectra of the light source at different time points during the battery life period (including the predicate device, K143663), Testing intensity change during the shelf life period (including the predicate device, K143663), Bite testing, Water proof testing, pH resistance testing, and Shelf life testing (Operation time, Package integrity, pH resistance and Bite) was done. The results of this performance testing conclude that the material and technological characteristics have not diminished the safety and effectiveness of the IntroMedic Co., Ltd. MiroCam® Capsule Endoscope System device when compared to the predicate device.

- Summary of Clinical Information (MiroView™ Express 3.0)

Saurin JC, et al. Endoscopy International Open 2018; 06: E616-621 “Multicenter prospective evaluation of the express view reading mode for small-bowel capsule endoscopy studies”

- 1) Patients: 83 patients with obscure gastrointestinal bleeding
- 2) Center: 10 in France
- 3) Method: Films were read in standard mode, then a second reading was performed in express view mode at a second center. For each lesion, the precise location, nature, and relevance were collected. A consensus reading and review were done by three experts, and considered to be the gold standard.
- 4) Results: The mean reading time of capsule films was 39.7 minutes (11-180 minutes) and 19.7 minutes (4-40 minutes) by standard and express view mode, respectively ($P < 1 \times 10^{-4}$). The consensus review identified a significant lesion in 44/83 patients (53.0%). Standard reading and express view reading had a 93.3% and 82.2% sensitivity, respectively (NS). Consensus review identified 70 significant images from which standard reading and express view reading detected 58 (82.9%) and 55 (78.6%), respectively. The informatics algorithm detected 66/70 images (94.3%) thus missing four small-bowel angiodysplasia.
- 5) Conclusion: The express view algorithm allows an important shortening of Intromedic capsule film reading time with a high sensitivity.

8. Conclusion

Based on Indications for Use, Substantial Equivalence, Performance testing of the devices, IntroMedic Co., Ltd. believes that the MiroCam® Capsule Endoscope System and the predicate devices selected are substantially equivalent and do not raise new issues of safety or effectiveness.
