



May 18, 2018

Liofilchem s. r. l.
% Laura Koeth
President
Laboratory Specialists, Inc.
26214 Center Ridge Road
Westlake, Ohio 44145

Re: K180886

Trade/Device Name: Liofilchem MIC Test Strip (MTS), Delafloxacin 0.002 - 32 µg/mL
Regulation Number: 21 CFR 866.1640
Regulation Name: Antimicrobial susceptibility test powder
Regulatory Class: Class II
Product Code: JWY
Dated: April 3, 2018
Received: April 4, 2018

Dear Laura Koeth:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801 and Part 809); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR

Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/>) and CDRH Learn (<http://www.fda.gov/Training/CDRHLearn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<http://www.fda.gov/DICE>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

 Ribhi Shawar -S

For

Uwe Scherf, M. Sc., Ph.D.

Director

Division of Microbiology Devices

Office of In Vitro Diagnostics

and Radiological Health

Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K180886

Device Name

Liofilchem MIC Test Strip (MTS), Delafloxacin 0.002-32 µg/mL

Indications for Use (Describe)

The Liofilchem® MIC Test Strip (MTS) is a quantitative method intended for the in vitro determination of antimicrobial susceptibility of bacteria. MTS consists of specialized paper impregnated with a pre-defined concentration gradient of an antimicrobial agent, which is used to determine the minimum inhibitory concentration (MIC) in µg/mL of antimicrobial agents against bacteria as tested on agar media using overnight incubation and manual reading procedures.

The Delafloxacin MTS at concentrations of 0.002-32 µg/mL should be interpreted at 16-20 hours of incubation.

Delafloxacin has been shown to be active both clinically and in vitro against the non-fastidious bacteria listed below according to the FDA drug label:

Gram-negative bacteria

Escherichia coli

Klebsiella pneumoniae

Enterobacter cloacae

Pseudomonas aeruginosa

Delafloxacin has been shown to be active in vitro only against the non-fastidious bacteria listed below according to the FDA drug label:

Klebsiella (Enterobacter) aerogenes

Klebsiella oxytoca

Proteus mirabilis

Type of Use (Select one or both, as applicable)

☒ Prescription Use (Part 21 CFR 801 Subpart D)

☐ Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASTaff@fda.hhs.gov

"Afl agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."