



September 18, 2018

Ethicon Endo-Surgery, LLC
% Rubina Dosani, M.S.
Sr. Regulatory Affairs Program Lead
4545 Creek Road
Cincinnati, Ohio 45242

Re: K181653

Trade/Device Name: Ethicon™ Circular Staplers, Ethicon™ XL Circular Staplers
Regulation Number: 21 CFR 878.4750
Regulation Name: Implantable Staple
Regulatory Class: Class II
Product Code: GDW
Dated: June 21, 2018
Received: June 22, 2018

Dear Ms. Dosani:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for

devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/CombinationProducts/GuidanceRegulatoryInformation/ucm597488.htm>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/>) and CDRH Learn (<http://www.fda.gov/Training/CDRHLearn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<http://www.fda.gov/DICE>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

David Krause -S

for Binita S. Ashar, M.D., M.B.A., F.A.C.S.
Director
Division of Surgical Devices
Office of Device Evaluation
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K181653

Device Name

Ethicon™ Circular Staplers, Ethicon™ XL Circular Staplers

Indications for Use (Describe)

The Ethicon™ Circular Staplers have applications throughout the alimentary tract for end-to-end, end-to side, and side-to-side anastomoses.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary

Submitter Information: Ethicon Endo-Surgery, LLC
475 Calle Street
Guaynabo, PR 00969

Application Correspondent

Rubina Dosani
Sr. Regulatory Affairs Program Lead
Ethicon Endo-Surgery, LLC.
Telephone: (513) 337-3566
Fax: (513) 337-2314
Email: rdosani@its.jnj.com

Date Prepared June 21, 2018

Device Trade Name: Ethicon™ Circular Staplers,
Ethicon™ Circular Staplers XL Sealed
Device Common Name: Circular Stapler
Classification Regulation: 21 CFR 878.4750; Implantable Staple
Device Class: II
Panel: 79, General & Plastic Surgery
Classification (Product) Code: GDW

Legally Marketed Predicate Device: Ethicon Endo-Surgery® Curved Intraluminal
Staplers (cleared under K983536)

Device Description

The Ethicon™ Circular Staplers are sterile, single use devices that simultaneously staple and cut tissue to create an anastomosis. The devices deliver 2 concentric rows of staples on the outside of the cut line. The Ethicon™ Circular Staplers are available in two device configurations: CDH and ECS and two shaft lengths; a standard 26 cm shaft and an XL 37 cm shaft. Each configuration is available in 4 endeffector sizes: 21 mm, 25 mm, 29 mm, 33 mm. These configurations function in the same manner - to compress tissue and to produce the same closed staple on tissue. They differ primarily in the length of the curved shaft and the shaft's ability to maintain a pneumatic seal specifically for use in laparoscopic surgical procedures. Each device has a detachable anvil that allows a surgeon to place the anvil in the desired location.

Indications for Use

The Ethicon™ Circular Staplers have applications throughout the alimentary tract for end-to-end, end-to side, and side-to-side anastomoses. The predicate device also has the same indications for use.

Technological Characteristics

The Ethicon™ Circular Stapler is substantially equivalent to the predicate Ethicon Endo-Surgery® Curved Intraluminal Staplers with respect to the device function and design. The subject device has a similar design as the predicate with the addition of shorter staple height range and lower force to fire. Similar to the predicate, a rotatable adjustment knob enables the compression of tissue and selection of a target staple height based on the tissue compression within the green zone. The device is manually powered; it is not powered by an outside energy source. Neither the subject device or predicate device uses software.

Performance Data

Performance data demonstrate that the subject device is substantially equivalent to the predicate device and the differences between the devices were found not to affect safety or performance. The following bench and animal tests were performed to demonstrate substantial equivalence to the predicate:

- staple line integrity and staple form quality equivalency
- leak onset pressure equivalency
- force to fire
- formed staple height
- biocompatibility
- acute hemostasis evaluation
- device removal acceptability
- healing of stapled anastomosis.

The conclusions of the testing demonstrate that the subject device performs substantially equivalent to the predicate device and does not raise new questions of safety and effectiveness.

This submission does not include data from Clinical Studies.