



October 3, 2018

Mighty Oak Medical Inc.
Mr. Mark A. Wylie
Vice President of Quality and Regulatory
750 West Hampden Avenue, Suite 120
Englewood, Colorado 80110

Re: K181883

Trade/Device Name: FIREFLY® Midline Navigation Guide
Regulation Number: 21 CFR 888.3070
Regulation Name: Thoracolumbosacral pedicle screw system
Regulatory Class: Class II
Product Code: PQC
Dated: July 11, 2018
Received: July 13, 2018

Dear Mr. Wylie:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's

requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/CombinationProducts/GuidanceRegulatoryInformation/ucm597488.htm>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/>) and CDRH Learn (<http://www.fda.gov/Training/CDRHLearn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<http://www.fda.gov/DICE>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Ronald P. Jean -S

for Mark N. Melkerson
Director
Division of Orthopedic Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

DEPARTMENT OF HEALTH AND HUMAN SERVICES

Food and Drug Administration

Indications for Use

Form Approved: OMB No. 0910-0120

Expiration Date: January 31, 2017

See PRA Statement below.

510(k) Number (if known)

K181883

Device Name

FIREFLY® Midline Navigation Guide

Indications for Use (Describe)

The FIREFLY® Midline Navigation Guide can be used with any 510(k) cleared, legally marketed, pedicle screw spinal system (for its cleared indications for use) and its respective compatible components for non-cervical open, posterior spinal fixation procedures (T10-S2) intended for fusion, with the additional conditions listed below:

- Pedicle screw's shank is straight along its longitudinal axis (i.e. not curved)
- Pedicle screw's major and minor thread diameters are centered about the longitudinal axis
- Pedicle screw's longitudinal axis matches the direction of insertion
- Pedicle screw is intended to be inserted into a pilot hole
- Pedicle screw's diameter is larger than the pilot hole created with FIREFLY®
- Patient's pedicle must be dimensionally adequate to accommodate a pedicle screw, as determined on preoperative scan
- Compatible pedicle screw spinal system instruments may be used with the FIREFLY® Midline Screw Navigation Guide
 - Pedicle sounding probes (a.k.a. feeler/ball-tip probes) may be used with the FIREFLY® Midline Navigation Guide to confirm pedicle integrity
 - All other pedicle screw spinal system components and accessories (including non-guided taps) are to be used, after removal of the FIREFLY® Midline Navigation Guide, as directed by the pedicle screw spinal system's instructions for use

This device is intended for single use only.

Type of Use (Select one or both, as applicable)



Prescription Use (Part 21 CFR 801 Subpart D)



Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer Paperwork
Reduction Act (PRA) Staff
PRAStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(K) SUMMARY



FIREFLY® Midline Navigation Guide

Submitter:

Mighty Oak Medical
750 W. Hampden Ave., Suite 120
Englewood, CO 80110
(720) 398-9703

Contact: Mark A. Wylie, VP of Quality and Regulatory

Date Prepared: 11July 2018

Device

Trade Name: FIREFLY® Midline Navigation Guide

Common Name: Pedicle Screw Placement Guide

Device Classification: Class II

Regulation, Name: 21 CFR 888.3070, Pedicle screw spinal system

Device Product Code: PQC

Predicate Device(s):

FIREFLY® Midline Navigation Guide

Primary Predicate:

K162419 (S.E. 10/28/2016)

Description

The FIREFLY® Midline Navigation Guide is intended to assist in the accurate placement of pedicle screws. It consists of single-use components designed for treatment of a specific patient.

The FIREFLY® Midline Navigation Guide uses Patient-Specific Guides that fit on the patient's anatomy to guide surgical instruments in line with trajectories chosen presurgically, by the surgeon, based on the patient's CT imaging data. Navigation Guides are intended to guide instruments to create pilot holes in the pedicles for placing pedicle screws following the Approved Surgical Plan.

Patient-Specific Bone Models are also provided.

The primary purpose of the 510(k) is to receive clearance from the Food and Drug Administration ("FDA" or "the Agency") regarding the above requests.

The FIREFLY® Midline Navigation Guide design is the similar to the originally cleared FIREFLY® Pedicle Screw Navigation Guide in **K143222** and **K162419**.

Indications for Use

The FIREFLY® Midline Navigation Guide can be used with any 510(k) cleared, legally marketed, pedicle screw spinal system (for its cleared indications for use) and its respective compatible components for non-cervical open, posterior spinal fixation procedures (T10-S2) intended for fusion, with the additional conditions listed below:

- *Pedicle screw's shank is straight along its longitudinal axis (i.e. not curved)*
- *Pedicle screw's major and minor thread diameters are centered about the longitudinal axis*
- *Pedicle screw's longitudinal axis matches the direction of insertion*
- *Pedicle screw is intended to be inserted into a pilot hole*
- *Pedicle screw's diameter is larger than the pilot hole created with FIREFLY®*
- *Patient's pedicle must be dimensionally adequate to accommodate a pedicle screw, as determined on preoperative scan*
- *Compatible pedicle screw spinal system instruments may be used with the FIREFLY® Midline Screw Navigation Guide*
 - o *Pedicle sounding probes (a.k.a. feeler/ball-tip probes) may be used with the FIREFLY® Midline Navigation Guide to confirm pedicle integrity*
 - o *All other pedicle screw spinal system components and accessories (including non-guided taps) are to be used, after removal of the FIREFLY® Midline Navigation Guide, as directed by the pedicle screw spinal system's instructions for use*

This device is intended for single use only.

Materials

The patient-contacting components of the FIREFLY® Midline Navigation Guide are manufactured from epoxy resin (Accura ABS White SL 7810).

Performance Data

Cadaveric accuracy testing of the FIREFLY® Midline Navigation Guide was performed. The results demonstrated that the acceptance criteria were met and that the FIREFLY® Midline Navigation Guide's performance is adequate to perform as intended.

Technological Characteristics

The subject FIREFLY® Midline Navigation Guide possesses the same technological characteristics as the predicate device. These include:

- Performance
- Manufacturing process
- Sterilization
- Biocompatible materials
- Basic design

Technological characteristics which are different have been supported with descriptive information and/or performance data. Therefore the fundamental scientific technology of FIREFLY® is the same as the previously cleared device.

Conclusion

The FIREFLY® Midline Navigation Guide possesses the same intended use and technological characteristics as the predicate device. Therefore the FIREFLY® Midline Navigation Guide is substantially equivalent for its intended use.