



November 1, 2018

Bayer Medical Care Inc.
Katie Dillon
Regulatory Affairs Manager
1 Bayer Drive
Indianola, Pennsylvania 15051

Re: K182276

Trade/Device Name: MEDRAD MRXperion MR Injection System and MR Injection System Syringe Kit

Regulation Number: 21 CFR 870.1650

Regulation Name: Angiographic Injector And Syringe

Regulatory Class: Class II

Product Code: DXT

Dated: September 28, 2018

Received: October 1, 2018

Dear Katie Dillon:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal

statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/CombinationProducts/GuidanceRegulatoryInformation/ucm597488.htm>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/>) and CDRH Learn (<http://www.fda.gov/Training/CDRHLearn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<http://www.fda.gov/DICE>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Michael John -S
2018.11.01 14:44:18 -04'00'

for Bram D. Zuckerman, M.D.
Director
Division of Cardiovascular Devices
Office of Device Evaluation
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K182276

Device Name

MEDRAD® MRXperion MR Injection System and MR Injection System Syringe Kit

Indications for Use (Describe)

The MEDRAD® MRXperion MR Injection System is a syringe-based fluid delivery system indicated for delivery of contrast media and saline during MR applications. It is intended to be used for the specific purpose of injecting intravenous MR contrast media and saline into the human vascular system for diagnostic studies in magnetic resonance imaging (MRI) applications with MRI scanners that have a magnetic field strength between 0.7 and 3.0 Tesla. Only trained healthcare professionals are intended to operate this device.

Type of Use (Select one or both, as applicable)

☒ Prescription Use (Part 21 CFR 801 Subpart D)

☐ Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

K182276



510(k) Summary

Submitter: Bayer Medical Care, Inc.
1 Bayer Drive
Indianola, PA 15051

Contact Person: Katie Dillon
Regulatory Affairs Manager
Phone: (412) 406-3212
Email: katie.dillon@bayer.com

Date Prepared: October 30, 2018

Device Trade Name: MEDRAD MRXperion MR Injection System
MEDRAD MRXperion MR Injection
System Syringe Kit

Common Name: Angiographic Injector and Syringe

Classification Name: Injector and Syringe, Angiographic [21 CFR
870.1650]

Product Code: DXT

Classification: Class II

Predicate Device: The subject device is substantially
equivalent to the following device:
MEDRAD MRXperion MR Injection
System and MR Injection System
Syringe Kit, K143538

**Device Description:**

The MEDRAD MRXperion MR Injection System is a software-controlled medical device used for the administration of intravenous MR contrast media and saline into the human vascular system for diagnostic studies in magnetic resonance imaging (MRI) procedures. Commonly referred to as an automated injection system, it is designed to allow a user to fill disposable syringes and then to perform an injection with a user-programmed volume and flow rate.

Indications for Use:

The MEDRAD® MRXperion MR Injection System is a syringe-based fluid delivery system indicated for delivery of contrast media and saline during MR applications. It is intended to be used for the specific purpose of injecting intravenous MR contrast media and saline into the human vascular system for diagnostic studies in magnetic resonance imaging (MRI) applications with MRI scanners that have a magnetic field strength between 0.7 and 3.0 Tesla. Only trained healthcare professionals are intended to operate this device.

Comparison to the Predicate Device:

A new control room unit (CRU) has been developed as an alternative to the existing CRU used with the MEDRAD MRXperion MR Injection System. This design change does not involve any changes to the MEDRAD MRXperion MR Injection System Syringe Kit. Table 1 provides a comparison of the proposed MEDRAD MRXperion MR Injection System with the predicate device, the MEDRAD MRXperion MR Injection System cleared per K143538.

The fundamental scientific technology, principle of operation, and indications for use are unchanged from the predicate device, the MEDRAD MRXperion MR Injection System cleared per K143538. The differences between the proposed device and the predicate device do not introduce new issues of safety and effectiveness and do not change the indications for use or result in a different fundamental scientific technology.



K182276

Table 1. Comparison of MEDRAD MRXperion MR Injection System (proposed device) with predicate device.

PARAMETER	PREDICATE DEVICE MRXperion MR Injection System (K143538)	PROPOSED DEVICE MRXperion MR Injection System (K182276)	RATIONALE FOR CHANGE
Indications for Use	The MEDRAD® MRXperion MR Injection System is a syringe- based fluid delivery system indicated for delivery of contrast media and saline during MR applications. It is intended to be used for the specific purpose of injecting intravenous MR contrast media and saline into the human vascular system for diagnostic studies in magnetic resonance imaging (MRI) applications with MRI scanners that have a magnetic field strength between 0.7 and 3.0 Tesla. Only trained healthcare professionals are intended to operate this device.	Same	No change
Fill Volume (Syringe A)	0.5 ml to max. syringe volume in: *0.1 ml increments between 0.5 and 31 ml * 1 ml increments for 31 ml and above	Same	No change
Fill Volume (Syringe B)	1 ml to max. syringe volume in 1 ml	Same	No change



K182276

PARAMETER	PREDICATE DEVICE MRXperion MR Injection System (K143538)	PROPOSED DEVICE MRXperion MR Injection System (K182276)	RATIONALE FOR CHANGE
	increments		
Fill Speed (low speed)	1.0 to 10.0 ml/s in 0.5 ml/s increments	Same	No change
Fill Speed (high speed)	1.0 to 10.0 ml/s in 0.5 ml/s increments	Same	No change
Flow Rate	0.01 to 10 ml/s in increments of: * 0.01 ml/s between 0.01 and 3.1 ml/s * 0.1 ml/s between 3.1 and 10 ml/s	Same	No change
Delay	N/A. Addressed by Pause Phase.	Same	No change
Pause Phase	1 to 1200 s in 1 s increments	Same	No change
Programmable Pressure Limit (PSI/kPa)	100/690 150/1035 200/1380 250/1725 300/2070 325/2240	Same	No change
Keep Vein Open (KVO)	Yes; 0.25 ml, time adjustable. Defaults at every 30 sec.	Same	No change
Protocol Memory	60 protocols of up to 6 phases each	Same	No change
Injection History Memory	Previous 20 successful injections	Same	No change
Control Room Unit	Windows-based Control Room Unit (CRU) with separate hardware components.	Windows-based Control Room Unit with integrated hardware components	Integration of hardware components for aesthetic improvement,



PARAMETER	PREDICATE DEVICE MRXperion MR Injection System (K143538)	PROPOSED DEVICE MRXperion MR Injection System (K182276)	RATIONALE FOR CHANGE
			modular design and serviceability. No impact on safety and effectiveness.
Information Display (Control Room)	Color LCD	Same	No change
Programming Keys (Control Room)	Software-generated via an LCD resistive touch screen	Software-generated via LCD capacitive touch screen	Touch screen calibration not required for capacitive touch screen. No impact on safety and effectiveness.
Programming Keys (Scan Room)	Dedicated keys on injector head	Same	No change
Multi-Phase	6 phases per injection	Same	No change
Safety Stop Mechanism	Software stops and electromechanical switch	Same	No change
Syringe System	Dual Syringes	Same	No change
Syringe Docking	Non-rotational orientation	Same	No change
Docking	Manual and Automatic	Same	No change
Fill Control	Manual and Automatic	Same	No change
Prime Control	Manual and Automatic	Same	No change
Retract Control	Manual and Automatic	Same	No change
Check for Air Confirmation	Operator visual inspection; user confirmed	Same	No change
Start/Stop Switch	Control Room and Scan	Same	No change



K182276

PARAMETER	PREDICATE DEVICE MRXperion MR Injection System (K143538)	PROPOSED DEVICE MRXperion MR Injection System (K182276)	RATIONALE FOR CHANGE
	Room		
Hand Switch	Control Room (optional)	Same	No change
Communication	Fiber Optic	Same	No change
MR Compatibility	0.7T to 3.0T	Same	No change
eGFR Calculator	Yes	Same	No change
Weight-Based Dosing Calculator	Yes	Same	No change
Informatics Compatibility	Yes	Same	No change
Power management	The Base supplies power to the Injector Head and the main processor via an AC/DC power supply-module. A rechargeable lithium ion battery is used to power the motors in the base during an injection.	Same	No change
MRXperion MR Injection System Syringe Kit	There are no changes to the MRXperion MR Injection System Syringe Kit.		No change

Performance Data:

Verification and validation testing was conducted to assess the impact of the modification on the MEDRAD MRXperion MR Injection System. The following testing was successfully completed:

- Device performance testing included verification of fluid delivery, flow rates, volumes, and pressures. Testing also verified that the device was not affected by environmental conditions such as atmospheric conditions and handling. All



K182276

testing passed and the demonstrated product performance met all prior established acceptance criteria.

- Safety and Compatibility testing included verification of configurations and specifications, circuitry, compliance with IEC 60601-1 and EMC requirements, electrical safety controls, ability to detect failures in communication and controls, programming keys, and sensors, and safe operation. All testing passed and the demonstrated product performance demonstrated met all prior established acceptance criteria.
- Reliability testing was performed using statistical methods to demonstrate the capability to sequentially and repeatedly meet system performance requirements. Testing verified there was no degradation to performance when the MEDRAD MRXperion MR Injection System and Informatics processes were run simultaneously. All testing passed and the demonstrated performance met all prior established acceptance criteria.
- Simulated Use and Human Factors testing was performed by using the device and disposables in a simulated clinical environment to validate the clinical user needs were met by the design per EN 62366-1: 2015 and FDA Guidance "Applying Human Factors and Usability Engineering to Medical Devices." Testing demonstrated that no new or different questions of safety or effectiveness were raised.
- Cleaning and disinfection validation was performed per FDA Guidance "Reprocessing Medical Devices in Health Care Settings: Validation Methods and Labeling," and confirmed that the injection system meets its cleaning and disinfection requirements.

All test results demonstrate that the design and materials of the MEDRAD MRXperion MR Injection System meet the established performance criteria and will perform as intended. The results of the design verification and validation, including human factors engineering evaluation, demonstrate that the subject device is substantially equivalent to its predicate device.

No performance standards have been established under Section 514 of the Food, Drug and Cosmetic Act for angiographic injectors and syringes.

**Clinical Testing:**

No clinical testing was required or performed to support this Special 510(k) Premarket Notification.

Conclusion:

Bayer Medical Care Inc. considers the MEDRAD MRXperion MR Injection System to be substantially equivalent to the predicate device listed above. The fundamental scientific technology, principle of operation, and indications for use are unchanged from the predicate device. The differences between the proposed device and the predicate device do not introduce new issues of safety and effectiveness.