



July 3, 2019

Nobio Ltd.
% Shoshanna Friedman
CEO
ProMedoss, Inc.
3521 Hatwynn Rd.
Charlotte, North Carolina 28269

Re: K182580

Trade/Device Name: Novidia Flowable Composite
Regulation Number: 21 CFR 872.3690
Regulation Name: Tooth Shade Resin Material
Regulatory Class: Class II
Product Code: EBF
Dated: June 3, 2019
Received: June 4, 2019

Dear Shoshanna Friedman:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal

statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Srinivas Nandkumar, Ph.D.
Acting Assistant Director
DHT1B: Division of Dental Devices
OHT1: Office of Ophthalmic, Anesthesia,
Respiratory, ENT and Dental Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K182580

Device Name

Novidia™ Flowable Composite

Indications for Use (Describe)

The Novidia™ Flowable Composite is indicated for:

1. Class III and V restorations
2. Restoration of minimally invasive cavity preparations (including small, non-stress bearing occlusal restorations)
3. As base/liner under direct or indirect restorations
4. Repair of small defects in esthetic indirect restorations
5. Pit and fissure sealing
6. Blocking out of undercuts
7. Repair of resin and acrylic temporary materials

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(K) SUMMARY

[as required by section 807.92(c)]

Novidia™ Flowable Composite

510(k) Number K182580

5.1 SUBMITTER

Applicant's Name:

Nobio Ltd.
8 Hamatechet St.
POB 50502
Kadima, Israel 6092000
Phone: +972-3-9059966

Contact Person:

Shoshana (Shosh) Friedman
Regulatory Affairs Consultant
3521 Hatwynn Rd.
Charlotte, NC 28269
Phone: 704-430-8695
s.friedman@promedoss.com

Date Prepared:

June 3, 2019

5.2 DEVICE

Trade Name:

Novidia™ Flowable Composite

Classification: **Name:** Material, Tooth Shade, Resin

Product Code: EBF

Regulation No: 872.3690

Class: 2

Review Panel: Dental

5.3 PREDICATE DEVICE

Filtek™ Supreme Ultra Flowable Restorative, manufactured by 3M Company, cleared under K100235.

Additionally, we are using the following reference devices as examples of FDA-cleared devices that incorporate quaternary ammonium in their formulation:

- Clearfil Protect Bond cleared under K033938
- Orthodontic Acrylic cleared under K141439 and Orthodontic Acrylic 2 cleared under K163482.

5.4 DEVICE DESCRIPTION

Novidia™ Flowable Composite is a low viscosity, visible light activated, radiopaque, flowable composite indicated for minimally invasive cavity preparations as well as various other indications.

The resin matrix of the Novidia™ Flowable Composite contains Urethane dimethacrylate (UDMA), Bisphenol-a glycidyl dimethacrylate (Bis-GMA) and triethylene glycol dimethacrylate (TEGDMA).

The inorganic filler of the Novidia™ Flowable Composite is a mix of particles of alumino-silicate-based glasses, silica dioxide and pigments. A small percentage of the silica-based filler particles are covalently bound to quaternary ammonium residues (QASi), added to maintain the integrity of the restoration.

Note: *“Clinical studies demonstrating that the presence of QASi in this device improves clinical outcomes have not been conducted”.*

The Novidia™ Flowable Composite is provided in syringes (2 g) and in single-dose capsules (0.2 g) and is available in three shades (A1, A2 and A3).

5.5 INDICATIONS FOR USE

Novidia™ Flowable Composite is indicated for use in:

- 1) Class III and V restorations
- 2) Restoration of minimally invasive cavity preparations (including small, non-stress-bearing occlusal restorations)
- 3) As base/liner under direct or indirect restorations
- 4) Repair of small defects in esthetic indirect restorations
- 5) Pit and fissure sealant
- 6) Blocking out of undercuts
- 7) Repair of resin and acrylic temporary materials

5.6 SUBSTANTIAL EQUIVALENCE

The Novidia™ Flowable Composite has the same indications as the Filtek™ Supreme Ultra Flowable Restorative.

The technological characteristics of the Novidia™ Flowable Composite are substantially equivalent to these of the predicate device and other methacrylate-based products currently on the market. Table 5-1 below shows a comparison of Novidia™ Flowable Composite and the predicate device.

Table 5-1: Comparison of Novidia™ Flowable Composite and Filtek™ Supreme Ultra Flowable Restorative

Feature	Novidia™ Flowable Composite	Filtek Supreme Ultra Flowable Restorative
510(k) Number		K100235
Classification	EBF	EBF
Indications	<ul style="list-style-type: none"> • Class III and V restorations • Restoration of minimally invasive cavity preparations (including small, non-stress-bearing occlusal restorations) • As base/liner under direct or indirect restorations • Repair of small defects in esthetic indirect restorations • Pit and fissure sealant • Blocking out of undercuts • Repair of resin and acrylic temporary materials 	<ul style="list-style-type: none"> • Class III and V restorations • Restoration of minimally invasive cavity preparations (including small, non-stress-bearing occlusal restorations) • Base/liner under direct restorations • Repair of small defects in esthetic indirect restorations • Pit and fissure sealant • Undercut blockout • Repair of resin and acrylic temporary materials
Composition	Methacrylate resins, photo-initiators, inorganic fillers	Methacrylate resins, photo-initiators, inorganic fillers
Packaging	Syringe and single-dose capsule	Syringe and single-dose capsule
Flexural Strength	ISO 4049:2009	ISO 4049:2009
Intensity for curing	≥550 mW/cm ²	400 -1000 mW/cm ²
Wavelength for curing	430-490 nm	400-500 nm
Curing time	20 sec.	Opaque shades 40 sec; All other shades 20 sec.
Radio-opacity	ISO 4049: 2009	ISO 4049: 2009
Depth of Cure	ISO 4049: 2009	ISO 4049: 2009
Water Sorption	ISO 4049:2009	ISO 4049:2009
Water Solubility	ISO 4049:2009	ISO 4049:2009
Spontaneous Polymerization Sensitivity at Ambient Light	ISO 4049:2009	ISO 4049:2009

5.7 PERFORMANCE DATA

Non-Clinical Performance Testing:

Non-clinical and biological testing was completed to assess the performance and biocompatibility of the Novidia™ Flowable Composite and to support substantial equivalence. The data provided in this 510(k) submission shows that the composite is biocompatible based on the biocompatibility assessment conducted as per ISO 10993 and ISO 7405 and performs as intended based on the bench testing per ISO 4049 and FDA guidance document “Dental Composite Resin Devices - Premarket Notification [510(k)] Submissions”. The list of these tests is provided in Table 5-2.

Table 5-2: List of Tests Completed on Novidia™ Flowable Composite

Biocompatibility
Cytotoxicity
Oral Mucosal Irritation Test
Acute Systemic Toxicity
Material Mediated Pyrogenicity
Bacterial Reverse Mutation
Mouse Lymphoma Assay
Biological Risk Assessment
Bench Testing
Flexural strength
Elastic modulus
Compression strength
Shade and color stability
Polymerization conversion degree
Viscosity
Spontaneous polymerization sensitivity at ambient light
Water solubility
Water sorption
Depth of cure
Radio-opacity
Knoop hardness

An additional test that was performed is “Preservation of Surface Integrity”. This study used the atomic force microscopy (AFM) method to test the hypothesis that incorporation of QASi into the Novidia Flowable Composite prevents the increase of surface roughness (a surrogate measure of material degradation) caused by biofilm. The study concluded that the Novidia Flowable Composite with 1.2% (wt/wt) QASi filler is resistant to microbial degradation evidenced by preserved surface roughness in presence of continuous microbial challenge and therefore supports the thesis that 1.2% QASi filler improves the integrity of the dental composite restorations and preserves its functionality over time. Clinical studies demonstrating that the addition of 1.2% wt/wt QASi results in enhanced clinical outcomes have not been conducted.

Animal and Clinical Performance Testing:

Animal and clinical performance data was not included.

5.8 CONCLUSION

Nobio Ltd. believes that Novidia™ Flowable Composite is substantially equivalent to the Filtek™ Supreme Ultra Flowable Restorative and other legally marketed products. It does not introduce new indications for use, has similar technological characteristics and does not introduce any new safety or effectiveness concerns.