



October 21, 2019

Maquet Critical Care AB
% Mark Dinger
Sr. Regulatory Affairs Specialist
Maquet Medical Systems USA
45 Barbour Pond Drive
Wayne, New Jersey 07470

Re: K182862
Trade/Device Name: Servo Guard
Regulation Number: 21 CFR 868.5260
Regulation Name: Breathing Circuit Bacterial Filter
Regulatory Class: Class II
Product Code: CAH
Dated: September 19, 2019
Received: September 20, 2019

Dear Mark Dinger:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal

statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

For Elizabeth Claverie, M.S.
Assistant Director for THT4B2
Acting Assistant Director for THT4B1
DHT4B: Division of Infection Control
and Plastic Surgery Devices
OHT4: Office of Surgical
and Infection Control Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K182862

Device Name

Servo Guard

Indications for Use (Describe)

The Servo Guard is a bacterial/viral filter for applications in respiratory care. The filter is intended to be used in patient circuits to reduce the spread of viruses and bacteria from patient to patient/personnel and to protect the patient circuits and ventilators from contamination. The filter shall only be used with the ventilators Servo-u, Servo-n, Servo-i and Servo-s.

Type of Use (Select one or both, as applicable)

☒ Prescription Use (Part 21 CFR 801 Subpart D)

☐ Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) SUMMARY
as required by section 21 CFR 807.92

Submitter Name & Address

Maquet Critical Care AB
Röntgenvägen 2
SE-171 54 Solna, Sweden
Phone: (011) 46 10 335 7300

Contact Persons for this submission:

Jerker Åberg
Regulatory Affairs Manager
Phone: direct: (011) 46 10 335 7985
e-mail: jerker.aberg@getinge.com

Application Correspondent:

Mark Dinger
Sr. Regulatory Affairs Specialist
Maquet Medical Systems USA
45 Barbour Pond Drive
Wayne, NJ 07470
e-mail: mark.dinger@getinge.com
Phone: 973-709-7691
Fax: 973-909-9954

Date prepared: September 19th, 2019

Trade Name:

Servo Guard

Model no:

6595487

Device Classification

<i>Common Name</i>	<i>Classification Number</i>	<i>Class</i>	<i>Regulation Number</i>
Breathing Circuit Bacteria Filter	CAH	II	21 CFR 868.5260

Predicate Device Identification

<i>Legally marketed devices to which equivalence is being claimed</i>	<i>510(k) #</i>
Servo Guard Model number 64 81 290 EH88E	K030071

Indications for Use

The Servo Guard is a bacterial/viral filter for applications in respiratory care. The filter is intended to be used in patient circuits to reduce the spread of viruses and bacteria from patient to patient/personnel and to protect the patient circuits and ventilators from contamination. The filter shall only be used with the ventilators Servo-u, Servo-n, Servo-i and Servo-s.

Device Description

A Breathing Circuit Bacteria Filter is intended to protect from contamination as well as to reduce particles such as dust and microbiological matters from gases used in the patient circuit of a respiratory device. The bacterial/viral filter is a disposable accessory normally used on the expiratory limb of the ventilator breathing system to reduce possible cross contamination between patient and equipment and between patient and patient. It can also be used on the inspiratory pipe, in order to reduce dust particles to the patient.



Figure 1 Servo Guard filter with water trap

Technological Characteristics

Shown below are a comparison of the technological characteristic between the subject device and predicate.

	<i>Cleared Device</i>	<i>Changed Device</i>	<i>Conclusion</i>
Device	Servo Guard (model no 64 81 290 EH88E)	Servo Guard version (model no 6595487)	-
Manufacturer	Maquet Critical Care AB	Maquet Critical Care AB	-
Device Classification Name	Breathing circuit bacterial filter. CAH	Breathing circuit bacterial filter. CAH	-
510(k) Number	K030071	K182862	-
<i>Indications for Use according to the 510(k) Summary</i>	The Servo Guard is an efficient bacterial and viral filter for applications in respiratory care and anesthesia . The Servo Guard is a disposable single-use device that provides filtration for reducing possible cross contamination between patient and equipment	The Servo Guard is a bacterial/viral filter for applications in respiratory care. The filter is intended to be used in patient circuits to reduce the spread of viruses and bacteria from patient to patient/personnel and to protect the patient circuits and ventilators from contamination. The filter shall only be used with the ventilators Servo-u, Servo-n, Servo-i and Servo-s.	Similar The intended use of the changed device does not include anesthesia. Limitation of compatible ventilators.
<i>Physical properties</i>			
<i>Placement</i>	To be used on the expiratory or/and inspiratory limb of a breathing circuit.	To be used on the expiratory or/and inspiratory limb of a breathing circuit.	Identical
<i>Filter Volume</i>	150 ml	150 ml	Identical
<i>Bacterial/viral removal efficiency</i>	>99,999%	>99,999%	Identical
<i>Resistance</i>	1.5 cm H ₂ O@60 l/min	2.0 cmH ₂ O@60 l/min (dry) and 2.2 cmH ₂ O@60 l/min (wet)	different
<i>Connections</i>	ISO 5356-1 M22/F15 (male/female, inlet) ISO 5356-1 F22 (female, outlet)	ISO 5356-1 M22/F15 (male/female, inlet) ISO 5356-1 F22 (female, outlet)	Identical
<i>Weight</i>	60 g	75 g including water trap	Similar
<i>Recommended use</i>	Single use, replace every 24 hour	Single use, replace every 24 hour	Identical
<i>Type of filter</i>	mechanical	mechanical	Identical
<i>Shelf life</i>	1 year	3 years	Improved
<i>Dimensions</i>	L: H:W = 92 x 66 x 66 mm	L:H:W = 92 x 89 x 66 mm	Similar
<i>Patient range</i>	Neonatal to adult patient populations	Neonatal to adult patient populations	Identical

Summary of Non-clinical testing:

Test Method	Purpose	Acceptance Criteria	Result		
ISO 23328-1 Filter efficiency	Verify the filter efficiency for the Servo Guard filter	Servo Guard shall have a filter efficiency > 99.3 % measured in both flow directions	Filter efficiency: 99,329 %- 99,93%. Pass		
ASTM F2101-14 Standard Test Method for Evaluating the Bacterial Filtration Efficiency	Verify the bacterial and viral filtration efficiency for the Servo Guard filter	BFE/VFE filter efficiency shall be > 99.999 %.	BFE: 99.99951%- 99.999984 % VFE: 99.99960% - 99.999980 Pass		
ISO 18562-2: 2017 Biocompatibility evaluation of breathing gas pathways in healthcare applications-- Part 2: Tests for emissions of particulate matter	Verify that the amount of particulate matter are below the acceptance limit	<12 [µg/m3]	1.4 [µg/m3] Pass		
ISO 18562-3 Biocompatibility evaluation of breathing gas pathways in healthcare applications – Part 3: Tests for emissions of volatile organic compounds (VOCs)	Verify that the amount of Volatile substances are below the acceptance limit	The MOS (Margin Of Safety) value shall be higher than 1 for identified substances.	Substance	MOS	MOS
				Neonate	Pedriatic/ Adult
			Cyclohexane	35.0	83.3
			Ethylbenzene	7.8	32.5
			m+p-Xylene	4.2	10.0
			o-Xylene	15.0	35.7
			Tetrachloroethylene	4.7	11.1
			Trichlorofluoromethane	27.6	65.6
			Pass		

Conclusion:

The conclusions drawn from the clinical tests demonstrate that the device is as safe, as effective, and performs as well as or better than the legally marketed device (K030071).