



Weifang KM Electronics Co., LTD
Andy Pang
General Manager
No. 7999 Health East Street, High-tech District
Weifang, 261000 CN

April 15, 2019

Re: K182924

Trade/Device Name: Diode Laser Treatment System

Regulation Number: 21 CFR 878.4810

Regulation Name: Laser Surgical Instrument for use in General and Plastic Surgery and in Dermatology

Regulatory Class: Class II

Product Code: GEX

Dated: March 11, 2019

Received: March 13, 2019

Dear Andy Pang:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for

devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/CombinationProducts/GuidanceRegulatoryInformation/ucm597488.htm>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/>) and CDRH Learn (<http://www.fda.gov/Training/CDRHLearn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<http://www.fda.gov/DICE>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Neil R
Ogden -S

Digitally signed by Neil
R Ogden -S
Date: 2019.04.15
09:23:59 -04'00'

For Binita S. Ashar, M.D., M.B.A., F.A.C.S.
Director
Division of Surgical Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

Section 4 Indications for use

DEPARTMENT OF HEALTH AND HUMAN SERVICES Food and Drug Administration Indications for Use	Form Approved: OMB No. 0910-0120 Expiration Date: 06/30/2020 See PRA Statement below.
---	---

510(k) Number (if known)

K182924

Device Name

Diode Laser Treatment System

Indications for Use (Describe)

The Diode Laser Treatment System is intended for hair removal, permanent hair reduction on all skin types (Fitzpatrick skin type I-VI), including tanned skin.

Permanent hair reduction is defined as the long-term, stable reduction in the number of hairs regrowing when measured at 6, 9, and 12 months after the completion of a treatment regime.

Type of Use (Select one or both, as applicable)

☒ Prescription Use (Part 21 CFR 801 Subpart D)

☐ Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRAStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

Section 5 510(k) summary

I Submitter

Device submitter: Weifang KM Electronics Co., Ltd.

No.7999 Health East Street, High-tech District, Weifang City, Shandong,
China

Contact person: Andy PANG

General Manager

Phone: +86 136 0646 4486

Fax: +86-536-8885351

Email: 51050359@qq.com

Date of preparation: Oct 12th, 2018

II Device

Trade Name of Device: Diode Laser Treatment System

Models: KM200D, KM300D, KM600D, KM800D, KM900D

Common name: Powered Laser Surgical Instrument

Regulation Number: 21 CFR 878.4810

Regulatory Class: II

Product code: GEX

Review Panel: General & Plastic Surgery

III Predicate Devices

Trade Name of Device: Diode Laser Therapy Machine

Common name: Powered Laser Surgical Instrument

Regulation Number: 21 CFR 878.4810

Regulatory Class: II

Product code: GEX

Review Panel: General & Plastic Surgery

510(k) number: K161692

Trade Name of Device: Diode Laser Hair Removal System

Common name: Powered Laser Surgical Instrument

Regulation Number: 21 CFR 878.4810

Regulatory Class: II

Product code: GEX

Review Panel: General & Plastic Surgery

510(k) number: K141973

IV Device description

The Diode Laser Treatment System consists of the main unit and a hand piece. The system uses a diode laser as an active medium placed in an optical cavity to produce amplified beam at the wavelength of 808 nm. A microprocessor is used to control electronics for the front panel. A self-contained water cooling system is built into the power supply unit.

The device provides 36 working modes, which are six modes for men and six modes for women. The men or women mode includes face, armpit, arm, body, bikini, leg mode respectively for different treatment part. The diode laser operates in a pulsed mode with a fixed pulse width and fixed pulse duration of the pulse train for each mode. The number of pulses can be adjusted within the preset range.

V Indications for use

The Diode Laser Treatment System is intended for hair removal, permanent hair reduction on all skin types (Fitzpatrick skin type I-VI), including tanned skin.

Permanent hair reduction is defined as the long-term, stable reduction in the number of hairs regrowing when measured at 6, 9, and 12 months after the completion of a treatment regime.

VI Comparison of technological characteristics with the predicate devices

The Diode Laser Treatment System has the same technological characteristics and fundamental design as the predicate device. The subject device and the predicate devices are all designed for hair removal on different parts of the body. The differences between the subject device and predicate devices do not alter suitability of the proposed device for its intended use.

Device feature	Diode Laser Treatment System (subject device)	Diode Laser Therapy Machine FG 2000-B K161692	Diode Laser Hair Removal System K141973
Product code	GEX	GEX	GEX
Regulation number	21 CFR 878.4810	21 CFR 878.4810	21 CFR 878.4810
Indications for use	The Diode Laser Treatment System is intended for hair removal, permanent hair reduction on all skin types (Fitzpatrick skin	The Diode Laser Therapy Machine is intended for hair removal, permanent hair reduction on all skin types (Fitzpatrick skin	The Diode Laser Hair Removal System is intended for hair removal, permanent hair reduction on all skin types (Fitzpatrick skin

	type I-VI), including tanned skin. Permanent hair reduction is defined as the long-term, stable reduction in the number of hairs regrowing when measured at 6, 9, and 12 months after the completion of a treatment regime.	type I-VI), including tanned skin. Permanent hair reduction is defined as the long-term, stable reduction in the number of hairs regrowing when measured at 6, 9, and 12 months after the completion of a treatment regime.	type I-VI), including tanned skin. Permanent hair reduction is defined as the long-term, stable reduction in the number of hairs regrowing when measured at 6, 9, and 12 months after the completion of a treatment regime.
Operation principle	Melanin could absorb the energy from the laser, which would result in temperature rapid increase, to destroy surrounding hair follicles, and finally remove hair.	Melanin could absorb the energy from the laser, which would result in temperature rapid increase, to destroy surrounding hair follicles, and finally remove hair.	Melanin could absorb the energy from the laser, which would result in temperature rapid increase, to destroy surrounding hair follicles, and finally remove hair.
Laser type	Diode laser	Diode laser	Diode laser
Laser classification	Class IV	Class IV	Class IV
Laser wavelength	808nm	808nm	808nm
Spot Size	1.44 cm ²	1.44 cm ²	1.44 cm ²
Fluence	2-120J/cm ²	2-120J/cm ²	1-120J/cm ²
Frequency	1-10Hz	1-10Hz	1Hz, 2Hz, 3Hz, 10Hz
Pulse Duration	10-300ms	9-143ms	2.9-348ms
Power supply	100-240V AC, 50/60Hz	AC110V/50-60Hz	AC110V, 60Hz
Dimension	KM200D 60*42*38cm KM300D 60*42*35cm KM600D 60*42*35cm KM800D 60*42*40cm KM900D 55*42*30cm	42*63*54cm	38*54*120cm
Weight	35kg	30kg	55kg
Patient contact material	Sapphire in handpiece and handpiece tip (stainless steel)	Sapphire in handpiece and handpiece tip (stainless steel)	Sapphire in handpiece and handpiece tip (stainless steel)
Biocompatibility	Comply with ISO10993-1	Comply with ISO10993-1	Comply with ISO10993-1

Electrical Safety	Comply with IEC60601-1, IEC60601-2-22	Comply with IEC60601-1, IEC60601-2-22	Comply with IEC60601-1, IEC60601-2-22
EMC	Comply with IEC60601-1-2,	Comply with IEC60601-1-2,	Comply with IEC60601-1-2,
Laser safety	Comply with IEC60825-1, IEC60601-2-22	Comply with IEC60825-1, IEC60601-2-22	Comply with IEC60825-1, IEC60601-2-22

VII Performance data

The following performance data were provided in support of the substantial equivalence determination.

Biocompatibility testing

Biocompatibility of the Laser Treatment System was evaluated in accordance with ISO 10993-1:2009 for the body contact category of “Surface – Mucosal Membrane” with a contact duration of “Limited (< 24 hours)”. The following tests were performed, as recommended: Cytotoxicity, Irritation and Sensitization. All evaluation acceptance criteria were met

Electrical safety and electromagnetic compatibility (EMC)

Electrical safety and EMC testing were conducted on the Laser Treatment System. The system has been tested to comply with the following standards:

- IEC 60601-1:2012 Medical Electrical Equipment - Part 1: General Requirements For Basic Safety And Essential Performance;
- IEC 60601-2-22:2007, Medical Electrical Equipment - Part 2-22: Particular Requirements For Basic Safety And Essential Performance Of Surgical, Cosmetic, Therapeutic And Diagnostic Laser Equipment;
- IEC 60825-1: 2007, Safety of laser products - Part 1: Equipment classification and requirements.
- IEC 60601-1-2:2007, Medical electrical equipment- Part 1-2: General requirements for basic safety and essential performance- Collateral standard: Electromagnetic compatibility- Requirements and tests.

VIII Conclusion

The Diode Laser Treatment System is substantially equivalent to its predicate devices. The non-clinical testing demonstrates that the device is as safe, as effective and performs as well as the legally marketed device.