



January 28, 2019

Foosin Medical Supplies Inc., Ltd  
% Diana Hong  
General Manager  
Mid-Link Consulting Co., Ltd  
P.O. Box 120-119  
Shanghai, 200120 CN

Re: K183004

Trade/Device Name: WEGO-PGA Rapid  
Regulation Number: 21 CFR 878.4493  
Regulation Name: Absorbable Poly(Glycolide/L-Lactide) Surgical Suture  
Regulatory Class: Class II  
Product Code: GAM  
Dated: September 26, 2018  
Received: October 30, 2018

Dear Ms. Hong:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's

requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/CombinationProducts/GuidanceRegulatoryInformation/ucm597488.htm>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/>) and CDRH Learn (<http://www.fda.gov/Training/CDRHLearn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<http://www.fda.gov/DICE>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

  
Cynthia Chang -S

for Binita S. Ashar, M.D., M.B.A., F.A.C.S.  
Director  
Division of Surgical Devices  
Office of Device Evaluation  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)

K183004

Device Name

WEGO-PGA RAPID

Indications for Use (Describe)

WEGO-PGA RAPID sutures are intended for use in superficial soft tissue approximation of skin and mucosa where only short-term wound support is required. WEGO-PGA RAPID is not intended for use in ligation, ophthalmic, cardiovascular or neurological procedures.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

## 510(k) Summary

This 510(k) Summary of 510(k) safety and effectiveness information is being submitted in accordance with requirements of SMDA 1990 and Title 21, CFR Section 807.92.

The assigned 510(k) Number: K183004

1. Date of Preparation: 01/24/2019

2. Sponsor Identification

Foosin Medical Supplies Inc., Ltd

No.20, Xingshan Road, Weihai Torch Hi-tech Science Park, Weihai, Shandong, 264210, China

Establishment Registration Number: 3006562124

Contact Person: Lei Tang

Position: Quality Manager

Tel: +86-631-5661968

Fax: +86-631-5621680

Email: tang-lei@live.com

3. Designated Submission Correspondent

Ms. Diana Hong (Primary Contact Person)

Mr. Chengyu Wang (Alternative Contact Person)

**Mid-Link Consulting Co., Ltd**

P.O. Box 120-119, Shanghai, 200120, China

Tel: +86-21-22815850,

Fax: 360-925-3199

Email: [info@mid-link.net](mailto:info@mid-link.net)

4. Identification of Proposed Device

Trade Name: WEGO-PGA RAPID

Common Name: Absorbable Synthetic Suture with or without Needle

Size: USP 6-0, USP 5-0, USP 4-0, USP 3-0, USP 2-0, USP 0, USP 1 and USP 2

Regulatory Information

Classification Name: Suture, Absorbable, Synthetic, Polyglycolic Acid

Classification: II

Product Code: GAM

Regulation Number: 21 CFR 878.4493

Review Panel: General & Plastic Surgery

Indications for Use:

WEGO-PGA RAPID sutures are intended for use in superficial soft tissue approximation of skin and mucosa where only short-term wound support is required. WEGO-PGA RAPID is not intended for use in ligation, ophthalmic, cardiovascular or neurological procedures.

Device Description

WEGO-PGA RAPID sutures are synthetic, absorbable, braided, sterile surgical sutures composed of Polyglycolic Acid (PGA). The formula of the polymer is  $(C_2H_2O_2)_n$ . The characteristic rapid loss of strength is achieved by use of a polymer material with a lower molecular weight than regular WEGO-PGA suture. WEGO-PGA RAPID sutures are available undyed and dyed violet with D&C Violet No.2 (Colour Index number 60725). WEGO-PGA RAPID sutures are coated with polycaprolactone and calcium stearate.

The performance of this absorbable suture complies with United States Pharmacopeia (U.S.P.) monograph requirements for Absorbable Surgical Suture, USP 40<861>, USP 40<871> and USP 40 <881>.

The Synthetic Absorbable Suture is provided EO sterilized as a single use device.

5. Identification of Predicate Device

Predicate Device

510(k) Number: K944110

Product Name: VICRYL Rapide Suture

Manufacturer: ETHICON, INC

6. Non-Clinical Test Conclusion

Non clinical tests were conducted to verify that the proposed device met all design specifications as was Substantially Equivalent (SE) to the predicate device. The test results demonstrated that the proposed device complies with the following standards:

- ⑩ USP 40-NF 35:2017 <861> Sutures – Diameter;
- ⑩ USP 40-NF 35:2017 <871> Sutures - Needle Attachment;
- ⑩ USP 40-NF 35:2017 <881> Tensile Strength;
- ⑩ ISO 10993-3:2014 Biological Evaluation of Medical Devices, Part 3: Tests for Genotoxicity, Carcinogenicity and Reproductive Toxicity;
- ⑩ ISO 10993-5:2009 Biological evaluation of medical devices -- Part 5: Tests for in vitro cytotoxicity;
- ⑩ ISO 10993-6:2007 Biological evaluation of medical devices -- Part 6: Tests for local effects after implantation;
- ⑩ ISO 10993-10: 2010 Biological evaluation of medical devices -- Part 10: Tests for irritation and skin sensitization;
- ⑩ ISO 10993-11:2006 Biological evaluation of medical devices -- Part 11: Tests for systemic toxicity;
- ⑩ USP 40-NF 35:2017 <85> Bacterial Endotoxins Test;
- ⑩ ASTM F88/F88M-15 Standard Test Method for Seal Strength of Flexible Barrier Materials;
- ⑩ ASTM F1929-15 Standard Test Method for Detecting Seal Leaks in Porous Medical Package by Dye Penetration;
- ⑩ ISO 10993-7:2008 Biological evaluation of medical devices -- Part 7: Ethylene oxide sterilization residuals;
- ⑩ USP 39-NF33:2016 <151> Pyrogen Test
- ⑩ ASTM F756-13 Standard Practice For Assessment Of Hemolytic Properties Of Materials

The physical performance and stability of proposed device have been demonstrated by testing on samples before and after real-time-aging, to support the substantial equivalence between proposed device and the predicate, details are shown as follow:

## Clinical Test Conclusion

No clinical study is included in this submission.

## 5. Substantially Equivalent (SE) Comparison

Table 1 Comparison of Proposed Device and Predicated Device

<b>Item</b>	<b>Proposed Device</b>	<b>Predicate Device K944110</b>
Product Code	GAM	GAM
Regulation Number	21 CFR 878.4493	21 CFR 878.4493
Class	II	II
Intended Use	WEGO-PGA RAPID sutures are intended for use in superficial soft tissue approximation of skin and mucosa where only short-term wound support is required. WEGO-PGA RAPID is not intended for use in ligation, ophthalmic, cardiovascular or neurological procedures.	The Vicryl Rapide Suture is indicated only for use in superficial soft tissue approximation of skin and mucosa where only short-term wound support (7-10 days) is required. Vicryl Rapide Suture is not intended for use in ligation, ophthalmic, cardiovascular or neurological procedures.
Configuration	Suture and Needle	Suture and Needle
Sterility	EO Sterilized	EO Sterilized
Single Use	Yes	Yes
Based Material	Polyglycolic Acid (PGA)	90% glycolide and 10% L-lactide (PGLA)
Coating Material	Polycaprolactone and calcium stearate	Polycaprolactone and calcium stearate
Color	Dyed and undyed	Dyed and undyed
Color Additive Material	D&C violet No. 2	D&C violet No. 2
Absorbable / Non-absorbable	Absorbable	Absorbable

Braided / Monofilament	Braided	Braided
Suture Size	USP 6-0, USP 5-0, USP 4-0, USP 3-0, USP 2-0, USP 0, USP 1 and USP 2	Unknown
Length of Suture	25cm, 30cm, 45cm, 50cm, 60cm, 70cm,	Unknown
	75cm and 90cm	
Absorption Time	42 days	42 days

The based material of proposed device is PGA, which is different from that of predicate device. However, the performance and biocompatibility of proposed device has been demonstrated by various tests. The specifications of predicate devices are not known, but the proposed device has been demonstrate to comply the requirements listed in USP Monograph of Absorbable Surgical Suture.

#### 6. Substantially Equivalent (SE) Conclusion

Based on the comparison and analysis above, the proposed device is determined to be Substantially Equivalent (SE) to the predicate devices.