



November 30, 2018

Luminex Molecular Diagnostics, Inc.
Tina Ip
Senior Regulatory Affairs Associate
439 University Avenue
Toronto, M5g 1y8 Ca

Re: K183023

Trade/Device Name: xTAG Gastrointestinal Pathogen Panel (GPP), xTAG Data Analysis Software (TDAS GPP)

Regulation Number: 21 CFR 866.3990

Regulation Name: Gastrointestinal microorganism multiplex nucleic acid-based assay

Regulatory Class: Class II

Product Code: PCH

Dated: October 31, 2018

Received: November 1, 2018

Dear Tina Ip:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801 and Part 809); medical device reporting (reporting of medical device-related adverse events) (21 CFR

803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/CombinationProducts/GuidanceRegulatoryInformation/ucm597488.htm>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/>) and CDRH Learn (<http://www.fda.gov/Training/CDRHLearn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<http://www.fda.gov/DICE>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,


Steven R. Gitterman -S for

Uwe Scherf, Ph.D.
Director
Division of Microbiology Devices
Office of In Vitro Diagnostics
and Radiological Health
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

Device Name

xTAG® Gastrointestinal Pathogen Panel (GPP)

Indications for Use (Describe)

The xTAG® Gastrointestinal Pathogen Panel (GPP) is a multiplexed nucleic acid test intended for the simultaneous qualitative detection and identification of multiple viral, bacterial and parasitic nucleic acids in human stool specimens or human stool in Cary-Blair media from individuals with signs and symptoms of infectious colitis or gastroenteritis. The following pathogen types, subtypes and toxin genes are identified using the xTAG GPP:

Viruses

- Adenovirus 40/41
- Norovirus GI/GII
- Rotavirus A

Bacteria

- Campylobacter (C. jejuni, C. coli and C. lari only)
- Clostridium difficile (C. difficile) toxin A/B
- Escherichia coli (E. coli) O157
- Enterotoxigenic E. coli (ETEC) LT/ST
- Salmonella
- Shiga-like Toxin producing E. coli (STEC) stx 1/stx 2
- Shigella (S. boydii, S. sonnei, S. flexneri and S. dysenteriae)
- Vibrio cholerae (V. cholerae) cholera toxin gene (ctx)

Parasites

- Cryptosporidium (C. parvum and C. hominis only)
- Entamoeba histolytica (E. histolytica)
- Giardia (G. lamblia only, also known as G. intestinalis and G. duodenalis)

The detection and identification of specific gastrointestinal microbial nucleic acid from individuals exhibiting signs and symptoms of gastrointestinal infection aids in the diagnosis of gastrointestinal infection when used in conjunction with clinical evaluation, laboratory findings and epidemiological information. A gastrointestinal microorganism multiplex nucleic acid-based assay also aids in the detection and identification of acute gastroenteritis in the context of outbreaks.

xTAG GPP positive results are presumptive and must be confirmed by FDA-cleared tests or other acceptable reference methods.

The results of this test should not be used as the sole basis for diagnosis, treatment, or other patient management decisions. Confirmed positive results do not rule out co-infection with other organisms that are not detected by this test, and may not be the sole or definitive cause of patient illness. Negative xTAG GPP results in the setting of clinical illness compatible with gastroenteritis may be due to infection by pathogens that are not detected by this test or non-infectious causes such as ulcerative colitis, irritable bowel syndrome, or Crohn's disease.

xTAG GPP is not intended to monitor or guide treatment for C. difficile infections.

The xTAG GPP test is indicated for use with the Luminex® 100/200™ and MAGPIX® instruments with xPONENT® software.

Type of Use (Select one or both, as applicable)☒ Prescription Use (Part 21 CFR 801 Subpart D)☐ Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."