



Taiwan Aulisa Medical Devices Technologies, Inc.  
% Don Mizota  
Consultant  
725 Morninghome Road  
Danville, California 94526

Re: K183067

Trade/Device Name: Guardian Angel Rx GA2000 Digital Vital Sign Monitoring System  
Regulation Number: 21 CFR 870.2700  
Regulation Name: Oximeter  
Regulatory Class: Class II  
Product Code: DQA, DRG  
Dated: January 28, 2019  
Received: February 1, 2019

Dear Don Mizota:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part

801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/CombinationProducts/GuidanceRegulatoryInformation/ucm597488.htm>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/>) and CDRH Learn (<http://www.fda.gov/Training/CDRHLearn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<http://www.fda.gov/DICE>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Todd D. Courtney -

S

for Tina Kiang, Ph.D.

Acting Director

Division of Anesthesiology,

General Hospital, Respiratory,

Infection Control, and Dental Devices

Office of Device Evaluation

Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)

K183067

Device Name

Guardian Angel Rx GA2000 Digital Vital Sign Monitoring System

Indications for Use (Describe)

The Guardian Angel Rx GA2000 Digital Vital Sign Monitoring System is indicated for use in measuring and displaying functional oxygen saturation of arterial hemoglobin (SpO<sub>2</sub>) and pulse rate of adult and pediatric patients. It is indicated for spot-checking and/or continuous monitoring of patients during non-motion and under well-perfused conditions. The intended environments of use are hospitals, medical facilities, home care, and subacute environments. This system is a reusable device.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

**Section 5: 510(k) Summary**

This 510(k) Summary is being submitted in accordance with the requirements of 21 CFR 807.92

**5.1. General Information**

Date of Preparation: February 1, 2019

Submitted by: Taiwan Aulisa Medical Devices Technologies, Inc.  
Rm. 1052, Bldg. H, 10F., No.3-2, YuanQu St.,  
Nangang Dist., Taipei City 115 TW  
Phone: 886-2-2655-7297  
FAX: 886-2-2655-7260

Contact Person: Paul Liu  
Regulatory Affairs Supervisor  
Taiwan Aulisa Medical Devices Technologies, Inc.  
Rm. 1052, Bldg. H, 10F., No.3-2, YuanQu St.,  
Nangang Dist., Taipei City 115 TW  
Phone: 886-2-2655-7297  
FAX: 886-2-2655-7260  
Email: [paul.liu@aulisa.com](mailto:paul.liu@aulisa.com)

**5.2. Trade/Device Name**

Guardian Angel Rx GA2000 Digital Vital Sign Monitoring System

**5.3. Regulatory Information**

Regulation Number	Regulation Name	Regulation Class	Product Code
21 CFR 870.2700	Oximeter	Class II	DQA
21 CFR 870.2910	Radiofrequency Physiological Signal Transmitter and Receiver	Class II	DRG

**5.4. Predicate Devices**Primary predicate

K040589, Avant 9600 Digital Pulse Oximeter, Nonin Medical, Inc.

### Reference devices

K150361, ViSi Mobile Monitoring System, Sotera Wireless, Inc.

K162580, Guardian Angel GA1000 Digital Vital Sign Monitoring System, Taiwan Aulisa Medical Devices Technologies, Inc.

## **5.5. Device Description**

The subject device is a digital vital sign monitoring system that measures and displays a patient's pulse rate and oxygen saturation (SpO<sub>2</sub>). It is also equipped with an audio/video camera to monitor the patient in real time. In addition, the subject device provides visual and auditory alarms that alert the caregiver when a patient's pulse rate and/or SpO<sub>2</sub> falls outside of preset limits or when a technical error is detected. During a physiological alarm event, the pulse rate and SpO<sub>2</sub> data along with the audio/video data are recorded automatically by the subject device. The caregiver can review the historical data whenever needed.

The system consists of a self-contained wrist-worn Sensor Module (SM), a Receiver/Transponder (RT) with an embedded audio/video camera and a portable, table-top wireless Display Unit (DU).

It uses non-invasive red and infrared technology to measure the pulse rate and SpO<sub>2</sub>. The measurements are taken by the SM and are transmitted to the RT which delivers the measurements along with audio/video signals to the DU for display, wherein Bluetooth technology is used to transmit data between the SM and the RT, and data is transmitted from the RT to the DU via the Wi-Fi 802.11 band.

## **5.6. Intended Use**

The Guardian Angel Rx GA2000 Digital Vital Sign Monitoring System is indicated for use in measuring and displaying functional oxygen saturation of arterial hemoglobin (SpO<sub>2</sub>) and pulse rate of adult and pediatric patients. It is indicated for spot-checking and/or continuous monitoring of patients during non-motion and under well-perfused conditions. The intended environments of use are hospitals, medical facilities, home care, and subacute environments. This system is a reusable device.

## **5.7. Comparison with Predicates**

The subject device has similar intended use and technology characteristics to the primary predicate, K040589, Avant 9600 Digital Pulse Oximeter, except that the subject device is indicated for a narrower range of patient population and environment of use than the primary predicate. Further, the subject device wirelessly transmits data to a remote display whereas the primary predicate does not.

The subject device uses the same Sensor Module (SM) and a similar Display Unit (DU) as currently configured on our reference device, K162580, Guardian Angel GA1000 Digital Vital Sign Monitoring System (cleared under), and uses the same Bluetooth technology for data transmission. The subject device has a Receiver/Transponder (RT) included to transmit the data from the Sensor Module to the Display Unit using the 802.11 radio frequency band which

extends the data transmission range of the system. A camera and microphone is incorporated in the RT, which transmits audio/video data via the same 802.11 band. The 802.11-based telemetry technology is equivalent to that used in the Secondary Predicate: ViSi Mobile Monitoring System (cleared under K150361).

The comparison table for the subject device versus the primary predicate, Avant 9600 (K040589), and the secondary predicate ViSi Mobile Monitoring System (K150361), is shown in **Table 5.1**.

**Table 5.1 – Comparison with Predicate**

Item	Subject Device	Primary Predicate	Reference Device	Reference Device
Indication for use	<p>The Guardian Angel Rx GA2000 Digital Vital Sign Monitoring System is indicated for use in measuring, recording, and displaying functional oxygen saturation of arterial hemoglobin (SpO<sub>2</sub>), pulse rate, and audio video signals of adult and pediatric patients. It is indicated for spot-checking and/or continuous monitoring of patients during non-motion and under well-perfused conditions. The intended environments of use are hospitals, medical facilities, home care, and subacute environments. This system is a reusable device.</p>	<p>The Nonin Avant 9600 Digital Pulse Oximeter is a portable, tabletop device indicated for use in simultaneously measuring, displaying, and recording functional oxygen saturation of arterial hemoglobin (SpO<sub>2</sub>) and pulse rate of adult, pediatric, infant, and neonatal patients in hospitals, medical facilities, home care, and subacute environments. It may also be used in patient transport, sleep laboratories, and EMS environments. The Avant 9600 is intended for continuous monitoring and / or spot-checking of patients during both no motion and motion conditions, for patients who are well or poorly perfused.</p>	<p>The ViSi Mobile Monitoring System is intended for use by clinicians and medically qualified personnel for single or multi-parameter vital signs monitoring of adult patients (18 years or older). It is indicated for ECG (3 or 5 lead-wire), respiration rate (RESP), heart rate (HR), noninvasive blood pressure (NIBP), continuous noninvasive blood pressure (cNIBP), noninvasive monitoring of functional oxygen saturation of arterial hemoglobin (SpO<sub>2</sub>), pulse rate (PR), and skin temperature (TEMP) in hospital-based facilities; including, general medical-surgical floors, intermediate care floors, and emergency departments. The ViSi Mobile Monitoring System may be used as standalone devices or networked to ViSi Mobile Remote Viewers through wireless 802.11 communication.</p>	<p>Guardian Angel GA1000 Digital Vital Sign Monitoring System (K162580)</p> <p>The Guardian Angel GA1000 Digital Vital Sign Monitoring System is indicated for use in measuring and displaying functional oxygen saturation of arterial hemoglobin (SpO<sub>2</sub>) and pulse rate of adult and pediatric patients. It is indicated for spot-checking and / or continuous monitoring of patients during non-motion and under well-perfused conditions. The intended environment of use is hospital. This system is a reusable device.</p>
Patient population	Adult and pediatric	Adult and pediatric, infant, and neonatal patients	Adult	Adults and pediatrics

## Section 5

## 510(k) Summary

Item	Subject Device		Primary Predicate		Reference Device		Reference Device	
		The Guardian Angel Rx GA2000 Digital Vital Sign Monitoring System		Avant 9600 Digital Pulse Oximeter (K040589) with Reusable Flex sensors		ViSi Mobile Monitoring System (K150361)		Guardian Angel GA1000 Digital Vital Sign Monitoring System (K162580)
Environment of use	Hospitals, medical facilities, home care, and subacute environments		Hospitals, medical facilities, home care, and subacute environments		Hospital-based facilities		Hospitals, medical facilities, home care, and subacute environments	
Out-of-hospital transport	No		Yes		Not specified		No	
Motion	Non-motion		Non-motion Motion		Not specified		Non-motion	
Perfusion	Well-perfused		Well-perfused Poorly-perfused		Not specified		Well-perfused	
Single-use or reusable	Reusable		Reusable		Reusable		Reusable	
Measurement	Pulse rate and SpO <sub>2</sub>		Pulse rate and SpO <sub>2</sub>		Pulse rate, SpO <sub>2</sub> , ECG, blood pressure, respiration rate and skin temperature		Pulse rate and SpO <sub>2</sub>	
Technology of pulse oximetry	Red and Infrared technology		Red and Infrared technology		Red and Infrared technology		Red and Infrared technology	
LED wavelengths & output power of pulse oximetry	Red: 660 nm @ 1.8 mw nominal Infrared: 905 nm @ 2 mw nominal		Red: 660 nm @ 0.8 mw max. avg. Infrared: 910 nm @ 1.2 mw max. avg.		Red: 660 nm @ 6.5 mw max. Infrared: 905 nm @ 5.2 mw max.		Red: 660 nm @ 1.8 mw nominal Infrared: 905 nm @ 2 mw nominal	
Accuracy (No motion)	SpO <sub>2</sub>	± 3 digits (70-100%)	SpO <sub>2</sub>	± 2 digits (70-100%)	SpO <sub>2</sub>	± 2 digits (70-100%)	SpO <sub>2</sub>	± 3 digits (70-100%)
	Pulse Rate	± 3% (30-290 bpm)	Pulse Rate	± 3 digits (18-300 bpm)	Pulse Rate	± 3 digits (30-240 bpm)	Pulse Rate	± 3% (30-290 bpm)
Displayed range	SpO <sub>2</sub>	1-100%	SpO <sub>2</sub>	0-100%	SpO <sub>2</sub>	0-100%	SpO <sub>2</sub>	1-100%
	Pulse Rate	30-290 bpm	Pulse Rate	18-300 bpm	Pulse Rate	0-240 bpm	Pulse Rate	30-290 bpm
Display	10.1" LCD		LED		OLED		7" LCD	
Alarms	Visual and auditory alarms		Visual and auditory alarms		Visual and auditory alarms		Visual and auditory alarms	
Power Supply	Lithium battery, AC adaptor		NiMH battery, AC adaptor		Lithium battery, AC adaptor		Lithium battery, AC adaptor	
Wireless technology / Data transmission	Bluetooth 802.11		None		802.11		Bluetooth	
Biocompatibility	Skin (surface) contact Prolonged contact		Skin (surface) contact Prolonged contact		Skin (surface) contact Prolonged contact		Skin (surface) contact Prolonged contact	

## 5.8. Summary of Performance Testing

The following performance data were provided in support of the substantial equivalence determination.

### Biocompatibility testing

The biocompatibility evaluation for the subject device was conducted in accordance with ISO 10993-1 "Biological Evaluation of Medical Devices – Part 1: Evaluation and Testing Within a Risk Management Process," as recognized by FDA. The battery of testing included the following tests:

- Cytotoxicity
- Sensitization
- Irritation

### Electrical safety and electromagnetic compatibility (EMC)

Electrical safety and EMC testing were conducted on the subject device. The device complies with the IEC 60601-1, IEC 60601-1-11, IEC 60601-1-8 and ISO 80601-2-61 standards for safety and the IEC 60601-1-2 standard for EMC.

### Software Verification and Validation Testing

Software verification and validation testing were conducted and documentation was provided as recommended by FDA's Guidance for Industry and FDA Staff, Guidance for the Content of Premarket Submissions for Software Contained in Medical Devices."

### Clinical Study

Clinical data were collected, according to Clause 201.12.1.101.2 and Annex EE.2 of ISO 80601-2-61:2011 Standard, to verify the accuracy of the subject device on healthy adult subjects over the range of 70%-100% SpO<sub>2</sub> through controlled induced hypoxia. Over 200 data points were collected. The A<sub>rms</sub> is less than 3 digits, compliant with FDA guidance on Pulse Oximeters – Premarket notification submissions [510(k)].

## 5.9. Substantially Equivalent Conclusion

Based on the non-clinical testing and clinical data summarized in this 510(k) submission, the results demonstrate that the subject device is substantially equivalent to the predicate. The differences do not raise different questions of safety or effectiveness when compared to the predicate.