



October 18, 2019

HeartVista, Inc.
% Mr. James J. Rogers
FDA Regulatory Affairs, Quality Assurance, and Clinical Studies
4984 El Camino Real, Suite 102
LOS ALTOS CA 94022

Re: K183274

Trade/Device Name: RTHawk, HeartVista Cardiac Package
Regulation Number: 21 CFR 892.1000
Regulation Name: Magnetic resonance diagnostic device
Regulatory Class: Class II
Product Code: LNH
Dated: September 19, 2019
Received: September 20, 2019

Dear Mr. Rogers:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see

<https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

For

Thalia T. Mills, Ph.D.
Director
Division of Radiological Health
OHT7: Office of In Vitro Diagnostics
and Radiological Health
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K183274

Device Name

RTHawk, HeartVista Cardiac Package

Indications for Use (Describe)

RTHawk is an accessory to 1.5T and 3.0T whole-body magnetic resonance diagnostic devices (MRDD or MR). It is intended to operate alongside, and in parallel with, the existing MR console to acquire traditional, real-time and accelerated images. The HeartVista Cardiac Package is a collection of RTHawk Apps designed to acquire, reconstruct and display cardiovascular MR (CMR) images.

RTHawk produces static and dynamic transverse, coronal, sagittal, and oblique cross-sectional images that display the internal structures and/or functions of the entire body. The images produced reflect the spatial distribution of nuclei exhibiting magnetic resonance. The magnetic resonance properties that determine image appearance are proton density, spin-lattice relaxation time (T1), spin-spin relaxation time (T2) and flow. When interpreted by a trained physician, these images provide information that may assist in the determination of a diagnosis.

RTHawk is intended for use as an accessory to the following MRI systems:

Manufacturer: GE Healthcare (GEHC)

Field Strength: 1.5T and 3.0T

Scanner Software Versions: 12, 15, 16, 23, 24, 25, 26

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary

RTHawk; HeartVista Cardiac Package

510(k) Number: K183274

Submitted in accordance with the requirements of SMDA 1990 and 21CFR 807.92.

1.0 Medical Establishment Registration

Medical Establishment Registration No.: 3011767965

2.0 Contact Information

James Jochen Rogers
FDA Regulatory Affairs, Quality Assurance, and Clinical Studies
T: 724.713.2298
E: jr@heartvista.com

3.0 Establishment Name and Address

HeartVista, Inc.
4984 El Camino Real, Suite 102
Los Altos, CA 94022

4.0 Submission Date

November 19, 2018, September 19, 2019, October 15, 2019

5.0 Device Information

Trade/Proprietary Name: RTHawk, HeartVista Cardiac Package

Common Name: RTHawk, HeartVista Cardiac Package

Model Number(s):

- HeartVista Cardiac Package (HVCP)
- RTHawk

Regulation Number: 892.1000

Regulation Name: Magnetic resonance diagnostic device (MRDD)

Regulatory Class: Class II

Device Classification Name: System, Nuclear Magnetic Resonance Imaging

Classification Panel: Radiology

Classification Product Code(s): LNH

6.0 Predicate Device(s)

510(k) #	Device	510(k) Sponsor	510(k) Clearance Date
K170090	RTHawk (ver 2.3.2) HeartVista Cardiac Package	HeartVista	07/14/2017

7.0 Device Description

RTHawk is a software system designed from the ground up to provide a platform for efficient real-time MRI data acquisition, data transfer, image reconstruction, and interactive scan control and display of static and dynamic MR imaging data.

RTHawk is as an accessory to clinical 1.5T and 3.0T MR systems, operating alongside, and in parallel with, the MR scanner console with no permanent physical modifications to the MRI system required.

RTHawk is designed to run on a stand-alone linux-based computer workstation, color monitor, keyboard and mouse. It is designed to operate alongside, and in parallel with, the existing MR console with no hardware modifications required to be made to the MR system or console. This RTHawk Workstation is sourced by the Customer in conformance with HeartVista-provided specifications, and verified prior to installation.

A private ethernet network connects the RTHawk workstation to the MR scanner computer. When not in use, the RTHawk workstation may be detached from the MR scanner with no detrimental, residual impact upon MR scanner function, operation, or throughput.

The RTHawk application is written to run on top of the Linux operating system, much like application software for word processing, accounting, graphics, etc. Additional software is installed on the MR scanner computer, for receiving communications and control commands from RTHawk, and for directing MRI raw data to RTHawk for image reconstruction, display and processing.

RTHawk is an easy-to-use, yet fully functional, MR Operating System environment. The RTHawk operating system has been designed to provide a platform for the real-time acquisition, control, reconstruction, display, and storage of high-quality static and dynamic MRI images and data.

Data is continuously acquired and displayed. By user interaction or data feedback, fundamental scan parameters can be modified. Real-time and high-resolution image acquisition methods are used throughout RTHawk for scan plane localization, for tracking of patient motion, for detection

of transient events, for on-the-fly, sub-second latency adjustment of image acquisition parameters (e.g., scan plane, flip angle, field-of-view, etc.) and for image visualization. Additional features are provided to automate and facilitate the set of tasks performed during a typical cardiac exam.

Conventional MR scanners queue an entire scan ahead of time and provide for little or no modification to a scan already in progress. Conversely, the RTHawk software prepares scan waveforms just as they are needed. RTHawk's efficient management of pulse sequence waveforms and instructions for modifying those pulse sequence waveforms uses the entire scanning interval for preparation of the next sequence. Scan parameters may be manipulated in real time, while providing all checks necessary to assure patient safety.

RTHawk makes extensive use of spiral image acquisition techniques to maximize scan efficiency. While conventional scans acquire data line-by-line in a Cartesian grid, RTHawk collects data more efficiently in a spiral pattern. Spiral-pattern raw data must be reformatted for correct reconstruction and display, requiring additional computing resources and image correction procedures to reduce image artifacts and distortions, ensuring high-quality reconstructed images.

RTHawk implements the conventional MRI concept of Protocols. Protocols are pre-set by HeartVista, but new protocols can be created and modified by the end user.

RTHawk Apps (Applications) are comprised of a pulse sequence, predefined fixed and adjustable parameters, reconstruction pipeline(s), and a tailored graphical user interface containing image visualization and scan control tools. RTHawk Apps may provide real-time interactive scanning, conventional (traditional) batch-mode scanning, accelerated scanning, or calibration functions, in which data acquired may be used to tune or optimize other Apps.

The HeartVista Cardiac Package is a collection of RTHawk APPs that enables the performance of a comprehensive cardiovascular MR (CMR) study in a clinically feasible amount of time. These APPs are designed and optimized to acquire, reconstruct, and display CMR images, with features including:

- On-the-fly, sub-second latency adjustment of image acquisition parameters (e.g., scan plane, flip angle, field-of-view, etc.)
- Real-time imaging, enabling less reliance on ECG gating and artifact suppression techniques. Real-time imaging may be used for scan plane localization, instantaneous tracking of patient motion, and clinical user observation of transient events
- Scan automation tools including automatic pushbutton localization of standard cardiac views, automatic determination of inversion time, automatic detection of artifacts, and automated myocardial segmentation

- High spatial resolution imaging, including single breath-hold, multi-slice high-resolution GRE app offering near total heart coverage
- Free-breathing, multi-slice SSFP and GRE apps that rapidly acquire high-quality images - potentially useful for patients who suffer from arrhythmia or who cannot hold their breath
- Multi-slice dynamic SR GRE app with one heartbeat temporal resolution for time-course imaging.
- Continuous flow quantification

The conventional MRI concept of anatomy- and indication-specific Protocols is implemented within the HeartVista Cardiac Package. APPs within the HeartVista Cardiac Package are organized into basic Protocols pre-set by HeartVista. The clinical user may modify APP parameters from default values within their ranges. These modified APPs may be saved into new or existing user-created Protocols to create unique CMR-indicated protocols tailored to the user’s clinical interests.

RTHawk operates compatible MR scanners within the safety parameters listed below:

Safety Parameter	Safety Level
Magnetic Field strength	1.5T, 3.0T
Operating Modes IEC 60601-2-33 (2010-03)	FIRST LEVEL CONTROLLED OPERATING MODE
Safety Parameter Display	SAR, dB/dt
Max SAR	< 4W/kg whole-body
Max dB/dt	FIRST LEVEL CONTROLLED OPERATING MODE

8.0 Indications for Use

RTHawk is an accessory to 1.5T and 3.0T whole-body magnetic resonance diagnostic devices (MRDD or MR). It is intended to operate alongside, and in parallel with, the existing MR console to acquire traditional, real-time and accelerated images. The HeartVista Cardiac Package is a collection of RTHawk Apps designed to acquire, reconstruct and display cardiovascular MR (CMR) images.

RTHawk produces static and dynamic transverse, coronal, sagittal, and oblique cross-sectional images that display the internal structures and/or functions of the entire body. The images produced reflect the spatial distribution of nuclei exhibiting magnetic resonance. The magnetic resonance properties that determine image appearance are proton density, spin-lattice

relaxation time (T1), spin-spin relaxation time (T2) and flow. When interpreted by a trained physician, these images provide information that may assist in the determination of a diagnosis.

RTHawk is intended for use as an accessory to the following MRI systems:

Manufacturer: GE Healthcare (GEHC)
Field Strength: 1.5T and 3.0T
Scanner Software Versions: 12, 15, 16, 23, 24, 25, 26

9.0 Performance Data - Discussion of Non-Clinical Tests

Design controls quality assurance measures during the development of RTHawk include:

- Code reviews
- Design reviews
- Unit and integration level testing
- Verification testing, including System and Manual testing
- Safety testing, including SAR, dB/dt, and acoustic noise
- Performance testing, including SNR and uniformity
- Validation testing

RTHawk has been designed to comply with the FDA Recognized Consensus Standards listed in the table below, as applicable to device features and components:

Reference #	Title
IEC 60601-2-33 Ed 3.0 (2010-03)	Medical electrical equipment - Part 2-33: Particular requirements for the basic safety and essential performance of magnetic resonance equipment for medical diagnostic (radiology).
MS1-2008	Determination of Signal-to-Noise Ratio (SNR) in Diagnostic Magnetic Resonance Imaging
MS3-2008	Determination of Image Uniformity in Diagnostic Magnetic Resonance Images
MS4-2010	Acoustic Noise Measurement Procedure for Diagnostic Magnetic Resonance Imaging Devices
MS8-2008	Characterization of the Specific Absorption Rate (SAR) for Magnetic Resonance Imaging Systems
NEMA PS3.1 - 3.20 (2011)	Digital Imaging And Communications In Medicine (DICOM) Set.

ISO 14971:2007	Medical Devices - Application Of Risk Management To Medical Devices
ES60601-1:2005/(R)2012 +A1 +C1 +A2	(Consolidated Text) Medical Electrical Equipment - Part 1: General Requirements For Basic Safety And Essential Performance (IEC 60601-1:2005, Mod). Section 14 Programmable Electrical Medical Systems (PEMS)

Risk management, compliant with ISO 14971:2007, identified hazards, sequences of events, and resultant harms; developed, implemented, and tested risk-controlling mitigations; and evaluated residual risks.

10.0 Technological Characteristics Comparison to Predicate Device and Discussion

Both the subject device and the predicate device software are intended as an accessory to GEHC 1.5T and 3.0T MRI systems, and are intended to integrate and interact seamlessly with the operating system software within those MRI systems. Both devices support all coils available on the specific installation's MRI console. Neither device supports software-controlled patient table movements and shifts. Both devices support remote access to and imaging on the specific installation's MRI system.

The structure of the subject device software is identical to the predicate device software, and is comprised of the following functional modules:

- Acquisition - responsible for the transfer of MR raw data from the MR scanner to the HeartVista Workstation
- Analysis - contains the image post-processing tools
- Application - HeartVista APPs. Each APP is comprised of a pulse sequence, user parameters, a reconstruction pipeline, and a specific user interface
- Information System - the central repository of all relevant MRI system configuration, patient, study, scan, etc., parameters associated with the current patient study
- Reconstruction - responsible for the efficient processing of raw data to generate MR images via a flexible, pipelined topology
- Scan Control - responsible for the real-time network transfer of controlling orders for APPs, APPs parameters modifications, and dynamic information from the MR host in response to user or program requests
- Sequencer - creates and provides a specific set of pulse sequence waveforms to control the MR scanner
- Storage - obtains current patient and scan information, performs non-volatile local storage, exports images and data in DICOM format, and logs events.

- Visualization - implements all aspects of the user interface, including APP selection, controls to modify APP parameters, image display, graphical slice prescription, and image review, save, and export.

As with the predicate device, RTHawk Apps (Applications) are comprised of a pulse sequence, predefined fixed and adjustable parameters, reconstruction pipeline(s), and a tailored graphical user interface containing image visualization and scan control tools. RTHawk Apps provide real-time interactive, batch-mode, and accelerated scanning, as well as calibration functions, in which data acquired may be used to tune or optimize other Apps. Orthogonal, oblique, and double oblique imaging planes are fully supported. The HeartVista Cardiac Package is a collection of RTHawk Apps designed to acquire, reconstruct and display cardiovascular MR (CMR) images.

In this version of the HeartVista Cardiac Package, notable changes include:

- The operating system intended for use with the HeartVista Cardiac Package is now Ubuntu 18.04 ("Bionic"), released April 2018.
- Additional compatibility with GE Signa scanners with software revision DV 26.
- The addition of new CMR Apps including:
 - B0 Mapping
 - Cardiac Localizer
 - Cardiac T2 Map
 - Cardiac T2* Map Cartesian
 - Cartesian Shimming
 - FB DE GRE
 - FB MS Tagging GRE
 - Multi-Slice Cine Flow
 - Multi-Slice DE SSFP
 - Noise Measurement
 - Stack-of-Spiral Cine Flow
- Additional variations on existing APPs, including
 - Automating the determination and prescription of standard cardiac views (short-axis, fourchamber, etc.)
 - The optional addition of L1-based denoising ("Compressed Sensing") to some Apps
 - The ability to automatically detect artifacts in some Apps
 - Automation of the breath-holding process through automatic voice
 - Automated determination of an optimal TI value from Cine DE Cal data
- New features intended to address cybersecurity concerns
- The ability to automatically segment LV endocardium, LV epicardium, and RV endocardium

Instructions for use are included within the device labeling, and the information provided enables the user to operate the device in a safe and effective manner. The subject device implements FDA Unique Device Identifier (UDI) labeling requirements. Both devices have equivalent statements of Intended Use.

The table below summarizes a comparison of the revised technological characteristics to the predicate device:

Attribute	Predicate Device K170090 RTHawk (ver 2.3.2) HeartVista Cardiac Package	Modified Device K183274 RTHawk (ver 2.5.1) HeartVista Cardiac Package
Indications for Use	<p>RTHawk is an accessory to 1.5T and 3.0T whole-body magnetic resonance diagnostic devices (MRDD or MR). It is intended to operate alongside, and in parallel with, the existing MR console to acquire traditional, real-time, and accelerated images. The HeartVista Cardiac Package is a collection of RTHawk Apps designed to acquire, reconstruct and display cardiovascular MR (CMR) images.</p> <p>RTHawk produces static and dynamic transverse, coronal, sagittal, and oblique cross-sectional images that display the internal structures and/or functions of the entire body. The images produced reflect the spatial distribution of nuclei exhibiting magnetic resonance. The magnetic resonance properties that determine image appearance are proton density, spin-lattice relaxation time (T1), spin-spin relaxation time (T2) and flow. When interpreted by a trained physician, these images provide information</p>	<p>RTHawk is an accessory to 1.5T and 3.0T whole-body magnetic resonance diagnostic devices (MRDD or MR). It is intended to operate alongside, and in parallel with, the existing MR console to acquire traditional, real-time, and accelerated images. The HeartVista Cardiac Package is a collection of RTHawk Apps designed to acquire, reconstruct and display cardiovascular MR (CMR) images.</p> <p>RTHawk produces static and dynamic transverse, coronal, sagittal, and oblique cross-sectional images that display the internal structures and/or functions of the entire body. The images produced reflect the spatial distribution of nuclei exhibiting magnetic resonance. The magnetic resonance properties that determine image appearance are proton density, spin-lattice relaxation time (T1), spin-spin relaxation time (T2) and flow. When interpreted by a trained physician, these images provide information</p>

	that may assist in the determination of a diagnosis. RTHawk is intended for use as an accessory to the following MRI systems: Manufacturer: GE Healthcare (GEHC) Field Strength: 1.5T and 3.0T Scanner Software Versions: 12, 15, 16, 23, 24, 25	that may assist in the determination of a diagnosis. RTHawk is intended for use as an accessory to the following MRI systems: Manufacturer: GE Healthcare (GEHC) Field Strength: 1.5T and 3.0T Scanner Software Versions: 12, 15, 16, 23, 24, 25, 26
Magnetic Field Strength(s)	1.5T, 3.0T	1.5T, 3.0T
Imaging Planes	Transverse, Coronal, Sagittal, Oblique, Double Oblique	Transverse, Coronal, Sagittal, Oblique, Double Oblique
Time Course Imaging	Yes	Yes
Parameter Mapping	T1, T2*	T1, T2, T2*
Ventricular Function	Free-Breathing and Breath-Held	Free-Breathing and Breath-Held
MDE	Free-Breathing and Breath-Held	Free-Breathing and Breath-Held
Gated MRA	Yes	Yes
Black Blood Imaging	Yes	Yes
SPAMM Tagging	No	Yes
Flow Imaging	Real-Time	Real-Time, Multi-Slice, and 4D
Remote Scanning and Support	Yes	Yes
Automated Scan Planning	No	Yes

11.0 Conclusions

Based upon verification testing and compliance with voluntary standards, the Company believes that RTHawk, and the HeartVista Cardiac Package, are substantially equivalent to the predicate device, and do not raise any new questions of safety or effectiveness.