



3D Systems, Inc.
Kim Torluemke
Vice President, Quality & Regulatory
5381 South Alkire Circle
Littleton, Colorado 80127

August 21, 2019

Re: K190044

Trade/Device Name: VSP® Orthopedics System
Regulation Number: 21 CFR 888.3030
Regulation Name: Single/Multiple Component Metallic Bone Fixation Appliances And Accessories
Regulatory Class: Class II
Product Code: PBF
Dated: July 23, 2019
Received: July 24, 2019

Dear Kim Torluemke:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for

devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

For, Shumaya Ali, MPH
Assistant Director
DHT6C: Division of Stereotaxic, Trauma
and Restorative Devices
OHT6: Office of Orthopedic Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)
K190044

Device Name
VSP® Orthopedics System

Indications for Use (Describe)

The VSP® Orthopedics System is intended to be used as a surgical instrument to assist in preoperative planning and/or in guiding the marking of bone and/or in guiding surgical instruments in non-acute, non-joint replacing osteotomies for adult patients in the distal femur, tibia, and non-sacrum pelvis.

Type of Use (Select one or both, as applicable)

☒ Prescription Use (Part 21 CFR 801 Subpart D)

☐ Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

Medical Modeling[™]

510(K) SUMMARY

1. INTRODUCTION

This document contains the 510(k) summary for the VSP[®] Orthopedics System. The content of this summary is based on the requirements of 21 CFR 807.92.

2. SUBMITTER

Name: 3D Systems, Inc.

Address: 5381 South Alkire Circle
Littleton, CO 80127, USA
Phone: (720) 643-1001
Fax: (720) 643-1009

Official Contact: Kim Torluemke
Vice President, Quality and Regulatory, Healthcare

Date Prepared: July 23, 2019

3. DEVICE

Trade Name: VSP[®] Orthopedics System

Common Name: Patient specific orthopedic anatomical models, templates, guides, and surgical plans.

Classification Name: Orthopaedic Surgical Planning and Instrument Guides

Classification: Class II, 21 CFR 888.3030

Product Code: PBF

4. PREDICATE DEVICES

Predicate device:

- SurgiCase Orthopaedics System, Materialise N.V. (K163156)

Reference devices:

- VSP[®] System, 3D Systems Inc. (K120956, K133907)
- VSP[®] Cranial System, 3D Systems Inc. (K151285)

Medical Modeling™

5. DESCRIPTION OF THE DEVICE

The VSP® Orthopedics System is intended to assist a surgeon with pre-operative planning and transfer of the pre-operative plan to the surgery in orthopedic procedures. The system contains several physical and digital outputs including patient-specific anatomical models, templates, and guides (physical outputs); and patient-specific surgical plans and digital files (digital or documentation outputs).

Outputs of the VSP® Orthopedics System are designed with physician input and reviewed by the physician prior to finalization and distribution.

The VSP® Orthopedics System also contains Stainless Steel Drill Inserts (VSP® Orthopedics System Accessories) which are intended to be used by the physician to guide drilling activities during the surgical procedure. The inserts fit into a standard hole in the cutting / drill guides and can be used across all VSP® Orthopedics System guides and templates.

6. INDICATIONS FOR USE

The VSP® Orthopedics System is intended to be used as a surgical instrument to assist in preoperative planning and/or in guiding the marking of bone and/or in guiding surgical instruments in non-acute, non-joint replacing osteotomies for adult patients in the distal femur, tibia, and non-sacrum pelvis.

7. COMPARISON OF TECHNOLOGICAL CHARACTERISTICS WITH THE PREDICATE DEVICE

The intended use and technological characteristics of the subject device (VSP® Orthopedics System) are either identical or substantially equivalent to the predicate device (Surgicase Orthopaedics System), differing only in the expansion of anatomical regions included within the VSP® Orthopedics System. The potential impact on substantial equivalence of each technological difference was addressed by risk analysis and verification and validation testing.

8. SUMMARY OF PERFORMANCE TESTING

The testing outlined below was intended to show that the output of the design and development process demonstrated compliance with the device specifications. Bench and cadaveric testing was conducted to prove the subject device performs in accordance with its intended use and is substantially equivalent to the listed predicate device.

The following testing was conducted for the VSP® Orthopedics System:

- Equipment Qualification (IQ/OQ/PQ)
- Process Qualification (PQ)
- Mechanical Testing
- Software Validation
- Cleaning Validation
- Sterilization Validation

Medical Modeling[™]

- Biocompatibility Validation
- Packaging Validation
- Shelf Life Validation
- Drop Test Validation
- Debris Validation
- Cadaver Study

Summary

All design, process, and other verification and validation testing, which were conducted as a result of risk analyses and design impact assessments, showed conformity to pre-established specifications and acceptance criteria. The acceptance criteria was established in support of device performance, and testing demonstrated substantial equivalence of the system to the predicate device.

9. CONCLUSION

Based on a comparison of the intended use and technological characteristics, the VSP[®] Orthopedics System is substantially equivalent to the identified predicate device. Minor differences in the indications for use, technological and performance characteristics do not raise new or different questions of safety and effectiveness. Additionally, the non-clinical testing supports that the system performs in accordance with its intended use and is as safe, as effective, and performs as well as the predicate device.