



July 19, 2019

Spes Medica S.r.l.  
Giorgio Facco  
Regulatory Affairs & Quality Assurance  
Via Europa - zona industriale  
Battipaglia (SA), 84091 It

Re: K190050

Trade/Device Name: Tech Dots - Conductive gel  
Regulation Number: 21 CFR 882.1275  
Regulation Name: Electroconductive Media  
Regulatory Class: Class II  
Product Code: GYB  
Dated: May 23, 2019  
Received: June 19, 2019

Dear Giorgio Facco:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal

statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Vivek Pinto, PhD  
Assistant Director, Acute Injury Devices Team  
DHT5B: Division of Neuromodulation  
and Rehabilitation Devices  
OHT5: Office of Neurological  
and Physical Medicine Devices  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)  
K190050

Device Name  
Tech Dots - Conductive gel

### Indications for Use (Describe)

Tech Dots are intended for use in clinical and research EEG/EP recordings from humans. They are used with external electrodes as the conductor between the scalp and recessed electrodes to reduce impedance between the electrode surface and the skin

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

	<b>Tech Dots – Conductive gel</b>	REV.	5	Date REV.	19.07.2019
		<b>510(k) Summary</b>			
<b>Traditional 510(k)</b>	<i>Spes Medica S.r.l. Via Europa , Zona Industriale – 84091 Battipaglia (SA)</i>				
Page 1 of 6					

## 510(k) Summary

<b>Manufacturer's Name:</b>	Spes Medica S.r.l. via Europa (Zona Ind.le), 84091 Battipaglia (SA) – Italy
<b>Official Correspondent:</b>	Giorgio Facco Quality Assurance and Regulatory Affairs
<b>Telephone Number:</b>	0039 0828 614191
<b>Fax Number:</b>	0039 0828 341788
<b>Trade Names:</b>	Tech Dots - Conductive gel
<b>Common or Usual Name:</b>	Conductive gel
<b>Classification Name:</b>	Media, Electroconductive
<b>Device Class:</b>	Class II
<b>Product Code:</b>	GYB
<b>Classification Regulation:</b>	882.1275
<b>Predicate Device:</b>	Electro-Gel 510(k) number: K111717
<b>Device Description:</b>	<p>Tech Dots are conductive gel dots to be used with external electrodes as the conductor between skin and electrode and to reduce impedance between the electrode surface and the skin.</p> <p>A single Tech Dot has a <math>11 \pm 1</math> mm diameter, <math>2.5 \pm 0.5</math> mm high, and weights <math>0.14 \pm 0.01</math> g.</p> <p>It's characterized by clear colour, no crystallization, no flocculation, no adverse smell, brightness.</p> <p>Tech Dots function is of conductor between the electrode used and the patient's skin and of getting the impedance lower for a better recording of the signal. Tech Dots are for use with external electrodes only.</p> <p>Tech Dot is made of Potassium Chloride as conductor, combined with thickening agents and humectants, all in an aqueous solvent.</p> <p>The composition is the following: <i>Water, Glycerol (vegetable origin), Polyacrylate co-polymer (proprietary), Potassium chloride</i></p> <p>The pH range is <math>4.1 \pm 0.1</math>, and Impedance at 10Hz is <math>80 \pm 10</math> Ohm.</p>

	<b>Tech Dots – Conductive gel</b>	<b>REV.</b>	5	<b>Date REV.</b>	19.07.2019
		<b>510(k) Summary</b>			
<b>Traditional 510(k)</b>	<i>Spes Medica S.r.l. Via Europa, Zona Industriale – 84091 Battipaglia (SA)</i>				
Page 2 of 6					

The Conductivity is 2 mS/cm

Shelf life of TechDots is 3 years if stored properly in sealed aluminum pouch and at the limits of temperature indicated on the labeling

**Intended Use:**

Tech Dots are intended for use in clinical and research EEG/EP recordings from humans. They are used with external electrodes as the conductor between the scalp and recessed electrodes to reduce impedance between the electrode surface and the skin

**Technological Comparison:**

The Tech Dots consist in conductive gel dots laid on a siliconized PET support. The characteristics of Tech Dots are substantially equivalent to the predicate device. No new questions of safety or effectiveness are raised.

Tech Dots employ the same technological characteristics as the predicate device with just different design of packaging: the predicate device is provided in jar (dot).

To support the technological comparison the ingredients, pH, impedance, weight and conductivity of the TechDots were evaluated internally and compared to the predicate device.


Both devices are water based with salt as conductive material and with thickening agents (Glycerin is used for both the products). TechDots do not contain any preservatives.

The pH of the TechDot is comparable to the Predicate: the pH evaluated by Spes Medica is 4÷5. The impedance is  $80 \pm 10$  Ohm, lower than the Predicate Device


**Substantial Equivalence:**

Tech Dots are equivalent to the device cleared under K111717 as is presented below in Table.

It has been shown in this 510(k) submission that the differences between Tech Dots and the predicate device Electro-Gel do not raise any questions regarding its safety and effectiveness. The Tech Dots device is substantially equivalent to the predicate device as it has the same intended use and similar technological characteristics as the previously cleared predicate devices.

	<b>Tech Dots – Conductive gel</b>	<b>REV.</b>	5	<b>Date REV.</b>	19.07.2019
		<b>510(k) Summary</b>			
<b>Traditional 510(k)</b>	<i>Spes Medica S.r.l.Via Europa , Zona Industriale – 84091 Battipaglia (SA)</i>				
Page 3 of 6					

Manufacturer	Electro-Cap international, Inc	Spes Medica S.r.l.	Discussion Differences
Trade Name	Electro-Gel	Tech Dots	
510(k) number	K111717	K190050	
Product Code	GYB	GYB	
Indications for use	Electro-Gel is intended for use in clinical and research EEG/EP recordings from humans. It is used with external electrodes as the conductor between the scalp and the (recessed) electrodes. It also reduces impedance (resistance to alternating current) between the electrode surface and the skin.	Tech Dots are intended for use in clinical and research EEG/EP recordings from humans. They are used with external electrodes as the conductor between the scalp and recessed electrodes to reduce impedance between the electrode surface and the skin	Same as predicate device
Regulation Name	Media, Electroconductive	Media, Electroconductive	Same as predicate device
Regulation Number	882.1275	882.1275	Same as predicate device
Environment of use	Electrophysiological	Electrophysiological	Same as predicate device
Intended user	Neurologists	Neurologists	Same as predicate device
Target patient	Adult and children	Adult and children	Same as predicate device
Where used	Topically on intact skin	Topically on intact skin	Same as predicate device
Conductive material	Salt (NaCl)	Salt (NaCl)	Same as predicate device
Thickening agent	Aragum, Glycerin	Sodium Acrylates Copolymers, Glycerin	Equivalent to predicate device
Sterilization method	Provide non sterile	Provide non sterile	Same as predicate device
Shelf-life	1 year	3 years	More than predicate device
Chemical Safety	No OSHA PEL	No OSHA PEL	Same as predicate device
Preservative	Methylparaben and Propylparaben	No preservative	Equivalent to predicate
Biocompatibility	Test in accordance with ISO 10993	Test in accordance with ISO 10993	Same as predicate device
Cytotoxicity	Yes	Yes	Same as predicate device
Irritation	Yes	Yes	Same as predicate device
Sensitization	Yes	Yes	Same as predicate device
Single Use	Yes	Yes	Same as predicate device
pH	4.5÷6.0	4÷5	Comparable to predicate device

	<b>Tech Dots – Conductive gel</b>	<b>REV.</b>	5	<b>Date REV.</b>	19.07.2019
		<b>510(k) Summary</b>			
<b>Traditional 510(k)</b>	<i>Spes Medica S.r.l.Via Europa , Zona Industriale – 84091 Battipaglia (SA)</i>		Page 4 of 6		

Manufacturer	Electro-Cap international, Inc	Spes Medica S.r.l.	Discussion Differences
Trade Name	Electro-Gel	Tech Dots	
510(k) number	K111717	K190050	
Product Code	GYB	GYB	
Impedance	0.5K Ohm	80 ± 10 Ohm	Less than predicate device
Weight	16, 32 or 128 ounces	0.14 ± 0.01 g per Dot	Different packaging and shape
Characteristics	Salt Base Non-irritating Non toxic	Salt Base Non-irritating Non toxic	Equivalent to predicate device
Packaging	PE	Aluminum/PET/PE	Different ways of packaging. Both materials are validated.

	<b>Tech Dots – Conductive gel</b>	REV.	5	Date REV.	19.07.2019
		<b>510(k) Summary</b>			
<b>Traditional 510(k)</b>	<i>Spes Medica S.r.l. Via Europa, Zona Industriale – 84091 Battipaglia (SA)</i>				
Page 5 of 6					

### Summary of Performance Testing-Biocompatibility

Spes Medica Tech Dots are no invasive product, the Biocompatibility Evaluation testing summarized below was conducted on Tech Dots to demonstrate compliance of this product to the following standards:

- ISO 10993-5 Biological evaluation of medical devices -- Part 5: Tests for in vitro cytotoxicity
- ISO 10993-10 Biological evaluation of medical devices -- Part 10: Tests for irritation and skin sensitization

Component Name	Type of contact			Material
	Skin	Blood	Tissue	
Gel	Y	N	N	Water, Glycerol (vegetable origin), Polyacrylate co-polymer (proprietary), Potassium chloride

Contact duration: >24h, <30days

### Performance Testing-Bench Testing

Performance Testing was performed on device characteristics of Spes Medica Tech Dots. This performance mechanical testing consisted of

- **Aging test**  
 The aim of this test was to validate the shelf life of 3 years through an accelerated aging procedure according to the ASTM F1980-16 "Standard guide for accelerated aging of sterile barrier system for medical devices").  
 Pass/fail criteria was fixed at the beginning of the test and all the result of the parameters evaluated (colour, odour, crystallization, flocculation, brightness, pH) comply according to the pass/fail criteria: TechDots should be characterized by clear colour, no crystallization, no flocculation, no adverse smell, brightness.  
 Also the impedance was evaluated and was found out to comply according to the ANSI/AAMIEC12:2000/(R)2015.
- **Long term conductivity**  
 The aim of this test was to evaluate the electrical performances (in terms of AC Impedance and DC Offset Voltage) of the product TechDots over time.  
 Pass/fail criteria were set at the beginning of the test according to ANSI/AAMIEC12:2000/(R)2015 : the DC Offset voltage should not exceed 100mV and AC Impedance should not exceed 2000 Ohm.  
 The parameters of AC Impedance and DC Offset Voltage comply according to the ANSI/AAMIEC12:2000/(R)2015 limits even after 7 days testing.

Spes Medica Tech Dots are tested internally for pH, impedance on a regular basis



	<b>Tech Dots – Conductive gel</b>	<b>REV.</b>	5	<b>Date REV.</b>	19.07.2019
		<b>510(k) Summary</b>			
<b>Traditional 510(k)</b>	<i>Spes Medica S.r.l. Via Europa , Zona Industriale – 84091 Battipaglia (SA)</i>		Page 6 of 6		

**Conclusion**

All performance testing conducted as outlined above demonstrate that the device meets the performance and design specifications.