



December 20, 2019

Fine Advanced Compound Co., Ltd
% Chris Park
General Manager
Med.com
1809 Holland Dr
Somerset, New Jersey 08873

Re: K190112

Trade/Device Name: Non-Sterile Zirconia Block (Model: Finebase, Montblanc, Trione HT, Trione C, Trione HT+)
Regulation Number: 21 CFR 872.6660
Regulation Name: Porcelain Powder For Clinical Use
Regulatory Class: Class II
Product Code: EIH
Dated: November 27, 2019
Received: December 03, 2019

Dear Chris Park:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

for Srinivas "Nandu" Nandkumar, Ph.D.
Director
DHT1B: Division of Dental Devices
OHT1: Office of Ophthalmic, Anesthesia,
Respiratory, ENT and Dental Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

Device Name

Non-Sterile Zirconia Block

Model: name: Finebase, Montblanc, Trione HT, Trione C, Trione HT+

Indications for Use (Describe)

Non-Sterile Zirconia Block (Model: Finebase, Montblanc, Trione HT, Trione C, Trione HT+) are indicated for the production of artificial teeth in fixed or removable dentures, or for jacket crowns, facings, and veneers.

Type of Use (Select one or both, as applicable)

☒ Prescription Use (Part 21 CFR 801 Subpart D)

☐ Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary

This 510(k) Summary is being submitted in accordance with requirements of SDMA 1990 and Title 21, CFR Section 807.92

The assigned 510(k) Number: K190112

1. Date of Preparation: 11/20/2019
2. Sponsor Identification
FINE ADVANCED COMPOUND Co., Ltd
Address: < 7, Galgot-gil, Jinwi-myeon, Pyeongtaek-si, Gyeonggi-do, 451-862, Korea
Establishment Registration Number: Not yet registered for the Number
Contact Person: Jeong Kwang Ho
Position: General Manager
Tel: 82-31-358-9312
Fax: 82-31-376-3816
Email:<jkh4456@finematerial.co.kr>

3. Identification of Proposed Device
Trade Name: Non-Sterile Zirconia Block
Common Name: Dental restorative material, porcelain powder/blocks

Regulatory Information

Classification Name: Powder, Porcelain

Classification: Class II

Product Code: EIH Regulation Number: 872.6660

Review Panel: Dental

Indication for Use Statement:

Non-Sterile Zirconia Block (Model name: Finebase, Montblanc, Trione HT, Trione C, Trione HT+) are indicated for the production of artificial teeth in fixed or removable dentures, or for jacket crowns, facings, and veneers.

Device Description

Non-Sterile Zirconia Block (Model name: Finebase, Montblanc, Trione HT, Trione C, Trione HT+) is a manufacture unit that is cut, processed and sintered by CAD/CAM system of the computers designed for production of dental restoration.

4. Identification of Predicate Device(s)
Primary predicate
- 510(k) number: K180252
- Company: Crown Porcelain Dental Technology Co., Ltd.
- Model: Crown Dental Zirconia Blank & Crown Dental Zirconia Pre-Shaded Blank
Wieland Dental + Technik GmbH & Co. KG, ZENO Zr Disc
- Classification: Class II

5. Non-Clinical Test Conclusion

Non clinical tests were conducted to verify that the proposed device met all design specifications as was Substantially Equivalent (SE) to the predicate device, including

- Performance Test per ISO 6872 Fourth Edition 2015-06-01, Dentistry - Ceramic Materials.
- Density Test
- Cytotoxicity per ISO 10993-5:2009;
- Intracutaneous Reactivity Test per ISO 10993-10:2010;
- Sensitization Test per ISO 10993-10:2010
- Acute Systematic Toxicity per ISO 10993-11:2006;
- Genotoxicity Tests per ISO 10993-3:2014.

6. Clinical Test Conclusion

No clinical study is included in this submission.

7. Substantially Equivalent (SE) Comparison

Table 1 Comparison of Technology Characteristics

ITEM	Non-Sterile Zirconia Block (Model name: Finebase, Montblanc, Trione HT, Trione C, Trione HT+)	Primary Predicate K180252	Differences
Product Code	EIH	EIH	None
Regulation Number	872.6660	872.6660	Same
Indication for use	Non-Sterile Zirconia Block (Model name: Finebase, Montblanc, Trione HT, Trione C, Trione HT+) are indicated for the production of artificial teeth in fixed or removable dentures, or for jacket crowns, facings, and veneers.	Crown Dental Zirconia Blank & Crown Dental Zirconia Pre-Shaded Blank are indicated for the production of artificial teeth in fixed or removable dentures, or for jacket crowns, facings, and veneers.	Indication for use is same as predicates.
Feature	Colored	Colored	Same
Shape	Discs	Discs	Same
Type and Class per ISO 6872: 2015	Type II Class 5	Type II Class 5	Same
Sterility	Non-sterile	Non-sterile	Same
Crystal Morphology	Tetragonal	Tetragonal	Same
Density	6.00g/cm ³	6.00g/cm ³	Same
Sintering temperature	1500±50°C	1500±50°C	Same
Performance	Comply with ISO 6872	Comply with ISO 6872	Same
Contact Level	surface device with permanent contact	surface device with permanent contact	Same
Biocompatibility	Tested for Cytotoxicity, irritation, sensitization, accurate systematic toxicity, genotoxicity, no adverse react	Tested for Cytotoxicity, irritation, sensitization, accurate systematic toxicity, genotoxicity, no adverse react	Same

Subject device, Crown Dental Zirconia Blank & Crown Dental Zirconia Pre-Shaded Blank is very similar to the predicate device.

Substantially Equivalent (SE) Conclusion

Based on the comparison and analysis above, the proposed devices are determined to be Substantially Equivalent (SE) to the predicate devices.