



May 19, 2019

T.A.G. Medical Products Corporation, Ltd.
Anat Rozen
RA Manager
Gaaton 2513000, Israel

Re: K190125

Trade/Device Name: FiberStitch™ Implant, Curved with two Polyester Implants and 2-0 FiberWire®,
FiberStitch™ Implant, Straight with two Polyester Implants and 2-0 FiberWire®
Regulation Number: 21 CFR 888.3040
Regulation Name: Smooth or threaded metallic bone fixation fastener
Regulatory Class: Class II
Product Code: MBI
Dated: April 3, 2019
Received: April 8, 2019

Dear Anat Rozen:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part

801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/CombinationProducts/GuidanceRegulatoryInformation/ucm597488.htm>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/>) and CDRH Learn (<http://www.fda.gov/Training/CDRHLearn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<http://www.fda.gov/DICE>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Laurence D. Coyne, Ph.D.
Assistant Director
DHT6C: Division of Stereotaxic, Trauma
and Restorative Devices
OHT6: Office of Orthopedic Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

DEPARTMENT OF HEALTH AND HUMAN SERVICES
Food and Drug AdministrationForm Approved: OMB No. 0910-0120
Expiration Date: 06/30/2020
See PRA Statement below.**Indications for Use**

510(k) Number (if known)

K190125

Device Name

FiberStitch™ Implant, Curved with two Polyester Implants and 2-0 FiberWire®

FiberStitch™ Implant, Straight with two Polyester Implants and 2-0 FiberWire®

Indications for Use (Describe)

The FiberStitch™ Meniscal Repair Device is intended for use as a suture retention device to facilitate endoscopic soft tissue procedures.

The FiberStitch™ Meniscal Repair Device is indicated for use in meniscal repair procedures.

Type of Use (Select one or both, as applicable)

☒ Prescription Use (Part 21 CFR 801 Subpart D)☐ Over-The-Counter Use (21 CFR 801 Subpart C)**CONTINUE ON A SEPARATE PAGE IF NEEDED.**

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."



510(k) Summary

Pursuant to CFR 807.92, the following 510(k) Summary is provided:

- 1. Submitter Address:** Shlomi Dines
T.A.G. Medical Products Corporation, Ltd.
Gaaton 2513000, ISRAEL
www.tag-med.com

Mfg. Phone: Tel.: 972-4-9858400

Contact Person: Anat Rozen

Date: April 03, 2019
- 2. Device & Classification Name:** Suture Anchor, class II, 21 CFR 888.3040 Fastener, Fixation, Nondegradable, Soft Tissue, product code MBI
FiberStitch™ Implant, Curved with two Polyester Implants and 2-0 FiberWire®
FiberStitch™ Implant, Straight with two Polyester Implants and 2-0 FiberWire®
- 3. Predicate Devices:** K121861- Smith & Nephew - ULTRA FAST-FIX Meniscal Repair System
K111564- Biomet Sports Medicine - Maxfire Marxmen Meniscal Repair Device
K132043- Arthrex Inc. - Arthrex SpeedCinch
K943949- LOOK – Polyviolene Surgical Suture
- 4. Description:** The FiberStitch™ devices are an all-inside meniscal repair device. The devices include two non-absorbable polyester implants, pre-tied with #2-0 non-absorbable sutures and preloaded into a needle delivery system. The adjustable depth penetration limiter is preset to approximately 18mm from the tip of the needle. It can be adjusted down in 2 (mm) increments to approximately 10mm.
- 5. Intended Use:** The FiberStitch™ Meniscal Repair Device is intended for use as a suture retention device to facilitate endoscopic soft tissue procedures.
The FiberStitch™ Meniscal Repair Device is indicated for use in meniscal repair procedures.
- 6. Comparison of Technological Characteristics:** With respect to its indication for use, the FiberStitch™ Implant, Curved with two Polyester Implants and 2-0 FiberWire® and FiberStitch™ Implant, Straight with two Polyester Implants and 2-0 FiberWire® are substantially equivalent to its predicate devices in their design which allows for the devices to be endoscopically delivered from a single access point as the proposed FiberStitch™ Implant, Curved with two Polyester Implants and 2-0 FiberWire® and the FiberStitch™ Implant, Straight with two Polyester Implants and 2-0 FiberWire® devices. Sutures material is similar to Maxfire Marxmen Meniscal Repair Device. The clinical use is similar; devices are considered all-inside meniscal repair devices. Conversely, FiberStitch™ Implant, Curved with two Polyester Implants and 2-0 FiberWire® and the FiberStitch™ Implant, Straight with two Polyester Implants and 2-0 FiberWire® are different than their predicate devices in geometry and details of the operating mechanism while general dimensions are same. The deployment of the sutures in FiberStitch™ Implant, Curved with two Polyester Implants and 2-0 FiberWire® and the FiberStitch™ Implant, Straight with two Polyester Implants and 2-0 FiberWire® is accomplished by rotating the wheel while for the predicate



MEDICAL PRODUCTS

devices deployment of sutures is accomplished by pushing the dedicated trigger. Suture materials are different albeit they are biocompatible.

Nonclinical test discussion:

Nonclinical testing was completed to demonstrate that the FiberStitch™ Implant, Curved with two Polyester Implants and 2-0 FiberWire® and FiberStitch™ Implant, Straight with two Polyester Implants and 2-0 FiberWire® meets the established performance characteristics, and to verify that design requirements are satisfied. Testing included biocompatibility evaluation per ISO 10993-1, ethylene oxide sterilization validation, and package qualification. Device testing included surface/visual, dimensional, mechanical and functional testing. It was concluded that the FiberStitch™ Implant, Curved with two Polyester Implants and 2-0 FiberWire® and the FiberStitch™ Implant, Straight with two Polyester Implants and 2-0 FiberWire® are as safe and effective as the predicate devices.