



October 15, 2019

Avenir Performance Europeenne Medical (APEM)
% Om Singh
Senior Scientific Consultant
Technology Sciences Group, Inc.
1150 18th Street NW, Suite 1000
Washington, District of Columbia 20036

Re: K190149
Trade/Device Name: Vial Adapter Ø20 mm
Regulation Number: 21 CFR 880.5440
Regulation Name: Intravascular Administration Set
Regulatory Class: Class II
Product Code: LHI
Dated: September 12, 2019
Received: September 13, 2019

Dear Om Singh:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal

statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

for Geeta Pamidimukkala
Acting Assistant Director
DHT3C: Division of Drug Delivery and
General Hospital Devices,
and Human Factors
OHT3: Office of Gastrorenal, ObGyn,
General Hospital and Urology Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K190149

Device Name

Vial Adapter Ø20mm

Indications for Use (Describe)

Vial Adapter Ø20mm are indicated for the transfer and mixing of drugs contained in vials.

Type of Use (Select one or both, as applicable)

☒ Prescription Use (Part 21 CFR 801 Subpart D)

☐ Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."



510(K) Summary – K190149

1. Submitter: Avenir Performance Européenne Medical (APEM)
ZI route de Souppes
77570 Château Landon
France
Tel : +330164785100
Fax : +330164785109

Contact Person: Mrs. Fatoumata DAO
Quality Manager
Office phone: 0033164785106
Cell Phone: 0033643953528
Fax: 0033164785109
E-mail: raq@ape-medical.fr

Preparing Date: Oct 15, 2019

US Agent:

Contact: Om V. Singh, Ph.D. Senior Regulatory Consultant
Address: 1150 18th Street NW, Suite 1000
Washington, DC 20036, USA
Office Phone: +1 202-828-8983
Fax: +1 202-872-0745
Email: om.singh@tsgconsulting.com

2. Device name and classification:

Regulation Name: Intravascular Administration Set

Trade Name: Vial Adapter Ø20mm

Common/ Usual Name: Vial Adapter

Product Code: LHI

Regulation No.: 880.5440



Regulatory Class: II

Panel identification: General Hospital Panel

3. Predicate Device: Vial Adapter 15mm - K171796 (Medimop Medical Project Ltd.)

4. Device Description:

The vial adapter (VA) are sterile polycarbonate molded devices that allows easy transfer of fluids into and out of drug vials. It incorporates a siliconized hollow spike for puncturing the stopper in the neck of the vial and a luer fitting that allows connection of a syringe on opposite side. After puncturing the hollow spike seats securely around the ferrule of drug vial utilizing the “legs” of the vial adapter. The opposite side of the vial adapter contains a luer fitting for the connection of a standard Luer Lock syringe for the reconstitution and removal of the contents of the drug vial. The proposed VA is available in Ø20mm to accommodate Ø20 drug vials and is available in 3 configurations, no filter, in-line 5-micron or 15-micron disc filter sub- assembly for particulate filtration.

5. Indication for use:

Vial Adapter Ø20mm are indicated for the transfer and mixing of drugs contained in vials.

6. Intended use:

Vial Adapter are intended for use in healthcare facilities or in-home environment by the patient or care-giver to aid and support prescribed treatment and therapy.

7. Technological Characteristics and Substantial Equivalence:

The proposed device, Vial Adapter Ø20mm has similar indications for use and the same principle of operation as the predicate device, Vial Adapter (K171796). However, the following differences were noted in device design:

The proposed device is offered in different size, i.e., Ø20mm, however this difference is for a different diameter vial, but the intended use is the same. The size difference of proposed device led to the difference in packaging size from the predicate. However, the sterilization process is validated to a minimum SAL 10^{-6} . The proposed device can be equipped with an in-line 5-micron or 15-micron disc filter sub- assembly for particulate filtration. The performance testing of the filters met required acceptable criteria. These differences in device design do not raise different questions of safety and effectiveness.



The predicate and proposed device share the same principle of operation and fundamental technology when utilizing the proposed device. The following table compares the Vial Adapter Ø20 mm to the Predicate Device Vial Adapter 15 mm with respect to intended use, technological characteristics and principles of operation, providing more detailed information regarding the basis for the determination of substantial equivalence.

The vial size to be used with the subject device is different than that of the predicate, however this difference does not raise any different questions of safety or effectiveness.

Comparison of proposed Vial Adaptor Ø20 mm to the predicate device Vial Adapter 15 mm (K171796)

Parameter	Subject Device (K190149) Vial Adapter Ø20 mm	Predicate (K171796) Vial Adapter 15mm (Medimop Medical Project Ltd.)	Comparison
Regulation Number	21 CFR 880.5440	21 CFR 880.5440	Same
Regulation Name	Intravascular administration set	Intravascular administration set	Same
Regulatory Class	II	II	Same
Product Code	LHI	LHI	Same
Indications for Use	Vial Adapter Ø20 mm are indicated for the transfer and mixing of drugs contained in vials.	The Vial Adapter 15mm is indicated for the transfer and mixing of drugs contained in vials.	Size difference
Sterilization Method	Gamma Irradiation	Gamma Irradiation	Same
Sterility Assurance Level	SAL 10 ⁻⁶	SAL 10 ⁻⁶	Same
Single use	Yes	Yes	Same
Body Material	Polycarbonate	Polycarbonate	Same



Expiration Date	5 years	5 years	Same
Vial Adapter Size	20 mm	15 mm	Different
Labeling	Proposed device labeling (IFU) includes transfer and mixing instructions	Predicate device labeling (IFU) includes transfer and mixing instructions	Same
Piercing Spike	Plastic - Single Lumen	Plastic - Single Lumen	Same
Vial Adapter Fit (Vial Side)	Snap Fit to Vial “Tight Grip”	Snap Fit to Vial “Tight Grip” Feature	Same
Packaging Size	Dimensioned for VA 20mm	Dimensioned for VA 15mm	Different
Filtration	5-micron or 15-micron disc filter configuration	No filter	Different

8. Substantial Equivalence:

The proposed device, Vial Adapter Ø20mm, has the same intended use and substantially similar indications for use, technological characteristics and principles of operation as the identified predicate devices. Difference in the size of vial adapter Ø20mm is for a different diameter vial, but intended use is the same. Based on the size difference, the packaging size differ from the predicate device. The sterilization process is validated to a minimum SAL 10^{-6} . The configurations with in-line 5-micron or 15-micron disc filter sub-assembly for particulate filtration is different from the predicate. However, the performance testing of the filters met required acceptable criteria. There are no significant differences between the Vial Adapter Ø20mm and the predicate device that would adversely affect the use of the product. Any differences in technological characteristics do no raise new questions of safety or effectiveness. Further, testing demonstrates substantially equivalent performance between the device and the predicates. Accordingly, the Vial Adapter Ø20mm is substantially equivalent to the predicate device in design, function, material, and intended use.

9. Summary of Non-Clinical Testing:

The vial adapter was evaluated, and risk assessment was performed to ensure that the device modifications did not introduce any new potential risks.

The following tests were performed as a result of the risk assessment to ensure that all potential risks associated with the device design are mitigated to acceptable levels.

Non-Clinical/ Performance testing Summary

Performance Testing Summary

A: Packaging		Testing standard
A1	Seal integrity test by dye penetration	ISO 11607-1
A2	Sealing strength (resistance)	ISO 11607-1
A3	Peeling open characteristics test	ISO 11607-1
B: Mechanical tests and leaks		Testing standard
B1	VA Snapping force	Internal performance standards
B2	VA Unsnapping force	Internal performance standards
B3	Leak Spike/Vial	ISO 8871-5:2016(F)
B6	Spike ductility	Internal performance standards
D: Fluid flow		Testing standard
D1	Flow injection from syringe to Vial	Internal performance standards
D2	Flow aspiration from Vial to syringe	Internal performance standards
D3	Dead volume	Internal performance standards

E: Filter		Testing standard
E1	Filter integrity	Internal performance standards
E2	Filter holding force / VA	Internal performance standards
E3	Leak external filter / internal VA	Internal performance standards
E3	Particle Filtration Efficiency (Particulate)	Internal specification
F: Luer Lock		Testing standard
F1	Gauging	ISO 80369-7
F2	Liquid leak checking of the liquid-tightness of the cone Luer	ISO 80369-7
F3	Air leak checking of the air-tightness of the cone Luer	ISO 80369-7
F4	Separation force	ISO 80369-7
F5	Unscrewing torque force	ISO 80369-7
F6	Friendly assembly	ISO 80369-7
F7	Thread resistance	ISO 80369-7
F8	Cracks and change color	ISO 80369-7

Particulate Testing per USP <788>

All testing met the required acceptance criteria.

10. Clinical Testing:

There was no clinical testing required to support the medical device as the Indications for Use is equivalent to the predicate device. The substantial equivalence of the device is supported by the non-clinical testing.

11. Sterilization:



Vial Adapter Ø20mm are sterilized by Gamma irradiation in accordance to standard ISO 11137-1, ISO 11137-2 and AAMI TIR 33. The sterilization subcontractor is STERIS AST (Applied Sterilization Technologies). The sterilization process is validated to a minimum SAL 10^{-6} .

12. Biocompatibility:

LAL Testing via Bacterial Endotoxin

Vial Adapter Ø20mm was tested for bacterial endotoxin by a validated Limulus Amoebocyte Lysate (LAL) method.

The biocompatibility assessments were conducted according to ISO 10993-1

The following biocompatibility testing has been successfully completed.

- a. Cytotoxicity (Tested to ISO 10993-5)
- b. Sensitization (Tested to ISO 10993-10)
- c. ASTM Hemolysis (Tested to ASTM F756 and ISO 10993-4)
- d. Intracutaneous Reactivity (Tested to ISO 10993-10)
- e. Systemic Toxicity (Acute Systemic Injection) (Tested to ISO 10993-11)
- f. Rabbit Pyrogen Test (Tested to ISO 10993-11)

13. Conclusion:

The evaluation of the Vial Adapter Ø20mm through performance testing demonstrates that the device is substantially equivalent to the predicate device.