



May 9, 2019

ReFlow Medical, Inc.
Rebecca Pine
Official Correspondent
208 Avenida Fabricante #100
San Clemente, California 92672

Re: K190393
Trade/Device Name: Wingman 14C Crossing Catheter
Regulation Number: 21 CFR 870.1250
Regulation Name: Percutaneous Catheter
Regulatory Class: Class II
Product Code: DQY
Dated: April 9, 2019
Received: April 10, 2019

Dear Ms. Pine:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for

devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/CombinationProducts/GuidanceRegulatoryInformation/ucm597488.htm>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/>) and CDRH Learn (<http://www.fda.gov/Training/CDRHLearn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<http://www.fda.gov/DICE>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

for Kenneth J. Cavanaugh, Ph.D.
Director (Acting)
DHT2C: Division of Coronary
and Peripheral Intervention Devices
OHT2: Office of Cardiovascular Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K190393

Device Name

Wingman 14C Crossing Catheter

Indications for Use (Describe)

The Wingman 14C Crossing Catheter is intended to be used in conjunction with steerable guidewires to access discreet regions of the peripheral and/or coronary vasculature. It may be used to facilitate placement and exchange of guidewires and other interventional devices and provide a conduit for the delivery of saline solutions or diagnostic/therapeutic agents.

Type of Use (Select one or both, as applicable)

☒ Prescription Use (Part 21 CFR 801 Subpart D)

☐ Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary

I. SUBMITTER

ReFlow Medical, Inc.
208 Avenida Fabricante #100
San Clemente, CA 92672

Contact person: Rebecca K Pine
Phone: (760) 809-5178
Fax: (760) 290.3216
Date prepared: April 15, 2019

II. DEVICE

Name of the device: Wingman 14C Crossing Catheter
Common of usual name: Support Catheter
Classification name: Percutaneous Catheter
Regulatory Class: 2
Product Code: DQY

III. PREDICATE DEVICE

Wingman 14C Crossing Catheter (K170996)
This predicate has not been subject to a design-related recall

IV. DEVICE DESCRIPTION

The Wingman 14C Crossing Catheter is a device intended to provide additional support to a steerable guidewire when accessing discrete regions of the peripheral and/or coronary vasculature.

The device consists of a support catheter, with a concealed radiopaque beveled guide-tip, and activating handle. The through-lumen of the device can serve as a conduit for the delivery of diagnostic and therapeutic agents.

V. INDICATIONS FOR USE

The Wingman 14C Crossing Catheter is intended to be used in conjunction with steerable guidewires to access discreet regions of the peripheral and/or coronary vasculature. It may be used to facilitate placement and exchange of guidewires and other interventional devices and provide a conduit for the delivery of saline solutions or diagnostic/therapeutic agents.

VI. COMPARISON OF TECHNOLOGICAL CHARACTERISTICS WITH THE PREDICATE DEVICE

The technological characteristics of the Wingman 14C Crossing Catheter is highly analogous to the technological characteristics of the Wingman 14C Wingman Crossing Catheter previously cleared (K170996) version of the device.

At a high level, the subject and predicate devices are based on the following same technological elements:

- all delivered to the target site using an over-the-wire percutaneous technique
- all have a through lumen to allow passage and exchange of guidewires
- all have a smooth inner lumen to provide reduced friction for guidewire movement
- all have a polymer catheter shaft with specific geometry to control the torque and push movements associated with lesion crossing
- all use a specialized beveled distal tip to facilitate crossing of the lesion

The following technological differences exist between the subject and predicate devices:

- The extendable tip component of the subject Wingman 14C Crossing Catheter is coated with parylene

VII. PERFORMANCE DATA

The following performance data were provided in support of the substantial equivalence.

- Biocompatibility Testing
- Particulate Characterization Testing
- Simulated Use Testing
- Corrosion Resistance Testing
- Flexibility and Kink
- Flow Rate
- Dimensional Verification
- Tensile Strength
- Torque Strength
-

The Wingman 14C Crossing Catheter met all specified criteria and did not raise new safety or performance questions. Based on the performance testing the Wingman 14C Crossing Catheter was found to have a safety and effectiveness profile that is similar to the predicate device.

VIII. CONCLUSIONS

The design testing performed for the Wingman 14C Crossing Catheter demonstrated that the performance of the device is substantially equivalent to the legally marketed predicate devices.