



September 26, 2019

Shen Wei USA Inc.
Cheryl Reep
Compliance Director
33278 Central Ave Ste 102
Union City, California 94587

Re: K190466

Trade/Device Name: PF Nitrile Exam Glove with pH coating, Black-Gray & White-Black Colors,
Tested for Use with Chemotherapy Drugs and Fentanyl Citrate

Regulation Number: 21 CFR 880.6250

Regulation Name: Non-Powdered Patient Examination Glove

Regulatory Class: Class I

Product Code: LZA, LZC, QDO

Dated: September 3, 2019

Received: September 9, 2019

Dear Cheryl Reep:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal

statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

For Elizabeth Claverie-Williams, MS
Assistant Director for THT4B2
Acting Assistant Director for THT4B1
DHT4B: Division of Infection Control
and Plastic Surgery Devices
OHT4: Office of Surgical
and Infection Control Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K190466

Device Name

Powder-Free Nitrile Examination Glove with pH coating, Black-Gray and White-Black Colors, Tested for Use with Chemotherapy Drugs and Fentanyl Citrate

Indications for Use (Describe)

A disposable device intended for medical purpose that is worn on the examiner's hand to prevent contamination between patient and examiner. Tested for use with chemotherapy drugs and Fentanyl Citrate.

Chemotherapy Drug and Concentration	White-Black	Black-Gray
	Breakthrough	Detection Time in Minutes
Carmustine (BCNU), 3.3mg/ml	27.8	27.1
Cisplatin, 1.0mg/ml	>240	>240
Cyclophosphamide (Cytosan), 20mg/ml	>240	>240
Cyclosporin A, 100 mg/ml	>240	not tested
Dacarbazine (DTIC) 10g mg/ml	>240	>240
Doxorubicin Hydrochloride, 2.0 mg/ml	>240	>240
Etoposide (Toposar), 20mg/ml	>240	>240
Fluorouracil 50mg/ml	>240	>240
Paclitaxel (Taxol), 6.0mg/ml	>240	>240
Thiotepa, 10mg/ml	77.9	27.4
Vincristine Sulfate, 1.0mg/ml	>240	not tested
Fentanyl Citrate, 100mcg/2ml	>240	>240

WARNING: Do not use with Carmustine and Thiotepa.

Please note that Carmustine (BCNU) has low permeation of 27.8 minutes for White-Black color and 27.1 minutes for Black-Gray color; Thiotepa has 77.9 minutes for White-Black color and 27.4 minutes for Black-Gray color.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRAStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."