



June 6, 2019

bioMerieux, Inc.  
Cherece Jones  
Staff Regulatory Affairs Specialist  
595 Anglum Road  
Hazelwood, Missouri 63042

Re: K190616

Trade/Device Name: VITEK 2 AST-Gram Positive Dalbavancin ( $\leq 0.015$  -  $\geq 1$   $\mu\text{g/mL}$ )  
Regulation Number: 21 CFR 866.1645  
Regulation Name: Fully automated short-term incubation cycle antimicrobial susceptibility system  
Regulatory Class: Class II  
Product Code: LON, LTW, LTT  
Dated: March 8, 2019  
Received: March 11, 2019

Dear Cherece Jones:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's

requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801 and Part 809); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

for

Uwe Scherf, M.Sc., Ph.D.  
Director  
Division of Microbiology Devices  
OHT7: Office of In Vitro Diagnostics  
and Radiological Health  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health

Enclosure



## 510(k) SUMMARY

### VITEK<sup>®</sup> 2 AST-Gram Positive Dalbavancin ( $\leq 0.015$ - $\geq 1$ $\mu\text{g/mL}$ )

#### 510(k) Submission Information:

Submitter's Name: bioMérieux, Inc.  
Address: 595 Anglum Road  
Hazelwood, MO 63042  
Contact Person: Cherece L. Jones  
Staff Regulatory Affairs Specialist  
Phone Number: 314 -731-8684  
Fax Number: 314-731-8689  
Date of Preparation: March 8, 2019

#### B. Device Name:

Formal/Trade Name: VITEK<sup>®</sup> 2 AST-Gram Positive Dalbavancin ( $\leq 0.015$  -  $\geq 1$   $\mu\text{g/mL}$ )  
Classification Name: 21 CFR 866.1645  
Fully Automated Short-Term Incubation Cycle  
Antimicrobial Susceptibility System  
Product Code: LON  
Subsequent Product Codes: LTT LTW  
Common Name: VITEK<sup>®</sup> 2 AST-GP Dalbavancin ( $\leq 0.015$  -  $\geq 1$   $\mu\text{g/mL}$ )

C. Predicate Device: VITEK<sup>®</sup> 2 AST-GP Ceftaroline ( $\leq 0.06$  -  $\geq 4$   $\mu\text{g/mL}$ )  
(K141149)

#### D. 510(k) Summary:

VITEK<sup>®</sup> 2 AST-Gram Positive Dalbavancin is designed for antimicrobial susceptibility testing of Gram positive microorganisms and is intended for use with the VITEK<sup>®</sup> 2 and VITEK<sup>®</sup> 2 Compact Systems as a laboratory aid in the determination of in vitro susceptibility to antimicrobial agents. VITEK<sup>®</sup> 2 AST-Gram Positive Dalbavancin is a quantitative test. Dalbavancin has been shown to be active against most strains of the microorganisms listed below, according to the FDA label for this antimicrobial.

**Active both in vitro and in clinical infections**

*Staphylococcus aureus* (including methicillin-resistant isolates)

*Enterococcus faecalis* (vancomycin-susceptible isolates only)

*Streptococcus agalactiae*

The VITEK® 2 Gram-positive Susceptibility Card is intended for use with the VITEK® 2 Systems in clinical laboratories as an in vitro test to determine the susceptibility of *Staphylococcus* spp., *Enterococcus* spp., and *S. agalactiae* to antimicrobial agents when used as instructed.

The antimicrobial presented in VITEK® 2 AST-GP Cards is in concentrations equivalent by efficacy to standard method concentrations in mcg/ml. The VITEK® 2 AST Cards are essentially miniaturized versions of the doubling dilution technique for determining the minimum inhibitory concentration (MIC) microdilution methodology.

The isolate to be tested is diluted to a standardized concentration in 0.45 - 0.50% saline before being used to rehydrate the antimicrobial medium within the card. The VITEK® 2 automatically fills, seals and places the card into the incubator/reader. The VITEK® 2 Compact has a manual filling and sealing operation. The VITEK® 2 monitors the growth of each well in the card over a defined period of time (20-24 hours for Streptococci and 16-20 hours for Staphylococci and Enterococci). At the completion of the incubation cycle, a report is generated that contains the MIC value along with the interpretive category result for each antimicrobial contained on the card.

VITEK® 2 AST-GP Dalbavancin demonstrated substantially equivalent performance when compared with the CLSI broth microdilution reference method, as defined in the FDA Class II Special Controls Guidance Document: Antimicrobial Susceptibility Test (AST) Systems; Guidance for Industry and FDA (Issued August 28, 2009).

The Premarket Notification (510[k]) presents data in support of VITEK® 2 AST-GP Dalbavancin. An external evaluation was conducted with fresh and stock clinical isolates, as well as a set of challenge strains. The external evaluations were designed to confirm the acceptability of VITEK® 2 AST-GP Dalbavancin by comparing its performance with the CLSI broth microdilution reference method incubated at 20-24 hours for Streptococci and 16-20 hours for Staphylococci and Enterococci. The data is representative of performance on both the VITEK® 2 and VITEK® 2 Compact instrument platforms.

The VITEK® 2 AST-GP Dalbavancin demonstrated acceptable performance as presented in **Table 1** below:

**Table 1: VITEK® 2 AST-GP Dalbavancin Performance**

| <b>Overall Performance<br/>(with the reference<br/>method)</b> | <b>%EA</b> | <b>VME</b> | <b>ME</b> | <b>mE</b> | <b>%CA</b> | <b>VME</b> | <b>ME</b> | <b>mE</b> |
|--|------------|------------|-----------|-----------|------------|------------|-----------|-----------|
|  | 98.1       | N/A        | N/A       | N/A       | 100.0      | 0.0        | 0.0       | N/A       |

Reproducibility and Quality Control demonstrated acceptable results.