



May 22, 2019

Wenzhou Renhua Instruments Co., Ltd
Wang Saijie
General Manager
#.1-1704 Wanda Business Building Yongding Road,
Longwan
Wenzhou, 325024 CN

Re: K190902

Trade/Device Name: RH non-Automated Blood Pressure Meter, MODEL Max02 and RH non-Automated Blood Pressure Meter with Stethoscope, MODEL Max0201, Max0202, Max0203

Regulation Number: 21 CFR 870.1120

Regulation Name: Blood Pressure Cuff

Regulatory Class: Class II

Product Code: DXQ, LDE

Dated: March 29, 2019

Received: April 8, 2019

Dear Wang Saijie:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of

Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/CombinationProducts/GuidanceRegulatoryInformation/ucm597488.htm>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/>) and CDRH Learn (<http://www.fda.gov/Training/CDRHLearn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<http://www.fda.gov/DICE>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

for

Matthew Hillebrenner
Director (Acting)
Division of Cardiac Electrophysiology, Diagnostics
and Monitoring Devices
Office of Cardiovascular Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K190902

Device Name

RH non-Automated Blood Pressure Meter, MODEL Max02 and RH non-Automated Blood Pressure Meter with Stethoscope, MODEL Max0201, Max0202, Max0203

Indications for Use (Describe)

RH non-Automated Blood Pressure Meter is intended to be used by medical professionals or in the home for the measurement of systolic and diastolic pressure on children or adults. The device is intended to be manually attached to a patient and manually inflated along with a manual method for detecting Korotkoff sounds. There are a variety of stethoscopes can be an optioned accessory, depending on the model.

Type of Use (Select one or both, as applicable)

☐ Prescription Use (Part 21 CFR 801 Subpart D)

☒ Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRAStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary

Company Name: WENZHOU RENHUA INSTRUMENTS CO.,LTD

Company Address: #.1-1704 WANDA BUSINESS BUILDING YONGDING ROAD,
LONGWAN ,Wenzhou ,Zhejiang325024 ,CHINA

Telephone: +86-577-86893308

Fax: N/A

Contact Person: WANG SAIJIE

Summary Preparation Date: 2019.3.29

Device Trade Name: RH non-Automated Blood Pressure Meter, MODEL Max02 and RH non-Automated Blood Pressure Meter with Stethoscope, MODEL Max0201, Max0202, Max0203

Classification Name: Blood pressure cuff, Stethoscope, manual

Regulation Number: 21 CFR 870.1120/ 21 CFR 870.1875

Product Code: DXQ; LDE

Device Class: Class 2

Panel: Cardiovascular

PREDICATE DEVICE:

ANEROID SPHYGMOMANOMETER (K173246)

INDICATIONS FOR USE

RH non-Automated Blood Pressure Meter is intended to be used by medical professionals or in the home for the measurement of systolic and diastolic pressure on children or adults. The device is intended to be manually attached to a patient and manually inflated along with a manual method for detecting Korotkoff sounds. There are a variety of stethoscopes can be an optioned accessory, depending on the model.

DEVICE DESCRIPTION

RH non-Automated Blood Pressure Meter is a manual non-invasive aneroid sphygmomanometer which respectively uses an inflation cuff wrapped around the upper arm. The cuff is inflated and deflated by a manual inflation bulb. Besides a manometer (Aneroid gauge), the accessories of MODEL Max02, Max0201, Max0202, Max0203 also includes one selected cuff, one inflation bulb, instruction manual and a vinyl bag. Model Max0201, Max0202, Max0203 include additional Stethoscope (Single head, Dual head and Sprague Rappaport). It is conjunction with Stethoscope when use.

The PVC or nature latex inflation bulb and optional various size of cuffs which cuffs outside layer is

made of Nylon Cloth or Cotton Cloth can be selected according to user's need.

Substantial Equivalence Comparison

It is substantially equivalent to the predicate device (K173246), ANEROID SPHYGMOMANOMETER with respect to indications for use, device description, and technical characteristics.

All comparison table for applied devices are as following, and the substantial equivalence determination is based on the 510(k) Substantial Equivalence Decision-Making Process Flowchart which includes the comparison and discussion of indications for use, technology, and performance specifications.

comparison item	New device RH non-Automated Blood Pressure Meter, MODEL Max02 and RH non-Automated Blood Pressure Meter with Stethoscope, MODEL Max0201, Max0202, Max0203	Predicate device ANEROID SPHYGMOMANOMETER, BK2002/BK2002a (K173246)	Result of Comparasion
Applicant	WENZHOU RENHUA INSTRUMENTS CO.,LTD	Wenzhou Bokang Instruments Co., Ltd	Different
Classification	Class II	Class II	SAME
Regulation number	21CFR 870.1120	21CFR 870.1120	SAME
Product code	DXQ, LDE	DXQ, LDE	SAME
Indications for use	The device is intended to be used by medical professionals or at home for the measurement of systolic and diastolic pressure by	The device is intended to be used by medical professionals or at home for the measurement of systolic and diastolic pressure by	SAME

comparison item	New device RH non-Automated Blood Pressure Meter, MODEL Max02 and RH non-Automated Blood Pressure Meter with Stethoscope, MODEL Max0201, Max0202, Max0203	Predicate device ANEROID SPHYGMOMANOMETER, BK2002/BK2002a (K173246)	Result of Comparasion
	detecting Korotkoff sounds.	detecting Korotkoff sounds.	
Over-The-Counter Use	Yes	Yes	SAME
Where used	Home, Hospital, healthcare facility, ambulance etc.	Home, Hospital, healthcare facility, ambulance etc.	SAME
Target population	infants, children, young adults and adults	infants, children, young adults and adults	SAME
Anatomical sites	Upper Arm (leg for child)	Upper Arm (leg for child)	SAME
Measurement Method	Auscultatory Korotkoff sounds Method	Auscultatory Korotkoff sounds Method	SAME
Inflation	Manual by inflation bulb	Manual by inflation bulb	SAME
Deflation	Manual deflation via valve	Manual deflation via valve	SAME
Display	Aneroid Manometer	Aneroid Manometer	SAME
The monitor scale	From 0 to 300 mmHg with a minimum interval of 2 mmHg.	From 0 to 300 mmHg with a minimum interval of 2 mmHg.	SAME
Design of blood pressure meter	The device comprises tubing attached to a soft inelastic cuff with an	The device comprises tubing attached to a soft inelastic cuff with an	Same as_ <u>(BK2002/2002a)</u>

comparison item	New device RH non-Automated Blood Pressure Meter, MODEL Max02 and RH non-Automated Blood Pressure Meter with Stethoscope, MODEL Max0201, Max0202, Max0203	Predicate device ANEROID SPHYGMOMANOMETER, BK2002/BK2002a (K173246)	Result of Comparasion
	integrated inflatable bladder that is wrapped around the patient's limb and secured by hook and loop closure.	integrated inflatable bladder that is wrapped around the patient's limb and secured by hook and loop closure.	
Design of Stethoscope	Three types option: Single head Dual head Sprague Rappaport	One type option: Single head	Different
Materials	The manometer: aluminum and stainless steel materials. The tubing, inflation bulb: PVC or nature latex. Cuff: <u>Nylon cloth</u> or cotton cloth for outside layer. Cuff bladder: <u>PVC</u> or nature latex	The manometer: aluminum and stainless steel materials. The tubing, inflation bulb: PVC or nature latex. Cuff: <u>Nylon cloth</u> or cotton cloth for outside layer. Cuff bladder: <u>PVC</u> or nature latex	SAME
Accuracy	Pressure: +/- 3 mmHg of reading	Pressure: +/- 3 mmHg of reading	SAME
Compatibility	It can be used from 50°F to	It can be used from 50°F to	SAME

comparison item	New device RH non-Automated Blood Pressure Meter, MODEL Max02 and RH non-Automated Blood Pressure Meter with Stethoscope, MODEL Max0201, Max0202, Max0203	Predicate device ANEROID SPHYGMOMANOMETER, BK2002/BK2002a (K173246)	Result of Comparasion
with environment	104°F (10°C to 40°C) and 15% ~ 85%RH humidity.	104°F (10°C to 40°C) and 15% ~ 85%RH humidity.	
Cuff Size	20"x 5.5" (510mm*140mm) 21.7" x 6.3" (550mm*160mm) 24.4" x 6.9" (620mm*175mm) 28.3" x 8.3" (720mm*210mm) 13.4" x 4.15" (340mm*105mm) 10.2" x 3" (260mm*75 mm)	20"x 5.5" (510mm*140mm) 21.7" x 6.3" (550mm*160mm) 24.4" x 6.9" (620mm*175mm) 28.3" x 8.3" (720mm*210mm) 13.4" x 4.15" (340mm*105mm) 10.2" x 3" (260mm*75 mm)	SAME
Cuff Circumference	Fits arm circumferences 8.7" to 17.3" (220mm to 440 mm), The standard cuff should be available for use in measuring a child's leg blood pressure and for children with larger arms.	Fits arm circumferences 8.7" to 17.3" (220mm to 440 mm), The standard cuff should be available for use in measuring a child's leg blood pressure and for children with larger arms.	SAME
Cuff bladder Size	8.7"x 4.7"(220mm*120mm) 11.8"x 5.9"(300mm*150mm) 13.4"x 6.7"(340mm*170mm) 6.9"x 3.3"(175mm*85mm)	8.7"x 4.7"(220mm*120mm) 11.8"x 5.9"(300mm*150mm) 13.4"x 6.7"(340mm*170mm) 6.9"x 3.3"(175mm*85mm)	SAME

comparison item	New device RH non-Automated Blood Pressure Meter, MODEL Max02 and RH non-Automated Blood Pressure Meter with Stethoscope, MODEL Max0201, Max0202, Max0203	Predicate device ANEROID SPHYGMOMANOMETER, BK2002/BK2002a (K173246)	Result of Comparasion
	7.5''x 2''(190mm*50mm)	7.5''x 2''(190mm*50mm)	
Cuff Color	Blue, Pink, Black	Blue, Gray, Red, Green, Pink, Yellow, Black	Different
Contents (with accessories)	Aneroid gauge, Arm Cuff, Inflation Bulb, Vinyl storage pouch and Instruction Manual, Stethoscope (option)	Aneroid gauge, Arm Cuff, Inflation Bulb, Vinyl storage pouch and Instruction Manual, Stethoscope (option)	Same
Biocompatibility	biocompatible as requirement of ISO 10993-1 ISO 10993-5 ISO 10993-10	biocompatible as requirement of ISO 10993-1 ISO 10993-5 ISO 10993-10	Same model and same supplier
Performance	compatible as requirement of ISO 81060-1	compatible as requirement of ISO 81060-1	Same

Discussion

Comparing with the predicate device, ANEROID SPHYGMOMANOMETER (K173246), RH non-Automated Blood Pressure Meter, MODEL Max02 and RH non-Automated Blood Pressure Meter with Stethoscope, MODEL Max0201, Max0202, Max0203 have same product name, classification and product code, product physical design and operation, same Aneroid gauge, packaging materials, performance parameter ranges, physical properties, mechanical properties, the subject cuffs are the same like our own cleared under predicate K173426 and that the biocompatibility standards compliance/reports are the same. The differences between those two devices are as following,

- Comparing the predicate device ANEROID SPHYGMOMANOMETER (K173246), the design of new device is same as BK2002/BK2002a except there are no Gray, Red, Green and yellow cuff color included.
- Design of Stethoscope for new device, there are added the combination of “Dual head” Stethoscope (Max0202) and “Sprague Rappaport” Stethoscope (Max0203). The different type of Stethoscopes didn’t raise any safety or effectiveness issue.
- All performance specification of new devices for various cuff sizes were verified to comply with the ISO81060-1 standard. (Please see section 18 of this submission)

Non-Clinical Testing

Non-clinical testing included biocompatibility and bench testing. The tests listed in the ISO10993 and ISO81060-1, were included. Results of testing were acceptable.

Clinical Study

No clinical study is included in this submission.

CONCLUSION

RH non-Automated Blood Pressure Meter, MODEL Max02 and RH non-Automated Blood Pressure Meter with Stethoscope, MODEL Max0201, Max0202, Max0203 have the same intended use and similar technological characteristics as the above predicate devices. Moreover, information contained in this submission supplied demonstrates that any differences in their characteristics do not raise any new questions of safety or effectiveness. Thus, RH non-Automated Blood Pressure Meter, MODEL Max02 and RH non-Automated Blood Pressure Meter with Stethoscope, MODEL Max0201, Max0202, Max0203 are substantially equivalent to the predicate devices.