



SAMSUNG MEDISON CO., LTD.
% Ji Yea Lee
Regulatory Affairs Specialist
3366, Hanseo-ro, Nam-myeon,
Hongcheon-gun, Gangwon-do 25108
REPUBLIC OF KOREA

June 12, 2019

Re: K191115

Trade/Device Name: RS85 Diagnostic Ultrasound System
Regulation Number: 21 CFR 892.1550
Regulation Name: Ultrasonic pulsed doppler imaging system
Regulatory Class: Class II
Product Code: IYN, IYO, ITX
Dated: May 17, 2019
Received: May 21, 2019

Dear Ji Yea Lee:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for

devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

For

Thalia T. Mills, Ph.D.
Director
Division of Radiological Health
OHT7: Office of In Vitro Diagnostics
and Radiological Health
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K191115

Device Name

RS85 Diagnostic Ultrasound System

Indications for Use (Describe)

The ultrasound diagnostic system and probes are designed to obtain ultrasound images and analyze body fluid. The clinical applications include: Fetal/Obstetrics, Abdominal, Gynecology, Intraoperative, Pediatric, Small Organ, Neonatal Cephalic, Adult Cephalic, Trans-rectal, Trans-vaginal, Muscular-Skeletal (Conventional, Superficial), Urology, Cardiac Adult, Cardiac Pediatric, Trans-esophageal (Cardiac) and Peripheral vessel.

Type of Use (Select one or both, as applicable)

☒ Prescription Use (Part 21 CFR 801 Subpart D)

☐ Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

DIAGNOSTIC ULTRASOUND INDICATIONS FOR USE STATEMENT

510(k) No.:

Device Name: RS85 Diagnostic Ultrasound System

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation (*includes simultaneous B-mode)						
General (Track I only)	Specific (Tracks I & III)	B	M	PWD	CWD	Color Doppler*	Combined* (Spec.)	Other (Spec.)
Ophthalmic	Ophthalmic							
Fetal Imaging & Other	Fetal/Obstetrics (See Note 3)	P	P	P		P	Note 1	Notes 2, 3, 4, 7, 8, 9, 11, 14
	Abdominal (See Note 10)	P	P	P	P	P	Note 1	Notes 2, 4, 5, 6, 7, 8, 9, 11, 12, 14, 15
	Intra-operative (See Note 6)	P	P	P		P	Note 1	Note 7, 9, 11
	Intra-operative (Neuro.)	P	P	P		P	Note 1	Note 7, 9, 11
	Laparoscopic							
	Pediatric	P	P	P	P	P	Note 1	Notes 2, 7, 8, 9, 11
	Small Organ (See Note 5)	P	P	P		P	Note 1	Note 2, 5, 6, 7, 8, 9, 11, 12, 14
	Neonatal Cephalic	P	P	P		P	Note 1	Notes 8, 9, 11
	Adult Cephalic	P	P	P	P	P	Note 1	Notes 7
	Trans-rectal	P	P	P		P	Note 1	Notes 2, 7, 8, 9, 11, 12
	Trans-vaginal	P	P	P		P	Note 1	Notes 2, 7, 8, 9, 11, 12
	Trans-urethral							
	Trans-esoph. (non-Cardiac)							
	Musculo-skel. (Convent.)	P	P	P		P	Note 1	Note 2, 5, 6, 7, 8, 9, 11, 14
	Musculo-skel. (Superfic.)	P	P	P		P	Note 1	Note 2, 5, 6, 7, 8, 9, 11, 14
	Intra-luminal							
	Other (See Note 13)	P	P	P		P	Note 1	Notes 2, 7, 8, 9, 12
Cardiac	Cardiac Adult	P	P	P	P	P	Note 1	Notes 4, 7
	Cardiac Pediatric	P	P	P	P	P	Note 1	Notes 4, 7
	Trans-esophageal (Cardiac)	P	P	P	P	P	Note 1	Notes 4, 7
	Other (spec.)							
Peripheral Vessel	Peripheral vessel	P	P	P	P	P	Note 1	Note 2, 5, 6, 7, 8, 9, 11, 14
	Other (spec.)							

N= new indication; P= previously cleared by FDA K182932; E= added under Appendix E

Additional Comments:

Color Doppler includes Power (Amplitude) Doppler

Note 1: B+M, B+PW, B+C, B+PD, B+DPD, B+PPI, B+TD, B+CW, B+C+PW, B+PD+PW, B+DPD+PW, B+PPI+PW, B+TD+PW, B+C+M, Dual/Quad, B+C+CW, B+PD+CW, B+E, B+B/C

Note 2: Includes imaging for guidance of biopsy

Note 3: Includes infertility monitoring of follicle development

Note 4: Color M-mode

Note 5: For example: thyroid, parathyroid, breast, scrotum and penis in adult, pediatric and neonatal patients

Note 6: Abdominal organs and peripheral vessel

Note 7: Tissue Harmonic Imaging (THI)

Note 8: 3D imaging

Note 9: Spatial Compound Imaging

Note 10: Includes Renal, Gynecology/Pelvis

Note 11: Panoramic imaging

Note 12: ElastoScan

Note 13: Includes Urology/Prostate

Note 14: S-Fusion

Note 15: S-Shearwave

Concurrence of Center for Devices and Radiological Health (CDRH)
Prescription Use (Per 21 CFR 801.109)

DIAGNOSTIC ULTRASOUND INDICATIONS FOR USE STATEMENT

510(k) No.:

Device Name: CA1-7A for use with RS85

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation (*includes simultaneous B-mode)						
General (Track I only)	Specific (Tracks I & III)	B	M	PWD	CWD	Color Doppler*	Combined* (Spec.)	Other (Spec.)
Ophthalmic	Ophthalmic							
Fetal Imaging & Other	Fetal/Obstetrics (See Note 3)	P	P	P		P	Note 1	Notes 2,4,7,8,9,11,14
	Abdominal (See Note 10)	P	P	P		P	Note 1	Notes 2,6,7,8,9,11,14,15
	Intra-operative (See Note 6)							
	Intra-operative (Neuro.)							
	Laparoscopic							
	Pediatric	P	P	P		P	Note 1	Notes 2,6,7,8,9,11,14
	Small Organ (See Note 5)							
	Neonatal Cephalic							
	Adult Cephalic							
	Trans-rectal							
	Trans-vaginal							
	Trans-urethral							
	Trans-esoph. (non-Cardiac)							
	Musculo-skel. (Convent.)	P	P	P		P	Note 1	Notes 2,6,7,8,9,11,14
	Musculo-skel. (Superfic.)							
	Intra-luminal							
	Other (See Note 13)	P	P	P		P	Note 1	Notes 2,6,7,8,9,11,14
Cardiac	Cardiac Adult							
	Cardiac Pediatric							
	Trans-esophageal (Cardiac)							
	Other (spec.)							
Peripheral Vessel	Peripheral vessel	P	P	P		P	Note 1	Notes 2,6,7,8,9,11,14
	Other (spec.)							

N= new indication; P= previously cleared by FDA K182932; E= added under Appendix E

Additional Comments:

Color Doppler includes Power (Amplitude) Doppler

Note 1: B+M, B+PW, B+C, B+PD, B+DPD, B+PPI, B+TD, B+CW, B+C+PW, B+PD+PW, B+DPD+PW, B+PPI+PW, B+TD+PW, B+C+M, Dual/Quad, B+C+CW, B+PD+CW, B+E, B+B/C

Note 2: Includes imaging for guidance of biopsy

Note 3: Includes infertility monitoring of follicle development

Note 4: Color M-mode

Note 5: For example: thyroid, parathyroid, breast, scrotum and penis in adult, pediatric and neonatal patients

Note 6: Abdominal organs and peripheral vessel

Note 7: Tissue Harmonic Imaging (THI)

Note 8: 3D imaging

Note 9: Spatial Compound Imaging

Note 10: Includes Renal, Gynecology/Pelvis

Note 11: Panoramic imaging

Note 12: ElastoScan

Note 13: Includes Urology/Prostate

Note 14: S-Fusion

Note 15: S-Shearwave

Concurrence of Center for Devices and Radiological Health (CDRH)
Prescription Use (Per 21 CFR 801.109)

DIAGNOSTIC ULTRASOUND INDICATIONS FOR USE STATEMENT

510(k) No.:

Device Name: CF4-9 for use with RS85

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation (*includes simultaneous B-mode)						
General (Track I only)	Specific (Tracks I & III)	B	M	PWD	CWD	Color Doppler*	Combined* (Spec.)	Other (Spec.)
Ophthalmic	Ophthalmic							
Fetal Imaging & Other	Fetal/Obstetrics (See Note 3)							
	Abdominal (See Note 10)	P	P	P		P	Note 1	Notes 7, 8, 9, 11
	Intra-operative (See Note 6)							
	Intra-operative (Neuro.)							
	Laparoscopic							
	Pediatric	P	P	P		P	Note 1	Notes 7, 8, 9, 11
	Small Organ (See Note 5)							
	Neonatal Cephalic	P	P	P		P	Note 1	Notes 7, 8, 9, 11
	Adult Cephalic							
	Trans-rectal							
	Trans-vaginal							
	Trans-urethral							
	Trans-esoph. (non-Cardiac)							
	Musculo-skel. (Convent.)							
	Musculo-skel. (Superfic.)							
	Intra-luminal							
	Other (See Note 13)							
Cardiac	Cardiac Adult							
	Cardiac Pediatric							
	Trans-esophageal (Cardiac)							
	Other (spec.)							
Peripheral Vessel	Peripheral vessel	P	P	P		P	Note 1	Notes 8, 9, 11
	Other (spec.)							

N= new indication; P= previously cleared by FDA K182932; E= added under Appendix E

Additional Comments:

Color Doppler includes Power (Amplitude) Doppler

Note 1: B+M, B+PW, B+C, B+PD, B+DPD, B+PPI, B+TD, B+CW, B+C+PW, B+PD+PW, B+DPD+PW, B+PPI+PW, B+TD+PW, B+C+M, Dual/Quad, B+C+CW, B+PD+CW, B+E, B+B/C

Note 2: Includes imaging for guidance of biopsy

Note 3: Includes infertility monitoring of follicle development

Note 4: Color M-mode

Note 5: For example: thyroid, parathyroid, breast, scrotum and penis in adult, pediatric and neonatal patients

Note 6: Abdominal organs and peripheral vessel

Note 7: Tissue Harmonic Imaging (THI)

Note 8: 3D imaging

Note 9: Spatial Compound Imaging

Note 10: Includes Renal, Gynecology/Pelvis

Note 11: Panoramic imaging

Note 12: ElastoScan

Note 13: Includes Urology/Prostate

Note 14: S-Fusion

Note 15: S-Shearwave

Concurrence of Center for Devices and Radiological Health (CDRH)
Prescription Use (Per 21 CFR 801.109)

DIAGNOSTIC ULTRASOUND INDICATIONS FOR USE STATEMENT

510(k) No.:

Device Name: CA2-8A for use with RS85

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation (*includes simultaneous B-mode)						
General (Track I only)	Specific (Tracks I & III)	B	M	PWD	CWD	Color Doppler*	Combined* (Spec.)	Other (Spec.)
Ophthalmic	Ophthalmic							
Fetal Imaging & Other	Fetal/Obstetrics (See Note 3)	P	P	P		P	Note 1	Notes 2, 4, 7, 8, 9, 11
	Abdominal (See Note 10)	P	P	P		P	Note 1	Notes 2, 4, 7, 8, 9, 11
	Intra-operative (See Note 6)							
	Intra-operative (Neuro.)							
	Laparoscopic							
	Pediatric							
	Small Organ (See Note 5)							
	Neonatal Cephalic							
	Adult Cephalic							
	Trans-rectal							
	Trans-vaginal							
	Trans-urethral							
	Trans-esoph. (non-Cardiac)							
	Musculo-skel. (Convent.)							
	Musculo-skel. (Superfic.)							
	Intra-luminal							
	Other (See Note 13)							
Cardiac	Cardiac Adult							
	Cardiac Pediatric							
	Trans-esophageal (Cardiac)							
	Other (spec.)							
Peripheral Vessel	Peripheral vessel							
	Other (spec.)							

N= new indication; P= previously cleared by FDA K182932; E= added under Appendix E

Additional Comments:

Color Doppler includes Power (Amplitude) Doppler

Note 1: B+M, B+PW, B+C, B+PD, B+DPD, B+PPI, B+TD, B+CW, B+C+PW, B+PD+PW, B+DPD+PW, B+PPI+PW, B+TD+PW, B+C+M, Dual/Quad, B+C+CW, B+PD+CW, B+E, B+B/C

Note 2: Includes imaging for guidance of biopsy

Note 3: Includes infertility monitoring of follicle development

Note 4: Color M-mode

Note 5: For example: thyroid, parathyroid, breast, scrotum and penis in adult, pediatric and neonatal patients

Note 6: Abdominal organs and peripheral vessel

Note 7: Tissue Harmonic Imaging (THI)

Note 8: 3D imaging

Note 9: MultiVision (Spatial Compound Imaging)

Note 10: Includes Renal, Gynecology/Pelvis

Note 11: Panoramic imaging

Note 12: ElastoScan

Note 13: Includes Urology/Prostate

Note 14: S-Fusion

Note 15: S-Shearwave

Concurrence of Center for Devices and Radiological Health (CDRH)
Prescription Use (Per 21 CFR 801.109)

DIAGNOSTIC ULTRASOUND INDICATIONS FOR USE STATEMENT

510(k) No.:

Device Name: CA3-10A for use with RS85

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation (*includes simultaneous B-mode)						
General (Track I only)	Specific (Tracks I & III)	B	M	PWD	CWD	Color Doppler*	Combined* (Spec.)	Other (Spec.)
Ophthalmic	Ophthalmic							
Fetal Imaging & Other	Fetal/Obstetrics (See Note 3)	P	P	P		P	Note 1	Notes 2, 3, 7, 8, 9, 11
	Abdominal (See Note 10)	P	P	P		P	Note 1	Notes 2, 7, 8, 9, 11
	Intra-operative (See Note 6)							
	Intra-operative (Neuro.)							
	Laparoscopic							
	Pediatric	P	P	P		P	Note 1	Notes 2, 7, 8, 9, 11
	Small Organ (See Note 5)							
	Neonatal Cephalic							
	Adult Cephalic							
	Trans-rectal							
	Trans-vaginal							
	Trans-urethral							
	Trans-esoph. (non-Cardiac)							
	Musculo-skel. (Convent.)	P	P	P		P	Note 1	Notes 2, 7, 8, 9, 11
	Musculo-skel. (Superfic.)	P	P	P		P	Note 1	Notes 2, 7, 8, 9, 11
	Intra-luminal							
	Other (See Note 13)							
Cardiac	Cardiac Adult							
	Cardiac Pediatric							
	Trans-esophageal (Cardiac)							
	Other (spec.)							
Peripheral Vessel	Peripheral vessel	P	P	P		P	Note 1	Notes 2, 7, 8, 9, 11
	Other (spec.)							

N= new indication; P= previously cleared by FDA K182932; E= added under Appendix E

Additional Comments:

Color Doppler includes Power (Amplitude) Doppler

Note 1: B+M, B+PW, B+C, B+PD, B+DPD, B+PPI, B+TD, B+CW, B+C+PW, B+PD+PW, B+DPD+PW, B+PPI+PW, B+TD+PW, B+C+M, Dual/Quad, B+C+CW, B+PD+CW, B+E, B+B/C

Note 2: Includes imaging for guidance of biopsy

Note 3: Includes infertility monitoring of follicle development

Note 4: Color M-mode

Note 5: For example: thyroid, parathyroid, breast, scrotum and penis in adult, pediatric and neonatal patients

Note 6: Abdominal organs and peripheral vessel

Note 7: Tissue Harmonic Imaging (THI)

Note 8: 3D imaging

Note 9: MultiVision (Spatial Compound Imaging)

Note 10: Includes Renal, Gynecology/Pelvis

Note 11: Panoramic imaging

Note 12: ElastoScan

Note 13: Includes Urology/Prostate

Note 14: S-Fusion

Note 15: S-Shearwave

Concurrence of Center for Devices and Radiological Health (CDRH)
Prescription Use (Per 21 CFR 801.109)

DIAGNOSTIC ULTRASOUND INDICATIONS FOR USE STATEMENT

510(k) No.:

Device Name: E3-12A for use with RS85

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation (*includes simultaneous B-mode)						
General (Track I only)	Specific (Tracks I & III)	B	M	PWD	CWD	Color Doppler*	Combined* (Spec.)	Other (Spec.)
Ophthalmic	Ophthalmic							
Fetal Imaging & Other	Fetal/Obstetrics (See Note 3)	P	P	P		P	Note 1	Notes 2, 7, 8, 9
	Abdominal (See Note 10)	P	P	P		P	Note 1	Notes 2, 7, 8, 9, 12
	Intra-operative (See Note 6)							
	Intra-operative (Neuro.)							
	Laparoscopic							
	Pediatric							
	Small Organ (See Note 5)							
	Neonatal Cephalic							
	Adult Cephalic							
	Trans-rectal	P	P	P		P	Note 1	Notes 2, 7, 8, 9, 12
	Trans-vaginal	P	P	P		P	Note 1	Notes 2, 7, 8, 9, 12
	Trans-urethral							
	Trans-esoph. (non-Cardiac)							
	Musculo-skel. (Convent.)							
	Musculo-skel. (Superfic.)							
	Intra-luminal							
	Other (See Note 13)	P	P	P		P	Note 1	Notes 2, 7, 8, 9, 12
Cardiac	Cardiac Adult							
	Cardiac Pediatric							
	Trans-esophageal (Cardiac)							
	Other (spec.)							
Peripheral Vessel	Peripheral vessel							
	Other (spec.)							

N= new indication; P= previously cleared by FDA K182932; E= added under Appendix E

Additional Comments:

Color Doppler includes Power (Amplitude) Doppler

Note 1: B+M, B+PW, B+C, B+PD, B+DPD, B+PPI, B+TD, B+CW, B+C+PW, B+PD+PW, B+DPD+PW, B+PPI+PW, B+TD+PW, B+C+M, Dual/Quad, B+C+CW, B+PD+CW, B+E, B+B/C

Note 2: Includes imaging for guidance of biopsy

Note 3: Includes infertility monitoring of follicle development

Note 4: Color M-mode

Note 5: For example: thyroid, parathyroid, breast, scrotum and penis in adult, pediatric and neonatal patients

Note 6: Abdominal organs and peripheral vessel

Note 7: Tissue Harmonic Imaging (THI)

Note 8: 3D imaging

Note 9: Spatial Compound Imaging

Note 10: Includes Renal, Gynecology/Pelvis

Note 11: Panoramic imaging

Note 12: ElastoScan

Note 13: Includes Urology/Prostate

Note 14: S-Fusion

Note 15: S-Shearwave

Concurrence of Center for Devices and Radiological Health (CDRH)
Prescription Use (Per 21 CFR 801.109)

DIAGNOSTIC ULTRASOUND INDICATIONS FOR USE STATEMENT

510(k) No.:

Device Name: L3-12A for use with RS85

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation (*includes simultaneous B-mode)						
General (Track I only)	Specific (Tracks I & III)	B	M	PWD	CWD	Color Doppler*	Combined* (Spec.)	Other (Spec.)
Ophthalmic	Ophthalmic							
Fetal Imaging & Other	Fetal/Obstetrics (See Note 3)							
	Abdominal (See Note 10)	P	P	P		P	Note 1	Note 2, 5, 6, 7, 8, 9, 11, 14
	Intra-operative (See Note 6)							
	Intra-operative (Neuro.)							
	Laparoscopic							
	Pediatric							
	Small Organ (See Note 5)	P	P	P		P	Note 1	Note 2, 5, 6, 7, 8, 9, 11, 12, 14
	Neonatal Cephalic							
	Adult Cephalic							
	Trans-rectal							
	Trans-vaginal							
	Trans-urethral							
	Trans-esoph. (non-Cardiac)							
	Musculo-skel. (Convent.)	P	P	P		P	Note 1	Note 2, 5, 6, 7, 8, 9, 11, 14
	Musculo-skel. (Superfic.)	P	P	P		P	Note 1	Note 2, 5, 6, 7, 8, 9, 11, 14
	Intra-luminal							
	Other (See Note 13)							
Cardiac	Cardiac Adult							
	Cardiac Pediatric							
	Trans-esophageal (Cardiac)							
	Other (spec.)							
Peripheral Vessel	Peripheral vessel	P	P	P		P	Note 1	Note 2, 5, 6, 7, 8, 9, 11, 14
	Other (spec.)							

N= new indication; P= previously cleared by FDA K182932; E= added under Appendix E

Additional Comments:

Color Doppler includes Power (Amplitude) Doppler

Note 1: B+M, B+PW, B+C, B+PD, B+DPD, B+PPI, B+TD, B+CW, B+C+PW, B+PD+PW, B+DPD+PW, B+PPI+PW, B+TD+PW, B+C+M, Dual/Quad, B+C+CW, B+PD+CW, B+E, B+B/C

Note 2: Includes imaging for guidance of biopsy

Note 3: Includes infertility monitoring of follicle development

Note 4: Color M-mode

Note 5: For example: thyroid, parathyroid, breast, scrotum and penis in adult, pediatric and neonatal patients

Note 6: Abdominal organs and peripheral vessel

Note 7: Tissue Harmonic Imaging (THI)

Note 8: 3D imaging

Note 9: Spatial Compound Imaging

Note 10: Includes Renal, Gynecology/Pelvis

Note 11: Panoramic imaging

Note 12: ElastoScan

Note 13: Includes Urology/Prostate

Note 14: S-Fusion

Note 15: S-Shearwave

Concurrence of Center for Devices and Radiological Health (CDRH)
Prescription Use (Per 21 CFR 801.109)

DIAGNOSTIC ULTRASOUND INDICATIONS FOR USE STATEMENT

510(k) No.:

Device Name: EA2-11B for use with RS85

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation (*includes simultaneous B-mode)						
General (Track I only)	Specific (Tracks I & III)	B	M	PWD	CWD	Color Doppler*	Combined* (Spec.)	Other (Spec.)
Ophthalmic	Ophthalmic							
Fetal Imaging & Other	Fetal/Obstetrics (See Note 3)	P	P	P		P	Note 1	Note 2, 7, 8, 9
	Abdominal (See Note 12)	P	P	P		P	Note 1	Note 2, 6, 7, 8, 9, 12
	Intra-operative (See Note 6)							
	Intra-operative (Neuro.)							
	Laparoscopic							
	Pediatric							
	Small Organ (See Note 5)							
	Neonatal Cephalic							
	Adult Cephalic							
	Trans-rectal	P	P	P		P	Note 1	Note 2, 7, 8, 9, 12
	Trans-vaginal	P	P	P		P	Note 1	Note 2, 7, 8, 9, 12
	Trans-urethral							
	Trans-esoph. (non-Cardiac)							
	Musculo-skel. (Convent.)							
	Musculo-skel. (Superfic.)							
	Intra-luminal							
	Other (See Note 13)	P	P	P		P	Note 1	Note 2, 7, 8, 9, 12
Cardiac	Cardiac Adult							
	Cardiac Pediatric							
	Trans-esophageal (Cardiac)							
	Other (spec.)							
Peripheral Vessel	Peripheral vessel							
	Other (spec.)							

N= new indication; P= previously cleared by FDA K182932; E= added under Appendix E

Additional Comments:

Color Doppler includes Power (Amplitude) Doppler

Note 1: B+M, B+PW, B+C, B+PD, B+DPD, B+PPI, B+TD, B+CW, B+C+PW, B+PD+PW, B+DPD+PW, B+PPI+PW, B+TD+PW, B+C+M, Dual/Quad, B+C+CW, B+PD+CW, B+E, B+B/C

Note 2: Includes imaging for guidance of biopsy

Note 3: Includes infertility monitoring of follicle development

Note 4: Color M-mode

Note 5: For example: thyroid, parathyroid, breast, scrotum and penis in adult, pediatric and neonatal patients

Note 6: Abdominal organs and peripheral vessel

Note 7: Tissue Harmonic Imaging (THI)

Note 8: 3D imaging

Note 9: MultiVision (Spatial Compound Imaging)

Note 10: Includes Renal, Gynecology/Pelvis

Note 11: Panoramic imaging

Note 12: ElastoScan

Note 13: Includes Urology/Prostate

Note 14: S-Fusion

Note 15: S-Shearwave

Concurrence of Center for Devices and Radiological Health (CDRH)
Prescription Use (Per 21 CFR 801.109)

DIAGNOSTIC ULTRASOUND INDICATIONS FOR USE STATEMENT

510(k) No.:

Device Name: LA2-9A for use with RS85

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation (*includes simultaneous B-mode)						
General (Track I only)	Specific (Tracks I & III)	B	M	PWD	CWD	Color Doppler*	Combined* (Spec.)	Other (Spec.)
Ophthalmic	Ophthalmic							
Fetal Imaging & Other	Fetal/Obstetrics (See Note 3)							
	Abdominal (See Note 10)	P	P	P		P	Note 1	Note 2, 5, 6, 7, 8, 9, 11
	Intra-operative (See Note 6)							
	Intra-operative (Neuro.)							
	Laparoscopic							
	Pediatric							
	Small Organ (See Note 5)	P	P	P		P	Note 1	Note 2, 5, 6, 7, 8, 9, 11,12,15
	Neonatal Cephalic							
	Adult Cephalic							
	Trans-rectal							
	Trans-vaginal							
	Trans-urethral							
	Trans-esoph. (non-Cardiac)							
	Musculo-skel. (Convent.)	P	P	P		P	Note 1	Note 2, 5, 6, 7, 8, 9, 11,12,15
	Musculo-skel. (Superfic.)	P	P	P		P	Note 1	Note 2, 5, 6, 7, 8, 9, 11,12,15
	Intra-luminal							
	Other (See Note 13)							
Cardiac	Cardiac Adult							
	Cardiac Pediatric							
	Trans-esophageal (Cardiac)							
	Other (spec.)							
Peripheral Vessel	Peripheral vessel	P	P	P		P	Note 1	Note 2, 5, 6, 7, 8, 9, 11
	Other (spec.)							

N= new indication; P= previously cleared by FDA K182932; E= added under Appendix E

Additional Comments:

Color Doppler includes Power (Amplitude) Doppler

Note 1: B+M, B+PW, B+C, B+PD, B+DPD, B+PPI, B+TD, B+CW, B+C+PW, B+PD+PW, B+DPD+PW, B+PPI+PW, B+TD+PW, B+C+M, Dual/Quad, B+C+CW, B+PD+CW, B+E, B+B/C

Note 2: Includes imaging for guidance of biopsy

Note 3: Includes infertility monitoring of follicle development

Note 4: Color M-mode

Note 5: For example: thyroid, parathyroid, breast, scrotum and penis in adult, pediatric and neonatal patients

Note 6: Abdominal organs and peripheral vessel

Note 7: Tissue Harmonic Imaging (THI)

Note 8: 3D imaging

Note 9: Spatial Compound Imaging

Note 10: Includes Renal, Gynecology/Pelvis

Note 11: Panoramic imaging

Note 12: ElastoScan

Note 13: Includes Urology/Prostate

Note 14: S-Fusion

Note 15: S-Shearwave

Concurrence of Center for Devices and Radiological Health (CDRH)
Prescription Use (Per 21 CFR 801.109)

DIAGNOSTIC ULTRASOUND INDICATIONS FOR USE STATEMENT

510(k) No.:

Device Name: LA3-16A for use with RS85

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation (*includes simultaneous B-mode)						
General (Track I only)	Specific (Tracks I & III)	B	M	PWD	CWD	Color Doppler*	Combined* (Spec.)	Other (Spec.)
Ophthalmic	Ophthalmic							
Fetal Imaging & Other	Fetal/Obstetrics (See Note 3)							
	Abdominal (See Note 10)							
	Intra-operative (See Note 6)							
	Intra-operative (Neuro.)							
	Laparoscopic							
	Pediatric							
	Small Organ (See Note 5)	P	P	P		P	Note 1	Note 2, 5, 6, 7, 9, 11, 12
	Neonatal Cephalic							
	Adult Cephalic							
	Trans-rectal							
	Trans-vaginal							
	Trans-urethral							
	Trans-esoph. (non-Cardiac)							
	Musculo-skel. (Convent.)	P	P	P		P	Note 1	Note 2, 5, 6, 7, 9, 11
	Musculo-skel. (Superfic.)	P	P	P		P	Note 1	Note 2, 5, 6, 7, 9, 11
	Intra-luminal							
	Other (See Note 13)							
Cardiac	Cardiac Adult							
	Cardiac Pediatric							
	Trans-esophageal (Cardiac)							
	Other (spec.)							
Peripheral Vessel	Peripheral vessel	P	P	P		P	Note 1	Note 2, 5, 6, 7, 9, 11
	Other (spec.)							

N= new indication; P= previously cleared by FDA K182932; E= added under Appendix E

Additional Comments:

Color Doppler includes Power (Amplitude) Doppler

Note 1: B+M, B+PW, B+C, B+PD, B+DPD, B+PPI, B+TD, B+CW, B+C+PW, B+PD+PW, B+DPD+PW, B+PPI+PW, B+TD+PW, B+C+M, Dual/Quad, B+C+CW, B+PD+CW, B+E, B+B/C

Note 2: Includes imaging for guidance of biopsy

Note 3: Includes infertility monitoring of follicle development

Note 4: Color M-mode

Note 5: For example: thyroid, parathyroid, breast, scrotum and penis in adult, pediatric and neonatal patients

Note 6: Abdominal organs and peripheral vessel

Note 7: Tissue Harmonic Imaging (THI)

Note 8: 3D imaging

Note 9: MultiVision (Spatial Compound Imaging)

Note 10: Includes Renal, Gynecology/Pelvis

Note 11: Panoramic imaging

Note 12: ElastoScan

Note 13: Includes Urology/Prostate

Note 14: S-Fusion

Note 15: S-Shearwave

Concurrence of Center for Devices and Radiological Health (CDRH)
Prescription Use (Per 21 CFR 801.109)

DIAGNOSTIC ULTRASOUND INDICATIONS FOR USE STATEMENT

510(k) No.:

Device Name: LA3-16AI for use with RS85

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation (*includes simultaneous B-mode)						
General (Track I only)	Specific (Tracks I & III)	B	M	PWD	CWD	Color Doppler*	Combined* (Spec.)	Other (Spec.)
Ophthalmic	Ophthalmic							
Fetal Imaging & Other	Fetal/Obstetrics (See Note 3)							
	Abdominal (See Note 10)							
	Intra-operative (See Note 6)	P	P	P		P	Note 1	Note 7, 9, 11
	Intra-operative (Neuro.)	P	P	P		P	Note 1	Note 7, 9, 11
	Laparoscopic							
	Pediatric							
	Small Organ (See Note 5)							
	Neonatal Cephalic							
	Adult Cephalic							
	Trans-rectal							
	Trans-vaginal							
	Trans-urethral							
	Trans-esoph. (non-Cardiac)							
	Musculo-skel. (Convent.)	P	P	P		P	Note 1	Note 7, 9, 11
	Musculo-skel. (Superfic.)	P	P	P		P	Note 1	Note 7, 9, 11
	Intra-luminal							
	Other (See Note 13)							
Cardiac	Cardiac Adult							
	Cardiac Pediatric							
	Trans-esophageal (Cardiac)							
	Other (spec.)							
Peripheral Vessel	Peripheral vessel	P	P	P		P	Note 1	Note 7, 9, 11
	Other (spec.)							

N= new indication; P= previously cleared by FDA K182932; E= added under Appendix E

Additional Comments:

Color Doppler includes Power (Amplitude) Doppler

Note 1: B+M, B+PW, B+C, B+PD, B+DPD, B+PPI, B+TD, B+CW, B+C+PW, B+PD+PW, B+DPD+PW, B+PPI+PW, B+TD+PW, B+C+M, Dual/Quad, B+C+CW, B+PD+CW, B+E, B+B/C

Note 2: Includes imaging for guidance of biopsy

Note 3: Includes infertility monitoring of follicle development

Note 4: Color M-mode

Note 5: For example: thyroid, parathyroid, breast, scrotum and penis in adult, pediatric and neonatal patients

Note 6: Abdominal organs and peripheral vessel

Note 7: Tissue Harmonic Imaging (THI)

Note 8: 3D imaging

Note 9: MultiVision (Spatial Compound Imaging)

Note 10: Includes Renal, Gynecology/Pelvis

Note 11: Panoramic imaging

Note 12: ElastoScan

Note 13: Includes Urology/Prostate

Note 14: S-Fusion

Note 15: S-Shearwave

Concurrence of Center for Devices and Radiological Health (CDRH)
Prescription Use (Per 21 CFR 801.109)

DIAGNOSTIC ULTRASOUND INDICATIONS FOR USE STATEMENT

510(k) No.:

Device Name: LA4-18B for use with RS85

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation (*includes simultaneous B-mode)						
General (Track I only)	Specific (Tracks I & III)	B	M	PWD	CWD	Color Doppler*	Combined* (Spec.)	Other (Spec.)
Ophthalmic	Ophthalmic							
Fetal Imaging & Other	Fetal/Obstetrics (See Note 3)							
	Abdominal (See Note 10)							
	Intra-operative (See Note 6)							
	Intra-operative (Neuro.)							
	Laparoscopic							
	Pediatric							
	Small Organ (See Note 5)	P	P	P		P	Note 1	Note 2, 5, 6, 7, 8, 9, 11, 12, 14
	Neonatal Cephalic							
	Adult Cephalic							
	Trans-rectal							
	Trans-vaginal							
	Trans-urethral							
	Trans-esoph. (non-Cardiac)							
	Musculo-skel. (Convent.)	P	P	P		P	Note 1	Note 2, 5, 6, 7, 8, 9, 11, 14
	Musculo-skel. (Superfic.)	P	P	P		P	Note 1	Note 2, 5, 6, 7, 8, 9, 11, 14
	Intra-luminal							
	Other (See Note 13)							
Cardiac	Cardiac Adult							
	Cardiac Pediatric							
	Trans-esophageal (Cardiac)							
	Other (spec.)							
Peripheral Vessel	Peripheral vessel	P	P	P		P	Note 1	Note 2, 5, 6, 7, 8, 9, 11, 14
	Other (spec.)							

N= new indication; P= previously cleared by FDA K182932; E= added under Appendix E

Additional Comments:

Color Doppler includes Power (Amplitude) Doppler

Note 1: B+M, B+PW, B+C, B+PD, B+DPD, B+PPI, B+TD, B+CW, B+C+PW, B+PD+PW, B+DPD+PW, B+PPI+PW, B+TD+PW, B+C+M, Dual/Quad, B+C+CW, B+PD+CW, B+E, B+B/C

Note 2: Includes imaging for guidance of biopsy

Note 3: Includes infertility monitoring of follicle development

Note 4: Color M-mode

Note 5: For example: thyroid, parathyroid, breast, scrotum and penis in adult, pediatric and neonatal patients

Note 6: Abdominal organs and peripheral vessel

Note 7: Tissue Harmonic Imaging (THI)

Note 8: 3D imaging

Note 9: MultiVision (Spatial Compound Imaging)

Note 10: Includes Renal, Gynecology/Pelvis

Note 11: Panoramic imaging

Note 12: ElastoScan

Note 13: Includes Urology/Prostate

Note 14: S-Fusion

Note 15: S-Shearwave

Concurrence of Center for Devices and Radiological Health (CDRH)
Prescription Use (Per 21 CFR 801.109)

DIAGNOSTIC ULTRASOUND INDICATIONS FOR USE STATEMENT

510(k) No.:

Device Name: LM4-15B for use with RS85

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation (*includes simultaneous B-mode)						
General (Track I only)	Specific (Tracks I & III)	B	M	PWD	CWD	Color Doppler*	Combined* (Spec.)	Other (Spec.)
Ophthalmic	Ophthalmic							
Fetal Imaging & Other	Fetal/Obstetrics (See Note 3)							
	Abdominal (See Note 10)	P	P	P		P	Note 1	Note 2, 5, 6, 7, 8, 9, 11
	Intra-operative (See Note 6)							
	Intra-operative (Neuro.)							
	Laparoscopic							
	Pediatric							
	Small Organ (See Note 5)	P	P	P		P	Note 1	Note 2, 5, 6, 7, 8, 9, 11, 12
	Neonatal Cephalic							
	Adult Cephalic							
	Trans-rectal							
	Trans-vaginal							
	Trans-urethral							
	Trans-esoph. (non-Cardiac)							
	Musculo-skel. (Convent.)	P	P	P		P	Note 1	Note 2, 5, 6, 7, 8, 9, 11
	Musculo-skel. (Superfic.)	P	P	P		P	Note 1	Note 2, 5, 6, 7, 8, 9, 11
	Intra-luminal							
	Other (See Note 13)							
Cardiac	Cardiac Adult							
	Cardiac Pediatric							
	Trans-esophageal (Cardiac)							
	Other (spec.)							
Peripheral Vessel	Peripheral vessel	P	P	P		P	Note 1	Note 2, 5, 6, 7, 8, 9, 11
	Other (spec.)							

N= new indication; P= previously cleared by FDA K182932; E= added under Appendix E

Additional Comments:

Color Doppler includes Power (Amplitude) Doppler

Note 1: B+M, B+PW, B+C, B+PD, B+DPD, B+PPI, B+TD, B+CW, B+C+PW, B+PD+PW, B+DPD+PW, B+PPI+PW, B+TD+PW, B+C+M, Dual/Quad, B+C+CW, B+PD+CW, B+E, B+B/C

Note 2: Includes imaging for guidance of biopsy

Note 3: Includes infertility monitoring of follicle development

Note 4: Color M-mode

Note 5: For example: thyroid, parathyroid, breast, scrotum and penis in adult, pediatric and neonatal patients

Note 6: Abdominal organs and peripheral vessel

Note 7: Tissue Harmonic Imaging (THI)

Note 8: 3D imaging

Note 9: MultiVision (Spatial Compound Imaging)

Note 10: Includes Renal, Gynecology/Pelvis

Note 11: Panoramic imaging

Note 12: ElastoScan

Note 13: Includes Urology/Prostate

Note 14: S-Fusion

Note 15: S-Shearwave

Concurrence of Center for Devices and Radiological Health (CDRH)
Prescription Use (Per 21 CFR 801.109)

DIAGNOSTIC ULTRASOUND INDICATIONS FOR USE STATEMENT

510(k) No.:

Device Name: PM1-6A for use with RS85

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation (*includes simultaneous B-mode)						
General (Track I only)	Specific (Tracks I & III)	B	M	PWD	CWD	Color Doppler*	Combined* (Spec.)	Other (Spec.)
Ophthalmic	Ophthalmic							
Fetal Imaging & Other	Fetal/Obstetrics (See Note 3)							
	Abdominal (See Note 10)	P	P	P	P	P	Note 1	Note 7
	Intra-operative (See Note 6)							
	Intra-operative (Neuro.)							
	Laparoscopic							
	Pediatric							
	Small Organ (See Note 5)							
	Neonatal Cephalic							
	Adult Cephalic	P	P	P	P	P	Note 1	Note 7
	Trans-rectal							
	Trans-vaginal							
	Trans-urethral							
	Trans-esoph. (non-Cardiac)							
	Musculo-skel. (Convent.)							
	Musculo-skel. (Superfic.)							
	Intra-luminal							
	Other (See Note 13)							
Cardiac	Cardiac Adult	P	P	P	P	P	Note 1	Note 4, 7
	Cardiac Pediatric	P	P	P	P	P	Note 1	Note 4, 7
	Trans-esophageal (Cardiac)							
	Other (spec.)							
Peripheral Vessel	Peripheral vessel							
	Other (spec.)							

N= new indication; P= previously cleared by FDA K182932; E= added under Appendix E

Additional Comments:

Color Doppler includes Power (Amplitude) Doppler

Note 1: B+M, B+PW, B+C, B+PD, B+DPD, B+PPI, B+TD, B+CW, B+C+PW, B+PD+PW, B+DPD+PW, B+PPI+PW, B+TD+PW, B+C+M, Dual/Quad, B+C+CW, B+PD+CW, B+E, B+B/C

Note 2: Includes imaging for guidance of biopsy

Note 3: Includes infertility monitoring of follicle development

Note 4: Color M-mode

Note 5: For example: thyroid, parathyroid, breast, scrotum and penis in adult, pediatric and neonatal patients

Note 6: Abdominal organs and peripheral vessel

Note 7: Tissue Harmonic Imaging (THI)

Note 8: 3D imaging

Note 9: Spatial Compound Imaging

Note 10: Includes Renal, Gynecology/Pelvis

Note 11: Panoramic imaging

Note 12: ElastoScan

Note 13: Includes Urology/Prostate

Note 14: S-Fusion

Note 15: S-Shearwave

Concurrence of Center for Devices and Radiological Health (CDRH)
Prescription Use (Per 21 CFR 801.109)

DIAGNOSTIC ULTRASOUND INDICATIONS FOR USE STATEMENT

510(k) No.:

Device Name: PA3-8B for use with RS85

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation (*includes simultaneous B-mode)						
General (Track I only)	Specific (Tracks I & III)	B	M	PWD	CWD	Color Doppler*	Combined* (Spec.)	Other (Spec.)
Ophthalmic	Ophthalmic							
Fetal Imaging & Other	Fetal/Obstetrics (See Note 3)							
	Abdominal (See Note 10)	P	P	P	P	P	Note 1	Note 7
	Intra-operative (See Note 6)							
	Intra-operative (Neuro.)							
	Laparoscopic							
	Pediatric	P	P	P	P	P	Note 1	Note 7
	Small Organ (See Note 5)							
	Neonatal Cephalic	P	P	P	P	P	Note 1	Note 7
	Adult Cephalic	P	P	P	P	P	Note 1	Note 7
	Trans-rectal							
	Trans-vaginal							
	Trans-urethral							
	Trans-esoph. (non-Cardiac)							
	Musculo-skel. (Convent.)							
	Musculo-skel. (Superfic.)							
	Intra-luminal							
	Other (See Note 13)							
Cardiac	Cardiac Adult	P	P	P	P	P	Note 1	Note 4, 7
	Cardiac Pediatric	P	P	P	P	P	Note 1	Note 4, 7
	Trans-esophageal (Cardiac)							
	Other (spec.)							
Peripheral Vessel	Peripheral vessel							
	Other (spec.)							

N= new indication; P= previously cleared by FDA K182932; E= added under Appendix E

Additional Comments:

Color Doppler includes Power (Amplitude) Doppler

Note 1: B+M, B+PW, B+C, B+PD, B+DPD, B+PPI, B+TD, B+CW, B+C+PW, B+PD+PW, B+DPD+PW, B+PPI+PW, B+TD+PW, B+C+M, Dual/Quad, B+C+CW, B+PD+CW, B+E, B+B/C

Note 2: Includes imaging for guidance of biopsy

Note 3: Includes infertility monitoring of follicle development

Note 4: Color M-mode

Note 5: For example: thyroid, parathyroid, breast, scrotum and penis in adult, pediatric and neonatal patients

Note 6: Abdominal organs and peripheral vessel

Note 7: Tissue Harmonic Imaging (THI)

Note 8: 3D imaging

Note 9: MultiVision (Spatial Compound Imaging)

Note 10: Includes Renal, Gynecology/Pelvis

Note 11: Panoramic imaging

Note 12: ElastoScan

Note 13: Includes Urology/Prostate

Note 14: S-Fusion

Note 15: S-Shearwave

Concurrence of Center for Devices and Radiological Health (CDRH)
Prescription Use (Per 21 CFR 801.109)

DIAGNOSTIC ULTRASOUND INDICATIONS FOR USE STATEMENT

510(k) No.:

Device Name: PA4-12B for use with RS85

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation (*includes simultaneous B-mode)						
General (Track I only)	Specific (Tracks I & III)	B	M	PWD	CWD	Color Doppler*	Combined* (Spec.)	Other (Spec.)
Ophthalmic	Ophthalmic							
Fetal Imaging & Other	Fetal/Obstetrics (See Note 3)							
	Abdominal (See Note 10)	P	P	P	P	P	Note 1	Note 7
	Intra-operative (See Note 6)							
	Intra-operative (Neuro.)							
	Laparoscopic							
	Pediatric	P	P	P	P	P	Note 1	Note 7
	Small Organ (See Note 5)							
	Neonatal Cephalic	P	P	P	P	P	Note 1	Note 7
	Adult Cephalic							
	Trans-rectal							
	Trans-vaginal							
	Trans-urethral							
	Trans-esoph. (non-Cardiac)							
	Musculo-skel. (Convent.)							
	Musculo-skel. (Superfic.)							
	Intra-luminal							
	Other (See Note 13)							
Cardiac	Cardiac Adult	P	P	P	P	P	Note 1	Note 4, 7
	Cardiac Pediatric	P	P	P	P	P	Note 1	Note 4, 7
	Trans-esophageal (Cardiac)							
	Other (spec.)							
Peripheral Vessel	Peripheral vessel							
	Other (spec.)							

N= new indication; P= previously cleared by FDA K182932; E= added under Appendix E

Additional Comments:

Color Doppler includes Power (Amplitude) Doppler

Note 1: B+M, B+PW, B+C, B+PD, B+DPD, B+PPI, B+TD, B+CW, B+C+PW, B+PD+PW, B+DPD+PW, B+PPI+PW, B+TD+PW, B+C+M, Dual/Quad, B+C+CW, B+PD+CW, B+E, B+B/C

Note 2: Includes imaging for guidance of biopsy

Note 3: Includes infertility monitoring of follicle development

Note 4: Color M-mode

Note 5: For example: thyroid, parathyroid, breast, scrotum and penis in adult, pediatric and neonatal patients

Note 6: Abdominal organs and peripheral vessel

Note 7: Tissue Harmonic Imaging (THI)

Note 8: 3D imaging

Note 9: MultiVision (Spatial Compound Imaging)

Note 10: Includes Renal, Gynecology/Pelvis

Note 11: Panoramic imaging

Note 12: ElastoScan

Note 13: Includes Urology/Prostate

Note 14: S-Fusion

Note 15: S-Shearwave

Concurrence of Center for Devices and Radiological Health (CDRH)
Prescription Use (Per 21 CFR 801.109)

DIAGNOSTIC ULTRASOUND INDICATIONS FOR USE STATEMENT

510(k) No.:

Device Name: MMPT3-7 for use with RS85

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation (*includes simultaneous B-mode)						
General (Track I only)	Specific (Tracks I & III)	B	M	PWD	CWD	Color Doppler*	Combined* (Spec.)	Other (Spec.)
Ophthalmic	Ophthalmic							
Fetal Imaging & Other	Fetal/Obstetrics (See Note 3)							
	Abdominal (See Note 12)							
	Intra-operative (See Note 6)							
	Intra-operative (Neuro.)							
	Laparoscopic							
	Pediatric							
	Small Organ (See Note 5)							
	Neonatal Cephalic							
	Adult Cephalic							
	Trans-rectal							
	Trans-vaginal							
	Trans-urethral							
	Trans-esoph. (non-Cardiac)							
	Musculo-skel. (Convent.)							
	Musculo-skel. (Superfic.)							
	Intra-luminal							
	Other (spec.) (See Note 13)							
Cardiac	Cardiac Adult							
	Cardiac Pediatric							
	Trans-esophageal (Cardiac)	P	P	P	P	P	Note 1	Note 4, 7
	Other (spec.)							
Peripheral Vessel	Peripheral vessel							
	Other (spec.)							

N= new indication; P= previously cleared by FDA K182932; E= added under Appendix E

Additional Comments:

Color Doppler includes Power (Amplitude) Doppler

Note 1: B+M, B+PW, B+C, B+PD, B+DPD, B+PPI, B+TD, B+CW, B+C+PW, B+PD+PW, B+DPD+PW, B+PPI+PW, B+TD+PW, B+C+M, Dual/Quad, B+C+CW, B+PD+CW, B+E, B+B/C

Note 2: Includes imaging for guidance of biopsy

Note 3: Includes infertility monitoring of follicle development

Note 4: Color M-mode

Note 5: For example: thyroid, parathyroid, breast, scrotum and penis in adult, pediatric and neonatal patients

Note 6: Abdominal organs and peripheral vessel

Note 7: Tissue Harmonic Imaging (THI)

Note 8: 3D imaging

Note 9: MultiVision (Spatial Compound Imaging)

Note 10: Includes Renal, Gynecology/Pelvis

Note 11: Panoramic imaging

Note 12: ElastoScan

Note 13: Includes Urology/Prostate

Note 14: S-Fusion

Note 15: S-Shearwave

Concurrence of Center for Devices and Radiological Health (CDRH)
Prescription Use (Per 21 CFR 801.109)

DIAGNOSTIC ULTRASOUND INDICATIONS FOR USE STATEMENT

510(k) No.:

Device Name: CV1-8A for use with RS85

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation (*includes simultaneous B-mode)						
General (Track I only)	Specific (Tracks I & III)	B	M	PWD	CWD	Color Doppler*	Combined* (Spec.)	Other (Spec.)
Ophthalmic	Ophthalmic							
Fetal Imaging & Other	Fetal/Obstetrics (See Note 3)	P	P	P		P	Note 1	Note 2, 4, 7, 8, 9, 11
	Abdominal (See Note 10)	P	P	P		P	Note 1	Note 2, 4, 7, 8, 9
	Intra-operative (See Note 6)							
	Intra-operative (Neuro.)							
	Laparoscopic							
	Pediatric							
	Small Organ (See Note 5)							
	Neonatal Cephalic							
	Adult Cephalic							
	Trans-rectal							
	Trans-vaginal							
	Trans-urethral							
	Trans-esoph. (non-Cardiac)							
	Musculo-skel. (Convent.)							
	Musculo-skel. (Superfic.)							
	Intra-luminal							
	Other (See Note 13)							
Cardiac	Cardiac Adult							
	Cardiac Pediatric							
	Trans-esophageal (Cardiac)							
	Other (spec.)							
Peripheral Vessel	Peripheral vessel							
	Other (spec.)							

N= new indication; P= previously cleared by FDA K182932; E= added under Appendix E

Additional Comments:

Color Doppler includes Power (Amplitude) Doppler

Note 1: B+M, B+PW, B+C, B+PD, B+DPD, B+PPI, B+TD, B+CW, B+C+PW, B+PD+PW, B+DPD+PW, B+PPI+PW, B+TD+PW, B+C+M, Dual/Quad, B+C+CW, B+PD+CW, B+E, B+B/C

Note 2: Includes imaging for guidance of biopsy

Note 3: Includes infertility monitoring of follicle development

Note 4: Color M-mode

Note 5: For example: thyroid, parathyroid, breast, scrotum and penis in adult, pediatric and neonatal patients

Note 6: Abdominal organs and peripheral vessel

Note 7: Tissue Harmonic Imaging (THI)

Note 8: 3D imaging

Note 9: MultiVision (Spatial Compound Imaging)

Note 10: Includes Renal, Gynecology/Pelvis

Note 11: Panoramic imaging

Note 12: ElastoScan

Note 13: Includes Urology/Prostate

Note 14: S-Fusion

Note 15: S-Shearwave

Concurrence of Center for Devices and Radiological Health (CDRH)
Prescription Use (Per 21 CFR 801.109)

DIAGNOSTIC ULTRASOUND INDICATIONS FOR USE STATEMENT

510(k) No.:

Device Name: V5-9 for use with RS85

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation (*includes simultaneous B-mode)						
General (Track I only)	Specific (Tracks I & III)	B	M	PWD	CWD	Color Doppler*	Combined* (Spec.)	Other (Spec.)
Ophthalmic	Ophthalmic							
Fetal Imaging & Other	Fetal/Obstetrics (See Note 3)	P	P	P		P	Note 1	Note 2, 7, 8, 9
	Abdominal (See Note 10)	P	P	P		P	Note 1	Note 2, 7, 8, 9, 12
	Intra-operative (See Note 6)							
	Intra-operative (Neuro.)							
	Laparoscopic							
	Pediatric							
	Small Organ (See Note 5)							
	Neonatal Cephalic							
	Adult Cephalic							
	Trans-rectal	P	P	P		P	Note 1	Note 2, 7, 8, 9, 12
	Trans-vaginal	P	P	P		P	Note 1	Note 2, 7, 8, 9, 12
	Trans-urethral							
	Trans-esoph. (non-Cardiac)							
	Musculo-skel. (Convent.)							
	Musculo-skel. (Superfic.)							
	Intra-luminal							
	Other (See Note 13)	P	P	P		P	Note 1	Note 2, 7, 8, 9, 12
Cardiac	Cardiac Adult							
	Cardiac Pediatric							
	Trans-esophageal (Cardiac)							
	Other (spec.)							
Peripheral Vessel	Peripheral vessel							
	Other (spec.)							

N= new indication; P= previously cleared by FDA K182932; E= added under Appendix E

Additional Comments:

Color Doppler includes Power (Amplitude) Doppler

Note 1: B+M, B+PW, B+C, B+PD, B+DPD, B+PPI, B+TD, B+CW, B+C+PW, B+PD+PW, B+DPD+PW, B+PPI+PW, B+TD+PW, B+C+M, Dual/Quad, B+C+CW, B+PD+CW, B+E, B+B/C

Note 2: Includes imaging for guidance of biopsy

Note 3: Includes infertility monitoring of follicle development

Note 4: Color M-mode

Note 5: For example: thyroid, parathyroid, breast, scrotum and penis in adult, pediatric and neonatal patients

Note 6: Abdominal organs and peripheral vessel

Note 7: Tissue Harmonic Imaging (THI)

Note 8: 3D imaging

Note 9: Spatial Compound Imaging

Note 10: Includes Renal, Gynecology/Pelvis

Note 11: Panoramic imaging

Note 12: ElastoScan

Note 13: Includes Urology/Prostate

Note 14: S-Fusion

Note 15: S-Shearwave

Concurrence of Center for Devices and Radiological Health (CDRH)
Prescription Use (Per 21 CFR 801.109)

DIAGNOSTIC ULTRASOUND INDICATIONS FOR USE STATEMENT

510(k) No.:

Device Name: LV3-14A for use with RS85

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation (*includes simultaneous B-mode)						
General (Track I only)	Specific (Tracks I & III)	B	M	PWD	CWD	Color Doppler*	Combined* (Spec.)	Other (Spec.)
Ophthalmic	Ophthalmic							
Fetal Imaging & Other	Fetal/Obstetrics (See Note 3)							
	Abdominal (See Note 10)							
	Intra-operative (See Note 6)							
	Intra-operative (Neuro.)							
	Laparoscopic							
	Pediatric							
	Small Organ (See Note 5)	P	P	P		P	Note 1	Note 2, 5, 6, 7, 8, 9, 11
	Neonatal Cephalic							
	Adult Cephalic							
	Trans-rectal							
	Trans-vaginal							
	Trans-urethral							
	Trans-esoph. (non-Cardiac)							
	Musculo-skel. (Convent.)	P	P	P		P	Note 1	Note 2, 5, 6, 7, 8, 9, 11
	Musculo-skel. (Superfic.)	P	P	P		P	Note 1	Note 2, 5, 6, 7, 8, 9, 11
	Intra-luminal							
	Other (See Note 13)							
Cardiac	Cardiac Adult							
	Cardiac Pediatric							
	Trans-esophageal (Cardiac)							
	Other (spec.)							
Peripheral Vessel	Peripheral vessel	P	P	P		P	Note 1	Note 2, 5, 6, 7, 8, 9, 11
	Other (spec.)							

N= new indication; P= previously cleared by FDA K182932; E= added under Appendix E

Additional Comments:

Color Doppler includes Power (Amplitude) Doppler

Note 1: B+M, B+PW, B+C, B+PD, B+DPD, B+PPI, B+TD, B+CW, B+C+PW, B+PD+PW, B+DPD+PW, B+PPI+PW, B+TD+PW, B+C+M, Dual/Quad, B+C+CW, B+PD+CW, B+E, B+B/C

Note 2: Includes imaging for guidance of biopsy

Note 3: Includes infertility monitoring of follicle development

Note 4: Color M-mode

Note 5: For example: thyroid, parathyroid, breast, scrotum and penis in adult, pediatric and neonatal patients

Note 6: Abdominal organs and peripheral vessel

Note 7: Tissue Harmonic Imaging (THI)

Note 8: 3D imaging

Note 9: Spatial Compound Imaging

Note 10: Includes Renal, Gynecology/Pelvis

Note 11: Panoramic imaging

Note 12: ElastoScan

Note 13: Includes Urology/Prostate

Note 14: S-Fusion

Note 15: S-Shearwave

Concurrence of Center for Devices and Radiological Health (CDRH)
Prescription Use (Per 21 CFR 801.109)

DIAGNOSTIC ULTRASOUND INDICATIONS FOR USE STATEMENT

510(k) No.:

Device Name: CW6.0 for use with RS85

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation (*includes simultaneous B-mode)						
General (Track I only)	Specific (Tracks I & III)	B	M	PWD	CWD	Color Doppler*	Combined* (Spec.)	Other (Spec.)
Ophthalmic	Ophthalmic							
Fetal Imaging & Other	Fetal/Obstetrics (See Note 3)							
	Abdominal (See Note 10)							
	Intra-operative (See Note 6)							
	Intra-operative (Neuro.)							
	Laparoscopic							
	Pediatric							
	Small Organ (See Note 5)							
	Neonatal Cephalic							
	Adult Cephalic							
	Trans-rectal							
	Trans-vaginal							
	Trans-urethral							
	Trans-esoph. (non-Cardiac)							
	Musculo-skel. (Convent.)							
	Musculo-skel. (Superfic.)							
	Intra-luminal							
	Other (See Note 13)							
Cardiac	Cardiac Adult				P			
	Cardiac Pediatric				P			
	Trans-esophageal (Cardiac)							
	Other (spec.)							
Peripheral Vessel	Peripheral vessel				P			
	Other (spec.)							

N= new indication; P= previously cleared by FDA K182932; E= added under Appendix E

Additional Comments:

Color Doppler includes Power (Amplitude) Doppler

Note 1: B+M, B+PW, B+C, B+PD, B+DPD, B+PPI, B+TD, B+CW, B+C+PW, B+PD+PW, B+DPD+PW, B+PPI+PW, B+TD+PW, B+C+M, Dual/Quad, B+C+CW, B+PD+CW, B+E, B+B/C

Note 2: Includes imaging for guidance of biopsy

Note 3: Includes infertility monitoring of follicle development

Note 4: Color M-mode

Note 5: For example: thyroid, parathyroid, breast, scrotum and penis in adult, pediatric and neonatal patients

Note 6: Abdominal organs and peripheral vessel

Note 7: Tissue Harmonic Imaging (THI)

Note 8: 3D imaging

Note 9: Spatial Compound Imaging

Note 10: Includes Renal, Gynecology/Pelvis

Note 11: Panoramic imaging

Note 12: ElastoScan

Note 13: Includes Urology/Prostate

Note 14: S-Fusion

Note 15: S-Shearwave

Concurrence of Center for Devices and Radiological Health (CDRH)
Prescription Use (Per 21 CFR 801.109)

DIAGNOSTIC ULTRASOUND INDICATIONS FOR USE STATEMENT

510(k) No.:

Device Name: DP2B for use with RS85

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation (*includes simultaneous B-mode)						
General (Track I only)	Specific (Tracks I & III)	B	M	PWD	CWD	Color Doppler*	Combined* (Spec.)	Other (Spec.)
Ophthalmic	Ophthalmic							
Fetal Imaging & Other	Fetal/Obstetrics (See Note 3)							
	Abdominal (See Note 10)							
	Intra-operative (See Note 6)							
	Intra-operative (Neuro.)							
	Laparoscopic							
	Pediatric							
	Small Organ (See Note 5)							
	Neonatal Cephalic							
	Adult Cephalic							
	Trans-rectal							
	Trans-vaginal							
	Trans-urethral							
	Trans-esoph. (non-Cardiac)							
	Musculo-skel. (Convent.)							
	Musculo-skel. (Superfic.)							
	Intra-luminal							
	Other (See Note 13)							
Cardiac	Cardiac Adult				P			
	Cardiac Pediatric							
	Trans-esophageal (Cardiac)							
	Other (spec.)							
Peripheral Vessel	Peripheral vessel							
	Other (spec.)							

N= new indication; P= previously cleared by FDA K182932; E= added under Appendix E

Additional Comments:

Color Doppler includes Power (Amplitude) Doppler

Note 1: B+M, B+PW, B+C, B+PD, B+DPD, B+PPI, B+TD, B+CW, B+C+PW, B+PD+PW, B+DPD+PW, B+PPI+PW, B+TD+PW, B+C+M, Dual/Quad, B+C+CW, B+PD+CW, B+E, B+B/C

Note 2: Includes imaging for guidance of biopsy

Note 3: Includes infertility monitoring of follicle development

Note 4: Color M-mode

Note 5: For example: thyroid, parathyroid, breast, scrotum and penis in adult, pediatric and neonatal patients

Note 6: Abdominal organs and peripheral vessel

Note 7: Tissue Harmonic Imaging (THI)

Note 8: 3D imaging

Note 9: MultiVision (Spatial Compound Imaging)

Note 10: Includes Renal, Gynecology/Pelvis

Note 11: Panoramic imaging

Note 12: ElastoScan

Note 13: Includes Urology/Prostate

Note 14: S-Fusion

Note 15: S-Shearwave

Concurrence of Center for Devices and Radiological Health (CDRH)
Prescription Use (Per 21 CFR 801.109)

DIAGNOSTIC ULTRASOUND INDICATIONS FOR USE STATEMENT

510(k) No.:

Device Name: DP8B for use with RS85

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation (*includes simultaneous B-mode)						
General (Track I only)	Specific (Tracks I & III)	B	M	PWD	CWD	Color Doppler*	Combined* (Spec.)	Other (Spec.)
Ophthalmic	Ophthalmic							
Fetal Imaging & Other	Fetal/Obstetrics (See Note 3)							
	Abdominal (See Note 12)							
	Intra-operative (See Note 6)							
	Intra-operative (Neuro.)							
	Laparoscopic							
	Pediatric							
	Small Organ (See Note 5)							
	Neonatal Cephalic							
	Adult Cephalic							
	Trans-rectal							
	Trans-vaginal							
	Trans-urethral							
	Trans-esoph. (non-Cardiac)							
	Musculo-skel. (Convent.)							
	Musculo-skel. (Superfic.)							
	Intra-luminal							
	Other (See Note 13)							
Cardiac	Cardiac Adult				P			
	Cardiac Pediatric				P			
	Trans-esophageal (Cardiac)							
	Other (spec.)							
Peripheral Vessel	Peripheral vessel				P			
	Other (spec.)							

N= new indication; P= previously cleared by FDA K182932; E= added under Appendix E

Additional Comments:

Color Doppler includes Power (Amplitude) Doppler

Note 1: B+M, B+PW, B+C, B+PD, B+DPD, B+PPI, B+TD, B+CW, B+C+PW, B+PD+PW, B+DPD+PW, B+PPI+PW, B+TD+PW, B+C+M, Dual/Quad, B+C+CW, B+PD+CW, B+E, B+B/C

Note 2: Includes imaging for guidance of biopsy

Note 3: Includes infertility monitoring of follicle development

Note 4: Color M-mode

Note 5: For example: thyroid, parathyroid, breast, scrotum and penis in adult, pediatric and neonatal patients

Note 6: Abdominal organs and peripheral vessel

Note 7: Tissue Harmonic Imaging (THI)

Note 8: 3D imaging

Note 9: MultiVision (Spatial Compound Imaging)

Note 10: Includes Renal, Gynecology/Pelvis

Note 11: Panoramic imaging

Note 12: ElastoScan

Note 13: Includes Urology/Prostate

Note 14: S-Fusion

Note 15: S-Shearwave

Concurrence of Center for Devices and Radiological Health (CDRH)
Prescription Use (Per 21 CFR 801.109)

DIAGNOSTIC ULTRASOUND INDICATIONS FOR USE STATEMENT

510(k) No.:

Device Name: EV3-10B for use with RS85

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation (*includes simultaneous B-mode)						
General (Track I only)	Specific (Tracks I & III)	B	M	PWD	CWD	Color Doppler*	Combined* (Spec.)	Other (Spec.)
Ophthalmic	Ophthalmic							
Fetal Imaging & Other	Fetal/Obstetrics (See Note 3)	P	P	P		P	Note 1	Note 2, 7, 8, 9, 11
	Abdominal (See Note 10)	P	P	P		P	Note 1	Note 2, 7, 8, 9, 11, 12
	Intra-operative (See Note 6)							
	Intra-operative (Neuro.)							
	Laparoscopic							
	Pediatric							
	Small Organ (See Note 5)							
	Neonatal Cephalic							
	Adult Cephalic							
	Trans-rectal	P	P	P		P	Note 1	Note 2, 7, 8, 9, 11, 12
	Trans-vaginal	P	P	P		P	Note 1	Note 2, 7, 8, 9, 11, 12
	Trans-urethral							
	Trans-esoph. (non-Cardiac)							
	Musculo-skel. (Convent.)							
	Musculo-skel. (Superfic.)							
	Intra-luminal							
	Other (See Note 13)	P	P	P		P	Note 1	Note 2, 7, 8, 9, 11, 12
Cardiac	Cardiac Adult							
	Cardiac Pediatric							
	Trans-esophageal (Cardiac)							
	Other (spec.)							
Peripheral Vessel	Peripheral vessel							
	Other (spec.)							

N= new indication; P= previously cleared by FDA K182932; E= added under Appendix E

Additional Comments:

Color Doppler includes Power (Amplitude) Doppler

Note 1: B+M, B+PW, B+C, B+PD, B+DPD, B+PPI, B+TD, B+CW, B+C+PW, B+PD+PW, B+DPD+PW, B+PPI+PW, B+TD+PW, B+C+M, Dual/Quad, B+C+CW, B+PD+CW, B+E, B+B/C

Note 2: Includes imaging for guidance of biopsy

Note 3: Includes infertility monitoring of follicle development

Note 4: Color M-mode

Note 5: For example: thyroid, parathyroid, breast, scrotum and penis in adult, pediatric and neonatal patients

Note 6: Abdominal organs and peripheral vessel

Note 7: Tissue Harmonic Imaging (THI)

Note 8: 3D imaging

Note 9: MultiVision (Spatial Compound Imaging)

Note 10: Includes Renal, Gynecology/Pelvis

Note 11: Panoramic imaging

Note 12: ElastoScan

Note 13: Includes Urology/Prostate

Note 14: S-Fusion

Note 15: S-Shearwave

Concurrence of Center for Devices and Radiological Health (CDRH)
Prescription Use (Per 21 CFR 801.109)

510(K) Summary: K191115

In accordance with 21 CFR 807.92 the following summary of information is provided:

1. Date Prepared - April 25, 2019
2. Manufacturer
SAMSUNG MEDISON CO., LTD.
3366, Hanseo-ro, Nam-myeon,
Hongcheon-gun, Gangwon-do 25108,
REPUBLIC OF KOREA
3. Primary Contact Person
Ji Yea Lee
Regulatory Affairs Specialist
Phone: +82.2.2194.1594
Fax: +82. 31.8017.9576
Email: jiyea722.lee@samsungmedison.com

Secondary Contact Person
Scully Kim
Regulatory Affairs Specialist
Phone: +82.2.2194.1312
Fax: +82. 31.8017.9576
Email: scully.kim@samsungmedison.com

4. Proposed Device
 - Common/Usual Name: Diagnostic Ultrasound System and Accessories
 - Proprietary Name: RS85 Diagnostic Ultrasound System
 - Common Name: Diagnostic Ultrasound System
 - Classification Names: system, imaging, pulsed doppler, ultrasonic
 - Product Code: IYN, IYO, ITX
 - Regulation: 892.1550, 892.1560, 892.1570
5. Predicate Device
RS85 Diagnostic Ultrasound System (K182932)

6. Device Description
The RS85 is a general purpose, mobile, software controlled, diagnostic ultrasound system. Its function is to acquire ultrasound data and to display the data as 2D mode, M mode, Color Doppler imaging, Power Doppler imaging (including Directional Power Doppler mode; S-Flow), PW Spectral Doppler mode, CW Spectral Doppler mode, Harmonic imaging, Tissue Doppler imaging, Tissue Doppler Wave, 3D imaging mode (real-time 4D imaging mode), Elastoscanner⁺ Mode or as a combination of these modes. The RS85 also gives the operator the ability to measure anatomical structures and offers analysis packages that provide information that may aid in making a diagnosis by competent health care professionals. The RS85 has real time acoustic output display with two basic indices, a mechanical index and a thermal index, which are both automatically displayed.

7. Indications for Use
The ultrasound diagnostic system and probes are designed to obtain ultrasound images and analyze body

fluids.

The clinical applications include: Fetal/Obstetrics, Abdominal, Gynecology, Intraoperative, Pediatric, Small Organ, Neonatal Cephalic, Adult Cephalic, Trans-rectal, Trans-vaginal, Muscular-Skeletal (Conventional, Superficial), Urology, Cardiac Adult, Cardiac Pediatric, Trans-esophageal (Cardiac) and Peripheral vessel.

The Indications for use of the proposed device is the same as predicates.

8. Technology

The RS85 employs the same fundamental scientific technology as its predicate device(s).

9. Determination of Substantial Equivalence

The proposed RS85 is same to the predicate device with regards to intended use, imaging capabilities, technological characteristics and safety and effectiveness.

A comparison of the proposed device to the currently marketed predicates is provided in the table below.

<Change List>

RS85	Addition of V1.05	Remarks
Clinical application	No addition	
Operation of modes	No addition	
Applied transducers	No addition	
SW Features	Improvement of the cleared Software : Setup&Preference, Patient Management, Measurement, ScanMode(CEUS+), S-Fusion	
HW Features	No addition	

Device name (510(k) clearance)	RS85 (Under Review)	RS85 (K182932) Predicate
Manufacturer	Samsung Medison, Co.,LTD	Samsung Medison, Co.,LTD
Clinical Application	- Fetal/Obstetrics - Abdominal - Gynecology - Intra-operative - Pediatric - Small Organ, Neonatal Cephalic - Adult Cephalic - Trans-rectal - Trans-vaginal - Trans-esoph. (non-Cardiac) - Muscular-Skeletal (Conventional, Superficial) - Urology - Cardiac Adult - Cardiac Pediatric - Trans-esophageal (Cardiac) - Peripheral vessel	- Fetal/Obstetrics - Abdominal - Gynecology - Intra-operative - Pediatric - Small Organ, Neonatal Cephalic - Adult Cephalic - Trans-rectal - Trans-vaginal - Trans-esoph. (non-Cardiac) - Muscular-Skeletal (Conventional, Superficial) - Urology - Cardiac Adult - Cardiac Pediatric - Trans-esophageal (Cardiac) - Peripheral vessel
Scanhead Types	- Linear Array - Curved Linear Array - Endocavity - Phased Array - Static Probes	- Linear Array - Curved Linear Array - Endocavity - Phased Array - Static Probes
Scanhead Frequency	1.0 ~ 20.0 MHz	1.0 ~ 20.0 MHz
Acoustic Output Display & FDA Limits	- Display Feature for Higher Output–Track3 - MI Output Display - TI Output Display	- Display Feature for Higher Output–Track3 - MI Output Display - TI Output Display
Modes of Operation	- B-mode - M-mode - Pulsed wave (PW) Doppler - Continuous wave (CW) Doppler - Color Doppler - Power Amplitude Doppler - Tissue Harmonic Imaging - 3D/4D imaging mode - Combined modes	- B-mode - M-mode - Pulsed wave (PW) Doppler - Continuous wave (CW) Doppler - Color Doppler - Power Amplitude Doppler - Tissue Harmonic Imaging - 3D/4D imaging mode - Combined modes
#Transmit Channels	192	192
#Receive Channels	192	192
System Characteristics:	- Beamformer 192 - Mobile cart	- Beamformer 192 - Mobile cart

	- LCD Monitor (LED Backlight unit) - 256 gray shades on monitor - 100-240VAC, 1100VA, 50/60Hz	- LCD Monitor (LED Backlight unit) - 256 gray shades on monitor - 100-240VAC, 1100VA, 50/60Hz
Product Safety Certification	- IEC 60601-1 - UL 60601-1 - CSA C22.2 No.601.1 - IEC 60601-2-37	- IEC 60601-1 - UL 60601-1 - CSA C22.2 No.601.1 - IEC 60601-2-37
EMC Compliance	IEC 60601-1-2	IEC 60601-1-2
Acoustic Output Display Standard	Track 3	Track 3
Biocompatibility Compliance	ISO10993-1	ISO10993-1
Software Functionality	- Q Scan - ClearVision - MultiVision - Panoramic - Needle Mate - Auto IMT+ - Strain+ - Stress Echo - Elastoscan - E-Thyroid - E-Breast - E-Strain - S-Shearwave - S-Detect for Breast - S-Detect for Thyroid - HQ Vision - S-Fusion - Clear Track - Virtual Track - S-3D Arterial Analysis - ADVR - 3D Imaging - (Volume Data Acquisition) - 3D Imaging presentation - 3D Cine/4D Cine - 3D Rendering MPR - 3D XI MSV/Oblique View - Volume CT - 3D MagiCut - Volume Calculation - (VOCAL, XI VOCAL) - XI STIC - HDVI - Realistic Vue - Natural Vue	- Q Scan - ClearVision - MultiVision - Panoramic - Needle Mate - Auto IMT+ - Strain+ - Stress Echo - Elastoscan - E-Thyroid - E-Breast - E-Strain - S-Shearwave - S-Detect for Breast - S-Detect for Thyroid - HQ Vision - S-Fusion - Clear Track - Virtual Track - S-3D Arterial Analysis - ADVR - 3D Imaging - (Volume Data Acquisition) - 3D Imaging presentation - 3D Cine/4D Cine - 3D Rendering MPR - 3D XI MSV/Oblique View - Volume CT - 3D MagiCut - Volume Calculation - (VOCAL, XI VOCAL) - XI STIC - HDVI - Realistic Vue - Natural Vue

	- CEUS+ - MV-Flow - S-Shearwave Imaging	- CEUS+ - MV-Flow - S-Shearwave Imaging
Transducer	- CA1-7A - CF4-9 - CA2-8A - CA3-10A - E3-12A - L3-12A - EA2-11B - LA2-9A - LA3-16A - LA3-16AI - LA4-18B - LM4-15B - PM1-6A - PA3-8B - PA4-12B - MMPT3-7 - CV1-8A - V5-9 - LV3-14A - CW6.0 - DP2B - DP8B - EV3-10B	- CA1-7A - CF4-9 - CA2-8A - CA3-10A - E3-12A - L3-12A - EA2-11B - LA2-9A - LA3-16A - LA3-16AI - LA4-18B - LM4-15B - PM1-6A - PA3-8B - PA4-12B - MMPT3-7 - CV1-8A - V5-9 - LV3-14A - CW6.0 - DP2B - DP8B - EV3-10B
Biopsy Guide	- BP-KIT-029 - BP-KIT-030 - BP-KIT-043 - BP-KIT-053 - BP-KIT-054 - BP-KIT-055 - BP-KIT-057 - BP-KIT-058 - BP-KIT-059 - BP-KIT-060 - BP-KIT-067 - BP-KIT-068 - BP-KIT-069 - BP-KIT-071 - BP-KIT-075 - BP-KIT-083	- BP-KIT-029 - BP-KIT-030 - BP-KIT-043 - BP-KIT-053 - BP-KIT-054 - BP-KIT-055 - BP-KIT-057 - BP-KIT-058 - BP-KIT-059 - BP-KIT-060 - BP-KIT-067 - BP-KIT-068 - BP-KIT-069 - BP-KIT-071 - BP-KIT-075 - BP-KIT-083
Digital Storage/Transfer Station	- Foot Switch - ECG (defibrillation proof) - Gel Warmer - Clear Track/ Virtual Track System Stand - WLAN Card	- Foot Switch - ECG - Gel Warmer - Clear Track/ Virtual Track System Stand - WLAN Card

10. Summary of Non-Clinical Test

The proposed device has been evaluated for acoustic output, biocompatibility, cleaning and disinfection effectiveness as well as thermal, electrical, electromagnetic and mechanical safety, and has been found to conform with applicable medical device safety standards. The RS85 and its applications comply with voluntary standards.

Test	Standards and FDA Guidance
Risk Management	ISO 14971 Second edition 2007 Medical devices - Application of risk management to medical devices
Electrical Safety	The RS85 Ultrasound System with defibrillation-proof ECG electrode was evaluated per the following standards. ANSI AAMI ES60601-1:2005/(R)2012 and A1:2012, C1:2009/(R)2012 and A2:2010/(R)2012 Medical Electrical Equipment - Part 1: General Requirements for basic safety and essential performance.
Electromagnetic Compatibility	IEC60601-1-2: 2014(4th Edition) Medical Electrical Equipment -- Part 1-2: General Requirements For Basic Safety And Essential Performance -- Collateral Standard: Electromagnetic Disturbances -- Requirements And Tests
Biocompatibility	ISO 10993-1 Fourth edition 2009-10-15 Biological evaluation of medical devices - Part 1: Evaluation and testing within a risk management process.
Reprocessing Medical Devices	FDA Guidance: Reprocessing Medical Devices in Health Care Settings: Validation Methods and Labeling issued on March 17, 2015, revised June 9, 2017.
Software/Firmware-driven Functionality	All migrated software functionality was evaluated using the same test criteria as the predicate for all applicable imaging modes to ensure that migration into a new system design did not compromise image quality with respect to the intended use of each feature. Guidance for the Content of Premarket Submissions for Software Contained in Medical Devices issued on May 11, 2005
Ultrasound Safety	Information for Manufacturers Seeking Marketing Clearance of Diagnostic Ultrasound Systems and Transducers issued September 9, 2008
	IEC60601-2-37:2007 + A1:2015, Particular requirements for the safety of ultrasonic medical diagnostic and monitoring equipment
	NEMA UD 2-2004 (R2009) Acoustic Output Measurement Standard for Diagnostic Ultrasound Equipment Revision 3
	NEMA UD 3-2004 (R2009) Standard for Real-Time Display of Thermal and Mechanical Acoustic Output Indices on Diagnostic Ultrasound Equipment, Revision 2

11. Summary of Clinical Tests

The subject of this premarket submission, RS85, did not require clinical studies to support substantial equivalence.

12. Conclusion

SAMSUNG MEDISON considers the RS85 to be as safe, as effective, and performance is substantially equivalent to the predicate device.

END of 510(K) Summary