



Shanghai United Imaging Healthcare Co., Ltd.  
% Xin Gao  
Regulatory Affairs Specialist  
No. 2258 Chengbei Road  
Shanghai, Shanghai 201807  
CHINA

May 29, 2019

Re: K191157  
Trade/Device Name: uMR 780, uMR 790  
Regulation Number: 21 CFR 892.1000  
Regulation Name: Magnetic resonance diagnostic device  
Regulatory Class: Class II  
Product Code: LNH  
Dated: April 30, 2019  
Received: May 1, 2019

Dear Xin Gao:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for

devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/CombinationProducts/GuidanceRegulatoryInformation/ucm597488.htm>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/>) and CDRH Learn (<http://www.fda.gov/Training/CDRHLearn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<http://www.fda.gov/DICE>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

For

Thalia T. Mills, Ph.D.  
Director  
Division of Radiological Health  
OHT7: Office of In Vitro Diagnostics  
and Radiological Health  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)

K191157

Device Name

uMR 780, uMR 790

Indications for Use (Describe)

The uMR 780/uMR790 system is indicated for use as a magnetic resonance diagnostic device (MRDD) that produces sagittal, transverse, coronal, and oblique cross sectional images, and spectroscopic images, and that display internal anatomical structure and/or function of the head, body and extremities.

These images and the physical parameters derived from the images when interpreted by a trained physician yield information that may assist the diagnosis. Contrast agents may be used depending on the region of interest of the scan.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
PRASStaff@fda.hhs.gov

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

**510 (k) SUMMARY**

K191157

**1. Date of Prepared**

April 30, 2019

**2. Sponsor Identification**

**Shanghai United Imaging Healthcare Co.,Ltd.**

No.2258 Chengbei Rd. Jiading District, 201807, Shanghai, China

Contact Person: Xin GAO

Position: Regulatory Affairs Specialist

Tel: +86-021-67076888-5386

Fax: +86-021-67076889

Email: [xin.gao@united-imaging.com](mailto:xin.gao@united-imaging.com)

**3. Identification of Proposed Device(s)**

**Trade Name:** uMR 780, uMR 790

**Common Name:** Magnetic Resonance Diagnostic Device

**Model(s):** uMR 780, uMR 790

**Regulation Number:** 21 CFR 892.1000

**Regulation Name:** Magnetic Resonance Diagnostic Device

**Regulatory Class:** II

**Product Code:** LNH

**4. Identification of Predicate Device(s)**

**Predicate Device #1**

**510(k) Number:** K181370

**Device Name:** uMR 780

**Regulation Number:** 21 CFR 892.1000

**Regulation Name:** Magnetic Resonance Diagnostic Device

**Regulatory Class:** II

**Product Code:** LNH

**Predicate Device #2**

**510(k) Number:** K181371

**Device Name:** uMR 790

**Regulation Number:** 21 CFR 892.1000

**Regulation Name:** Magnetic Resonance Diagnostic Device

**Regulatory Class:** II

**Product Code:** LNH

## **Reference Device**

**510(k) Number:** K183186

**Device Name:** Head Coil - 12, Head Coil - 32, Carotid Coil - 8, Temporomandibular Joint Coil - 4, Infant Coil - 24, Cardiac Coil - 24, Foot & Ankle Coil - 24

**Regulation Number:** 21 CFR 892.1000

**Regulation Name:** Magnetic Resonance Diagnostic Device

**Regulatory Class:** II

**Product Code:** MOS

## **5. Device Description**

The uMR 780 is a 3.0T superconducting magnetic resonance diagnostic device with a 65cm size patient bore. It consists of components such as magnet, RF power amplifier, RF coils, gradient power amplifier, gradient coils, patient table, spectrometer, computer, equipment cabinets, power distribution system, internal communication system, and vital signal module etc. The uMR 780 Magnetic Resonance Diagnostic Device is designed to conform to NEMA and DICOM standards.

The uMR 790 is a 3.0T superconducting magnetic resonance diagnostic device with a 60cm size patient bore. It consists of components such as magnet, RF power amplifier, RF coils, gradient power amplifier, gradient coils, patient table, spectrometer, computer, equipment cabinets, power distribution system, internal communication system, and vital signal module etc. The uMR 790 Magnetic Resonance Diagnostic Device is designed to conform to NEMA and DICOM standards.

The modification performed on the previously cleared devices (K181370 for uMR 780 and K181371 for uMR 790) is due to the modification of magnet. The modification, which do not affect the intended use or alter the fundamental scientific technology of the device, is following:

- modify magnet with a new model number;

In addition, seven coils which have been previously cleared by FDA via K183186 (Head Coil - 12, Head Coil - 32, Carotid Coil - 8, Temporomandibular Joint Coil - 4, Infant Coil - 24, Cardiac Coil - 24, Foot & Ankle Coil – 24) are added in this submission.

## **6. Indications for Use**

The uMR 780/uMR 790 system is indicated for use as a magnetic resonance diagnostic device (MRDD) that produces sagittal, transverse, coronal, and oblique

cross sectional images, and spectroscopic images, and that display internal anatomical structure and/or function of the head, body and extremities.

These images and the physical parameters derived from the images when interpreted by a trained physician yield information that may assist the diagnosis. Contrast agents may be used depending on the region of interest of the scan.

## **7. Technological Characteristic**

The technology characteristics of the modified uMR 780 and uMR 790, reflected in this 510(k) submission, do not alter the scientific technology of the devices and are substantially equivalent to those of the predicate devices.

## **8. Non-Clinical Tests**

The following testing was conducted on the proposed devices:

- Signal to Noise Ratio
- Geometric Distortion
- Image Uniformity
- Magnetic Field Homogeneity
- Magnetic Field Decay
- IEC 60601-2-33 Ed. 3.2 B:2015, Medical Electrical Equipment - Part 2-33: Particular Requirements For The Basic Safety And Essential Performance Of Magnetic Resonance Equipment For Medical Diagnostic

The test results demonstrated that the device performs as expected and thus, it is substantially equivalent to the predicate devices to which it has been compared.

## **9. Clinical Tests**

No clinical testing was conducted on the proposed devices.

## **10. Conclusion**

In accordance with the Federal Food, Drug and Cosmetic Act, 21 CFR Part 807 and based on the information provided in this premarket notification, we concludes that the modified uMR 780 and uMR 790 Magnetic Resonance Diagnostic Device are substantially equivalent to the predicate devices. It does not introduce new indications for use, and has the same technological characteristics and does not introduce new potential hazards or safety risks.