



September 12, 2019

Medacta International SA  
% Chris Lussier  
Director, Quality and Regulatory  
Medacta USA  
3973 Delp Street  
Memphis, Tennessee 38118

Re: K191300

Trade/Device Name: MectaLock Ti Suture Anchor  
Regulation Number: 21 CFR 888.3040  
Regulation Name: Smooth or threaded metallic bone fixation fastener  
Regulatory Class: Class II  
Product Code: MBI  
Dated: July 30, 2019  
Received: July 31, 2019

Dear Mr. Lussier:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal

statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Laurence D. Coyne, Ph.D.  
Assistant Director  
DHT6C: Division of Restorative, Repair  
and Trauma Devices  
OHT6: Office of Orthopedic Devices  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health

Enclosure

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
Food and Drug Administration

Form Approved: OMB No. 0910-0120  
Expiration Date: 06/30/2020  
See PRA Statement below.

### Indications for Use

510(k) Number (if known)

K191300

Device Name

MectaLock Ti Suture Anchor

Indications for Use (Describe)

The MectaLock Ti Suture Anchors are intended for use in arthroscopic or open surgical approaches for fixation of suture (soft tissue) to bone in shoulder in the following procedure:

- Shoulder: cuff rotator repair and biceps tenodesis.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

#### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
PRASStaff@fda.hhs.gov

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

## 2.0 510(k) Summary

### I. Submitter

Medacta International SA  
Strada Regina  
6874 Castel San Pietro (CH)  
Switzerland  
Phone (+41) 91 696 60 60  
Fax (+41) 91 696 60 66

Contact Person: Stefano Baj, Regulatory Affairs Manager  
Date Prepared: May 13, 2019

### II. Device

Device Proprietary Name:	MectaLock TI Suture Anchor
Common or Usual Name:	Suture Anchor
Classification Name:	Fastener, Fixation, Nondegradable, Soft Tissue
Primary Product Code:	MBI
Regulation Number:	21 CFR 888.3040
Device Classification	II

### III. Predicate Device

Substantial equivalence is claimed to the following device:

Primary Predicate:

- ConMed Linvatec, SuperRevo® Herculine™ Suture Anchor (K041713).

Additional Predicates:

- Medacta International SA, M-ARS ACL: Anatomic Ribbon Surgery System (K171640)
- DePuy Mitek, FASTIN RC Anchor (K060664)

### IV. Device Description

The MectaLock TI Suture anchor is an implantable device used for the soft tissues refixation (i.e.: muscles, tendons, ligaments..) composed of an anchoring component (titanium alloy anchor) and two Ultra High Molecular Weight PolyEthylene non-absorbable braided sutures.

This assembly, which stays within the patient during all its lifetime, is provided mounted on a dedicated disposable driver which allows the surgeon to insert and place the MectaLock TI Suture anchor into the patient. The driver can be disposed immediately after the implant is been placed.

The MectaLock TI Suture anchor comes in two different sizes:  $\varnothing 5.0$  and  $\varnothing 6.5$ mm with a fixed length of 15mm, to cover the intended population and bone quality.

## **V. Indications for Use**

The MectaLock TI Suture Anchors are intended for use in arthroscopic or open surgical approaches for fixation of suture (soft tissue) to bone in shoulder in the following procedure:

- Shoulder: cuff rotator repair and biceps tenodesis.

## **VI. Comparison of Technological Characteristics**

The MectaLock TI Suture Anchor and the predicate device share the following characteristics:

- Materials (Anchor: Ti 6Al-4V ELI [ASTM F136 & ISO 5832-3] and Suture: Ultra-High molecular weight polyethylene)
- Provided Sterile
- External shape (Tip and Thread)
- Suture typology
- Device Usage

The MectaLock TI Suture Anchor is technologically different from the predicate device as follows:

- Diameters
- Inner shape (Eyelet and Driver connection)

The materials used in the MectaLock TI Suture Anchor product are:

- Anchor: Ti 6Al-4V ELI according to ASTM F136 & ISO 5832-3
- Suture: Ultra High Molecular Weight PolyEthylene
- Disposable driver: Stainless steel and Polycarbonate medical grade

All of these materials were chosen in alignment with the predicate device and in according with the most common equivalent products in the orthopedic field.

Due to the extensive history of use in currently marketed medical devices, additional biocompatibility testing was deemed unnecessary for the MectaLock TI Suture Anchor components.

### *Discussion*

The technological differences between the subject and predicate devices do not raise new questions of safety and effectiveness. The MectaLock TI Suture Anchor is the same or similar to the predicate device in terms of materials of construction, external shape (Tip and Thread), device usage, suture typology and disposable driver design and sterility.

Although there is a difference in the position of the eyelet (internal for the MectaLock TI Suture anchor and external for the predicate device), and in the Driver connection, the intended use and functionality of the component are the same.

The product range is shared with the reference device DePuy Mitek, FASTIN RC Anchor (K060664).

Based on the comparison of technological characteristics and performance data provided within this submission, the data supports the substantial equivalence of the MectaLock TI Suture Anchor to the identified predicate devices.

## **VII. Performance Data**

Based on the risk analysis, a design comparison and cadaver workshops were conducted to written protocols. The following performance tests are being provided in support of the substantial equivalence determination:

### Non-Clinical Studies

- *DESIGN VALIDATION*
  - Design Validation, according to Medacta Design Validation Protocol A1 (Cadaver Workshop) M07.85.003 and Evaluation form Titanium anchor. *Test Report A1*.
  - MR compatibility, *MR Safety Evaluation – Titanium Anchor*
- *CHARACTERIZATION TESTING*
  - Cyclic and load-to-failure properties of suture anchors – according to Empa Test report No. 18-06-25\_5214019237\_1e\_Anchor-test\_final.pdf, according to Medacta Protocol IL 07.09.488\_rev.0. *Test report A2*.
- *PYROGENICITY:*
  - Bacterial endotoxin test (LAL test) according to European Pharmacopoeia §2.6.14 (which is equivalent to USP chapter <85>)
  - Pyrogen test according to USP chapter <151> for pyrogenicity determination.
- *STERILIZATION:*
  - ISO 11135:2014 Sterilization of health-care products - Ethylene Oxide - Requirements for the development, validation and routine control of a sterilization process for medical devices.
  - ISO 10993-7:2008 Biological evaluation of medical devices -- Part 7: Ethylene oxide sterilization residuals.

### Clinical Studies:

- No clinical studies were conducted.

## **VIII. Conclusion**

The information provided above supports that the MectaLock TI Suture Anchor is as safe and effective as the predicate devices. Therefore, it is concluded that the MectaLock TI Suture Anchor is substantially equivalent to the predicate device.