



July 11, 2019

Viseon, Inc.  
Cora Sim  
Regulatory Affairs Manager  
13900 Alton Parkway, Suite 125  
Irvine, California 92618

Re: K191579  
Trade/Device Name: Voyant System  
Regulation Number: 21 CFR 888.1100  
Regulation Name: Arthroscope  
Regulatory Class: Class II  
Product Code: HRX  
Dated: June 13, 2019  
Received: June 14, 2019

Dear Cora Sim:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's

requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Long Chen, Ph.D.  
Assistant Director  
DHT4A: Division of General Surgery Devices  
OHT4: Office of Surgical  
and Infection Control Devices  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)  
K191579

Device Name  
Voyant System

Indications for Use (Describe)

The Voyant System is indicated to provide minimally invasive access, visualization, illumination, and magnification of the surgical area of the spine.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

---

**Voyant System 510(k) Summary – K191579**

In accordance with Title 21 of the Code of Federal Regulations, Part 807, and in particular 21 CFR §807.92, the following summary of information is provided:

**A. Submitted by:**

Cora Sim  
Regulatory Affairs Manager  
Viseon, Inc.  
13900 Alton Parkway, Suite 125  
Irvine, CA 92618  
Telephone: (949) 662-3959 ext. 106  
Facsimile: (949) 581-4761

Date Prepared: July 1, 2019

**B. Device Name:**

Trade or Proprietary Name:	<i>Voyant System</i>
Common or Usual Name:	Arthroscope
Classification Name:	Arthroscope
Device Class:	Class II
Classification:	21 CFR § 888.1100
Product Code:	HRX

**C. Predicate Device:**

The subject *Voyant System* is substantially equivalent to the predicate device, *Voyant System* (K181885).

**D. Device Description:**

The *Voyant System* is composed of the *Voyant Sheath* and *Voyant Imager*, and *Accessories* which includes an *Image Control Box (ICB)* and connecting cables. The *Voyant Sheath* and *Voyant Imager* are supplied sterile by ethylene oxide (EtO) while the *ICB* and connecting cables are supplied non-sterile and are reusable. The *Voyant System* is used in minimally invasive spine surgery to provide access and visualization of the surgical area of the spine.

The *Voyant System* is a sterile, single-use access and visualization device consisting of a *Sheath* for access to the spine and an *Imager* for illumination and visualization of the surgical area. The *Sheaths* are available in multiple diameters and lengths to accommodate a variety of patient anatomies and spinal surgical procedures. Likewise, the *Voyant Imager* is available in corresponding diameters to attach to the *Voyant Sheath*. The *Voyant Imager* connects to the *Image Control Box (ICB)* via an *HDMI* cable and live video image is displayed to a connected external *HD* monitor in the operating room. The *ICB* functions to power the *Voyant Imager*, process video data from the *Voyant Imager* to display live video image to the connected third-party *HD* Monitor.

The purpose of this submission is to implement design changes to the predicate *Voyant System* (K181885) to create a product line extension which includes the *Voyant Sheath*, *Voyant Imager*, and *Accessories*. These modifications to the *Voyant System* do not change the indications for use of the device, nor do they change the fundamental scientific technology of the device. Specifically, the following design and labeling modifications are described in this submission:

- Design modifications to the Voyant Sheath including additional diameters and lengths
- Material change of the patient contacting components of the *Voyant System*
- Design modifications to the Voyant Imager
- Design modifications to the Image Control Box (ICB)
- Labeling changes to include updated images and instructions based upon the modified subject device and additional warning based upon the improved performance of the subject device.

The models for the subject device are listed in the table below:

**Voyant System Models:**

	Model Number	Package Contents**		
		Voyant Sheath	Voyant Imager	
<b>Sterile, Single-use, Voyant System</b>	VSY-1640	16mm OD	40mm length	16mm OD
	VSY-1650		50mm length	
	VSY-1660		60mm length	
	VSY-1670		70mm length	
	VSY-1680		80mm length	
	VSY-1840	18mm OD	40mm length	18mm OD
	VSY-1850		50mm length	
	VSY-1860		60mm length	
	VSY-1870		70mm length	
	VSY-1880		80mm length	
	VSY-2040	20mm OD	40mm length	20mm OD
	VSY-2050		50mm length	
	VSY-2060		60mm length	
	VSY-2070		70mm length	
	VSY-2080		80mm length	
	VSY-2240	22mm OD	40mm length	22mm OD
	VSY-2250		50mm length	
	VSY-2260		60mm length	
	VSY-2270		70mm length	
	VSY-2280		80mm length	
VSY-2640	26mm OD	40mm length	26mm OD	
VSY-2650		50mm length		
VSY-2660		60mm length		
VSY-2670		70mm length		
VSY-2680		80mm length		
<b>Non-sterile, Reusable, Accessories</b>	ICB-02***	Voyant Image Control Box		

\*\*One Voyant Sheath of specified diameter (OD) and length and one Voyant Imager of corresponding diameter (OD) are packaged together

\*\*\*VLBL08-002 Operator's Manual

**E. Indications for Use:**

The *Voyant System* is indicated to provide minimally invasive access, visualization, illumination, and magnification of the surgical area of the spine.

**F. Comparison of Technological Characteristics with the Predicate Device:**

As was established in this submission, the subject *Voyant System* is substantially equivalent to the predicate *Voyant System* (K181885) previously cleared by FDA for commercial distribution in the United States. There have been no design changes to the devices previously cleared in the predicate 510(k). The subject device has been shown to be substantially equivalent and has equivalent technological characteristics to its predicate device through comparison in areas including design, labeling/intended use, principles of operation, materials, and function.

**Design Comparison Table for Substantial Equivalence:**

Feature	PREDICATE DEVICE	SUBJECT DEVICE
	<i>Voyant System</i>	<i>Voyant System</i>
<b>510(k) Number</b>	K181885	K191579
<b>Product Code</b>	HRX	Same
<b>Indications for Use</b>	The <i>Voyant System</i> is indicated to provide minimally invasive access, visualization, illumination, and magnification of the surgical area of the spine.	Same
<b>Intended Use</b>	Visualization of surgical area.	Same
<b>Design</b>	<p>The <i>Voyant System</i> is an access and visualization system designed to provide illumination and magnification of the surgical area during minimally invasive spine surgery.</p> <p>The system consists of the <i>Voyant Sheath</i> and the <i>Voyant Imager</i> and Accessories. The <i>Voyant Sheath</i> is a tubular retractor used to provide access to the surgical area. The <i>Voyant Imager</i> is an external camera with LED light source that attaches to the <i>Voyant Sheath</i>. Accessories to the <i>Voyant Imager</i> include an <i>Image Control Box (ICB)</i> with connecting cables.</p> <p>The <i>Image Control Box</i> is a portable accessory used to control the image output of the <i>Voyant Imager</i>, which is displayed on a connected operating room HD monitor.</p>	Same
<b>Patient Contacting Materials (Tubular Retractor)</b>	Anodized 6061-T6 Aluminum and medical grade plastic and adhesive	Parylene coated 6061-T6 Aluminum and medical grade adhesive

Feature	PREDICATE DEVICE	SUBJECT DEVICE
	<i>Voyant System</i>	<i>Voyant System</i>
<b>Labeling (Single use/Reusable)</b>	Voyant Sheath and Voyant Imager: Sterile, Single use Image Control Box and Connecting Cables: Non-sterile, Reusable	Same
<b>Surgical Site Access Device</b>	Tubular retractors	Same
<b>Image Acquisition</b>	Image acquisition is achieved through an integrated camera external to the surgical opening	Same
<b>Image Processing</b>	The image is digitally processed.	Same
<b>Image Display</b>	External monitor connection.	Same
<b>Illumination</b>	Illumination is achieved via direct transmission using an LED light source.	Same
<b>Sterilization</b>	Voyant Sheath and Voyant Imager: <ul style="list-style-type: none"> <li>• Supplied sterile by Ethylene Oxide (EtO) to a sterility assurance level of <math>10^{-6}</math>, for single use</li> </ul> Image Control Box: Supplied non-sterile and reusable.	Same
<b>Performance Characteristics</b>	Camera resolution: 1080p Sensor resolution: 3280 x 2464 pixels Camera sensor: 1/4" CMOS Tubular retractor compressive force: $\geq 100N$ Image resolution: $\geq 11$ lp/mm at 63mm Working length: 6-13cm	Camera resolution: Same Sensor resolution: Same Camera sensor: Same Tubular retractor compressive force: Same Image resolution: $\geq 11$ lp/mm at 80mm Working length: 4-13cm

**G. Performance Data**

Non-clinical performance verification testing was performed to demonstrate that the subject *Voyant System* is substantially equivalent to the predicate device. Testing included design verification testing, software validation, electrical safety and electromagnetic emissions and immunity testing. Testing was performed in accordance with the following standards in order to establish equivalence to the predicate device:

<b>Test/Document Description</b>	<b>Applicable Test Standard</b>
Electrical Safety	IEC 60601-1
Electromagnetic Emissions and Immunity	IEC 60601-1-2
Software	IEC 62304
Risk Management	EN ISO 14971
Design Verification Testing	N/A

The results demonstrate that the subject *Voyant System* is substantially equivalent to the predicate.

**H. Conclusions**

Based on the indications for use, technological characteristics, non-clinical performance data, and a comparison to the predicate device, the subject *Voyant System* has been shown to be substantially equivalent to the legally marketed predicate device for its intended use.