



Alphatec Spine, Inc.
Ms. Ruby Zheng
Regulatory Affairs Specialist
5818 El Camino Real
Carlsbad, California 92008

October 18, 2019

Re: K191723

Trade/Device Name: ATEC IOM Accessory Instruments
Regulation Number: 21 CFR 874.1820
Regulation Name: Surgical Nerve Stimulator/Locator
Regulatory Class: Class II
Product Code: PDQ, ETN
Dated: September 16, 2019
Received: September 17, 2019

Dear Ms. Ruby Zheng:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmnm.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part

801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

for

Jay Gupta
Assistant Director
DHT5A: Division of Neurosurgical,
Neurointerventional
and Neurodiagnostic Devices
OHT5: Office of Neurological
and Physical Medicine Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K191723

Device Name

ATEC IOM Accessory Instruments

Indications for Use (Describe)

The ATEC IOM Accessory Instruments are utilized in spine surgical procedures to assist in location of the nerves during or after preparation and placement of implants (intervertebral fusion cages and pedicle screw fixation devices) in open and percutaneous minimally invasive approaches.

Type of Use (Select one or both, as applicable)

☒ Prescription Use (Part 21 CFR 801 Subpart D)

☐ Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."



This 510(k) summary of safety and effectiveness is being submitted in accordance with the requirements of 21 CFR 807.92.

I. SUBMITTER:

Alphatec Spine, Inc.
5818 El Camino Real
Carlsbad, CA 92008
Phone: (760) 431-9286
Fax: (760) 431-0289

Contact Person: Ruby Zheng
Regulatory Affairs Specialist

Date Summary Prepared: September 16, 2019

II. DEVICE

Trade or Proprietary Name: ATEC IOM Accessory Instruments
Common Name: Surgical nerve stimulator/locator.
Classification Name: Neurosurgical Nerve Locator
Regulation Number: 21 CFR 874.1820
Classification: Class II
Product Code: PDQ, ETN

III. LEGALLY MARKETED PREDICATE DEVICES

Predicate Device(s):

510(k)	Product Name	Clearance Date
K182617	ATEC IOM Accessory Instruments	May 9, 2019
K132138	RhythmLink Disposable Concentric Stimulating Probe	April 11, 2014



IV. DEVICE DESCRIPTION

The *ATEC IOM Accessory Instruments* are surgical instruments that provide electrical stimulation to the body to locate and identify nerves in either open, minimally invasive, or percutaneous procedures. These surgical instruments are compatible with common FDA cleared neuromonitoring platforms as they are connected via a compatible clip or probe depending on the system. The neuromonitoring capability provides the surgeon with spinal nerve location, proximity, and integrity information. This information assists the surgeon during targeting, bone preparation, and placement of orthopedic implants such as intervertebral fusion devices (e.g., interbodies) and bone screws (e.g., pedicle screws). All sterile instruments are single use only and all reusable instruments are offered non-sterile to be steam sterilized by the end user.

V. INDICATIONS FOR USE

The *ATEC IOM Accessory Instruments* are utilized in spine surgical procedures to assist in location of the nerves during or after preparation and placement of implants (intervertebral fusion cages and pedicle screw fixation devices) in open and percutaneous minimally invasive approaches.

VI. TECHNOLOGICAL COMPARISON TO PREDICATES

The technological design features of the subject implants were compared to the predicates in intended use, indications for use, design, function and technology and it was demonstrated that they are substantially equivalent. See **Table 1** below.

Table 1: Summary of Technological Comparison to Predicates

Attribute	Primary Predicate K182617	Additional Predicate K132138	Subject Device
	ATEC IOM Instruments	RhythmLink Disposable Concentric Stimulating Probe	ATEC IOM Instruments
Indications for Use	The ATEC IOM Accessory Instruments are utilized in spine surgical procedures to assist in location of the nerves during or after preparation and placement of implants (intervertebral fusion cages and pedicle screw fixation devices) in open and percutaneous minimally invasive approaches.	The RhythmLink Disposable Concentric Stimulating Probe is used to perform Intended Use(s) localized stimulation of neural tissue and to locate, identify and monitor cranial motor nerves, peripheral nerve and spinal nerve roots during surgery. The RhythmLink Disposable Concentric Stimulating Probe is a single patient use device	The ATEC IOM Accessory Instruments are utilized in spine surgical procedures to assist in location of the nerves during or after preparation and placement of implants (intervertebral fusion cages and pedicle screw fixation devices) in open and percutaneous minimally invasive approaches.
Regulation Number, Product Code, & Classification	21 CFR 874.1820 PDQ, ETN Class II	21 CFR 874.1820 ETN Class II	21 CFR 874.1820 PDQ, ETN Class II
IEC 60601 Compliant	Yes Protected Pin Design via compatible clip/probe Clause 56.3(c)	Yes DIN 42 802 touch proof connectors Clause 56.3(c)	Yes Protected Pin Design via compatible clip/probe Clause 56.3(c)
Instrument Type (Description)	Drills/Taps, Awls, Probes, Screwdrivers, Dilators (Sleeves), Guidewires, and Needles	Probes	Drills/Taps, Awls, Probes, Screwdrivers, Dilators (Sleeves), Guidewires, and Needles
Biocompatibility Patient Contact Duration	Limited patient duration contact (≤ 24 hours)	Limited patient duration contact (≤ 24 hours)	Limited patient duration contact (≤ 24 hours)
Biocompatible	Yes	Yes	Yes
Surgical approach	Open or Percutaneous/Minimally Invasive	Not specified	Open or Percutaneous/Minimally Invasive



Attribute	Primary Predicate K182617	Additional Predicate K132138	Subject Device
	ATEC IOM Instruments	RhythmLink Disposable Concentric Stimulating Probe	ATEC IOM Instruments
Sterility	Non-sterile Non-sterile devices are provided with validated steam sterilization parameters to assure an SAL of 10^{-6}	Sterile via EtO	Sterile via EtO and non-sterile Non-sterile devices are provided with validated steam sterilization parameters to assure an SAL of 10^{-6}
Reusable/Single Use	Guidewires, Targeting Needles and Dilators – Single Use Awls, Drills/Taps, Probes, Dilators (Sleeves), and Screwdrivers – Reusable	Single Use	Guidewires, Targeting Needles and Dilators – Single Use SafeOp Ball Tip Probe – Single Use Awls, Drills/Taps, Probes, Dilators (Sleeves), and Screwdrivers – Reusable
Compatible with Common Neuromonitoring Consoles & Software	Yes	Yes	Yes
Minimum exposed surface area	8.6 mm ²	14.6 mm ²	8.6 mm ²

VII. PERFORMANCE DATA

Performance testing includes IEC 60601-1 testing, reliability testing, and functional testing on insulation effectiveness and electrical resistance. **Table 2** below summarizes the testing which was performed on the Subject Devices to show substantial equivalence to the predicate device. Testing results demonstrated the subject *ATEC IOM Accessory Instruments* are appropriate for neuromonitoring applications and do not introduce a new worst case compared to other legally marketed devices cleared by FDA.

Table 2: Summary of Performance Testing

Test	Test Method Summary	Results
Electrical Safety Testing and/or Evaluation	Evaluation and testing was performed on the subject devices in accordance with IEC 60601-1: 2005.	All samples passed acceptance criteria
Reliability Testing	Testing was performed to verify that there is no adverse effect on the safety or effectiveness of the subject devices based on the intended environment and storage conditions.	All samples passed acceptance criteria
Functional Performance Testing and Verification Analysis	<ul style="list-style-type: none">• Insulation Effectiveness• Electrical Resistance	All samples passed acceptance criteria

Clinical Information

Not applicable; determination of substantial equivalence is not based on an assessment of clinical performance data.

VIII. BIOCOMPATIBILITY DATA

A risk analysis was performed taking into account nature of body contact and duration to categorization the use of existing data, end-specific testing, and endpoint assessment to cover the identified test methods. Additionally, data was leveraged by other means (e.g., authorized use of Master File, predicate and reference devices, well known and characterized materials) to support the biocompatibility of the subject devices.

Biocompatibility testing conducted per ISO 10993-1.

- Cytotoxicity
- Sensitization
- Irritation/Intracutaneous
- Reactivity
- Acute Systemic Toxicity
- Material Mediated Pyrogenicity
- Hemocompatibility



In conclusion, the *ATEC IOM Accessory Instruments* are manufactured from the same materials as other legally US-marketed devices.

IX. CONCLUSION

Based upon the information provided in this 510(k) submission, it has been determined that the subject devices are substantially equivalent to legally marketed devices in regards to indications for use, intended use, design, technology, and performance.